

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033121\
 Data File : BN014156.D
 Acq On : 01 Apr 2021 00:52
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 01:25:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 12:15:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	115	0.00
2	1,4-Dioxane	0.400	0.414	-3.5	110	0.00
3	n-Nitrosodimethylamine	0.400	0.352	12.0	105	0.00
4 S	2-Fluorophenol	0.400	0.405	-1.3	109	0.00
5 S	Phenol-d6	0.400	0.403	-0.8	110	0.00
6	bis(2-Chloroethyl)ether	0.400	0.416	-4.0	113	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	118	0.00
8 S	Nitrobenzene-d5	0.400	0.369	7.8	110	0.00
9	Naphthalene	0.400	0.413	-3.2	118	0.01
10	Hexachlorobutadiene	0.400	0.413	-3.2	118	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.421	-5.2	121	0.00
12	2-Methylnaphthalene	0.400	0.424	-6.0	121	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	127	0.00
14 S	2,4,6-Tribromophenol	0.400	0.370	7.5	125	0.01
15 S	2-Fluorobiphenyl	0.400	0.395	1.3	124	0.00
16	Acenaphthylene	0.400	0.369	7.8	124	0.01
17	Acenaphthene	0.400	0.398	0.5	127	0.00
18	Fluorene	0.400	0.415	-3.7	131	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	141	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.335	16.3	133	0.00
21	4-Bromophenyl-phenylether	0.400	0.404	-1.0	140	0.00
22	Hexachlorobenzene	0.400	0.402	-0.5	139	0.00
23	Atrazine	0.400	0.366	8.5	135	0.00
24	Pentachlorophenol	0.400	0.325	18.8	131	0.00
25	Phenanthrene	0.400	0.409	-2.2	141	0.00
26	Anthracene	0.400	0.392	2.0	141	0.00
27 SURR	Fluoranthene-d10	0.400	0.419	-4.7	149	0.00
28	Fluoranthene	0.400	0.423	-5.7	151	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	155	0.00
30	Pyrene	0.400	0.417	-4.2	149	0.00
31 S	Terphenyl-d14	0.400	0.433	-8.2	156	0.00
32	Benzo(a)anthracene	0.400	0.409	-2.2	158	0.00
33	Chrysene	0.400	0.423	-5.7	155	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.433	-8.2	174	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.415	-3.7	157	0.00
36 I	Perylene-d12	0.400	0.400	0.0	157	0.00
37	Benzo(b)fluoranthene	0.400	0.419	-4.7	162	0.00
38	Benzo(k)fluoranthene	0.400	0.418	-4.5	163	0.00
39 C	Benzo(a)pyrene	0.400	0.423	-5.7	167	0.00
40	Dibenzo(a,h)anthracene	0.400	0.398	0.5	156	0.00
41	Benzo(g,h,i)perylene	0.400	0.390	2.5	150	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4