

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033121\
 Data File : BN014170.D
 Acq On : 01 Apr 2021 09:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

mohammad
 4/1/2021 10:51:44 AM

Quant Time: Apr 01 09:48:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 12:15:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.684	152	5768	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	21030	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	10797	0.40	ng	0.00
19) Phenanthrene-d10	17.068	188	20253	0.40	ng	0.00
29) Chrysene-d12	21.250	240	19354	0.40	ng	0.00
36) Perylene-d12	23.469	264	20743	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	6179	0.42	ng	0.00
5) Phenol-d6	6.855	99	7663	0.41	ng	0.00
8) Nitrobenzene-d5	8.832	82	5373	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.057	152	12844	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1202	0.35	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	16234	0.40	ng	0.00
27) Fluoranthene-d10	19.099	212	22224	0.42	ng	0.00
31) Terphenyl-d14	19.700	244	17500	0.39	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.231	88	3278	0.44	ng	95
3) n-Nitrosodimethylamine	3.569	42	1997	0.35	ng	# 87
6) bis(2-Chloroethyl)ether	7.121	93	6205	0.41	ng	100
9) Naphthalene	10.505	128	21286	0.41	ng	100
10) Hexachlorobutadiene	10.796	225	3882	0.41	ng	# 99
12) 2-Methylnaphthalene	12.129	142	13825	0.41	ng	100
16) Acenaphthylene	14.029	152	15310	0.37	ng	99
17) Acenaphthene	14.384	154	11715	0.39	ng	100
18) Fluorene	15.374	166	14033	0.39	ng	99
20) 4,6-Dinitro-2-methylph...	15.450	198	824	0.33	ng	# 86
21) 4-Bromophenyl-phenylether	16.264	248	4059	0.39	ng	# 89
22) Hexachlorobenzene	16.374	284	4646	0.40	ng	99
23) Atrazine	16.532	200	2519	0.35	ng	99
24) Pentachlorophenol	16.715	266	1111	0.31	ng	98
25) Phenanthrene	17.104	178	22184	0.41	ng	100
26) Anthracene	17.189	178	17877	0.39	ng	100
28) Fluoranthene	19.129	202	23664	0.42	ng	100
30) Pyrene	19.491	202	24378	0.39	ng	100
32) Benzo(a)anthracene	21.239	228	21703	0.40	ng	100
33) Chrysene	21.282	228	25231	0.43	ng	100
34) Bis(2-ethylhexyl)phtha...	21.165	149	6567	0.40	ng	100
35) Indeno(1,2,3-cd)pyrene	25.694	276	33813	0.55	ng	99
37) Benzo(b)fluoranthene	22.805	252	29016m	0.40	ng	
38) Benzo(k)fluoranthene	22.849	252	28213	0.39	ng	98
39) Benzo(a)pyrene	23.373	252	26523	0.42	ng	# 89
40) Dibenzo(a,h)anthracene	25.707	278	28446	0.44	ng	99
41) Benzo(g,h,i)perylene	26.371	276	30117	0.44	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033121\
Data File : BN014170.D
Acq On : 01 Apr 2021 09:17
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM

Quant Time: Apr 01 09:48:07 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 31 12:15:38 2021
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.684 min Scan# 609
 Delta R.T. 0.000 min
 Lab File: BN014170.D
 Acq: 01 Apr 2021 09:17




Tgt Ion: 152 Resp: 5768

Ion	Ratio	Lower	Upper
152	100		
150	159.2	121.2	181.8
115	61.9	46.2	69.2





#2
 1,4-Dioxane
 Concen: 0.44 ng
 RT: 3.231 min Scan# 56
 Delta R.T. -0.008 min
 Lab File: BN014170.D
 Acq: 01 Apr 2021 09:17



Tgt Ion: 88 Resp: 3278


Ion	Ratio	Lower	Upper
88	100		
58	73.5	63.1	94.7
43	30.8	26.2	39.2



Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

mohammad
 4/1/2021 10:51:44 AM







#3
n-Nitrosodimethylamine
Concen: 0.35 ng
RT: 3.569 min Scan# 98
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion: 42 Resp: 1997
Ion Ratio Lower Upper
42 100
74 104.0 94.6 141.8
44 13.7 4.2 6.2#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM



#4
2-Fluorophenol
Concen: 0.42 ng
RT: 5.284 min Scan# 311
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17


Tgt Ion: 112 Resp: 6179
Ion Ratio Lower Upper
112 100
64 60.6 48.3 72.5
63 32.3 25.4 38.0





#5
Phenol-d6
Concen: 0.41 ng
RT: 6.855 min Scan# 506
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion: 99 Resp: 7663
Ion Ratio Lower Upper
99 100
42 22.9 18.2 27.4
71 33.9 26.7 40.1

Manual Integrations
APPROVED



mohammad
4/1/2021 10:51:44 AM




#6
bis(2-Chloroethyl)ether
Concen: 0.41 ng
RT: 7.121 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion: 93 Resp: 6205
Ion Ratio Lower Upper
93 100
63 81.4 65.1 97.7
95 33.2 26.6 39.8





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:	136	Resp:	21030
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.1	13.7
54	5.7	4.3	6.5
68	4.5	3.4	5.2

Manual Integrations
APPROVED



mohammad
4/1/2021 10:51:44 AM




#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17



Tgt Ion:	82	Resp:	5373
Ion	Ratio	Lower	Upper
82	100		
128	42.6	33.5	50.3
54	51.8	40.3	60.5









#9
Naphthalene
Concen: 0.41 ng
RT: 10.505 min Scan# 860
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:	128	Resp:	21286
Ion Ratio	Lower	Upper	
128	100		
129	11.5	9.2	13.8
127	13.5	10.8	16.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.796 min Scan# 883
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:	225	Resp:	3882
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.6	51.6	77.4









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.057 min Scan# 1063
Delta R.T. 0.004 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:152 Resp: 12844
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM



#12
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.129 min Scan# 1079
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17


Tgt Ion:142 Resp: 13825
Ion Ratio Lower Upper
142 100
141 88.7 71.1 106.7
115 35.9 29.0 43.6





#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:164 Resp: 10797
Ion Ratio Lower Upper
164 100
162 98.6 81.4 122.2
160 43.5 37.5 56.3

Manual Integrations
APPROVED



mohammad
4/1/2021 10:51:44 AM




#14
2,4,6-Tribromophenol
Concen: 0.35 ng
RT: 15.805 min Scan# 1424
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17



Tgt Ion:330 Resp: 1202
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.4 31.2 46.8









#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:172 Resp: 16234
Ion Ratio Lower Upper
172 100
171 34.4 26.9 40.3
170 22.7 17.8 26.8

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.029 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:152 Resp: 15310
Ion Ratio Lower Upper
152 100
151 18.9 15.4 23.2
153 13.0 10.3 15.5









#17
Acenaphthene
Concen: 0.39 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
154	100		
153	111.4	89.0	133.4
152	53.5	43.1	64.7




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM



#18
Fluorene
Concen: 0.39 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17


Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.0	77.9	116.9
167	13.5	10.6	16.0





#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.068 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:188 Resp: 20253
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.9 5.2 7.8

Manual Integrations
APPROVED



mohammad
4/1/2021 10:51:44 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.33 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion:198 Resp: 824
Ion Ratio Lower Upper
198 100
51 75.7 47.8 71.6#
105 47.2 35.4 53.2





#21
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:248 Resp: 4059
Ion Ratio Lower Upper
248 100
250 105.2 77.0 115.6
141 39.4 39.7 59.5#

Manual Integrations
APPROVED



mohammad
4/1/2021 10:51:44 AM




#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.374 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion:284 Resp: 4646
Ion Ratio Lower Upper
284 100
142 35.8 29.2 43.8
249 30.9 24.7 37.1






#23
Atrazine
Concen: 0.35 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion:	200	Resp:	2519
Ion	Ratio	Lower	Upper
200	100		
173	25.3	20.9	31.3
215	49.2	39.5	59.3




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED



mohammad
4/1/2021 10:51:44 AM




#24
Pentachlorophenol
Concen: 0.31 ng
RT: 16.715 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion:	266	Resp:	1111
Ion	Ratio	Lower	Upper
266	100		
264	63.5	50.2	75.2
268	63.2	52.5	78.7









#25
Phenanthrene
Concen: 0.41 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17





Tgt Ion:178 Resp: 22184
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.6
179 15.4 12.4 18.6



#26
Anthracene
Concen: 0.39 ng
RT: 17.189 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion:178 Resp: 17877
Ion Ratio Lower Upper
178 100
176 18.3 14.7 22.1
179 15.3 12.3 18.5




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
4/1/2021 10:51:44 AM




#27
Fluoranthene-d10
Concen: 0.42 ng
RT: 19.099 min Scan# 1823
Delta R.T. 0.003 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17





Tgt Ion:212 Resp: 22224
Ion Ratio Lower Upper
212 100
106 12.7 10.1 15.1
104 7.0 5.5 8.3



#28
Fluoranthene
Concen: 0.42 ng
RT: 19.129 min Scan# 1829
Delta R.T. 0.003 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion:202 Resp: 23664
Ion Ratio Lower Upper
202 100
101 11.1 8.8 13.2
203 17.0 13.8 20.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
4/1/2021 10:51:44 AM







#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.003 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:240 Resp: 19354
Ion Ratio Lower Upper
240 100
120 10.5 8.4 12.6
236 26.3 21.0 31.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM



#30
Pyrene
Concen: 0.39 ng
RT: 19.491 min Scan# 1902
Delta R.T. 0.003 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17


Tgt Ion:202 Resp: 24378
Ion Ratio Lower Upper
202 100
200 20.4 16.2 24.4
203 17.5 13.9 20.9





#31
Terphenyl-d14
Concen: 0.39 ng
RT: 19.700 min Scan# 1944
Delta R.T. 0.003 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:244 Resp: 17500
Ion Ratio Lower Upper
244 100
212 7.6 6.0 9.0
122 11.2 8.6 13.0




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM



#32
Benzo(a)anthracene
Concen: 0.40 ng
RT: 21.239 min Scan# 2122
Delta R.T. 0.003 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:228 Resp: 21703
Ion Ratio Lower Upper
228 100
226 25.9 20.6 31.0
229 20.2 16.2 24.2









#33
Chrysene
Concen: 0.43 ng
RT: 21.282 min Scan# 2126
Delta R.T. -0.008 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:	228	Resp:	25231
Ion Ratio	Lower	Upper	
228	100		
226	30.7	24.4	36.6
229	19.2	15.5	23.3

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.40 ng
RT: 21.165 min Scan# 2115
Delta R.T. -0.008 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:	149	Resp:	6567
Ion Ratio	Lower	Upper	
149	100		
167	26.9	21.4	32.2
279	4.0	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.55 ng
RT: 25.694 min Scan# 3192
Delta R.T. 0.006 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:	276	Resp:	33813
Ion Ratio	Lower	Upper	
276	100		
138	23.0	19.4	29.2
277	24.9	19.8	29.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
4/1/2021 10:51:44 AM




#36
Perylene-d12
Concen: 0.40 ng
RT: 23.469 min Scan# 2542
Delta R.T. 0.003 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17

Tgt Ion:	264	Resp:	20743
Ion Ratio	Lower	Upper	
264	100		
260	24.4	19.8	29.8
265	33.7	30.7	46.1









#37
Benzo(b)fluoranthene
Concen: 0.40 ng m
RT: 22.805 min Scan# 2348
Delta R.T. 0.003 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17





Tgt Ion:252 Resp: 29016
Ion Ratio Lower Upper
252 100
253 26.6 20.6 30.8
125 17.8 9.4 14.2#



#38
Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 22.849 min Scan# 2361
Delta R.T. 0.006 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion:252 Resp: 28213
Ion Ratio Lower Upper
252 100
253 26.3 20.6 30.8
125 10.9 9.5 14.3




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
4/1/2021 10:51:44 AM




#39
Benzo(a)pyrene
Concen: 0.42 ng
RT: 23.373 min Scan# 2514
Delta R.T. 0.006 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17





Tgt Ion:252 Resp: 26523
Ion Ratio Lower Upper
252 100
253 29.1 21.9 32.9
125 24.6 10.8 16.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.44 ng
RT: 25.707 min Scan# 3196
Delta R.T. 0.006 min
Lab File: BN014170.D
Acq: 01 Apr 2021 09:17




Tgt Ion:278 Resp: 28446
Ion Ratio Lower Upper
278 100
139 16.3 13.4 20.0
279 26.9 21.3 31.9




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED



mohammad
4/1/2021 10:51:44 AM



#41
 Benzo(g,h,i)perylene
 Concen: 0.44 ng
 RT: 26.371 min Scan# 3390
 Delta R.T. 0.003 min
 Lab File: BN014170.D
 Acq: 01 Apr 2021 09:17



Tgt Ion:	276	Resp:	30117
Ion Ratio	Lower	Upper	
276	100		
277	24.2	20.1	30.1
138	20.8	17.0	25.4



Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

mohammad
 4/1/2021 10:51:44 AM