



Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN033121\  
 Data File : BN014138.D  
 Acq On : 31 Mar 2021 11:02  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN032621.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Mar 31 12:15:38 2021






AutoFind: Scans 2452, 2453, 2454; Background Corrected with Scan 2445

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 40.3         | 169413     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 2711       | PASS                |
| 69             | 198             | 0.00            | 100             | 38.9         | 163341     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 982        | PASS                |
| 127            | 198             | 10              | 80              | 48.8         | 204949     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.2          | 664        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 420245     | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 28528      | PASS                |
| 275            | 198             | 10              | 60              | 24.9         | 104491     | PASS                |
| 365            | 198             | 1               | 100             | 3.5          | 14686      | PASS                |
| 441            | 198             | 0.01            | 100             | 10.9         | 45605      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 323051     | PASS                |
| 443            | 442             | 15              | 24              | 17.6         | 56931      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN033121\  
Data File : BN014138.D  
Acq On : 31 Mar 2021 11:02  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Mar 31 11:48:08 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Mar 26 11:50:50 2021  
Response via : Initial Calibration



TIC: BN014138.D\data.ms

**(77) Benzidine**

19.316min (+ 0.018) 0.00 ng




response 2779815

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.45  |
| 183.00 | 13.20  | 11.61  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN033121\  
 Data File : BN014138.D  
 Acq On : 31 Mar 2021 11:02  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Quant Time: Mar 31 11:48:08 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Mar 26 11:50:50 2021  
 Response via : Initial Calibration



TIC: BN014138.D\data.ms

**(70) Pentachlorophenol (C)**

16.716min (+ 0.012) 22567.84 ng

response 494659

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 61.52  |
| 264.00 | 61.60  | 57.63  |
| 0.00   | 0.00   | 0.00   |

# DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 3/31/2021     | BNA_N            | <u>BN014138.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 1348607          | 20.55              |
| DDD           | 32258            | 20.11              |
| DDE           | 2713             | 19.6               |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 34971         | 1383578          | 2.53               |
|               |                  |                    |