

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040121\
 Data File : BN014172.D
 Acq On : 01 Apr 2021 10:31
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 11:05:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 12:15:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	6376	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	22848	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	11888	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	22856	0.40	ng	0.00
29) Chrysene-d12	21.247	240	20025	0.40	ng	# 0.00
36) Perylene-d12	23.469	264	19863	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	6744	0.41	ng	0.00
5) Phenol-d6	6.855	99	8359	0.40	ng	0.00
8) Nitrobenzene-d5	8.832	82	5862	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.057	152	14147	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1311	0.34	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	17921	0.40	ng	0.00
27) Fluoranthene-d10	19.096	212	24475	0.41	ng	0.00
31) Terphenyl-d14	19.697	244	18876	0.41	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.231	88	3379	0.41	ng	100
3) n-Nitrosodimethylamine	3.569	42	2028	0.32	ng	# 96
6) bis(2-Chloroethyl)ether	7.120	93	6771	0.41	ng	100
9) Naphthalene	10.505	128	23440	0.42	ng	100
10) Hexachlorobutadiene	10.796	225	4206	0.41	ng	# 100
12) 2-Methylnaphthalene	12.129	142	15173	0.41	ng	99
16) Acenaphthylene	14.029	152	16711	0.37	ng	99
17) Acenaphthene	14.384	154	13106	0.40	ng	100
18) Fluorene	15.374	166	15889	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.450	198	928	0.33	ng	97
21) 4-Bromophenyl-phenylether	16.264	248	4700	0.40	ng	92
22) Hexachlorobenzene	16.373	284	5196	0.40	ng	99
23) Atrazine	16.531	200	2795	0.35	ng	99
24) Pentachlorophenol	16.714	266	1341	0.33	ng	95
25) Phenanthrene	17.103	178	25006	0.41	ng	100
26) Anthracene	17.189	178	20020	0.39	ng	100
28) Fluoranthene	19.126	202	26349	0.41	ng	100
30) Pyrene	19.488	202	26609	0.41	ng	100
32) Benzo(a)anthracene	21.237	228	22197	0.40	ng	99
33) Chrysene	21.290	228	26116	0.43	ng	97
34) Bis(2-ethylhexyl)phtha...	21.173	149	6245	0.37	ng	100
35) Indeno(1,2,3-cd)pyrene	25.691	276	31433	0.49	ng	98
37) Benzo(b)fluoranthene	22.846	252	26902	0.38	ng	99
38) Benzo(k)fluoranthene	22.846	252	26902	0.39	ng	99
39) Benzo(a)pyrene	23.370	252	25680	0.43	ng	# 85
40) Dibenzo(a,h)anthracene	25.708	278	26385	0.43	ng	100
41) Benzo(g,h,i)perylene	26.375	276	28182	0.43	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040121\
 Data File : BN014172.D
 Acq On : 01 Apr 2021 10:31
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 11:05:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 12:15:38 2021
 Response via : Initial Calibration




Abundance Scan 610 (7.692 min): BN014120.D\data.ms (-602) (#1)





1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.692 min Scan# 610
Delta R.T. 0.008 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	152	Resp:	6376
Ion	Ratio	Lower	Upper
152	100		
150	151.5	121.2	181.8
115	57.2	46.2	69.2







Abundance Scan 56 (3.231 min): BN014120.D\data.ms (-52) (#2)



1,4-Dioxane
Concen: 0.41 ng
RT: 3.231 min Scan# 56
Delta R.T. -0.008 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	88	Resp:	3379
Ion	Ratio	Lower	Upper
88	100		
58	79.4	63.1	94.7
43	32.8	26.2	39.2









n-Nitrosodimethylamine
Concen: 0.32 ng
RT: 3.569 min Scan# 98
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 2028
Ion Ratio Lower Upper
42 100
74 114.3 94.6 141.8
44 7.4 4.2 6.2#



2-Fluorophenol
Concen: 0.41 ng
RT: 5.284 min Scan# 311
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion: 112 Resp: 6744
Ion Ratio Lower Upper
112 100
64 60.3 48.3 72.5
63 31.6 25.4 38.0









Phenol-d6
 Concen: 0.40 ng
 RT: 6.855 min Scan# 506
 Delta R.T. -0.000 min
 Lab File: BN014172.D
 Acq: 01 Apr 2021 10:31




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	99	Resp:	8359
Ion	Ratio	Lower	Upper
99	100		
42	21.8	18.2	27.4
71	33.1	26.7	40.1



bis(2-Chloroethyl)ether
 Concen: 0.41 ng
 RT: 7.120 min Scan# 539
 Delta R.T. -0.000 min
 Lab File: BN014172.D
 Acq: 01 Apr 2021 10:31




Tgt Ion:	93	Resp:	6771
Ion	Ratio	Lower	Upper
93	100		
63	81.4	65.1	97.7
95	32.9	26.6	39.8




Abundance Scan 857 (10.467 min): BN014120.D\data.ms (-857) (-)

Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4






Tgt Ion:	136	Resp:	22848
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.1	13.7
54	5.8	4.3	6.5
68	4.5	3.4	5.2




Abundance Scan 728 (8.832 min): BN014120.D\data.ms (-724) (-)


Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31



Tgt Ion:	82	Resp:	5862
Ion Ratio	Lower	Upper	
82	100		
128	40.1	33.5	50.3
54	51.3	40.3	60.5






Abundance Scan 860 (10.505 min): BN014120.D\data.ms (-855) (-)




Naphthalene
Concen: 0.42 ng
RT: 10.505 min Scan# 860
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	128	Resp:	23440
Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.2	13.8
127	13.4	10.8	16.2






Abundance Scan 883 (10.796 min): BN014120.D\data.ms (-879) (-)




Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.796 min Scan# 883
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	225	Resp:	4206
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.6	51.6	77.4





Abundance Scan 1062 (12.053 min): BN014120.D\data.ms (-1019) (-)




2-Methylnaphthalene-d10
Concen: 0.41 ng
RT: 12.057 min Scan# 1063
Delta R.T. 0.004 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 14147
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1



Abundance Scan 1063 (12.057 min): BN014172.D\data.ms (-1017) (-)





Abundance Scan 1078 (12.124 min): BN014120.D\data.ms (-1012) (-)





2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.129 min Scan# 1079
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:142 Resp: 15173
Ion Ratio Lower Upper
142 100
141 88.2 71.1 106.7
115 35.8 29.0 43.6



Abundance Scan 1079 (12.129 min): BN014172.D\data.ms (-1057) (-)









Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.321 min Scan# 1307
 Delta R.T. -0.000 min
 Lab File: BN014172.D
 Acq: 01 Apr 2021 10:31




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	164	Resp:	11888
Ion	Ratio	Lower	Upper
164	100		
162	98.8	81.4	122.2
160	44.2	37.5	56.3




2,4,6-Tribromophenol
 Concen: 0.34 ng
 RT: 15.805 min Scan# 1424
 Delta R.T. -0.000 min
 Lab File: BN014172.D
 Acq: 01 Apr 2021 10:31

Tgt Ion:	330	Resp:	1311
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	37.8	31.2	46.8





Abundance Scan 1198 (12.938 min): BN014120.D\data.ms (-1198) (-)




2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	17921
Ion	Ratio	Lower	Upper
172	100		
171	34.9	26.9	40.3
170	23.0	17.8	26.8



Abundance Scan 1198 (12.938 min): BN014172.D\data.ms (-1190) (-)





Abundance Scan 1284 (14.029 min): BN014120.D\data.ms (-1216) (-)





Acenaphthylene
Concen: 0.37 ng
RT: 14.029 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	152	Resp:	16711
Ion	Ratio	Lower	Upper
152	100		
151	18.8	15.4	23.2
153	13.0	10.3	15.5



Abundance Scan 1284 (14.029 min): BN014172.D\data.ms (-1268) (-)









Acenaphthene
Concen: 0.40 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:154 Resp: 13106
Ion Ratio Lower Upper
154 100
153 111.1 89.0 133.4
152 54.0 43.1 64.7




Fluorene
Concen: 0.40 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:166 Resp: 15889
Ion Ratio Lower Upper
166 100
165 98.2 77.9 116.9
167 13.6 10.6 16.0






Abundance Scan 1526 (17.055 min): BN014120.D\data.ms (-1510) (-)




Phenanthrene-d10
Concen: 0.40 ng
RT: 17.067 min Scan# 1527
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	188	Resp:	22856
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	6.8	5.2	7.8

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4






Abundance Scan 1395 (15.437 min): BN014120.D\data.ms (-1320) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.33 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	198	Resp:	928
Ion Ratio	Lower	Upper	
198	100		
51	60.2	47.8	71.6
105	40.8	35.4	53.2






Abundance Scan 1460 (16.252 min): BN014120.D\data.ms (-1425) (-)




4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.264 min Scan# 1461
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	248	Resp:	4700
Ion	Ratio	Lower	Upper
248	100		
250	101.8	77.0	115.6
141	41.2	39.7	59.5






Abundance Scan 1469 (16.361 min): BN014120.D\data.ms (-1423) (-)




Hexachlorobenzene
Concen: 0.40 ng
RT: 16.373 min Scan# 1470
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	284	Resp:	5196
Ion	Ratio	Lower	Upper
284	100		
142	36.5	29.2	43.8
249	31.6	24.7	37.1





Abundance Scan 1482 (16.519 min): BN014120.D\data.ms (-1426) (-)




Atrazine
Concen: 0.35 ng
RT: 16.531 min Scan# 1483
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	2795
Ion Ratio	Lower	Upper	
200	100		
173	24.9	20.9	31.3
215	49.3	39.5	59.3



Abundance Scan 1483 (16.531 min): BN014172.D\data.ms (-1442) (-)





Abundance Scan 1497 (16.702 min): BN014120.D\data.ms (-1424) (-)




Pentachlorophenol
Concen: 0.33 ng
RT: 16.714 min Scan# 1498
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31


Tgt Ion:	266	Resp:	1341
Ion Ratio	Lower	Upper	
266	100		
264	70.2	50.2	75.2
268	66.0	52.5	78.7



Abundance Scan 1498 (16.714 min): BN014172.D\data.ms (-1490) (-)



Abundance Scan 1529 (17.091 min): BN014120.D\data.ms (-1525) (-)





Phenanthrene
Concen: 0.41 ng
RT: 17.103 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31


Tgt Ion:	178	Resp:	25006
Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.6
179	15.6	12.4	18.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Abundance Scan 1530 (17.103 min): BN014172.D\data.ms



Abundance Scan 1530 (17.103 min): BN014172.D\data.ms (-1505) (-)





Abundance Scan 1536 (17.177 min): BN014120.D\data.ms (-1529) (-)




Anthracene
Concen: 0.39 ng
RT: 17.189 min Scan# 1537
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	178	Resp:	20020
Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.7	22.1
179	15.5	12.3	18.5


Abundance Scan 1537 (17.189 min): BN014172.D\data.ms



Abundance Scan 1537 (17.189 min): BN014172.D\data.ms (-1529) (-)





Abundance Scan 1821 (19.081 min): BN014120.D\data.ms (-1821) (-)




Fluoranthene-d10
Concen: 0.41 ng
RT: 19.096 min Scan# 1824
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 24475
Ion Ratio Lower Upper
212 100
106 12.6 10.1 15.1
104 7.0 5.5 8.3



Abundance Scan 1824 (19.096 min): BN014172.D\data.ms (-1784) (-)





Abundance Scan 1827 (19.111 min): BN014120.D\data.ms (-1827) (-)




Fluoranthene
Concen: 0.41 ng
RT: 19.126 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31


Tgt Ion:202 Resp: 26349
Ion Ratio Lower Upper
202 100
101 10.9 8.8 13.2
203 17.2 13.8 20.6



Abundance Scan 1830 (19.126 min): BN014172.D\data.ms (-1790) (-)





Abundance Scan 2122 (21.226 min): BN014120.D\data.ms (-2120) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.247 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	240	Resp:	20025
Ion	Ratio	Lower	Upper
240	100		
120	13.5	8.4	12.6#
236	27.5	21.0	31.6



Abundance Scan 2124 (21.247 min): BN014172.D\data.ms (-2105) (-)





Abundance Scan 1900 (19.473 min): BN014120.D\data.ms (-1890) (-)




Pyrene
Concen: 0.41 ng
RT: 19.488 min Scan# 1903
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31


Tgt Ion:	202	Resp:	26609
Ion	Ratio	Lower	Upper
202	100		
200	20.4	16.2	24.4
203	17.4	13.9	20.9



Abundance Scan 1903 (19.488 min): BN014172.D\data.ms (-1863) (-)





Abundance Scan 1942 (19.682 min): BN014120.D\data.ms (-1934) (-)




Terphenyl-d14
Concen: 0.41 ng
RT: 19.697 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 18876
Ion Ratio Lower Upper
244 100
212 7.7 6.0 9.0
122 12.1 8.6 13.0



Abundance Scan 1945 (19.697 min): BN014172.D\data.ms (-1915) (-)





Abundance Scan 2121 (21.215 min): BN014120.D\data.ms (-2195) (-)




Benzo(a)anthracene
Concen: 0.40 ng
RT: 21.237 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31


Tgt Ion:228 Resp: 22197
Ion Ratio Lower Upper
228 100
226 26.3 20.6 31.0
229 19.9 16.2 24.2



Abundance Scan 2123 (21.237 min): BN014172.D\data.ms (-2104) (-)




Abundance Scan 2125 (21.258 min): BN014120.D\data.ms (-2125) (-)





Chrysene
Concen: 0.43 ng
RT: 21.290 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.5	24.4	36.6
229	19.9	15.5	23.3





Abundance Scan 2115 (21.152 min): BN014120.D\data.ms (-2109) (-)



Bis(2-ethylhexyl)phthalate
Concen: 0.37 ng
RT: 21.173 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31




Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.9	21.4	32.2
279	4.5	3.5	5.3






Abundance Scan 3176 (25.633 min): BN014120.D\data.ms (-3135) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.49 ng
RT: 25.691 min Scan# 3193
Delta R.T. 0.003 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
276	100		
138	23.0	19.4	29.2
277	25.0	19.8	29.6






Abundance Scan 2533 (23.432 min): BN014120.D\data.ms (-2536) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.469 min Scan# 2544
Delta R.T. 0.003 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.7	19.8	29.8
265	34.4	30.7	46.1






Abundance Scan 2340 (22.771 min): BN014120.D\data.ms (-2340) (-)




Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 22.846 min Scan# 2362
Delta R.T. 0.044 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	252	Resp:	26902
Ion Ratio	Lower	Upper	
252	100		
253	25.9	20.6	30.8
125	11.0	9.4	14.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4






Abundance Scan 2352 (22.812 min): BN014120.D\data.ms (-2346) (-)




Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 22.846 min Scan# 2362
Delta R.T. 0.003 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Tgt Ion:	252	Resp:	26902
Ion Ratio	Lower	Upper	
252	100		
253	25.9	20.6	30.8
125	11.0	9.5	14.3





Abundance Scan 2505 (23.336 min): BN014120.D\data.ms (-2439) (-)




Benzo(a)pyrene
Concen: 0.43 ng
RT: 23.370 min Scan# 2515
Delta R.T. 0.003 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 25680
Ion Ratio Lower Upper
252 100
253 29.4 21.9 32.9
125 29.1 10.8 16.2#



Abundance Scan 2515 (23.370 min): BN014172.D\data.ms (-2456) (-)





Abundance Scan 3179 (25.643 min): BN014120.D\data.ms (-3140) (-)




Dibenzo(a,h)anthracene
Concen: 0.43 ng
RT: 25.708 min Scan# 3198
Delta R.T. 0.007 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31


Tgt Ion:278 Resp: 26385
Ion Ratio Lower Upper
278 100
139 16.6 13.4 20.0
279 26.5 21.3 31.9



Abundance Scan 3198 (25.708 min): BN014172.D\data.ms (-3173) (-)




Abundance Scan 3372 (26.303 min): BN014120.D\data.ms (-3349) (-)



Benzo(g,h,i)perylene
Concen: 0.43 ng
RT: 26.375 min Scan# 3393
Delta R.T. 0.007 min
Lab File: BN014172.D
Acq: 01 Apr 2021 10:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Abundance Scan 3393 (26.375 min): BN014172.D\data.ms



Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.7	20.1	30.1
138	20.9	17.0	25.4

Abundance Scan 3393 (26.375 min): BN014172.D\data.ms (-3333) (-)

