

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040121\
 Data File : BN014172.D
 Acq On : 01 Apr 2021 10:31
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 11:05:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 12:15:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.412	-3.0	93	0.00
3	n-Nitrosodimethylamine	0.400	0.321	19.8	82	0.00
4 S	2-Fluorophenol	0.400	0.412	-3.0	96	0.00
5 S	Phenol-d6	0.400	0.403	-0.8	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.409	-2.2	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	0.00
8 S	Nitrobenzene-d5	0.400	0.377	5.8	92	0.00
9	Naphthalene	0.400	0.416	-4.0	97	0.00
10	Hexachlorobutadiene	0.400	0.407	-1.7	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.407	-1.7	96	0.00
12	2-Methylnaphthalene	0.400	0.411	-2.7	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.400	0.343	14.2	87	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	95	0.00
16	Acenaphthylene	0.400	0.367	8.3	93	0.00
17	Acenaphthene	0.400	0.398	0.5	96	0.00
18	Fluorene	0.400	0.400	0.0	96	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.326	18.5	91	0.00
21	4-Bromophenyl-phenylether	0.400	0.400	0.0	96	0.00
22	Hexachlorobenzene	0.400	0.396	1.0	95	0.00
23	Atrazine	0.400	0.348	13.0	89	0.00
24	Pentachlorophenol	0.400	0.328	18.0	92	0.00
25	Phenanthrene	0.400	0.408	-2.0	98	0.00
26	Anthracene	0.400	0.388	3.0	97	0.00
27 SURR	Fluoranthene-d10	0.400	0.409	-2.2	101	0.00
28	Fluoranthene	0.400	0.411	-2.7	102	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	108	0.00
30	Pyrene	0.400	0.407	-1.7	102	0.00
31 S	Terphenyl-d14	0.400	0.409	-2.2	103	0.00
32	Benzo(a)anthracene	0.400	0.400	0.0	108	0.00
33	Chrysene	0.400	0.426	-6.5	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.369	7.8	103	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.494	-23.5	130	0.00
36 I	Perylene-d12	0.400	0.400	0.0	121	0.00
37	Benzo(b)fluoranthene	0.400	0.406	-1.5	121	0.00
38	Benzo(k)fluoranthene	0.400	0.393	1.8	118	0.00
39 C	Benzo(a)pyrene	0.400	0.427	-6.7	131	0.00
40	Dibenzo(a,h)anthracene	0.400	0.427	-6.7	130	0.00
41	Benzo(g,h,i)perylene	0.400	0.429	-7.2	128	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4