

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040221\
 Data File : BN014189.D
 Acq On : 02 Apr 2021 13:24
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Apr 02 14:34:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 02 14:32:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.684	152	4593	0.40	ng	0.00
7) Naphthalene-d8	10.454	136	16502	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	9399	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	19796	0.40	ng	0.00
29) Chrysene-d12	21.247	240	21319	0.40	ng	0.00
36) Perylene-d12	23.466	264	20828	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	9506	0.81	ng	0.00
5) Phenol-d6	6.855	99	11971	0.80	ng	0.00
8) Nitrobenzene-d5	8.832	82	9743	0.87	ng	0.00
11) 2-Methylnaphthalene-d10	12.053	152	28272	1.13	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	2590	0.86	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	28025	0.79	ng	0.00
27) Fluoranthene-d10	19.096	212	61865	1.19	ng	0.00
31) Terphenyl-d14	19.697	244	37508	0.76	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	4393	0.74	ng	97
3) n-Nitrosodimethylamine	3.561	42	3434	0.75	ng	# 91
6) bis(2-Chloroethyl)ether	7.112	93	9887	0.83	ng	98
9) Naphthalene	10.505	128	34887	0.86	ng	100
10) Hexachlorobutadiene	10.796	225	6471	0.87	ng	# 99
12) 2-Methylnaphthalene	12.124	142	23468	0.88	ng	99
16) Acenaphthylene	14.029	152	29982	0.83	ng	100
17) Acenaphthene	14.372	154	21524	0.83	ng	100
18) Fluorene	15.361	166	27091	0.86	ng	100
20) 4,6-Dinitro-2-methylph...	15.450	198	2061	0.84	ng	# 73
21) 4-Bromophenyl-phenylether	16.264	248	8718	0.86	ng	99
22) Hexachlorobenzene	16.373	284	9613	0.85	ng	99
23) Atrazine	16.532	200	5881	0.85	ng	98
24) Pentachlorophenol	16.714	266	1859	0.53	ng	97
25) Phenanthrene	17.104	178	45055	0.85	ng	100
26) Anthracene	17.189	178	38484	0.86	ng	100
28) Fluoranthene	19.126	202	53462	0.96	ng	100
30) Pyrene	19.488	202	53419	0.77	ng	100
32) Benzo(a)anthracene	21.237	228	53997	0.91	ng	100
33) Chrysene	21.279	228	55785	0.85	ng	99
34) Bis(2-ethylhexyl)phtha...	21.173	149	21349	1.18	ng	98
35) Indeno(1,2,3-cd)pyrene	25.687	276	67649	1.00	ng	100
37) Benzo(b)fluoranthene	22.799	252	60892	0.83	ng	98
38) Benzo(k)fluoranthene	22.843	252	57313	0.80	ng	98
39) Benzo(a)pyrene	23.367	252	53228	0.84	ng	98
40) Dibenzo(a,h)anthracene	25.705	278	57271	0.88	ng	99
41) Benzo(g,h,i)perylene	26.369	276	59444	0.86	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040221\
Data File : BN014189.D
Acq On : 02 Apr 2021 13:24
Operator : CG/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Quant Time: Apr 02 14:34:27 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 02 14:32:17 2021
Response via : Initial Calibration






Abundance Scan 609 (7.684 min): BN014190.D\data.ms (-602) (#1)




1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.684 min Scan# 609
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:152 Resp: 4593
Ion Ratio Lower Upper
152 100
150 156.1 126.2 189.4
115 60.1 49.8 74.6







Abundance Scan 56 (3.231 min): BN014190.D\data.ms (-53) (#2)



1,4-Dioxane
Concen: 0.74 ng
RT: 3.231 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion: 88 Resp: 4393
Ion Ratio Lower Upper
88 100
58 81.2 63.2 94.8
43 34.6 25.1 37.7









n-Nitrosodimethylamine
Concen: 0.75 ng
RT: 3.561 min Scan# 97
Delta R.T. -0.008 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion: 42 Resp: 3434
Ion Ratio Lower Upper
42 100
74 113.6 98.8 148.2
44 7.1 0.0 0.0#



2-Fluorophenol
Concen: 0.81 ng
RT: 5.284 min Scan# 311
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion: 112 Resp: 9506
Ion Ratio Lower Upper
112 100
64 60.0 48.3 72.5
63 31.3 25.8 38.6









Phenol-d6
 Concen: 0.80 ng
 RT: 6.855 min Scan# 506
 Delta R.T. 0.000 min
 Lab File: BN014189.D
 Acq: 02 Apr 2021 13:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	99	Resp:	11971
Ion	Ratio	Lower	Upper
99	100		
42	21.6	17.8	26.6
71	33.3	26.6	40.0




bis(2-Chloroethyl)ether
 Concen: 0.83 ng
 RT: 7.112 min Scan# 538
 Delta R.T. 0.000 min
 Lab File: BN014189.D
 Acq: 02 Apr 2021 13:24

Tgt Ion:	93	Resp:	9887
Ion	Ratio	Lower	Upper
93	100		
63	82.2	64.2	96.4
95	33.6	26.6	40.0






Abundance Scan 856 (10.454 min): BN014190.D\data.ms (-856#7-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.454 min Scan# 856
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:	136	Resp:	16502
Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.1	13.7
54	10.5	8.2	12.4
68	8.0	6.1	9.1






Abundance Scan 728 (8.832 min): BN014190.D\data.ms (-724#8)




Nitrobenzene-d5
Concen: 0.87 ng
RT: 8.832 min Scan# 728
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	82	Resp:	9743
Ion	Ratio	Lower	Upper
82	100		
128	45.6	36.8	55.2
54	47.2	38.7	58.1






Abundance Scan 860 (10.505 min): BN014190.D\data.ms (-8539) (-)




Naphthalene
Concen: 0.86 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:	128	Resp:	34887
Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.1	13.7
127	13.2	10.6	15.8






Abundance Scan 883 (10.796 min): BN014190.D\data.ms (-8710) (-)




Hexachlorobutadiene
Concen: 0.87 ng
RT: 10.796 min Scan# 883
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	225	Resp:	6471
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.1	50.8	76.2





Abundance Scan 1062 (12.053 min): BN014190.D\data.ms (-1017) (-)




2-Methylnaphthalene-d10
Concen: 1.13 ng
RT: 12.053 min Scan# 1062
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 28272
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1



Abundance Scan 1062 (12.053 min): BN014189.D\data.ms (-1017) (-)





Abundance Scan 1078 (12.124 min): BN014190.D\data.ms (-1017) (-)




2-Methylnaphthalene
Concen: 0.88 ng
RT: 12.124 min Scan# 1078
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:142 Resp: 23468
Ion Ratio Lower Upper
142 100
141 88.8 71.6 107.4
115 35.9 29.2 43.8




Abundance Scan 1078 (12.124 min): BN014189.D\data.ms (-1055) (-)






Abundance Scan 1307 (14.321 min): BN014190.D\data.ms (-1307) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:	164	Resp:	9399
Ion	Ratio	Lower	Upper
164	100		
162	96.9	77.4	116.2
160	43.5	34.2	51.4




Abundance Scan 1424 (15.805 min): BN014190.D\data.ms (-1414) (-)

2,4,6-Tribromophenol
Concen: 0.86 ng
RT: 15.805 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	330	Resp:	2590
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	36.3	29.8	44.6





Abundance Scan 1198 (12.938 min): BN014190.D\data.ms (-1190) (-)




2-Fluorobiphenyl
Concen: 0.79 ng
RT: 12.938 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	172	Resp:	28025
Ion	Ratio	Lower	Upper
172	100		
171	33.7	27.3	40.9
170	22.3	18.1	27.1



Abundance Scan 1198 (12.938 min): BN014189.D\data.ms (-1190) (-)





Abundance Scan 1284 (14.029 min): BN014190.D\data.ms (-1215) (-)





Acenaphthylene
Concen: 0.83 ng
RT: 14.029 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	152	Resp:	29982
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.1	22.7
153	13.1	10.5	15.7



Abundance Scan 1284 (14.029 min): BN014189.D\data.ms (-1268) (-)









Acenaphthene
Concen: 0.83 ng
RT: 14.372 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	154	Resp:	21524
Ion	Ratio	Lower	Upper
154	100		
153	111.4	89.2	133.8
152	54.2	43.8	65.6



Fluorene
Concen: 0.86 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	166	Resp:	27091
Ion	Ratio	Lower	Upper
166	100		
165	97.8	78.5	117.7
167	13.4	10.9	16.3









#129 (-)
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:	188	Resp:	19796
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.1	9.5	14.3




#200 (-)
4,6-Dinitro-2-methylphenol
Concen: 0.84 ng
RT: 15.450 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	198	Resp:	2061
Ion	Ratio	Lower	Upper
198	100		
51	45.2	59.4	89.0#
105	30.4	29.8	44.6






Abundance Scan 1461 (16.264 min): BN014190.D\data.ms (-1424) (-)




4-Bromophenyl-phenylether
Concen: 0.86 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:	248	Resp:	8718
Ion	Ratio	Lower	Upper
248	100		
250	107.2	85.6	128.4
141	30.4	26.1	39.1






Abundance Scan 1470 (16.373 min): BN014190.D\data.ms (-1426) (-)



Hexachlorobenzene
Concen: 0.85 ng
RT: 16.373 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24


Tgt Ion:	284	Resp:	9613
Ion	Ratio	Lower	Upper
284	100		
142	35.3	28.6	42.8
249	31.2	25.3	37.9





Abundance Scan 1483 (16.532 min): BN014190.D\data.ms (-1425) (-)

Atrazine
Concen: 0.85 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8



Tgt Ion:	200	Resp:	5881
Ion	Ratio	Lower	Upper
200	100		
173	20.6	17.6	26.4
215	51.0	41.4	62.0




Abundance Scan 1483 (16.532 min): BN014189.D\data.ms (-1442) (-)





Abundance Scan 1498 (16.714 min): BN014190.D\data.ms (-1421) (-)


Pentachlorophenol
Concen: 0.53 ng
RT: 16.714 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24



Tgt Ion:	266	Resp:	1859
Ion	Ratio	Lower	Upper
266	100		
264	62.2	51.8	77.8
268	60.9	47.7	71.5




Abundance Scan 1498 (16.714 min): BN014189.D\data.ms (-1490) (-)






Abundance Scan 1530 (17.104 min): BN014190.D\data.ms (-1525) (-)

Phenanthrene
Concen: 0.85 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:	178	Resp:	45055
Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.2	22.8
179	15.3	12.3	18.5




Abundance Scan 1537 (17.189 min): BN014190.D\data.ms (-1526) (-)

Anthracene
Concen: 0.86 ng
RT: 17.189 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	178	Resp:	38484
Ion	Ratio	Lower	Upper
178	100		
176	18.1	14.6	22.0
179	15.3	12.4	18.6






Abundance Scan 1824 (19.096 min): BN014190.D\data.ms (-1824) (-)




Fluoranthene-d10
Concen: 1.19 ng
RT: 19.096 min Scan# 1824
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:	212	Resp:	61865
Ion	Ratio	Lower	Upper
212	100		
106	12.4	9.9	14.9
104	6.9	5.6	8.4






Abundance Scan 1830 (19.126 min): BN014190.D\data.ms (-1828) (-)



Fluoranthene
Concen: 0.96 ng
RT: 19.126 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	202	Resp:	53462
Ion	Ratio	Lower	Upper
202	100		
101	10.7	8.5	12.7
203	17.3	13.9	20.9



Abundance Scan 2124 (21.247 min): BN014190.D\data.ms (-2129) (-)

Chrysene-d12

Concen: 0.40 ng

RT: 21.247 min Scan# 2124

Delta R.T. 0.000 min

Lab File: BN014189.D

Acq: 02 Apr 2021 13:24

Tgt Ion:240 Resp: 21319

Ion Ratio Lower Upper

240 100

120 8.8 7.3 10.9

236 26.2 20.8 31.2

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8

Abundance Scan 2124 (21.247 min): BN014189.D\data.ms

Ref 50

Raw 50

Sub 50

m/z-->

Abundance Scan 2124 (21.247 min): BN014189.D\data.ms (-2105) (-)

Ref 50

Raw 50

Sub 50

m/z-->


Abundance Scan 2124 (21.247 min): BN014189.D\data.ms (-2105) (-)

Ref 50

Raw 50

Sub 50

m/z-->



Abundance Scan 1903 (19.488 min): BN014190.D\data.ms (-1939) (-)

Pyrene

Concen: 0.77 ng

RT: 19.488 min Scan# 1903

Delta R.T. 0.000 min

Lab File: BN014189.D

Acq: 02 Apr 2021 13:24

Tgt Ion:202 Resp: 53419

Ion Ratio Lower Upper

202 100

200 20.5 16.3 24.5

203 17.5 13.9 20.9

Abundance Scan 1903 (19.488 min): BN014189.D\data.ms

Ref 50

Raw 50

Sub 50

m/z-->


Abundance Scan 1903 (19.488 min): BN014189.D\data.ms (-1863) (-)

Ref 50

Raw 50

Sub 50


m/z-->






Abundance Scan 1945 (19.697 min): BN014190.D\data.ms (-1934) (-)

Terphenyl-d14
Concen: 0.76 ng
RT: 19.697 min Scan# 1945
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:	244	Resp:	37508
Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.1	9.1
122	11.1	9.0	13.4




Abundance Scan 2123 (21.237 min): BN014190.D\data.ms (-2139) (-)

Benzo(a)anthracene
Concen: 0.91 ng
RT: 21.237 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	228	Resp:	53997
Ion	Ratio	Lower	Upper
228	100		
226	25.5	20.6	30.8
229	20.2	16.2	24.2





Abundance Scan 2127 (21.279 min): BN014190.D\data.ms (-2127) (-)




Chrysene
Concen: 0.85 ng
RT: 21.279 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	228	Resp:	55785
Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.3	36.5
229	19.1	15.4	23.0



Abundance Scan 2127 (21.279 min): BN014189.D\data.ms (-2118) (-)





Abundance Scan 2117 (21.173 min): BN014190.D\data.ms (-2117) (-)




Bis(2-ethylhexyl)phthalate
Concen: 1.18 ng
RT: 21.173 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	149	Resp:	21349
Ion	Ratio	Lower	Upper
149	100		
167	26.6	20.6	30.8
279	4.0	3.4	5.2




Abundance Scan 2117 (21.173 min): BN014189.D\data.ms (-2108) (-)





Abundance Scan 3192 (25.687 min): BN014190.D\data.ms (-3135) (-)

Indeno(1,2,3-cd)pyrene
Concen: 1.00 ng
RT: 25.687 min Scan# 3192
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24


Instrument :
BNA_N
ClientSampled :
SSTDIC0.8



Tgt Ion:	276	Resp:	67649
Ion	Ratio	Lower	Upper
276	100		
138	23.6	19.0	28.4
277	24.9	19.9	29.9




Abundance Scan 3192 (25.687 min): BN014189.D\data.ms (-3163) (-)





Abundance Scan 2543 (23.466 min): BN014190.D\data.ms (-2525) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.466 min Scan# 2543
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24




Tgt Ion:	264	Resp:	20828
Ion	Ratio	Lower	Upper
264	100		
260	24.3	20.1	30.1
265	31.2	25.0	37.4



Abundance Scan 2543 (23.466 min): BN014189.D\data.ms (-2499) (-)





Abundance Scan 2348 (22.798 min): BN014190.D\data.ms (-2348) (-)




Benzo(b)fluoranthene
Concen: 0.83 ng
RT: 22.799 min Scan# 2348
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	252	Resp:	60892
Ion	Ratio	Lower	Upper
252	100		
253	22.5	18.8	28.2
125	10.2	8.9	13.3



Abundance Scan 2348 (22.799 min): BN014189.D\data.ms (-2290) (-)





Abundance Scan 2361 (22.843 min): BN014190.D\data.ms (-2361) (-)




Benzo(k)fluoranthene
Concen: 0.80 ng
RT: 22.843 min Scan# 2361
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Tgt Ion:	252	Resp:	57313
Ion	Ratio	Lower	Upper
252	100		
253	22.8	18.9	28.3
125	10.0	8.5	12.7




Abundance Scan 2361 (22.843 min): BN014189.D\data.ms (-2273) (-)





Abundance Scan 2514 (23.367 min): BN014190.D\data.ms (-2509) (-)

Benzo(a)pyrene
Concen: 0.84 ng
RT: 23.367 min Scan# 2514
Delta R.T. 0.000 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8



Tgt Ion:	252	Resp:	53228
Ion	Ratio	Lower	Upper
252	100		
253	23.1	19.6	29.4
125	11.8	9.9	14.9






Abundance Scan 2514 (23.367 min): BN014189.D\data.ms (-2456) (-)




Abundance Scan 3196 (25.701 min): BN014190.D\data.ms (-3149) (-)

Dibenzo(a,h)anthracene
Concen: 0.88 ng
RT: 25.705 min Scan# 3197
Delta R.T. 0.004 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24


Tgt Ion:	278	Resp:	57271
Ion	Ratio	Lower	Upper
278	100		
139	15.7	13.1	19.7
279	24.6	19.9	29.9



Abundance Scan 3197 (25.705 min): BN014189.D\data.ms (-3173) (-)




Abundance Scan 3392 (26.372 min): BN014190.D\data.ms (-3341) (-)




Benzo(g,h,i)perylene
Concen: 0.86 ng
RT: 26.369 min Scan# 3391
Delta R.T. -0.003 min
Lab File: BN014189.D
Acq: 02 Apr 2021 13:24

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Abundance Scan 3391 (26.369 min): BN014189.D\data.ms



Tgt Ion:	276	Resp:	59444
Ion	Ratio	Lower	Upper
276	100		
277	24.2	19.9	29.9
138	20.3	16.4	24.6



Abundance Scan 3391 (26.369 min): BN014189.D\data.ms (-3334) (-)

