

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040221\
 Data File : BN014192.D
 Acq On : 02 Apr 2021 15:12
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Apr 02 15:42:05 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 02 14:32:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.684	152	4894	0.40	ng	0.00
7) Naphthalene-d8	10.454	136	16784	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	9735	0.40	ng	-0.01
19) Phenanthrene-d10	17.055	188	20240	0.40	ng	0.00
29) Chrysene-d12	21.247	240	22246	0.40	ng	0.00
36) Perylene-d12	23.466	264	20391	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	39742	3.16	ng	0.00
5) Phenol-d6	6.855	99	51824	3.25	ng	0.00
8) Nitrobenzene-d5	8.819	82	42171	3.70	ng	-0.01
11) 2-Methylnaphthalene-d10	12.053	152	120751	4.73	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	12879	4.12	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	117349	3.21	ng	0.00
27) Fluoranthene-d10	19.096	212	270026	5.10	ng	0.00
31) Terphenyl-d14	19.697	244	161316	3.15	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	18553	2.95	ng	98
3) n-Nitrosodimethylamine	3.545	42	17053	3.52	ng	# 91
6) bis(2-Chloroethyl)ether	7.112	93	42188	3.32	ng	99
9) Naphthalene	10.505	128	147262	3.55	ng	99
10) Hexachlorobutadiene	10.796	225	27060	3.57	ng	# 100
12) 2-Methylnaphthalene	12.124	142	99557	3.67	ng	99
16) Acenaphthylene	14.029	152	141661	3.80	ng	100
17) Acenaphthene	14.371	154	94194	3.50	ng	98
18) Fluorene	15.361	166	118949	3.65	ng	99
20) 4,6-Dinitro-2-methylph...	15.450	198	11471	4.56	ng	# 47
21) 4-Bromophenyl-phenylether	16.264	248	38075	3.66	ng	97
22) Hexachlorobenzene	16.373	284	40765	3.51	ng	99
23) Atrazine	16.532	200	27694	3.89	ng	98
24) Pentachlorophenol	16.714	266	11304	3.12	ng	97
25) Phenanthrene	17.103	178	195598	3.60	ng	99
26) Anthracene	17.189	178	174855	3.83	ng	100
28) Fluoranthene	19.126	202	231674	4.08	ng	100
30) Pyrene	19.488	202	231810	3.19	ng	100
32) Benzo(a)anthracene	21.237	228	238647	3.87	ng	99
33) Chrysene	21.279	228	241392	3.54	ng	100
34) Bis(2-ethylhexyl)phtha...	21.173	149	96840	5.15	ng	97
35) Indeno(1,2,3-cd)pyrene	25.691	276	284092	4.02	ng	100
37) Benzo(b)fluoranthene	22.798	252	254841	3.54	ng	97
38) Benzo(k)fluoranthene	22.843	252	246197	3.50	ng	97
39) Benzo(a)pyrene	23.367	252	225038	3.65	ng	95
40) Dibenzo(a,h)anthracene	25.701	278	240525	3.79	ng	98
41) Benzo(g,h,i)perylene	26.368	276	245286	3.64	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040221\
Data File : BN014192.D
Acq On : 02 Apr 2021 15:12
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Apr 02 15:42:05 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 02 14:32:17 2021
Response via : Initial Calibration






Abundance Scan 609 (7.684 min): BN014190.D\data.ms (-602) (#1)




1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.684 min Scan# 609
Delta R.T. 0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	152	Resp:	4894
Ion	Ratio	Lower	Upper
152	100		
150	156.3	126.2	189.4
115	61.6	49.8	74.6







Abundance Scan 56 (3.231 min): BN014190.D\data.ms (-53) (#2)



1,4-Dioxane
Concen: 2.95 ng
RT: 3.231 min Scan# 56
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Tgt Ion:	88	Resp:	18553
Ion	Ratio	Lower	Upper
88	100		
58	80.8	63.2	94.8
43	33.3	25.1	37.7









n-Nitrosodimethylamine
 Concen: 3.52 ng
 RT: 3.545 min Scan# 95
 Delta R.T. -0.024 min
 Lab File: BN014192.D
 Acq: 02 Apr 2021 15:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion: 42 Resp: 17053
 Ion Ratio Lower Upper
 42 100
 74 113.5 98.8 148.2
 44 7.6 0.0 0.0#



2-Fluorophenol
 Concen: 3.16 ng
 RT: 5.284 min Scan# 311
 Delta R.T. -0.000 min
 Lab File: BN014192.D
 Acq: 02 Apr 2021 15:12

Tgt Ion: 112 Resp: 39742
 Ion Ratio Lower Upper
 112 100
 64 60.2 48.3 72.5
 63 31.7 25.8 38.6









Phenol-d6
 Concen: 3.25 ng
 RT: 6.855 min Scan# 506
 Delta R.T. -0.000 min
 Lab File: BN014192.D
 Acq: 02 Apr 2021 15:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	99	Resp:	51824
Ion	Ratio	Lower	Upper
99	100		
42	22.0	17.8	26.6
71	33.6	26.6	40.0




bis(2-Chloroethyl)ether
 Concen: 3.32 ng
 RT: 7.112 min Scan# 538
 Delta R.T. -0.000 min
 Lab File: BN014192.D
 Acq: 02 Apr 2021 15:12

Tgt Ion:	93	Resp:	42188
Ion	Ratio	Lower	Upper
93	100		
63	80.8	64.2	96.4
95	33.0	26.6	40.0






Abundance Scan 856 (10.454 min): BN014190.D\data.ms (-856#7-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.454 min Scan# 856
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	136	Resp:	16784
Ion	Ratio	Lower	Upper
136	100		
137	10.8	9.1	13.7
54	10.1	8.2	12.4
68	7.7	6.1	9.1





Abundance Scan 728 (8.832 min): BN014190.D\data.ms (-724#8)




Nitrobenzene-d5
Concen: 3.70 ng
RT: 8.819 min Scan# 727
Delta R.T. -0.013 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Tgt Ion:	82	Resp:	42171
Ion	Ratio	Lower	Upper
82	100		
128	30.9	36.8	55.2#
54	62.5	38.7	58.1#



Abundance Scan 727 (8.819 min): BN014192.D\data.ms (-712) (-)






Abundance Scan 860 (10.505 min): BN014190.D\data.ms (-8539) (-)




Naphthalene
Concen: 3.55 ng
RT: 10.505 min Scan# 860
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	128	Resp:	147262
Ion	Ratio	Lower	Upper
128	100		
129	10.9	9.1	13.7
127	12.9	10.6	15.8






Abundance Scan 883 (10.796 min): BN014190.D\data.ms (-8710) (-)




Hexachlorobutadiene
Concen: 3.57 ng
RT: 10.796 min Scan# 883
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Tgt Ion:	225	Resp:	27060
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.7	50.8	76.2





Abundance Scan 1062 (12.053 min): BN014190.D\data.ms (-1017) (-)




2-Methylnaphthalene-d10
Concen: 4.73 ng
RT: 12.053 min Scan# 1062
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 120751
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1



Abundance Scan 1062 (12.053 min): BN014192.D\data.ms (-1017) (-)





Abundance Scan 1078 (12.124 min): BN014190.D\data.ms (-1017) (-)




2-Methylnaphthalene
Concen: 3.67 ng
RT: 12.124 min Scan# 1078
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Tgt Ion:142 Resp: 99557
Ion Ratio Lower Upper
142 100
141 88.9 71.6 107.4
115 35.4 29.2 43.8




Abundance Scan 1078 (12.124 min): BN014192.D\data.ms (-1056) (-)





Abundance Scan 1307 (14.321 min): BN014190.D\data.ms (-1307) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. -0.013 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:	164	Resp:	9735
Ion	Ratio	Lower	Upper
164	100		
162	104.1	77.4	116.2
160	47.8	34.2	51.4






Abundance Scan 1306 (14.308 min): BN014192.D\data.ms (-1291) (-)




Abundance Scan 1424 (15.805 min): BN014190.D\data.ms (-1414) (-)

2,4,6-Tribromophenol
Concen: 4.12 ng
RT: 15.805 min Scan# 1424
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Tgt Ion:	330	Resp:	12879
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	37.3	29.8	44.6



Abundance Scan 1424 (15.805 min): BN014192.D\data.ms (-1416) (-)





Abundance Scan 1198 (12.938 min): BN014190.D\data.ms (-1190) (-)




2-Fluorobiphenyl
Concen: 3.21 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	172	Resp:	117349
Ion	Ratio	Lower	Upper
172	100		
171	33.2	27.3	40.9
170	21.8	18.1	27.1



Abundance Scan 1198 (12.938 min): BN014192.D\data.ms (-1190) (-)





Abundance Scan 1284 (14.029 min): BN014190.D\data.ms (-1215) (-)




Acenaphthylene
Concen: 3.80 ng
RT: 14.029 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Tgt Ion:	152	Resp:	141661
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.1	22.7
153	13.1	10.5	15.7




Abundance Scan 1284 (14.029 min): BN014192.D\data.ms (-1268) (-)






Abundance Scan 1311 (14.372 min): BN014190.D\data.ms (-1303) (-)

Acenaphthene
Concen: 3.50 ng
RT: 14.371 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	154	Resp:	94194
Ion	Ratio	Lower	Upper
154	100		
153	110.0	89.2	133.8
152	52.9	43.8	65.6







Abundance Scan 1389 (15.361 min): BN014190.D\data.ms (-1312) (-)

Fluorene
Concen: 3.65 ng
RT: 15.361 min Scan# 1389
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12



Tgt Ion:	166	Resp:	118949
Ion	Ratio	Lower	Upper
166	100		
165	97.2	78.5	117.7
167	13.4	10.9	16.3









#129 (-)
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	188	Resp:	20240
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.1	9.5	14.3




#120 (-)
4,6-Dinitro-2-methylphenol
Concen: 4.56 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Tgt Ion:	198	Resp:	11471
Ion	Ratio	Lower	Upper
198	100		
51	19.4	59.4	89.0#
105	19.0	29.8	44.6#






Abundance Scan 1461 (16.264 min): BN014190.D\data.ms (-1424) (-)




4-Bromophenyl-phenylether
Concen: 3.66 ng
RT: 16.264 min Scan# 1461
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDIC3.2

Tgt Ion:	248	Resp:	38075
Ion	Ratio	Lower	Upper
248	100		
250	105.9	85.6	128.4
141	28.2	26.1	39.1






Abundance Scan 1470 (16.373 min): BN014190.D\data.ms (-1422) (-)



Hexachlorobenzene
Concen: 3.51 ng
RT: 16.373 min Scan# 1470
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Tgt Ion:	284	Resp:	40765
Ion	Ratio	Lower	Upper
284	100		
142	35.9	28.6	42.8
249	30.9	25.3	37.9






Abundance Scan 1483 (16.532 min): BN014190.D\data.ms (-1425) (-)

Atrazine
Concen: 3.89 ng
RT: 16.532 min Scan# 1483
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	200	Resp:	27694
Ion	Ratio	Lower	Upper
200	100		
173	19.8	17.6	26.4
215	50.7	41.4	62.0






Abundance Scan 1498 (16.714 min): BN014190.D\data.ms (-1421) (-)


Pentachlorophenol
Concen: 3.12 ng
RT: 16.714 min Scan# 1498
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12



Tgt Ion:	266	Resp:	11304
Ion	Ratio	Lower	Upper
266	100		
264	62.2	51.8	77.8
268	62.4	47.7	71.5





Abundance Scan 1530 (17.104 min): BN014190.D\data.ms (-1525) (-)




Phenanthrene
Concen: 3.60 ng
RT: 17.103 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 195598
Ion Ratio Lower Upper
178 100
176 18.6 15.2 22.8
179 15.3 12.3 18.5



Abundance Scan 1530 (17.103 min): BN014192.D\data.ms (-1505) (-)





Abundance Scan 1537 (17.189 min): BN014190.D\data.ms (-1526) (-)




Anthracene
Concen: 3.83 ng
RT: 17.189 min Scan# 1537
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Tgt Ion:178 Resp: 174855
Ion Ratio Lower Upper
178 100
176 18.1 14.6 22.0
179 15.3 12.4 18.6



Abundance Scan 1537 (17.189 min): BN014192.D\data.ms (-1529) (-)





Abundance Scan 1824 (19.096 min): BN014190.D\data.ms (-1824) (-)




Fluoranthene-d10
Concen: 5.10 ng
RT: 19.096 min Scan# 1824
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	212	Resp:	270026
Ion	Ratio	Lower	Upper
212	100		
106	12.5	9.9	14.9
104	6.9	5.6	8.4



Abundance Scan 1824 (19.096 min): BN014192.D\data.ms (-1784) (-)





Abundance Scan 1830 (19.126 min): BN014190.D\data.ms (-1828) (-)




Fluoranthene
Concen: 4.08 ng
RT: 19.126 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Tgt Ion:	202	Resp:	231674
Ion	Ratio	Lower	Upper
202	100		
101	10.8	8.5	12.7
203	17.4	13.9	20.9



Abundance Scan 1830 (19.126 min): BN014192.D\data.ms (-1790) (-)



Abundance Scan 2124 (21.247 min): BN014190.D\data.ms (-2129) (-)

Chrysene-d12

Concen: 0.40 ng

RT: 21.247 min Scan# 2124

Delta R.T. -0.000 min

Lab File: BN014192.D


Acq: 02 Apr 2021 15:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2






Tgt Ion:240 Resp: 22246

Ion Ratio Lower Upper

240 100

120 8.6 7.3 10.9

236 25.8 20.8 31.2



Abundance Scan 1903 (19.488 min): BN014190.D\data.ms (-1839) (-)

Pyrene


Concen: 3.19 ng

RT: 19.488 min Scan# 1903

Delta R.T. -0.000 min

Lab File: BN014192.D

Acq: 02 Apr 2021 15:12






Tgt Ion:202 Resp: 231810

Ion Ratio Lower Upper

202 100

200 20.4 16.3 24.5


203 17.5 13.9 20.9






Abundance Scan 1945 (19.697 min): BN014190.D\data.ms (-1934) (-)

Terphenyl-d14
Concen: 3.15 ng
RT: 19.697 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:	244	Resp:	161316
Ion	Ratio	Lower	Upper
244	100		
212	7.2	6.1	9.1
122	10.7	9.0	13.4




Abundance Scan 2123 (21.237 min): BN014190.D\data.ms (-2139) (-)

Benzo(a)anthracene
Concen: 3.87 ng
RT: 21.237 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Tgt Ion:	228	Resp:	238647
Ion	Ratio	Lower	Upper
228	100		
226	25.2	20.6	30.8
229	20.1	16.2	24.2






Abundance Scan 2127 (21.279 min): BN014190.D\data.ms (-2123) (-)




Chrysene
Concen: 3.54 ng
RT: 21.279 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	228	Resp:	241392
Ion	Ratio	Lower	Upper
228	100		
226	30.6	24.3	36.5
229	19.0	15.4	23.0






Abundance Scan 2117 (21.173 min): BN014190.D\data.ms (-2134) (-)



Bis(2-ethylhexyl)phthalate
Concen: 5.15 ng
RT: 21.173 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Tgt Ion:	149	Resp:	96840
Ion	Ratio	Lower	Upper
149	100		
167	27.1	20.6	30.8
279	4.0	3.4	5.2






Abundance Scan 3192 (25.687 min): BN014190.D\data.ms (-3135) (-)

Indeno(1,2,3-cd)pyrene
Concen: 4.02 ng
RT: 25.691 min Scan# 3193
Delta R.T. 0.003 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	276	Resp:	284092
Ion	Ratio	Lower	Upper
276	100		
138	23.8	19.0	28.4
277	24.9	19.9	29.9






Abundance Scan 2543 (23.466 min): BN014190.D\data.ms (-2525) (-)

Perylene-d12
Concen: 0.40 ng
RT: 23.466 min Scan# 2543
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12




Tgt Ion:	264	Resp:	20391
Ion	Ratio	Lower	Upper
264	100		
260	24.4	20.1	30.1
265	30.7	25.0	37.4






Abundance Scan 2348 (22.798 min): BN014190.D\data.ms (-230) (-)

Benzo(b)fluoranthene
Concen: 3.54 ng
RT: 22.798 min Scan# 2348
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:	252	Resp:	254841
Ion	Ratio	Lower	Upper
252	100		
253	22.1	18.8	28.2
125	9.6	8.9	13.3



Abundance Scan 2361 (22.843 min): BN014190.D\data.ms (-236) (-)

Benzo(k)fluoranthene
Concen: 3.50 ng
RT: 22.843 min Scan# 2361
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Tgt Ion:	252	Resp:	246197
Ion	Ratio	Lower	Upper
252	100		
253	21.8	18.9	28.3
125	9.5	8.5	12.7





Abundance Scan 2514 (23.367 min): BN014190.D\data.ms (-2509) (-)

Benzo(a)pyrene
Concen: 3.65 ng
RT: 23.367 min Scan# 2514
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:	252	Resp:	225038
Ion	Ratio	Lower	Upper
252	100		
253	22.0	19.6	29.4
125	10.6	9.9	14.9





Abundance Scan 2514 (23.367 min): BN014192.D\data.ms (-2456) (-)




Abundance Scan 3196 (25.701 min): BN014190.D\data.ms (-3149) (-)

Dibenzo(a,h)anthracene
Concen: 3.79 ng
RT: 25.701 min Scan# 3196
Delta R.T. -0.000 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12


Tgt Ion:	278	Resp:	240525
Ion	Ratio	Lower	Upper
278	100		
139	15.3	13.1	19.7
279	23.7	19.9	29.9



Abundance Scan 3196 (25.701 min): BN014192.D\data.ms (-3173) (-)




Abundance Scan 3392 (26.372 min): BN014190.D\data.ms (-3341) (-)




Benzo(g,h,i)perylene
Concen: 3.64 ng
RT: 26.368 min Scan# 3391
Delta R.T. -0.003 min
Lab File: BN014192.D
Acq: 02 Apr 2021 15:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Abundance Scan 3391 (26.368 min): BN014192.D\data.ms



Tgt Ion:	276	Resp:	245286
Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.9	29.9
138	20.5	16.4	24.6



Abundance Scan 3391 (26.368 min): BN014192.D\data.ms (-3334) (-)

