

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040221\
 Data File : BN014194.D
 Acq On : 02 Apr 2021 17:00
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN040221

Manual Integrations
 APPROVED

mohammad
 4/5/2021 11:56:04 AM

Quant Time: Apr 02 18:05:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 02 18:03:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.684	152	4550	0.40	ng	0.00
7) Naphthalene-d8	10.454	136	16935	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	9788	0.40	ng	-0.01
19) Phenanthrene-d10	17.055	188	20724	0.40	ng	0.00
29) Chrysene-d12	21.247	240	21690	0.40	ng	0.00
36) Perylene-d12	23.463	264	21359	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	4931	0.40	ng	0.00
5) Phenol-d6	6.855	99	6254	0.40	ng	0.00
8) Nitrobenzene-d5	8.819	82	4887	0.38	ng	-0.01
11) 2-Methylnaphthalene-d10	12.053	152	14866	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1236	0.34	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	14721	0.40	ng	0.00
27) Fluoranthene-d10	19.096	212	32264	0.39	ng	0.00
31) Terphenyl-d14	19.697	244	20706	0.42	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.223	88	2210	0.39	ng	99
3) n-Nitrosodimethylamine	3.561	42	1463	0.34	ng	# 97
6) bis(2-Chloroethyl)ether	7.112	93	5089	0.41	ng	99
9) Naphthalene	10.505	128	18297	0.40	ng	100
10) Hexachlorobutadiene	10.797	225	3407	0.40	ng	# 99
12) 2-Methylnaphthalene	12.124	142	12372	0.41	ng	98
16) Acenaphthylene	14.029	152	15434	0.38	ng	100
17) Acenaphthene	14.372	154	11287	0.40	ng	100
18) Fluorene	15.361	166	14189	0.39	ng	100
20) 4,6-Dinitro-2-methylph...	15.450	198	942	0.33	ng	92
21) 4-Bromophenyl-phenylether	16.252	248	4595	0.39	ng	# 64
22) Hexachlorobenzene	16.373	284	5139	0.40	ng	100
23) Atrazine	16.532	200	2984	0.36	ng	100
24) Pentachlorophenol	16.714	266	781	0.28	ng	96
25) Phenanthrene	17.104	178	23916	0.39	ng	100
26) Anthracene	17.189	178	19916	0.38	ng	100
28) Fluoranthene	19.126	202	27695	0.38	ng	100
30) Pyrene	19.488	202	28133	0.40	ng	100
32) Benzo(a)anthracene	21.226	228	27384	0.39	ng	97
33) Chrysene	21.279	228	29190	0.40	ng	99
34) Bis(2-ethylhexyl)phtha...	21.162	149	10275	0.33	ng	98
35) Indeno(1,2,3-cd)pyrene	25.684	276	34666	0.40	ng	99
37) Benzo(b)fluoranthene	22.799	252	31611m	0.40	ng	
38) Benzo(k)fluoranthene	22.843	252	31184	0.41	ng	99
39) Benzo(a)pyrene	23.367	252	28702	0.41	ng	# 94
40) Dibenzo(a,h)anthracene	25.701	278	29173	0.39	ng	99
41) Benzo(g,h,i)perylene	26.365	276	29920	0.39	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040221\
Data File : BN014194.D
Acq On : 02 Apr 2021 17:00
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM

Quant Time: Apr 02 18:05:06 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 02 18:03:30 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.684 min Scan# 609
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00





Tgt Ion: 152 Resp: 4550
Ion Ratio Lower Upper
152 100
150 154.7 126.2 189.4
115 61.3 49.8 74.6



#2
1,4-Dioxane
Concen: 0.39 ng
RT: 3.223 min Scan# 55
Delta R.T. -0.008 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00




Tgt Ion: 88 Resp: 2210
Ion Ratio Lower Upper
88 100
58 80.1 63.2 94.8
43 31.1 25.1 37.7



Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED

mohammad
4/5/2021 11:56:04 AM







#3
n-Nitrosodimethylamine
Concen: 0.34 ng
RT: 3.561 min Scan# 97
Delta R.T. -0.008 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion: 42 Resp: 1463
Ion Ratio Lower Upper
42 100
74 120.3 98.8 148.2
44 7.0 0.0 0.0#

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.284 min Scan# 311
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion: 112 Resp: 4931
Ion Ratio Lower Upper
112 100
64 59.6 48.3 72.5
63 32.2 25.8 38.6









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.855 min Scan# 506
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion: 99 Resp: 6254
Ion Ratio Lower Upper
99 100
42 22.6 17.8 26.6
71 33.8 26.6 40.0

Instrument :
BNA_N
ClientSampled :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.41 ng
RT: 7.112 min Scan# 538
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00


Tgt Ion: 93 Resp: 5089
Ion Ratio Lower Upper
93 100
63 81.5 64.2 96.4
95 33.7 26.6 40.0





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.454 min Scan# 856
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00



Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Tgt Ion:	136	Resp:	16935
Ion Ratio	Lower	Upper	
136	100		
137	11.6	9.1	13.7
54	10.0	8.2	12.4
68	7.6	6.1	9.1

Manual Integrations
APPROVED



mohammad
4/5/2021 11:56:04 AM




#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.819 min Scan# 727
Delta R.T. -0.013 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00



Tgt Ion:	82	Resp:	4887
Ion Ratio	Lower	Upper	
82	100		
128	31.3	36.8	55.2#
54	62.3	38.7	58.1#









#9
Naphthalene
Concen: 0.40 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:	128	Resp:	18297
Ion Ratio	Lower	Upper	
128	100		
129	11.5	9.1	13.7
127	13.5	10.6	15.8

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM



#10
Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.797 min Scan# 883
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00





Tgt Ion:	225	Resp:	3407
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.5	50.8	76.2








#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.053 min Scan# 1062
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:152 Resp: 14866
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.124 min Scan# 1078
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00


Tgt Ion:142 Resp: 12372
Ion Ratio Lower Upper
142 100
141 87.9 71.6 107.4
115 35.7 29.2 43.8




Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED




mohammad
4/5/2021 11:56:04 AM




#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. -0.013 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00





Tgt Ion:164 Resp: 9788
Ion Ratio Lower Upper
164 100
162 104.0 77.4 116.2
160 49.1 34.2 51.4



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.805 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00




Tgt Ion:330 Resp: 1236
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.6 29.8 44.6



Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED

mohammad
4/5/2021 11:56:04 AM







#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.938 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:	172	Resp:	14721
Ion Ratio	Lower	Upper	
172	100		
171	34.2	27.3	40.9
170	22.4	18.1	27.1

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.029 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00


Tgt Ion:	152	Resp:	15434
Ion Ratio	Lower	Upper	
152	100		
151	19.1	15.1	22.7
153	12.9	10.5	15.7





#17
Acenaphthene
Concen: 0.40 ng
RT: 14.372 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00




Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Tgt Ion	Ratio	Lower	Upper
154	100		
153	111.5	89.2	133.8
152	54.3	43.8	65.6

Manual Integrations
APPROVED



mohammad
4/5/2021 11:56:04 AM




#18
Fluorene
Concen: 0.39 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00



Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.4	78.5	117.7
167	13.5	10.9	16.3









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	11.5	9.5	14.3




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.33 ng
RT: 15.450 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion	Ratio	Lower	Upper
198	100		
51	79.9	59.4	89.0
105	42.8	29.8	44.6









#21
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.012 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00





Tgt Ion:248 Resp: 4595
Ion Ratio Lower Upper
248 100
250 89.4 85.6 128.4
141 88.2 26.1 39.1#



#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.373 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00




Tgt Ion:284 Resp: 5139
Ion Ratio Lower Upper
284 100
142 35.6 28.6 42.8
249 32.0 25.3 37.9



Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED

mohammad
4/5/2021 11:56:04 AM







#23
Atrazine
Concen: 0.36 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:	200	Resp:	2984
Ion Ratio	Lower	Upper	
200	100		
173	21.4	17.6	26.4
215	51.8	41.4	62.0

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM




#24
Pentachlorophenol
Concen: 0.28 ng
RT: 16.714 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:	266	Resp:	781
Ion Ratio	Lower	Upper	
266	100		
264	64.7	51.8	77.8
268	65.2	47.7	71.5









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00





Tgt Ion:178 Resp: 23916
Ion Ratio Lower Upper
178 100
176 18.8 15.2 22.8
179 15.2 12.3 18.5



#26
Anthracene
Concen: 0.38 ng
RT: 17.189 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00




Tgt Ion:178 Resp: 19916
Ion Ratio Lower Upper
178 100
176 18.3 14.6 22.0
179 15.4 12.4 18.6




Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED




mohammad
4/5/2021 11:56:04 AM




#27
Fluoranthene-d10
Concen: 0.39 ng
RT: 19.096 min Scan# 1824
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00





Tgt Ion:212 Resp: 32264
Ion Ratio Lower Upper
212 100
106 12.4 9.9 14.9
104 7.0 5.6 8.4



#28
Fluoranthene
Concen: 0.38 ng
RT: 19.126 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00




Tgt Ion:202 Resp: 27695
Ion Ratio Lower Upper
202 100
101 10.8 8.5 12.7
203 17.2 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED

mohammad
4/5/2021 11:56:04 AM







#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.247 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion	Ratio	Lower	Upper
240	100		
120	8.6	7.3	10.9
236	26.5	20.8	31.2

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM



#30
Pyrene
Concen: 0.40 ng
RT: 19.488 min Scan# 1903
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.2	16.3	24.5
203	17.4	13.9	20.9









#31
Terphenyl-d14
Concen: 0.42 ng
RT: 19.697 min Scan# 1945
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:244 Resp: 20706
Ion Ratio Lower Upper
244 100
212 7.6 6.1 9.1
122 10.9 9.0 13.4

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM



#32
Benzo(a)anthracene
Concen: 0.39 ng
RT: 21.226 min Scan# 2122
Delta R.T. -0.011 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:228 Resp: 27384
Ion Ratio Lower Upper
228 100
226 27.5 20.6 30.8
229 19.4 16.2 24.2









#33
Chrysene
Concen: 0.40 ng
RT: 21.279 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:	228	Resp:	29190
Ion Ratio	Lower	Upper	
228	100		
226	29.9	24.3	36.5
229	19.3	15.4	23.0

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.33 ng
RT: 21.162 min Scan# 2116
Delta R.T. -0.011 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:	149	Resp:	10275
Ion Ratio	Lower	Upper	
149	100		
167	26.6	20.6	30.8
279	4.2	3.4	5.2









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.40 ng
RT: 25.684 min Scan# 3191
Delta R.T. -0.003 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00





Tgt Ion: 276 Resp: 34666
Ion Ratio Lower Upper
276 100
138 22.9 19.0 28.4
277 24.7 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.463 min Scan# 2542
Delta R.T. -0.003 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00




Tgt Ion: 264 Resp: 21359
Ion Ratio Lower Upper
264 100
260 24.7 20.1 30.1
265 30.3 25.0 37.4



Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED

mohammad
4/5/2021 11:56:04 AM







#37
Benzo(b)fluoranthene
Concen: 0.40 ng m
RT: 22.799 min Scan# 2348
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:252 Resp: 31611
Ion Ratio Lower Upper
252 100
253 24.2 18.8 28.2
125 14.1 8.9 13.3#

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM



#38
Benzo(k)fluoranthene
Concen: 0.41 ng
RT: 22.843 min Scan# 2361
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion:252 Resp: 31184
Ion Ratio Lower Upper
252 100
253 24.0 18.9 28.3
125 10.6 8.5 12.7









#39
Benzo(a)pyrene
Concen: 0.41 ng
RT: 23.367 min Scan# 2514
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.2	19.6	29.4
125	18.0	9.9	14.9

Instrument :
BNA_N
ClientSampleId :
ICVBN040221




Manual Integrations
APPROVED


mohammad
4/5/2021 11:56:04 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.39 ng
RT: 25.701 min Scan# 3196
Delta R.T. 0.000 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00

Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.5	13.1	19.7
279	25.3	19.9	29.9







#41
Benzo(g,h,i)perylene
Concen: 0.39 ng
RT: 26.365 min Scan# 3390
Delta R.T. -0.007 min
Lab File: BN014194.D
Acq: 02 Apr 2021 17:00



Tgt Ion:	276	Resp:	29920
Ion Ratio	Lower	Upper	
276	100		
277	24.1	19.9	29.9
138	20.8	16.4	24.6



Instrument :
BNA_N
ClientSampleId :
ICVBN040221

Manual Integrations
APPROVED

mohammad
4/5/2021 11:56:04 AM