


Data Path : Z:\HPCHEM1\BNA_N\Data\BN040318\
Data File : BN000699.D
Acq On : 03 Apr 2018 13:43
Operator : JU/SJ
Sample : DFTPP91
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP91

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
Title : SVOA CALIBRATION
Last Update : Wed Apr 04 06:48:50 2018



AutoFind: Scans 2498, 2499, 2500; Background Corrected with Scan 2493

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.0	68077	PASS
68	69	0.00	2	1.8	1441	PASS
69	198	0.00	100	43.5	82290	PASS
70	69	0.00	2	0.7	559	PASS
127	198	10	80	51.7	97842	PASS
197	198	0.00	2	0.5	911	PASS
198	198	100	100	100.0	189077	PASS
199	198	5	9	6.8	12928	PASS
275	198	10	60	24.3	46005	PASS
365	198	1	100	3.2	5978	PASS
441	443	0.01	100	79.2	21770	PASS
442	198	50	100	72.5	137152	PASS
443	442	15	24	20.1	27502	PASS

m/z	Abundance
38.00	712.0
39.00	4223.0
40.00	584.0
43.90	331.0
49.00	588.0
50.00	14599.0
51.10	55664.0
52.00	3055.0
55.00	666.0
56.00	1703.0
57.00	4137.0
61.00	730.0
62.00	829.0
63.10	2474.0
64.00	436.0
65.00	1415.0
68.10	1050.0
69.00	67856.0
70.00	410.0
73.00	833.0
74.00	6940.0
75.00	11053.0
76.10	3865.0
77.10	78512.0
78.00	5940.0
79.00	5230.0
80.00	3908.0
81.00	5669.0
82.00	1385.0
83.00	1387.0
84.00	315.0
85.00	1224.0
86.00	1739.0
87.00	631.0
87.90	334.0
91.00	1384.0
91.90	1285.0
93.00	8760.0
94.00	761.0
95.10	316.0
96.00	478.0
98.00	6617.0
99.00	6280.0
99.90	498.0
101.00	3095.0
103.00	954.0
104.00	1962.0
105.00	1856.0
106.00	620.0
107.00	23000.0
108.00	3741.0
109.10	754.0
110.00	42688.0
111.00	7050.0
112.00	930.0
116.00	1287.0
117.00	19352.0
118.00	1551.0
122.00	1417.0
123.00	2118.0
124.00	1083.0
125.00	991.0
126.20	418.0
127.00	75608.0
128.00	5495.0
129.00	27888.0
130.00	2433.0
130.90	481.0
132.00	308.0
134.00	773.0
135.00	2329.0
136.00	1040.0
137.00	1395.0
137.90	367.0
140.00	345.0
141.00	3220.0
142.00	1427.0
143.00	680.0
146.00	699.0
147.00	1902.0
148.00	4757.0
149.00	1178.0
150.00	344.0
151.00	369.0
153.00	1061.0
154.00	894.0
155.00	1901.0
156.00	3040.0
157.10	564.0
158.00	795.0
159.00	621.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

160.00	1108.0
161.00	1795.0
161.90	439.0
165.00	1472.0
166.00	1123.0
167.00	7098.0
168.00	5387.0
169.00	737.0
171.00	304.0
171.90	646.0
173.00	823.0
174.00	1552.0
175.00	2685.0
176.00	768.0
177.00	1153.0
178.00	381.0
179.00	5036.0
180.00	3661.0
181.00	1570.0
183.90	405.0
185.00	2511.0
186.00	19200.0
187.00	5713.0
188.00	693.0
188.90	1165.0
191.00	670.0
192.00	1756.0
193.00	1790.0
194.10	370.0
196.00	3596.0
197.10	913.0
198.00	133504.0
199.00	9987.0
200.10	640.0
201.50	574.0
203.10	979.0
204.00	4591.0
205.00	8068.0
206.10	33608.0
207.10	5086.0
208.00	1310.0
209.00	410.0
210.10	605.0
211.00	1265.0
215.10	332.0
216.00	826.0
217.00	8684.0
218.00	1012.0
221.10	4613.0
221.80	1351.0
223.00	2005.0
224.00	18544.0
225.00	4769.0
226.00	701.0
227.00	7854.0
228.00	1094.0
229.00	1800.0
231.00	604.0
234.00	566.0
235.00	619.0
236.00	333.0
237.00	899.0
239.00	349.0
241.00	369.0
242.00	957.0
243.10	909.0
244.10	13485.0
245.10	1960.0
246.00	2516.0
247.00	530.0
249.00	504.0
253.00	318.0
255.00	66976.0
256.00	9849.0
257.10	858.0
258.00	4036.0
259.00	708.0
265.00	1711.0
273.00	2125.0
274.00	5094.0
275.10	29024.0
276.00	4046.0
277.00	2357.0
278.00	555.0
282.90	311.0
285.00	386.0
293.10	502.0
296.00	7929.0
297.00	1237.0
303.00	918.0
314.00	431.0
314.90	843.0
316.00	407.0
323.10	2495.0
324.10	464.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

327.00	485.0
334.10	1562.0
335.00	523.0
346.00	539.0
352.00	655.0
353.10	503.0
354.10	799.0
365.00	3232.0
366.00	557.0
372.10	1104.0
373.00	309.0
383.10	377.0
402.00	433.0
403.10	662.0
421.00	653.0
422.00	636.0
423.00	4143.0
424.10	1398.0
441.10	10902.0
442.10	66944.0
443.10	13932.0
444.00	1324.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

m/z	Abundance
38.00	1038.0
39.10	5679.0
40.00	530.0
41.10	408.0
49.10	750.0
50.00	20760.0
51.00	80208.0
52.00	4050.0
55.00	714.0
56.00	2492.0
57.00	5516.0
61.00	1154.0
62.00	1266.0
63.00	3704.0
64.00	465.0
65.00	1730.0
68.00	1853.0
69.00	95840.0
70.00	691.0
73.00	971.0
74.00	10138.0
75.00	15422.0
76.00	5705.0
77.10	115104.0
78.00	8175.0
79.00	7357.0
80.00	5849.0
81.00	8040.0
82.00	2023.0
83.00	1809.0
85.00	2093.0
86.00	2327.0
87.00	1028.0
88.00	481.0
91.00	1813.0
92.00	2230.0
93.00	12915.0
94.00	875.0
95.00	413.0
96.00	646.0
97.00	459.0
98.00	9629.0
99.00	8599.0
100.00	768.0
101.00	4364.0
103.00	1311.0
104.00	2909.0
105.00	3008.0
106.00	908.0
107.00	35096.0
108.00	5289.0
109.10	1170.0
110.00	63720.0
111.00	9626.0
112.00	1298.0
113.00	353.0
116.00	1828.0
117.00	27888.0
118.00	2152.0
118.90	342.0
120.00	393.0
122.00	2307.0
123.00	3317.0
124.00	1553.0
125.00	1578.0
126.20	624.0
127.00	114328.0
128.00	9164.0
129.00	43496.0
130.00	3972.0
131.00	742.0
132.00	372.0
134.00	1185.0
135.00	3609.0
136.00	1539.0
137.00	2260.0
138.00	487.0
139.90	424.0
141.00	5446.0
142.00	1844.0
142.90	1354.0
144.00	416.0
144.90	312.0
145.90	972.0
147.00	2893.0
148.00	6942.0
149.00	1506.0
150.00	487.0
151.00	743.0
151.80	362.0
152.20	477.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

153.00	1787.0
154.00	1420.0
155.00	2941.0
156.00	5041.0
157.00	945.0
158.00	1024.0
159.00	818.0
160.00	1859.0
161.00	2718.0
162.00	655.0
163.90	350.0
165.00	1975.0
166.00	1612.0
167.00	11759.0
168.00	7403.0
169.00	1203.0
170.00	372.0
170.90	615.0
172.00	839.0
173.00	1304.0
174.00	2390.0
175.00	4515.0
176.00	1260.0
177.00	2000.0
178.00	768.0
179.00	8123.0
180.00	5900.0
181.00	2641.0
182.00	462.0
184.00	691.0
185.00	4049.0
186.00	30672.0
187.00	8818.0
188.00	941.0
189.00	1913.0
190.00	579.0
191.10	968.0
192.00	2694.0
193.10	3006.0
194.00	567.0
195.10	426.0
196.00	6561.0
197.10	1821.0
198.00	219264.0
199.00	14828.0
200.00	1183.0
201.50	923.0
203.00	1409.0
204.00	7702.0
205.00	13429.0
206.10	57184.0
207.10	8065.0
208.00	1810.0
209.00	755.0
210.20	977.0
211.00	2219.0
215.00	598.0
216.00	1288.0
217.00	14566.0
218.00	1926.0
221.10	8631.0
221.80	1973.0
223.00	3347.0
224.10	30744.0
225.10	7763.0
226.00	907.0
227.00	13459.0
228.00	1951.0
229.00	2770.0
230.00	388.0
231.10	1132.0
234.00	840.0
235.00	942.0
236.00	618.0
237.00	1221.0
238.90	549.0
240.00	407.0
241.00	731.0
242.10	1639.0
243.10	1920.0
244.10	24096.0
245.10	3319.0
246.00	5149.0
247.00	1137.0
249.00	860.0
252.00	304.0
253.00	460.0
254.20	1125.0
255.00	118664.0
256.00	18096.0
257.00	1478.0
258.00	6788.0
259.00	1197.0
265.00	2984.0
266.00	371.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

271.00	388.0
273.00	3708.0
274.10	9483.0
275.10	53352.0
276.10	7086.0
277.00	4612.0
278.00	862.0
283.00	545.0
284.10	436.0
285.10	702.0
293.00	1049.0
294.00	312.0
295.10	360.0
296.00	15001.0
297.00	2026.0
302.00	308.0
303.10	1696.0
304.00	525.0
308.00	313.0
314.10	705.0
315.00	1491.0
316.00	847.0
321.10	481.0
323.10	4584.0
324.00	773.0
327.00	791.0
328.00	589.0
332.00	333.0
333.00	438.0
334.10	2950.0
335.00	783.0
341.00	534.0
346.00	1213.0
352.00	1380.0
353.10	963.0
354.00	1508.0
355.00	317.0
365.00	6768.0
366.00	1016.0
370.90	420.0
372.00	2344.0
373.10	706.0
383.00	737.0
390.00	340.0
402.00	1007.0
403.10	1547.0
404.00	512.0
421.00	1202.0
422.10	1473.0
423.10	8603.0
424.10	2712.0
425.10	324.0
441.10	23984.0
442.10	154816.0
443.10	30848.0
444.00	2840.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

m/z	Abundance
36.90	338.0
38.00	715.0
39.10	5035.0
39.90	431.0
41.00	429.0
44.00	406.0
49.00	618.0
50.00	17856.0
51.00	68360.0
52.00	3504.0
55.00	644.0
56.00	2371.0
57.00	4991.0
61.00	1056.0
62.00	1180.0
63.10	3259.0
63.90	472.0
65.00	1745.0
68.10	1422.0
69.00	83176.0
70.00	578.0
73.00	817.0
74.00	8721.0
75.00	13770.0
76.10	4980.0
77.10	101520.0
78.10	6688.0
79.00	6402.0
80.00	4896.0
81.00	7372.0
82.00	1686.0
83.00	1750.0
84.00	374.0
85.00	1966.0
86.00	2105.0
87.00	1109.0
88.00	338.0
91.00	1685.0
92.00	2028.0
93.00	11549.0
93.90	678.0
95.10	336.0
96.00	565.0
97.10	357.0
98.00	9094.0
99.00	7392.0
99.90	646.0
101.00	4058.0
103.00	1422.0
104.00	2457.0
105.00	2388.0
106.00	887.0
107.00	30768.0
108.00	4857.0
109.10	969.0
110.00	56520.0
111.00	8417.0
112.00	1139.0
112.90	375.0
116.00	1638.0
117.00	25864.0
118.00	1993.0
120.00	457.0
122.00	2071.0
123.00	3089.0
124.00	1334.0
125.00	1303.0
126.20	756.0
127.00	103592.0
128.00	8167.0
129.00	39352.0
130.00	3348.0
131.00	788.0
131.90	312.0
134.00	1221.0
135.00	3161.0
136.00	1455.0
137.00	2031.0
138.00	439.0
139.90	480.0
141.00	5329.0
142.00	1818.0
143.00	1206.0
144.00	336.0
144.90	418.0
146.00	1040.0
147.00	2837.0
148.00	6281.0
149.00	1597.0
150.00	414.0
151.00	741.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

151.80	378.0
153.00	1808.0
154.00	1412.0
155.00	3357.0
156.00	4478.0
157.10	918.0
158.00	927.0
158.90	797.0
160.00	1664.0
161.00	2873.0
162.00	851.0
163.00	309.0
163.90	357.0
165.00	2214.0
166.00	1709.0
167.00	11053.0
168.00	6264.0
169.00	1073.0
170.00	403.0
170.90	437.0
172.00	911.0
173.00	1217.0
174.00	2303.0
175.00	4419.0
175.90	1120.0
177.00	1924.0
177.90	623.0
179.00	7657.0
180.00	5597.0
181.00	2359.0
182.00	500.0
184.00	636.0
185.10	3776.0
186.00	29440.0
187.00	8192.0
188.00	875.0
189.00	2021.0
190.00	340.0
191.00	981.0
192.00	2457.0
193.10	2922.0
194.10	583.0
194.90	351.0
196.00	6243.0
198.00	214464.0
199.00	13970.0
200.00	1133.0
201.50	1005.0
203.00	1448.0
204.00	8195.0
205.10	13226.0
206.10	57016.0
207.10	7416.0
208.00	2005.0
209.00	693.0
210.10	955.0
211.00	1953.0
215.00	650.0
216.00	1270.0
217.00	14885.0
218.00	1878.0
221.00	8008.0
221.90	2041.0
223.00	3382.0
224.10	30264.0
225.10	7915.0
226.10	904.0
227.00	13325.0
228.00	2057.0
229.00	2751.0
230.10	364.0
231.10	1007.0
234.00	737.0
235.00	918.0
236.00	691.0
237.00	1142.0
239.00	507.0
240.00	438.0
241.10	789.0
242.00	1692.0
243.10	1876.0
244.10	24848.0
245.10	3181.0
246.00	4655.0
247.00	1024.0
249.00	890.0
253.00	603.0
255.00	122632.0
256.00	18160.0
257.00	1463.0
258.00	7192.0
259.00	1189.0
265.00	3036.0
266.00	415.0
272.00	339.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

273.00	3490.0
274.00	9808.0
275.10	55640.0
276.10	7561.0
277.00	4732.0
278.00	674.0
283.00	554.0
284.00	420.0
285.10	965.0
293.00	1164.0
295.00	320.0
296.10	15599.0
297.00	2081.0
303.00	1712.0
304.00	463.0
314.00	748.0
315.00	1750.0
316.10	964.0
321.00	565.0
323.10	4880.0
324.10	923.0
327.00	955.0
327.90	509.0
332.00	400.0
333.00	502.0
334.00	3514.0
335.00	875.0
341.00	571.0
346.10	1199.0
352.10	1691.0
353.00	1192.0
354.00	1720.0
355.00	396.0
365.00	7935.0
366.00	1036.0
371.00	547.0
372.10	2924.0
373.00	698.0
383.00	835.0
390.00	472.0
402.00	1169.0
403.10	1634.0
404.00	720.0
421.00	1440.0
422.10	1514.0
423.10	10537.0
424.10	2674.0
425.00	353.0
441.10	30424.0
442.10	189696.0
443.10	37728.0
444.10	3596.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91

Scan 2493 (18.101 min): BN000699.D
DFTPP91

m/z	Abundance
40.00	362.0
41.10	324.0
44.00	317.0
207.00	513.0

Instrument :
BNA_N
ClientSampleId :
DFTPP91