

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040323\
 Data File : BN024730.D
 Acq On : 03 Apr 2023 15:29
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN040323

Quant Time: Apr 04 03:14:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 04 03:11:39 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.413	0.447	-8.2	91	0.00
3	n-Nitrosodimethylamine	0.693	0.749	-8.1	90	0.00
4 S	2-Fluorophenol	0.910	0.960	-5.5	90	0.00
5 S	Phenol-d6	1.145	1.242	-8.5	94	0.00
6	bis(2-Chloroethyl)ether	1.125	1.175	-4.4	87	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.01
8 S	Nitrobenzene-d5	0.312	0.330	-5.8	92	0.00
9	Naphthalene	1.034	1.068	-3.3	86	0.00
10	Hexachlorobutadiene	0.206	0.214	-3.9	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.513	0.520	-1.4	85	0.00
12	2-Methylnaphthalene	0.695	0.707	-1.7	86	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.172	0.163	5.2	87	0.00
15 S	2-Fluorobiphenyl	1.561	1.687	-8.1	87	0.00
16	Acenaphthylene	1.666	1.640	1.6	86	0.00
17	Acenaphthene	1.157	1.218	-5.3	87	0.00
18	Fluorene	1.584	1.606	-1.4	83	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.034	0.024	29.4#	83	0.01
21	4-Bromophenyl-phenylether	0.240	0.239	0.4	85	0.00
22	Hexachlorobenzene	0.282	0.295	-4.6	86	0.00
23	Atrazine	0.146	0.138	5.5	87	0.00
24	Pentachlorophenol	0.100	0.079	21.0	81	0.00
25	Phenanthrene	1.155	1.170	-1.3	85	0.00
26	Anthracene	1.001	0.955	4.6	85	0.00
27 SURR	Fluoranthene-d10	0.847	0.833	1.7	84	0.00
28	Fluoranthene	1.229	1.202	2.2	85	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
30	Pyrene	1.683	1.845	-9.6	85	0.00
31 S	Terphenyl-d14	1.156	1.244	-7.6	83	0.00
32	Benzo(a)anthracene	1.282	1.358	-5.9	86	0.00
33	Chrysene	1.415	1.502	-6.1	83	0.00
34	Bis(2-ethylhexyl)phthalate	0.574	0.596	-3.8	86	0.00
35 I	Perylene-d12	1.000	1.000	0.0	96	0.01
36	Indeno(1,2,3-cd)pyrene	1.518	1.605	-5.7	91	0.00
37	Benzo(b)fluoranthene	1.625	1.694	-4.2	89	0.00
38	Benzo(k)fluoranthene	1.622	1.606	1.0	86	0.00
39 C	Benzo(a)pyrene	1.380	1.433	-3.8	94	0.00
40	Dibenzo(a,h)anthracene	1.196	1.251	-4.6	89	0.00
41	Benzo(g,h,i)perylene	1.400	1.410	-0.7	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN040323