

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040323\
 Data File : BN024730.D
 Acq On : 03 Apr 2023 15:29
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBN040323

Quant Time: Apr 04 03:14:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 04 03:11:39 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.433	-8.2	91	0.00
3	n-Nitrosodimethylamine	0.400	0.433	-8.2	90	0.00
4 S	2-Fluorophenol	0.400	0.422	-5.5	90	0.00
5 S	Phenol-d6	0.400	0.434	-8.5	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.418	-4.5	87	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.01
8 S	Nitrobenzene-d5	0.400	0.423	-5.7	92	0.00
9	Naphthalene	0.400	0.413	-3.2	86	0.00
10	Hexachlorobutadiene	0.400	0.417	-4.2	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.406	-1.5	85	0.00
12	2-Methylnaphthalene	0.400	0.407	-1.7	86	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.400	0.378	5.5	87	0.00
15 S	2-Fluorobiphenyl	0.400	0.432	-8.0	87	0.00
16	Acenaphthylene	0.400	0.394	1.5	86	0.00
17	Acenaphthene	0.400	0.421	-5.2	87	0.00
18	Fluorene	0.400	0.405	-1.3	83	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.422	-5.5	83	0.01
21	4-Bromophenyl-phenylether	0.400	0.400	0.0	85	0.00
22	Hexachlorobenzene	0.400	0.419	-4.7	86	0.00
23	Atrazine	0.400	0.378	5.5	87	0.00
24	Pentachlorophenol	0.400	0.318	20.5	81	0.00
25	Phenanthrene	0.400	0.405	-1.3	85	0.00
26	Anthracene	0.400	0.382	4.5	85	0.00
27 SURR	Fluoranthene-d10	0.400	0.393	1.8	84	0.00
28	Fluoranthene	0.400	0.391	2.3	85	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	94	0.00
30	Pyrene	0.400	0.439	-9.7	85	0.00
31 S	Terphenyl-d14	0.400	0.430	-7.5	83	0.00
32	Benzo(a)anthracene	0.400	0.424	-6.0	86	0.00
33	Chrysene	0.400	0.424	-6.0	83	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.415	-3.7	86	0.00
35 I	Perylene-d12	0.400	0.400	0.0	96	0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.423	-5.7	91	0.00
37	Benzo(b)fluoranthene	0.400	0.417	-4.2	89	0.00
38	Benzo(k)fluoranthene	0.400	0.396	1.0	86	0.00
39 C	Benzo(a)pyrene	0.400	0.415	-3.7	94	0.00
40	Dibenzo(a,h)anthracene	0.400	0.418	-4.5	89	0.00
41	Benzo(g,h,i)perylene	0.400	0.403	-0.8	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN040323