

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040323\
 Data File : BN024742.D
 Acq On : 03 Apr 2023 23:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 04 03:58:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 04 03:11:39 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	117	0.00
2	1,4-Dioxane	0.400	0.498	-24.5	127	0.00
3	n-Nitrosodimethylamine	0.400	0.504	-26.0#	128	0.00
4 S	2-Fluorophenol	0.400	0.583	-45.7#	151	0.00
5 S	Phenol-d6	0.400	0.718	-79.5#	189	0.00
6	bis(2-Chloroethyl)ether	0.400	0.429	-7.2	108	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	119	0.00
8 S	Nitrobenzene-d5	0.400	0.439	-9.7	117	0.00
9	Naphthalene	0.400	0.418	-4.5	107	0.00
10	Hexachlorobutadiene	0.400	0.409	-2.2	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	106	0.00
12	2-Methylnaphthalene	0.400	0.418	-4.5	108	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.432	-8.0	125	0.00
15 S	2-Fluorobiphenyl	0.400	0.443	-10.7	112	0.00
16	Acenaphthylene	0.400	0.407	-1.7	112	-0.01
17	Acenaphthene	0.400	0.428	-7.0	111	-0.01
18	Fluorene	0.400	0.419	-4.7	108	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	120	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.468	-17.0	131	0.00
21	4-Bromophenyl-phenylether	0.400	0.403	-0.8	107	0.00
22	Hexachlorobenzene	0.400	0.411	-2.7	106	0.00
23	Atrazine	0.400	0.397	0.8	114	0.00
24	Pentachlorophenol	0.400	0.374	6.5	120	0.00
25	Phenanthrene	0.400	0.412	-3.0	108	0.00
26	Anthracene	0.400	0.398	0.5	110	0.00
27 SURR	Fluoranthene-d10	0.400	0.400	0.0	107	0.00
28	Fluoranthene	0.400	0.406	-1.5	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	131	0.00
30	Pyrene	0.400	0.414	-3.5	111	0.00
31 S	Terphenyl-d14	0.400	0.433	-8.2	116	0.00
32	Benzo(a)anthracene	0.400	0.439	-9.7	125	0.00
33	Chrysene	0.400	0.431	-7.7	118	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.412	-3.0	118	0.00
35 I	Perylene-d12	0.400	0.400	0.0	138	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.407	-1.7	125	-0.03
37	Benzo(b)fluoranthene	0.400	0.408	-2.0	125	0.00
38	Benzo(k)fluoranthene	0.400	0.388	3.0	121	0.00
39 C	Benzo(a)pyrene	0.400	0.434	-8.5	140	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.399	0.3	122	-0.01
41	Benzo(g,h,i)perylene	0.400	0.377	5.8	114	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4