

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040323\
 Data File : BN024726.D
 Acq On : 03 Apr 2023 12:23
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Apr 04 03:06:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 03 15:20:44 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.997	152	10335	0.400	ng	0.00
7) Naphthalene-d8	10.802	136	29762	0.400	ng	0.00
13) Acenaphthene-d10	14.633	164	16028	0.400	ng	0.00
19) Phenanthrene-d10	17.376	188	32213	0.400	ng	0.00
29) Chrysene-d12	21.565	240	23046	0.400	ng	0.00
35) Perylene-d12	24.035	264	15761	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.541	112	17368	0.739	ng	0.00
5) Phenol-d6	7.152	99	22007	0.744	ng	0.00
8) Nitrobenzene-d5	9.158	82	17176	0.740	ng	0.00
11) 2-Methylnaphthalene-d10	12.391	152	28216	0.740	ng	0.00
14) 2,4,6-Tribromophenol	16.122	330	4790	0.694	ng	0.00
15) 2-Fluorobiphenyl	13.254	172	46624	0.746	ng	0.00
27) Fluoranthene-d10	19.404	212	50116	0.735	ng	0.00
31) Terphenyl-d14	19.998	244	49612	0.745	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.403	88	8172	0.766	ng	98
3) n-Nitrosodimethylamine	3.728	42	13924	0.778	ng	# 97
6) bis(2-Chloroethyl)ether	7.412	93	22025	0.758	ng	99
9) Naphthalene	10.856	128	57457	0.746	ng	99
10) Hexachlorobutadiene	11.144	225	11472	0.750	ng	# 100
12) 2-Methylnaphthalene	12.463	142	38412	0.743	ng	99
16) Acenaphthylene	14.355	152	47570	0.713	ng	99
17) Acenaphthene	14.686	154	33835	0.733	ng	99
18) Fluorene	15.680	166	46547	0.733	ng	99
20) 4,6-Dinitro-2-methylph...	15.750	198	1711	0.629	ng	# 64
21) 4-Bromophenyl-phenylether	16.569	248	14149	0.733	ng	99
22) Hexachlorobenzene	16.681	284	17109	0.753	ng	99
23) Atrazine	16.830	200	8304	0.706	ng	97
24) Pentachlorophenol	17.028	266	5293	0.660	ng	100
25) Phenanthrene	17.413	178	69059	0.742	ng	100
26) Anthracene	17.512	178	58234	0.722	ng	99
28) Fluoranthene	19.433	202	73120	0.739	ng	100
30) Pyrene	19.800	202	71707	0.740	ng	100
32) Benzo(a)anthracene	21.547	228	53231	0.721	ng	100
33) Chrysene	21.601	228	61102	0.749	ng	99
34) Bis(2-ethylhexyl)phtha...	21.467	149	22392	0.677	ng	99
36) Indeno(1,2,3-cd)pyrene	26.605	276	44811	0.749	ng	100
37) Benzo(b)fluoranthene	23.281	252	48260	0.754	ng	96
38) Benzo(k)fluoranthene	23.328	252	47508	0.749	ng	95
39) Benzo(a)pyrene	23.924	252	40244	0.740	ng	# 93
40) Dibenzo(a,h)anthracene	26.626	278	35561	0.755	ng	98
41) Benzo(g,h,i)perylene	27.401	276	41246	0.748	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040323\
Data File : BN024726.D
Acq On : 03 Apr 2023 12:23
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Apr 04 03:06:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 03 15:20:44 2023
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.997 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN024726.D
 Acq: 03 Apr 2023 12:23


Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8





Tgt Ion:152 Resp: 10335
 Ion Ratio Lower Upper
 152 100
 150 155.2 123.0 184.4
 115 61.2 48.7 73.1




#2
 1,4-Dioxane
 Concen: 0.766 ng
 RT: 3.403 min Scan# 17
 Delta R.T. 0.000 min
 Lab File: BN024726.D
 Acq: 03 Apr 2023 12:23



Tgt Ion: 88 Resp: 8172
 Ion Ratio Lower Upper
 88 100
 43 40.8 32.8 49.2
 58 88.7 73.0 109.6









#3
 n-Nitrosodimethylamine
 Concen: 0.778 ng
 RT: 3.728 min Scan# 61
 Delta R.T. -0.007 min
 Lab File: BN024726.D
 Acq: 03 Apr 2023 12:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion: 42 Resp: 13924
 Ion Ratio Lower Upper
 42 100
 74 90.3 73.9 110.9
 44 6.9 7.6 11.4#



#4
 2-Fluorophenol
 Concen: 0.739 ng
 RT: 5.541 min Scan# 313
 Delta R.T. 0.000 min
 Lab File: BN024726.D
 Acq: 03 Apr 2023 12:23

Tgt Ion: 112 Resp: 17368
 Ion Ratio Lower Upper
 112 100
 64 65.4 52.1 78.1
 63 35.2 28.7 43.1









#5
Phenol-d6
Concen: 0.744 ng
RT: 7.152 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 22007
Ion Ratio Lower Upper
99 100
42 29.3 23.7 35.5
71 35.0 28.2 42.4



#6
bis(2-Chloroethyl)ether
Concen: 0.758 ng
RT: 7.412 min Scan# 572
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion: 93 Resp: 22025
Ion Ratio Lower Upper
93 100
63 87.1 69.0 103.6
95 32.7 26.1 39.1









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.802 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	29762
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.6
54	14.1	11.4	17.0
68	8.7	7.0	10.6



#8
Nitrobenzene-d5
Concen: 0.740 ng
RT: 9.158 min Scan# 776
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:	82	Resp:	17176
Ion	Ratio	Lower	Upper
82	100		
128	36.5	29.4	44.2
54	62.1	51.3	76.9









#9
Naphthalene
Concen: 0.746 ng
RT: 10.856 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 57457
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 13.1 10.6 15.8



#10
Hexachlorobutadiene
Concen: 0.750 ng
RT: 11.144 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:225 Resp: 11472
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.4 50.5 75.7









#11
2-Methylnaphthalene-d10
Concen: 0.740 ng
RT: 12.391 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 28216
Ion Ratio Lower Upper
152 100
151 20.9 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.743 ng
RT: 12.463 min Scan# 1249
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:142 Resp: 38412
Ion Ratio Lower Upper
142 100
141 88.7 70.5 105.7
115 38.1 30.9 46.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.633 min Scan# 1545
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 16028
Ion Ratio Lower Upper
164 100
162 102.2 82.2 123.4
160 48.5 39.4 59.0



#14
2,4,6-Tribromophenol
Concen: 0.694 ng
RT: 16.122 min Scan# 1679
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:330 Resp: 4790
Ion Ratio Lower Upper
330 100
332 97.7 77.6 116.4
141 37.0 29.6 44.4









#15
2-Fluorobiphenyl
Concen: 0.746 ng
RT: 13.254 min Scan# 1416
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 46624
Ion Ratio Lower Upper
172 100
171 34.9 28.8 43.2
170 23.6 19.5 29.3



#16
Acenaphthylene
Concen: 0.713 ng
RT: 14.355 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:152 Resp: 47570
Ion Ratio Lower Upper
152 100
151 19.5 15.2 22.8
153 12.9 10.4 15.6









#17
Acenaphthene
Concen: 0.733 ng
RT: 14.686 min Scan# 1550
Delta R.T. -0.011 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:154 Resp: 33835
Ion Ratio Lower Upper
154 100
153 113.8 92.5 138.7
152 58.2 47.2 70.8



#18
Fluorene
Concen: 0.733 ng
RT: 15.680 min Scan# 1643
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:166 Resp: 46547
Ion Ratio Lower Upper
166 100
165 100.1 79.1 118.7
167 13.5 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.376 min Scan# 1780
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188 Resp: 32213
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.0 9.2 13.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.629 ng
RT: 15.750 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:198 Resp: 1711
Ion Ratio Lower Upper
198 100
51 94.3 118.9 178.3#
105 60.5 64.6 97.0#









#21
4-Bromophenyl-phenylether
Concen: 0.733 ng
RT: 16.569 min Scan# 1715
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 14149
Ion Ratio Lower Upper
248 100
250 104.7 83.6 125.4
141 35.2 30.3 45.5



#22
Hexachlorobenzene
Concen: 0.753 ng
RT: 16.681 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:284 Resp: 17109
Ion Ratio Lower Upper
284 100
142 36.1 29.4 44.2
249 30.7 24.8 37.2





#23

Atrazine

Concen: 0.706 ng

RT: 16.830 min Scan# 1736

Delta R.T. 0.000 min

Lab File: BN024726.D

Acq: 03 Apr 2023 12:23

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:200 Resp: 8304

Ion Ratio Lower Upper

200 100

173 26.9 23.7 35.5

215 45.4 37.0 55.4



#24

Pentachlorophenol

Concen: 0.660 ng

RT: 17.028 min Scan# 1752

Delta R.T. 0.000 min

Lab File: BN024726.D

Acq: 03 Apr 2023 12:23




Tgt Ion:266 Resp: 5293


Ion Ratio Lower Upper

266 100

264 62.2 49.4 74.0

268 63.6 51.0 76.4









#25
Phenanthrene
Concen: 0.742 ng
RT: 17.413 min Scan# 1783
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 69059
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.2
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.722 ng
RT: 17.512 min Scan# 1791
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:178 Resp: 58234
Ion Ratio Lower Upper
178 100
176 18.6 14.6 21.8
179 15.2 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.735 ng
RT: 19.404 min Scan# 2162
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 50116
Ion Ratio Lower Upper
212 100
106 16.2 13.0 19.4
104 9.2 7.4 11.0



#28
Fluoranthene
Concen: 0.739 ng
RT: 19.433 min Scan# 2169
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:202 Resp: 73120
Ion Ratio Lower Upper
202 100
101 12.6 10.1 15.1
203 17.1 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.565 min Scan# 2479

Delta R.T. 0.000 min

Lab File: BN024726.D

Acq: 03 Apr 2023 12:23

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 23046

Ion Ratio Lower Upper

240 100

120 15.9 10.6 16.0

236 28.3 22.3 33.5



#30

Pyrene

Concen: 0.740 ng

RT: 19.800 min Scan# 2256

Delta R.T. 0.004 min

Lab File: BN024726.D

Acq: 03 Apr 2023 12:23




Tgt Ion:202 Resp: 71707


Ion Ratio Lower Upper

202 100

200 21.0 16.6 25.0

203 17.8 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.745 ng
 RT: 19.998 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN024726.D
 Acq: 03 Apr 2023 12:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 49612
 Ion Ratio Lower Upper
 244 100
 212 8.8 7.4 11.0
 122 0.0 0.0 0.0



#32
 Benzo(a)anthracene
 Concen: 0.721 ng
 RT: 21.547 min Scan# 2477
 Delta R.T. 0.000 min
 Lab File: BN024726.D
 Acq: 03 Apr 2023 12:23

Tgt Ion:228 Resp: 53231
 Ion Ratio Lower Upper
 228 100
 226 27.5 22.2 33.2
 229 19.5 15.8 23.8









#33
Chrysene
Concen: 0.749 ng
RT: 21.601 min Scan# 2483
Delta R.T. 0.000 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:228 Resp: 61102
Ion Ratio Lower Upper
228 100
226 31.2 24.7 37.1
229 19.1 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.677 ng
RT: 21.467 min Scan# 2468
Delta R.T. 0.009 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Tgt Ion:149 Resp: 22392
Ion Ratio Lower Upper
149 100
167 26.6 21.0 31.6
279 2.9 2.8 4.2





#35

Perylene-d12

Concen: 0.400 ng

RT: 24.035 min Scan# 31

Delta R.T. -0.009 min

Lab File: BN024726.D

Acq: 03 Apr 2023 12:23

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:264 Resp: 15761

Ion Ratio Lower Upper

264 100

260 26.8 20.6 31.0

265 72.5 62.5 93.7



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.749 ng

RT: 26.605 min Scan# 4009

Delta R.T. -0.032 min

Lab File: BN024726.D

Acq: 03 Apr 2023 12:23




Tgt Ion:276 Resp: 44811


Ion Ratio Lower Upper

276 100

138 29.2 23.5 35.3

277 24.7 19.9 29.9





#37

Benzo(b)fluoranthene

Concen: 0.754 ng

RT: 23.281 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN024726.D

Acq: 03 Apr 2023 12:23

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 48260

Ion Ratio Lower Upper

252 100

253 23.8 20.4 30.6

125 14.0 13.4 20.0



#38

Benzo(k)fluoranthene

Concen: 0.749 ng

RT: 23.328 min Scan# 2888

Delta R.T. -0.003 min

Lab File: BN024726.D

Acq: 03 Apr 2023 12:23




Tgt Ion:252 Resp: 47508


Ion Ratio Lower Upper

252 100

253 23.7 21.0 31.4

125 14.1 13.3 19.9









#39
Benzo(a)pyrene
Concen: 0.740 ng
RT: 23.924 min Scan# 30
Delta R.T. -0.012 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 40244
Ion Ratio Lower Upper
252 100
253 24.8 22.5 33.7
125 16.5 16.5 24.7#



#40
Dibenzo(a,h)anthracene
Concen: 0.755 ng
RT: 26.626 min Scan# 4016
Delta R.T. -0.026 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23


Tgt Ion:278 Resp: 35561
Ion Ratio Lower Upper
278 100
139 20.5 16.9 25.3
279 26.0 21.7 32.5





#41
Benzo(g,h,i)perylene
Concen: 0.748 ng
RT: 27.401 min Scan# 41
Delta R.T. -0.032 min
Lab File: BN024726.D
Acq: 03 Apr 2023 12:23

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:276 Resp: 41246
Ion Ratio Lower Upper
276 100
277 24.9 20.0 30.0
138 25.8 21.6 32.4

