


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040324\
Data File : BN030601.D
Acq On : 02 Apr 2024 00:14
Operator : MA/JU
Sample : DFTPP210
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP210

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN032724.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Apr 02 02:09:33 2024



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.4	157477	PASS
68	69	0.00	2	1.6	2703	PASS
69	198	0.00	100	36.2	165509	PASS
70	69	0.00	2	0.4	695	PASS
127	198	10	80	46.7	213739	PASS
197	198	0.00	2	0.2	990	PASS
198	198	100	100	100.0	457813	PASS
199	198	5	9	6.6	30288	PASS
275	198	10	60	25.6	117080	PASS
365	198	1	100	3.5	16147	PASS
441	198	0.01	100	11.4	51992	PASS
442	198	50	100	68.9	315413	PASS
443	442	15	24	18.9	59469	PASS

m/z	Abundance
40.00	2921.0
41.00	505.0
42.00	166.0
43.00	577.0
44.00	4394.0
45.00	318.0
48.90	375.0
51.00	195.0
55.00	600.0
56.10	292.0
57.00	436.0
63.90	291.0
64.90	188.0
65.90	403.0
67.10	219.0
67.90	211.0
69.10	339.0
69.90	222.0
72.90	754.0
75.00	264.0
77.00	474.0
78.00	161.0
81.00	160.0
82.00	172.0
83.00	488.0
83.90	529.0
85.90	231.0
87.90	190.0
91.10	187.0
96.00	806.0
97.10	247.0
102.80	191.0
105.00	222.0
132.90	705.0
134.90	158.0
146.90	267.0
167.00	186.0
177.00	265.0
190.90	623.0
191.80	188.0
192.90	374.0
207.00	5812.0
208.00	1182.0
208.90	764.0
266.90	170.0
281.00	808.0
282.10	178.0
283.10	189.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

m/z	Abundance
36.00	257.0
37.00	864.0
38.00	2511.0
39.00	13804.0
40.00	2698.0
41.00	512.0
41.90	171.0
42.90	257.0
44.00	3967.0
45.00	531.0
46.90	264.0
49.00	1793.0
50.00	49392.0
51.00	192000.0
52.00	9887.0
53.00	518.0
54.90	1061.0
56.00	5532.0
57.00	12627.0
58.00	755.0
59.00	196.0
60.00	226.0
61.00	2364.0
62.00	2905.0
63.00	7465.0
64.00	1280.0
65.00	3803.0
66.00	687.0
67.00	402.0
68.00	3642.0
69.00	201984.0
70.00	1080.0
71.00	234.0
73.00	1960.0
74.00	21464.0
75.00	33008.0
76.00	12055.0
77.00	237376.0
78.00	16864.0
79.00	15414.0
80.00	12320.0
81.00	16704.0
82.00	4298.0
82.90	4078.0
84.00	853.0
85.00	2973.0
86.00	4904.0
87.00	1837.0
88.00	1130.0
89.00	513.0
91.00	3698.0
92.00	3834.0
93.00	28520.0
94.00	1911.0
95.00	477.0
96.00	1479.0
97.00	580.0
98.00	21688.0
99.00	18832.0
99.90	1562.0
101.00	9364.0
102.00	621.0
103.00	3293.0
103.90	6084.0
105.00	5493.0
106.00	2024.0
107.00	73312.0
108.00	11054.0
109.00	1965.0
110.00	130992.0
111.00	20448.0
112.00	2255.0
113.00	1013.0
114.10	168.0
115.00	352.0
116.00	4179.0
116.90	68528.0
117.90	5287.0
118.80	591.0
120.00	869.0
120.90	458.0
122.00	4823.0
123.00	7673.0
124.00	3767.0
124.90	2964.0
126.10	1622.0
127.00	240256.0
128.00	18664.0
129.00	96600.0
130.00	8809.0
131.00	1753.0
131.90	1075.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

133.00	944.0
133.90	3131.0
135.00	7484.0
136.00	3386.0
137.00	5249.0
138.00	973.0
139.00	645.0
140.00	1216.0
141.00	11952.0
142.00	3670.0
142.90	2693.0
144.00	864.0
144.90	675.0
145.90	2334.0
147.00	6901.0
147.90	16001.0
149.00	3607.0
150.00	1044.0
151.00	1619.0
151.60	849.0
152.90	3713.0
154.00	2499.0
155.00	7436.0
156.00	10510.0
157.00	1723.0
157.90	2193.0
159.00	1690.0
159.90	4000.0
161.00	6182.0
161.90	1846.0
163.00	538.0
164.00	845.0
165.00	4761.0
166.00	4086.0
167.00	26392.0
168.00	17352.0
169.00	2567.0
170.00	777.0
170.90	1058.0
172.00	1982.0
173.00	3255.0
173.90	5487.0
175.00	10344.0
176.00	2789.0
177.00	4847.0
178.00	1559.0
178.90	20040.0
180.00	12229.0
181.00	5774.0
182.00	996.0
183.00	508.0
183.90	1132.0
185.00	8905.0
186.00	70752.0
187.00	20352.0
188.00	2439.0
189.00	4308.0
190.00	736.0
191.00	2828.0
192.00	6447.0
193.00	6438.0
193.90	1536.0
195.00	709.0
196.00	12764.0
198.00	491584.0
198.90	34920.0
200.00	2950.0
201.50	2219.0
202.90	3533.0
204.00	17288.0
205.00	32024.0
206.00	129480.0
207.00	23016.0
208.00	5150.0
209.00	2257.0
210.20	2103.0
211.00	5284.0
211.90	853.0
213.00	325.0
215.00	1382.0
216.00	2795.0
217.00	33744.0
217.90	4062.0
218.90	611.0
221.00	21880.0
223.00	7497.0
224.00	69064.0
225.00	18296.0
226.00	2266.0
227.00	30808.0
228.00	4516.0
229.00	6646.0
230.00	914.0
231.00	2638.0
231.90	458.0
232.90	397.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

234.00	1717.0
235.00	2505.0
236.00	1407.0
237.00	3063.0
237.90	472.0
239.00	1286.0
240.00	966.0
240.90	1605.0
242.00	3661.0
243.00	3661.0
244.00	57232.0
245.10	6484.0
246.00	11229.0
247.00	2275.0
247.80	361.0
249.00	2018.0
250.00	472.0
250.90	421.0
252.00	477.0
253.00	1325.0
255.00	270336.0
256.00	38536.0
256.90	3120.0
258.00	15411.0
259.00	2631.0
259.90	422.0
260.90	360.0
262.00	179.0
262.90	207.0
264.00	399.0
265.00	6325.0
265.90	1058.0
267.00	205.0
269.90	360.0
271.00	728.0
272.00	808.0
273.00	8021.0
274.00	20600.0
275.00	123496.0
276.00	16808.0
277.00	10658.0
278.00	2303.0
279.00	327.0
281.00	957.0
282.00	567.0
283.00	1267.0
284.00	630.0
285.00	1610.0
285.90	323.0
288.80	385.0
289.90	337.0
290.80	167.0
292.10	488.0
293.00	2298.0
293.90	514.0
295.00	883.0
296.00	34136.0
297.00	5042.0
297.90	401.0
300.90	306.0
302.10	553.0
303.00	3909.0
304.00	924.0
305.00	271.0
307.90	446.0
308.90	286.0
310.00	585.0
313.00	310.0
314.00	1467.0
315.00	3521.0
316.00	2326.0
317.00	441.0
319.90	258.0
321.00	1019.0
321.90	609.0
323.00	10169.0
324.00	1851.0
324.90	234.0
325.90	249.0
326.90	1943.0
327.90	1171.0
328.90	156.0
332.00	845.0
333.00	1175.0
334.00	6336.0
335.00	2020.0
335.90	222.0
339.00	157.0
339.80	233.0
341.00	1020.0
341.90	388.0
346.00	2143.0
346.90	440.0
350.90	223.0
352.00	3026.0
352.90	2207.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

354.00	3137.0
355.00	880.0
359.10	377.0
365.00	15682.0
365.90	2265.0
367.00	166.0
369.90	296.0
370.90	749.0
372.00	5583.0
373.00	1309.0
373.90	156.0
376.80	173.0
382.90	1512.0
383.90	585.0
384.90	170.0
389.90	690.0
391.00	548.0
391.90	437.0
400.90	443.0
401.90	1830.0
402.90	2882.0
403.90	1289.0
404.90	269.0
414.90	205.0
421.00	2605.0
422.00	2384.0
423.00	18792.0
424.00	5304.0
425.00	531.0
441.00	47688.0
442.00	281280.0
443.00	54576.0
444.00	5676.0
444.90	347.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

m/z	Abundance
35.00	151.0
37.00	950.0
38.00	2251.0
39.00	13363.0
40.00	2810.0
41.00	599.0
43.00	298.0
44.00	3515.0
44.90	566.0
47.00	186.0
49.00	1685.0
50.00	47000.0
51.00	182144.0
52.00	8952.0
52.90	383.0
55.00	1308.0
56.00	5275.0
57.00	12143.0
58.00	479.0
58.90	327.0
59.90	292.0
61.00	2569.0
62.00	2469.0
63.00	7554.0
64.00	1320.0
65.00	3594.0
65.90	627.0
66.90	457.0
68.00	3422.0
69.00	191104.0
70.00	1147.0
71.00	175.0
71.90	190.0
73.00	1970.0
74.00	20616.0
75.00	31872.0
76.00	11379.0
77.00	228096.0
78.00	15983.0
79.00	14575.0
80.00	11302.0
81.00	16560.0
82.00	4187.0
83.00	3534.0
84.00	593.0
85.00	2586.0
86.00	5028.0
87.00	2381.0
87.90	894.0
88.80	372.0
91.00	3837.0
92.00	4905.0
93.00	26872.0
94.00	1712.0
95.00	624.0
96.00	1528.0
97.10	663.0
98.00	22248.0
99.00	18520.0
100.00	1322.0
101.00	8496.0
102.00	626.0
102.90	3170.0
104.00	5584.0
105.00	5463.0
106.00	1577.0
107.00	70992.0
108.00	11024.0
109.00	2242.0
110.00	127848.0
111.00	19152.0
112.00	2441.0
113.00	736.0
114.90	414.0
116.00	4037.0
117.00	65032.0
118.00	5141.0
119.00	714.0
120.00	895.0
121.00	441.0
121.90	4558.0
123.00	6934.0
123.90	3530.0
124.90	3029.0
126.10	1288.0
127.00	250176.0
128.00	18920.0
129.00	97136.0
130.00	8603.0
131.00	1613.0
131.90	950.0
133.00	827.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

134.00	2502.0
135.00	7867.0
136.00	3308.0
137.00	4655.0
138.00	915.0
138.90	488.0
140.00	1117.0
140.90	13295.0
142.00	4336.0
143.00	3000.0
143.90	811.0
144.90	660.0
146.00	2606.0
147.00	6887.0
147.90	16648.0
149.00	3412.0
149.90	1080.0
151.00	1938.0
151.80	1137.0
152.90	4062.0
154.00	3303.0
155.00	7309.0
156.00	10294.0
157.00	2034.0
157.90	2347.0
158.90	1952.0
160.00	4392.0
161.00	6142.0
162.00	1819.0
162.90	749.0
164.00	616.0
165.00	4579.0
166.00	4165.0
167.00	28608.0
168.00	16036.0
169.00	2671.0
169.80	641.0
170.80	1025.0
171.90	2261.0
172.90	2533.0
174.00	5624.0
175.00	10424.0
176.00	2603.0
176.90	5055.0
178.00	1564.0
179.00	19768.0
180.00	14228.0
181.00	5846.0
182.00	1085.0
182.90	540.0
183.90	1508.0
185.00	8824.0
186.00	73824.0
187.00	20880.0
188.00	2003.0
189.00	4630.0
189.80	732.0
191.00	2569.0
192.00	6640.0
193.00	7238.0
194.00	1571.0
195.00	916.0
196.00	14797.0
197.00	2969.0
197.90	540672.0
199.00	34160.0
199.90	3132.0
201.50	2312.0
202.90	3929.0
204.00	19200.0
205.00	31768.0
206.00	140736.0
207.10	22792.0
208.00	5617.0
209.00	2072.0
210.10	2273.0
211.00	5650.0
212.90	362.0
213.80	161.0
215.00	1836.0
215.90	3397.0
217.00	36304.0
218.00	4609.0
219.00	426.0
221.00	21728.0
221.90	5433.0
223.00	8752.0
224.00	78272.0
225.00	20368.0
226.00	2287.0
227.00	33800.0
228.00	4934.0
229.00	6938.0
230.00	1069.0
231.00	2764.0
231.90	521.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

233.00	641.0
233.90	2096.0
234.90	2286.0
235.90	1859.0
237.00	2621.0
238.00	483.0
238.90	1163.0
239.90	1040.0
241.00	1901.0
241.90	3930.0
243.00	4634.0
244.00	63064.0
245.00	7783.0
246.00	12793.0
247.00	2090.0
248.00	649.0
249.00	2291.0
250.00	404.0
250.90	524.0
252.00	488.0
253.00	1180.0
255.00	316416.0
256.00	45648.0
257.00	3213.0
258.00	18768.0
259.00	2911.0
259.90	532.0
260.90	528.0
263.00	224.0
263.90	473.0
265.00	7523.0
265.90	1131.0
266.90	164.0
269.90	333.0
271.00	841.0
272.00	803.0
273.00	9503.0
274.00	25296.0
275.00	133760.0
276.00	17456.0
277.00	12108.0
278.00	2437.0
279.00	491.0
281.00	831.0
282.00	403.0
283.00	1398.0
283.90	803.0
285.00	1992.0
286.00	367.0
288.00	195.0
288.90	480.0
289.90	288.0
291.00	348.0
292.00	528.0
293.00	2770.0
294.00	733.0
295.10	1233.0
296.00	42752.0
297.00	5773.0
298.10	337.0
301.10	605.0
301.90	683.0
303.00	4857.0
304.00	1083.0
305.00	213.0
308.00	493.0
309.00	348.0
309.90	600.0
312.00	283.0
312.90	410.0
314.00	2150.0
314.90	4676.0
315.90	2742.0
317.00	587.0
319.80	163.0
321.00	1103.0
322.00	678.0
323.00	13049.0
324.00	2437.0
324.90	210.0
326.00	215.0
327.00	2548.0
327.90	1169.0
328.90	179.0
332.00	893.0
333.00	1018.0
334.00	8943.0
335.00	2301.0
336.10	176.0
338.90	220.0
339.90	263.0
341.00	2105.0
341.90	352.0
345.90	2877.0
346.90	721.0
351.00	278.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

352.00	3889.0
353.00	2918.0
354.00	4276.0
355.00	940.0
358.90	282.0
364.90	19064.0
366.00	2941.0
366.80	215.0
370.00	520.0
371.00	1122.0
372.00	6386.0
373.00	1665.0
373.90	186.0
376.90	236.0
383.00	2016.0
383.90	580.0
389.90	739.0
391.00	845.0
392.00	551.0
396.90	192.0
401.10	524.0
402.00	2919.0
402.90	4264.0
404.00	1664.0
405.00	317.0
410.00	167.0
414.90	294.0
421.00	3375.0
422.00	3438.0
423.00	22256.0
424.00	5444.0
425.00	866.0
441.00	61000.0
442.00	360384.0
443.00	69208.0
444.00	6129.0
445.00	313.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

m/z	Abundance
36.00	273.0
37.00	513.0
38.00	1173.0
39.00	6947.0
40.00	2672.0
40.90	557.0
43.00	302.0
44.00	4631.0
45.00	445.0
47.00	198.0
48.10	177.0
49.00	1209.0
50.00	24472.0
51.00	98872.0
52.00	4725.0
52.90	329.0
54.90	613.0
55.90	2819.0
56.90	6532.0
57.90	274.0
59.00	294.0
60.00	174.0
60.90	1332.0
62.00	1595.0
63.00	4289.0
63.90	642.0
65.00	2149.0
65.90	520.0
67.00	333.0
68.00	1678.0
69.00	104456.0
70.00	525.0
70.90	159.0
73.00	1516.0
74.00	10905.0
75.00	17568.0
76.00	6346.0
77.00	127768.0
78.00	9044.0
79.00	8763.0
79.90	6594.0
81.00	9086.0
81.90	2190.0
83.00	1962.0
84.00	521.0
84.90	1319.0
86.00	2831.0
86.90	1303.0
87.90	513.0
88.80	280.0
91.00	2297.0
92.00	2635.0
93.00	15740.0
93.90	1117.0
94.80	302.0
96.00	1158.0
97.10	332.0
98.00	11799.0
99.00	9873.0
99.90	864.0
100.90	5442.0
101.80	278.0
102.90	2033.0
103.90	3790.0
105.00	3724.0
105.90	1010.0
107.00	39512.0
108.00	6249.0
109.00	1232.0
110.00	74168.0
111.00	11022.0
112.00	1385.0
113.00	477.0
114.10	173.0
114.90	453.0
116.00	2401.0
117.00	39080.0
118.00	2869.0
118.90	465.0
120.00	526.0
120.90	275.0
121.90	2630.0
123.00	4120.0
123.90	2362.0
124.90	1898.0
126.10	848.0
127.00	150784.0
128.00	11697.0
129.00	57056.0
130.00	5088.0
131.00	959.0
132.00	611.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

132.90	753.0
133.90	1609.0
135.00	4898.0
136.00	1897.0
137.00	2765.0
137.90	574.0
138.90	248.0
140.00	823.0
141.00	8216.0
142.00	2736.0
142.90	1757.0
143.90	524.0
145.00	386.0
145.90	1241.0
147.00	4324.0
148.00	9454.0
149.00	2220.0
149.90	640.0
151.10	1071.0
151.90	745.0
152.90	2469.0
153.90	1826.0
155.00	4691.0
156.00	6283.0
157.10	1418.0
157.90	1507.0
159.00	1003.0
159.90	2496.0
161.00	4101.0
162.00	878.0
162.90	444.0
163.90	495.0
164.90	3236.0
165.90	2592.0
167.00	17856.0
168.00	10977.0
169.00	1579.0
170.00	479.0
170.90	716.0
171.90	1489.0
173.00	2015.0
174.00	3650.0
175.00	6234.0
176.00	2146.0
177.00	2930.0
178.00	920.0
179.00	11710.0
180.00	8227.0
181.00	4016.0
181.90	550.0
183.00	282.0
184.00	1129.0
185.00	5877.0
186.00	48264.0
187.00	13441.0
188.00	1340.0
189.00	2789.0
189.90	436.0
191.00	2303.0
191.90	4350.0
193.00	4628.0
194.00	1105.0
195.10	651.0
196.00	10009.0
197.90	341184.0
198.90	21784.0
200.00	1627.0
201.60	1401.0
202.90	2363.0
204.00	12244.0
205.00	21776.0
206.00	91224.0
207.00	16480.0
208.00	3999.0
209.00	1848.0
210.10	1513.0
211.00	3339.0
213.00	285.0
214.90	962.0
216.00	1917.0
216.90	23352.0
218.00	3080.0
219.00	489.0
221.00	14394.0
221.90	2935.0
223.00	5465.0
224.00	53688.0
225.00	13326.0
226.00	1476.0
227.00	22384.0
228.00	2894.0
228.90	4489.0
230.00	607.0
231.00	1530.0
231.90	355.0
232.80	398.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

233.90	1428.0
234.90	1873.0
236.00	1165.0
237.00	1885.0
237.90	284.0
239.00	781.0
240.00	681.0
240.90	1265.0
242.00	2506.0
243.00	3061.0
244.00	40816.0
245.00	5301.0
246.00	8390.0
246.90	1763.0
247.90	386.0
248.90	1561.0
249.80	205.0
250.80	575.0
251.90	376.0
252.90	773.0
255.00	219840.0
256.00	29680.0
257.00	2555.0
257.90	12069.0
259.00	1895.0
259.90	368.0
260.90	365.0
264.00	353.0
265.00	4983.0
265.90	679.0
267.10	186.0
270.00	339.0
270.90	539.0
272.00	499.0
273.00	6556.0
274.00	17408.0
275.00	93984.0
276.00	13416.0
277.00	8220.0
277.90	1424.0
279.00	276.0
281.10	825.0
281.90	342.0
283.00	1047.0
284.10	731.0
284.90	1307.0
286.00	232.0
289.00	413.0
289.70	212.0
290.00	237.0
290.90	332.0
291.90	380.0
293.00	2351.0
293.90	461.0
295.00	545.0
296.00	30288.0
296.90	4042.0
298.00	209.0
300.80	350.0
301.90	433.0
303.00	2581.0
303.90	732.0
307.90	500.0
309.10	171.0
309.80	324.0
311.10	153.0
312.90	347.0
313.90	1184.0
315.00	3436.0
316.00	1825.0
317.00	338.0
319.80	193.0
320.90	954.0
322.00	451.0
323.00	8883.0
324.00	1600.0
324.90	162.0
326.90	1804.0
328.00	835.0
328.90	179.0
332.00	572.0
333.00	876.0
334.00	5989.0
335.00	1858.0
336.00	158.0
339.90	171.0
341.00	1317.0
341.90	271.0
345.90	2260.0
346.90	463.0
350.80	247.0
351.90	3078.0
353.00	2178.0
354.00	3123.0
355.00	598.0
358.90	225.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

364.10	251.0
364.90	13694.0
366.00	2455.0
367.00	162.0
369.90	362.0
370.90	816.0
372.00	5748.0
373.00	1270.0
373.90	190.0
382.90	1620.0
383.90	281.0
389.90	583.0
390.90	541.0
391.90	428.0
400.80	264.0
402.00	2401.0
403.00	3208.0
403.90	1296.0
404.90	367.0
414.90	179.0
420.90	2893.0
422.00	2771.0
423.00	19304.0
424.00	4765.0
425.00	489.0
441.00	47288.0
442.00	304576.0
443.00	54624.0
444.00	4734.0
444.90	239.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210