

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040422\
 Data File : BN019323.D
 Acq On : 04 Apr 2022 17:24
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBN040422

Quant Time: Apr 05 03:29:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 04 17:40:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	83	0.00
2	1,4-Dioxane	0.400	0.474	-18.5	88	0.00
3	n-Nitrosodimethylamine	0.400	0.466	-16.5	82	0.00
4 S	2-Fluorophenol	0.400	0.460	-15.0	80	0.00
5 S	Phenol-d6	0.400	0.423	-5.7	78	0.00
6	bis(2-Chloroethyl)ether	0.400	0.448	-12.0	79	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	83	0.00
8 S	Nitrobenzene-d5	0.400	0.403	-0.8	77	0.00
9	Naphthalene	0.400	0.454	-13.5	82	0.00
10	Hexachlorobutadiene	0.400	0.470	-17.5	84	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.440	-10.0	81	0.00
12	2-Methylnaphthalene	0.400	0.454	-13.5	83	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.400	0.387	3.3	73	0.00
15 S	2-Fluorobiphenyl	0.400	0.442	-10.5	82	0.00
16	Acenaphthylene	0.400	0.411	-2.7	79	0.00
17	Acenaphthene	0.400	0.437	-9.2	83	0.00
18	Fluorene	0.400	0.443	-10.7	82	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.310	22.5	67	0.01
21	4-Bromophenyl-phenylether	0.400	0.426	-6.5	81	-0.01
22	Hexachlorobenzene	0.400	0.452	-13.0	85	0.00
23	Atrazine	0.400	0.380	5.0	71	0.01
24	Pentachlorophenol	0.400	0.285	28.8#	53	0.00
25	Phenanthrene	0.400	0.431	-7.7	82	0.00
26	Anthracene	0.400	0.402	-0.5	78	0.00
27 SURR	Fluoranthene-d10	0.400	0.431	-7.7	81	0.00
28	Fluoranthene	0.400	0.429	-7.2	80	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	83	0.00
30	Pyrene	0.400	0.462	-15.5	81	0.00
31 S	Terphenyl-d14	0.400	0.443	-10.7	79	0.00
32	Benzo(a)anthracene	0.400	0.383	4.3	71	0.00
33	Chrysene	0.400	0.440	-10.0	78	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.366	8.5	66	0.00
35 I	Perylene-d12	0.400	0.400	0.0	81	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.397	0.8	70	0.00
37	Benzo(b)fluoranthene	0.400	0.430	-7.5	81	0.00
38	Benzo(k)fluoranthene	0.400	0.396	1.0	72	0.00
39 C	Benzo(a)pyrene	0.400	0.412	-3.0	75	0.00
40	Dibenzo(a,h)anthracene	0.400	0.388	3.0	66	0.00
41	Benzo(g,h,i)perylene	0.400	0.438	-9.5	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN040422