

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040422\
 Data File : BN019329.D
 Acq On : 04 Apr 2022 22:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 05 03:30:50 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 04 17:40:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	122	0.00
2	1,4-Dioxane	0.400	0.429	-7.2	117	0.00
3	n-Nitrosodimethylamine	0.400	0.427	-6.7	111	0.00
4 S	2-Fluorophenol	0.400	0.436	-9.0	112	0.00
5 S	Phenol-d6	0.400	0.396	1.0	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.442	-10.5	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.381	4.8	106	0.00
9	Naphthalene	0.400	0.454	-13.5	120	0.00
10	Hexachlorobutadiene	0.400	0.472	-18.0	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.454	-13.5	122	0.00
12	2-Methylnaphthalene	0.400	0.460	-15.0	122	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.400	0.380	5.0	105	0.00
15 S	2-Fluorobiphenyl	0.400	0.450	-12.5	122	0.00
16	Acenaphthylene	0.400	0.422	-5.5	118	0.00
17	Acenaphthene	0.400	0.456	-14.0	126	0.00
18	Fluorene	0.400	0.452	-13.0	122	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.313	21.8	97	0.00
21	4-Bromophenyl-phenylether	0.400	0.434	-8.5	119	-0.01
22	Hexachlorobenzene	0.400	0.455	-13.7	123	-0.01
23	Atrazine	0.400	0.380	5.0	102	0.00
24	Pentachlorophenol	0.400	0.385	3.8	103	0.00
25	Phenanthrene	0.400	0.436	-9.0	119	0.00
26	Anthracene	0.400	0.409	-2.2	114	0.00
27 SURR	Fluoranthene-d10	0.400	0.435	-8.7	117	0.00
28	Fluoranthene	0.400	0.436	-9.0	118	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	121	0.00
30	Pyrene	0.400	0.453	-13.2	116	0.00
31 S	Terphenyl-d14	0.400	0.437	-9.2	114	0.00
32	Benzo(a)anthracene	0.400	0.418	-4.5	114	0.00
33	Chrysene	0.400	0.422	-5.5	110	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.359	10.3	96	0.00
35 I	Perylene-d12	0.400	0.400	0.0	108	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.397	0.8	93	0.00
37	Benzo(b)fluoranthene	0.400	0.445	-11.2	111	0.00
38	Benzo(k)fluoranthene	0.400	0.418	-4.5	101	0.00
39 C	Benzo(a)pyrene	0.400	0.430	-7.5	104	0.00
40	Dibenzo(a,h)anthracene	0.400	0.382	4.5	87	0.00
41	Benzo(g,h,i)perylene	0.400	0.419	-4.7	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4