

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040523\
 Data File : BN024771.D
 Acq On : 05 Apr 2023 17:22
 Operator : CG/JU
 Sample : PB151694BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB151694BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 04/06/2023
 Supervised By :Jagrut Upadhyay 04/06/2023

Quant Time: Apr 05 17:51:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 05 11:04:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	7729	0.400	ng	0.00
7) Naphthalene-d8	10.771	136	20202	0.400	ng	0.00
13) Acenaphthene-d10	14.598	164	10830	0.400	ng	0.00
19) Phenanthrene-d10	17.336	188	24492	0.400	ng	0.00
29) Chrysene-d12	21.531	240	20282	0.400	ng	# 0.00
35) Perylene-d12	23.976	264	19651	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.499	112	6594	0.369	ng	0.00
5) Phenol-d6	7.117	99	8374	0.372	ng	0.00
8) Nitrobenzene-d5	9.127	82	5613	0.345	ng	0.00
11) 2-Methylnaphthalene-d10	12.355	152	9413	0.352	ng	0.00
14) 2,4,6-Tribromophenol	16.119	330	920m	0.189	ng	0.03
15) 2-Fluorobiphenyl	13.219	172	14602	0.353	ng	0.00
27) Fluoranthene-d10	19.372	212	19007	0.342	ng	0.00
31) Terphenyl-d14	19.962	244	18229	0.396	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040523\
Data File : BN024771.D
Acq On : 05 Apr 2023 17:22
Operator : CG/JU
Sample : PB151694BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB151694BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023

Quant Time: Apr 05 17:51:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 05 11:04:25 2023
Response via : Initial Calibration

