

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040523\
 Data File : BN024762.D
 Acq On : 05 Apr 2023 08:53
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 05 11:04:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 05 11:04:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	0.00
2	1,4-Dioxane	0.440	0.510	-15.9	128	0.00
3	n-Nitrosodimethylamine	0.745	0.849	-14.0	131	0.00
4 S	2-Fluorophenol	0.924	1.049	-13.5	131	0.00
5 S	Phenol-d6	1.165	1.292	-10.9	130	0.00
6	bis(2-Chloroethyl)ether	1.182	1.375	-16.3	132	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
8 S	Nitrobenzene-d5	0.322	0.355	-10.2	130	0.00
9	Naphthalene	1.030	1.195	-16.0	133	0.00
10	Hexachlorobutadiene	0.196	0.226	-15.3	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.530	0.598	-12.8	131	0.00
12	2-Methylnaphthalene	0.716	0.810	-13.1	132	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
14 S	2,4,6-Tribromophenol	0.180	0.183	-1.7	125	0.00
15 S	2-Fluorobiphenyl	1.528	1.815	-18.8	131	0.00
16	Acenaphthylene	1.681	1.864	-10.9	131	0.00
17	Acenaphthene	1.138	1.331	-17.0	132	0.00
18	Fluorene	1.562	1.804	-15.5	129	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.050	5.7	139	0.00
21	4-Bromophenyl-phenylether	0.236	0.261	-10.6	128	0.00
22	Hexachlorobenzene	0.267	0.307	-15.0	130	0.00
23	Atrazine	0.161	0.164	-1.9	124	0.00
24	Pentachlorophenol	0.073	0.062	15.1	144	0.00
25	Phenanthrene	1.161	1.308	-12.7	130	0.00
26	Anthracene	1.013	1.086	-7.2	129	0.00
27 SURR	Fluoranthene-d10	0.907	0.992	-9.4	128	0.00
28	Fluoranthene	1.318	1.438	-9.1	128	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	128	0.00
30	Pyrene	1.422	1.674	-17.7	128	0.00
31 S	Terphenyl-d14	0.907	1.141	-25.8#	134	0.00
32	Benzo(a)anthracene	1.352	1.586	-17.3	129	0.00
33	Chrysene	1.417	1.692	-19.4	131	0.00
34	Bis(2-ethylhexyl)phthalate	0.727	0.757	-4.1	128	0.00
35 I	Perylene-d12	1.000	1.000	0.0	131	0.00
36	Indeno(1,2,3-cd)pyrene	1.754	1.897	-8.2	125	0.00
37	Benzo(b)fluoranthene	1.517	1.689	-11.3	129	0.00
38	Benzo(k)fluoranthene	1.504	1.647	-9.5	125	0.00
39 C	Benzo(a)pyrene	1.451	1.623	-11.9	130	0.00
40	Dibenzo(a,h)anthracene	1.380	1.502	-8.8	126	0.00
41	Benzo(g,h,i)perylene	1.495	1.624	-8.6	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4