



Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN040523\  
 Data File : BN024751.D  
 Acq On : 04 Apr 2023 17:52  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN040523.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Apr 05 11:04:25 2023






AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2456

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 53.9         | 287791     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 4094       | PASS                |
| 69             | 198             | 0.00            | 100             | 44.3         | 236646     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 1310       | PASS                |
| 127            | 198             | 10              | 80              | 48.8         | 260459     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.1          | 695        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 534229     | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 36101      | PASS                |
| 275            | 198             | 10              | 60              | 24.7         | 131893     | PASS                |
| 365            | 198             | 1               | 100             | 3.7          | 19721      | PASS                |
| 441            | 198             | 0.01            | 100             | 8.9          | 47283      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 460907     | PASS                |
| 443            | 442             | 15              | 24              | 18.9         | 87227      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN040523\  
 Data File : BN024751.D  
 Acq On : 04 Apr 2023 17:52  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Quant Time: Apr 05 06:18:07 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Apr 05 06:18:03 2023  
 Response via : Initial Calibration



TIC: BN024751.D\data.ms

**(70) Pentachlorophenol (C)**

16.992min ( 0.000) 21217.10 ng




response 645379

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 68.66  |
| 264.00 | 61.60  | 64.14  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN040523\  
Data File : BN024751.D  
Acq On : 04 Apr 2023 17:52  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Apr 05 06:18:07 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Apr 05 06:18:03 2023  
Response via : Initial Calibration



TIC: BN024751.D\data.ms

(77) Benzidine

19.592min ( 0.000) 0.00 ng

response 4056365

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.14  |
| 183.00 | 13.20  | 11.40  |
| 0.00   | 0.00   | 0.00   |

# DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 4/5/2023      | BNA_N            | <u>BN024751.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 1659226          | 20.822             |
| DDD           | 45482            | 20.38              |
| DDE           | 527              | 19.869             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 46009         | 1705235          | 2.70               |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP