

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040725\
 Data File : BN036849.D
 Acq On : 07 Apr 2025 10:22
 Operator : RC/JU
 Sample : Q1731-04DL 5X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RMW-03B-90-040325DL

Quant Time: Apr 07 10:46:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 04 17:30:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.695	152	1423	0.400	ng	0.00
7) Naphthalene-d8	10.487	136	3367	0.400	ng	# 0.01
13) Acenaphthene-d10	14.334	164	2092	0.400	ng	0.00
19) Phenanthrene-d10	17.086	188	4259	0.400	ng	# 0.01
29) Chrysene-d12	21.277	240	3832	0.400	ng	0.00
35) Perylene-d12	23.519	264	3500	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.283	112	163	0.049	ng	0.00
5) Phenol-d6	6.879	99	106	0.026	ng	0.01
8) Nitrobenzene-d5	8.864	82	318	0.087	ng	0.02
11) 2-Methylnaphthalene-d10	12.101	152	469	0.094	ng	0.03
14) 2,4,6-Tribromophenol	15.845	330	110	0.116	ng	0.01
15) 2-Fluorobiphenyl	12.978	172	982	0.081	ng	0.02
27) Fluoranthene-d10	19.122	212	1303	0.119	ng	0.00
31) Terphenyl-d14	19.722	244	1067	0.116	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.218	88	2931	1.857	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040725\
Data File : BN036849.D
Acq On : 07 Apr 2025 10:22
Operator : RC/JU
Sample : Q1731-04DL 5X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RMW-03B-90-040325DL

Quant Time: Apr 07 10:46:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 04 17:30:43 2025
Response via : Initial Calibration

