

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040822\
 Data File : BN019402.D
 Acq On : 08 Apr 2022 23:24
 Operator : CG/JU
 Sample : N2333-06 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-10

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 04/11/2022
 Supervised By : mohammad ahmed 04/11/2022

Quant Time: Apr 11 04:12:00 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 04 17:40:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.826	152	4965	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	14668	0.400	ng	# 0.00
13) Acenaphthene-d10	14.463	164	9376	0.400	ng	0.00
19) Phenanthrene-d10	17.205	188	19467	0.400	ng	0.00
29) Chrysene-d12	21.387	240	16471	0.400	ng	# 0.00
35) Perylene-d12	23.741	264	13373	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	454	0.034	ng	0.00
5) Phenol-d6	7.009	99	226	0.014	ng	0.02
8) Nitrobenzene-d5	9.003	82	572	0.043	ng	0.02
11) 2-Methylnaphthalene-d10	12.227	152	1444	0.056	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	298	0.052	ng	0.01
15) 2-Fluorobiphenyl	13.095	172	2103	0.054	ng	0.00
27) Fluoranthene-d10	19.236	212	4153	0.066	ng	0.00
31) Terphenyl-d14	19.830	244	2444	0.068	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.297	149	28945	0.639	ng	Qvalue 100
37) Benzo(b)fluoranthene	23.027	252	1065m	0.020	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040822\
Data File : BN019402.D
Acq On : 08 Apr 2022 23:24
Operator : CG/JU
Sample : N2333-06 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-10

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 04/11/2022
Supervised By : mohammad ahmed 04/11/2022

Quant Time: Apr 11 04:12:00 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 04 17:40:00 2022
Response via : Initial Calibration

