




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040822\
Data File : BN019388.D
Acq On : 08 Apr 2022 12:38
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 11 04:40:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019388.D\data.ms

(77) Benzidine




19.445min (-0.021) 0.00 ng

response 1261015

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.68
183.00	13.20	12.06
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040822\
Data File : BN019388.D
Acq On : 08 Apr 2022 12:38
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 11 04:40:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019388.D\data.ms

(70) Pentachlorophenol (C)

16.863min (-0.009) 19527.49 ng

response 302969

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	65.38
264.00	61.60	65.68
0.00	0.00	0.00