

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040824\
 Data File : BN030779.D
 Acq On : 08 Apr 2024 21:26
 Operator : MA/JU
 Sample : P2002-03
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RMW-06B-44-040424

Quant Time: Apr 09 04:26:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN033024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 30 00:04:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.689	152	5303	0.400	ng	#-0.01
7) Naphthalene-d8	10.464	136	17112	0.400	ng	#-0.02
13) Acenaphthene-d10	14.327	164	2746	0.400	ng	#-0.01
19) Phenanthrene-d10	17.078	188	20833	0.400	ng	-0.01
29) Chrysene-d12	21.276	240	4425	0.400	ng	# 0.00
35) Perylene-d12	23.513	264	283	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.285	112	1733	0.150	ng	-0.01
5) Phenol-d6	6.830	99	4715	0.318	ng	-0.04
8) Nitrobenzene-d5	8.841	82	2636	0.219	ng	-0.01
11) 2-Methylnaphthalene-d10	12.058	152	5680	0.217	ng	-0.02
14) 2,4,6-Tribromophenol	15.812	330	1107	1.067	ng	-0.02
15) 2-Fluorobiphenyl	12.948	172	8829	0.832	ng	-0.02
27) Fluoranthene-d10	19.108	212	3769	0.067	ng	-0.02
31) Terphenyl-d14	19.711	244	10614	1.038	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.233	88	6937	1.084	ng	# 93
32) Benzo(a)anthracene	21.258	228	446	0.028	ng	# 53
33) Chrysene	21.312	228	722	0.043	ng	# 76
34) Bis(2-ethylhexyl)phtha...	21.222	149	1506	0.222	ng	# 66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040824\
Data File : BN030779.D
Acq On : 08 Apr 2024 21:26
Operator : MA/JU
Sample : P2002-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RMW-06B-44-040424

Quant Time: Apr 09 04:26:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN033024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 00:04:36 2024
Response via : Initial Calibration

