

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
 Data File : BN014251.D
 Acq On : 09 Apr 2021 13:19
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Apr 09 13:52:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 12:42:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	6574	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	23466	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	13165	0.40	ng	0.00
19) Phenanthrene-d10	17.091	188	29138	0.40	ng	0.00
29) Chrysene-d12	21.279	240	30282	0.40	ng	0.00
36) Perylene-d12	23.517	264	25465	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	24061	1.37	ng	0.00
5) Phenol-d6	6.879	99	30156	1.26	ng	0.00
8) Nitrobenzene-d5	8.857	82	23197	1.43	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	60011	1.51	ng	0.00
14) 2,4,6-Tribromophenol	15.831	330	8119	1.19	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	80020	1.75	ng	0.00
27) Fluoranthene-d10	19.126	212	134923	1.56	ng	0.00
31) Terphenyl-d14	19.732	244	105567	1.55	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.279	88	12567	1.73	ng	98
3) n-Nitrosodimethylamine	3.609	42	6604	1.65	ng	# 92
6) bis(2-Chloroethyl)ether	7.145	93	25187	1.56	ng	99
9) Naphthalene	10.543	128	94232	1.60	ng	99
10) Hexachlorobutadiene	10.834	225	18841	1.65	ng	# 99
12) 2-Methylnaphthalene	12.160	142	64164	1.51	ng	99
16) Acenaphthylene	14.067	152	83056	1.39	ng	100
17) Acenaphthene	14.410	154	60785	1.63	ng	99
18) Fluorene	15.399	166	76767	1.58	ng	99
20) 4,6-Dinitro-2-methylph...	15.463	198	5571	1.94	ng	# 62
21) 4-Bromophenyl-phenylether	16.288	248	24324	1.53	ng	98
22) Hexachlorobenzene	16.398	284	30948	1.78	ng	100
23) Atrazine	16.556	200	17168	1.09	ng	98
24) Pentachlorophenol	16.738	266	9129	1.26	ng	99
25) Phenanthrene	17.128	178	128211	1.66	ng	100
26) Anthracene	17.225	178	107326	1.42	ng	100
28) Fluoranthene	19.156	202	148658	1.64	ng	100
30) Pyrene	19.518	202	146482	1.51	ng	100
32) Benzo(a)anthracene	21.268	228	138376	1.44	ng	100
33) Chrysene	21.322	228	151611	1.62	ng	100
34) Bis(2-ethylhexyl)phtha...	21.215	149	43258	0.73	ng	98
35) Indeno(1,2,3-cd)pyrene	25.769	276	173142	1.52	ng	100
37) Benzo(b)fluoranthene	22.850	252	149844	1.80	ng	97
38) Benzo(k)fluoranthene	22.894	252	150906	1.88	ng	97
39) Benzo(a)pyrene	23.421	252	127184	1.67	ng	95
40) Dibenzo(a,h)anthracene	25.787	278	146103	1.85	ng	99
41) Benzo(g,h,i)perylene	26.454	276	147214	1.84	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
Data File : BN014251.D
Acq On : 09 Apr 2021 13:19
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Apr 09 13:52:08 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 12:42:56 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.716 min Scan# 613
 Delta R.T. 0.000 min
 Lab File: BN014251.D
 Acq: 09 Apr 2021 13:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:	152	Resp:	6574
Ion	Ratio	Lower	Upper
152	100		
150	154.6	124.8	187.2
115	60.8	48.6	73.0



1,4-Dioxane
 Concen: 1.73 ng
 RT: 3.279 min Scan# 62
 Delta R.T. -0.008 min
 Lab File: BN014251.D
 Acq: 09 Apr 2021 13:19

Tgt Ion:	88	Resp:	12567
Ion	Ratio	Lower	Upper
88	100		
58	67.5	52.5	78.7
43	27.1	21.8	32.8









n-Nitrosodimethylamine
 Concen: 1.65 ng
 RT: 3.609 min Scan# 103
 Delta R.T. -0.024 min
 Lab File: BN014251.D
 Acq: 09 Apr 2021 13:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion: 42 Resp: 6604
 Ion Ratio Lower Upper
 42 100
 74 131.6 111.2 166.8
 44 3.6 12.5 18.7#



2-Fluorophenol
 Concen: 1.37 ng
 RT: 5.316 min Scan# 315
 Delta R.T. 0.000 min
 Lab File: BN014251.D
 Acq: 09 Apr 2021 13:19

Tgt Ion: 112 Resp: 24061
 Ion Ratio Lower Upper
 112 100
 64 53.5 43.4 65.0
 63 29.6 23.5 35.3









Phenol-d6
 Concen: 1.26 ng
 RT: 6.879 min Scan# 509
 Delta R.T. 0.000 min
 Lab File: BN014251.D
 Acq: 09 Apr 2021 13:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	99	Resp:	30156
Ion	Ratio	Lower	Upper
99	100		
42	18.7	15.3	22.9
71	34.2	27.5	41.3



bis(2-Chloroethyl)ether
 Concen: 1.56 ng
 RT: 7.145 min Scan# 542
 Delta R.T. 0.000 min
 Lab File: BN014251.D
 Acq: 09 Apr 2021 13:19


Tgt Ion:	93	Resp:	25187
Ion	Ratio	Lower	Upper
93	100		
63	73.6	58.3	87.5
95	33.3	27.0	40.4






Abundance Scan 859 (10.492 min): BN014249.D\data.ms (-8547-)

Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6






Tgt Ion:	136	Resp:	23466
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	6.0	4.8	7.2
68	5.5	4.4	6.6




Abundance Scan 730 (8.857 min): BN014249.D\data.ms (-72778)

Nitrobenzene-d5
Concen: 1.43 ng
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Tgt Ion:	82	Resp:	23197
Ion	Ratio	Lower	Upper
82	100		
128	38.8	29.7	44.5
54	43.6	36.5	54.7









Naphthalene
Concen: 1.60 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	128	Resp:	94232
Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.0	13.6
127	12.8	10.5	15.7




Hexachlorobutadiene
Concen: 1.65 ng
RT: 10.834 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Tgt Ion:	225	Resp:	18841
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.3	51.9	77.9



Abundance Scan 1069 (12.084 min): BN014249.D\data.ms (-1015) (-)





2-Methylnaphthalene-d10
Concen: 1.51 ng
RT: 12.084 min Scan# 1069
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Tgt Ion:152 Resp: 60011
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Abundance Scan 1069 (12.084 min): BN014251.D\data.ms



Abundance Scan 1069 (12.084 min): BN014251.D\data.ms (-1024) (-)





Abundance Scan 1086 (12.160 min): BN014249.D\data.ms (-1016) (-)




2-Methylnaphthalene
Concen: 1.51 ng
RT: 12.160 min Scan# 1086
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Tgt Ion:142 Resp: 64164
Ion Ratio Lower Upper
142 100
141 87.8 70.6 106.0
115 34.8 28.6 42.8

Abundance Scan 1086 (12.160 min): BN014251.D\data.ms




Abundance Scan 1086 (12.160 min): BN014251.D\data.ms (-1064) (-)






Abundance Scan 1309 (14.346 min): BN014249.D\data.ms (-1293) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion	Ratio	Lower	Upper
164	100		
162	100.8	79.8	119.8
160	46.2	36.6	54.8




Abundance Scan 1426 (15.831 min): BN014249.D\data.ms (-1418) (-)

2,4,6-Tribromophenol
Concen: 1.19 ng
RT: 15.831 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Tgt Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	32.2	25.4	38.0






Abundance Scan 1200 (12.963 min): BN014249.D\data.ms (-1195) (-)




2-Fluorobiphenyl
Concen: 1.75 ng
RT: 12.963 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	172	Resp:	80020
Ion	Ratio	Lower	Upper
172	100		
171	35.0	28.6	42.8
170	24.0	19.7	29.5






Abundance Scan 1287 (14.067 min): BN014249.D\data.ms (-1216) (-)



Acenaphthylene
Concen: 1.39 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Tgt Ion:	152	Resp:	83056
Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.7	23.5
153	13.0	10.6	16.0





Abundance Scan 1314 (14.410 min): BN014249.D\data.ms (-1310) (-)

Acenaphthene
Concen: 1.63 ng
RT: 14.410 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	154	Resp:	60785
Ion	Ratio	Lower	Upper
154	100		
153	109.5	88.1	132.1
152	54.0	43.8	65.6





Abundance Scan 1314 (14.410 min): BN014251.D\data.ms (-1306) (-)




Abundance Scan 1392 (15.399 min): BN014249.D\data.ms (-1315) (-)

Fluorene
Concen: 1.58 ng
RT: 15.399 min Scan# 1392
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Tgt Ion:	166	Resp:	76767
Ion	Ratio	Lower	Upper
166	100		
165	98.3	78.0	117.0
167	13.5	10.6	16.0




Abundance Scan 1392 (15.399 min): BN014251.D\data.ms (-1376) (-)






Abundance Scan 1529 (17.091 min): BN014249.D\data.ms (-1514) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.091 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.9	5.5	8.3






Abundance Scan 1397 (15.463 min): BN014249.D\data.ms (-1324) (-)


4,6-Dinitro-2-methylphenol
Concen: 1.94 ng
RT: 15.463 min Scan# 1397
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19



Tgt Ion	Ratio	Lower	Upper
198	100		
51	47.6	76.0	114.0#
105	48.4	53.5	80.3#





Abundance Scan 1463 (16.288 min): BN014249.D\data.ms (-1424) (-)




4-Bromophenyl-phenylether
Concen: 1.53 ng
RT: 16.288 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	248	Resp:	24324
Ion Ratio	Lower	Upper	
248	100		
250	97.1	76.9	115.3
141	49.0	41.4	62.0



Abundance Scan 1463 (16.288 min): BN014251.D\data.ms (-1455) (-)





Abundance Scan 1472 (16.398 min): BN014249.D\data.ms (-1422) (-)




Hexachlorobenzene
Concen: 1.78 ng
RT: 16.398 min Scan# 1472
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Tgt Ion:	284	Resp:	30948
Ion Ratio	Lower	Upper	
284	100		
142	32.7	26.2	39.4
249	29.7	23.8	35.8




Abundance Scan 1472 (16.398 min): BN014251.D\data.ms (-1464) (-)





Abundance Scan 1485 (16.556 min): BN014249.D\data.ms (-1429) (-)

Atrazine
Concen: 1.09 ng
RT: 16.556 min Scan# 1485
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion	Ratio	Lower	Upper
200	100		
173	24.3	20.9	31.3
215	48.3	39.2	58.8





Abundance Scan 1485 (16.556 min): BN014251.D\data.ms (-1444) (-)




Abundance Scan 1500 (16.738 min): BN014249.D\data.ms (-1429) (-)

Pentachlorophenol
Concen: 1.26 ng
RT: 16.738 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.5	50.3	75.5
268	63.9	51.8	77.8



Abundance Scan 1500 (16.738 min): BN014251.D\data.ms (-1492) (-)






Abundance Scan 1532 (17.128 min): BN014249.D\data.ms (-1525) (-)




Phenanthrene
Concen: 1.66 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	178	Resp:	128211
Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	15.3	12.3	18.5






Abundance Scan 1540 (17.225 min): BN014249.D\data.ms (-1526) (-)



Anthracene
Concen: 1.42 ng
RT: 17.225 min Scan# 1540
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Tgt Ion:	178	Resp:	107326
Ion	Ratio	Lower	Upper
178	100		
176	18.4	15.0	22.4
179	15.3	12.3	18.5






Abundance Scan 1830 (19.126 min): BN014249.D\data.ms (-1797) (-)

Fluoranthene-d10
Concen: 1.56 ng
RT: 19.126 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:	212	Resp:	134923
Ion Ratio	Lower	Upper	
212	100		
106	11.4	9.1	13.7
104	6.4	5.1	7.7



Abundance Scan 1836 (19.156 min): BN014249.D\data.ms (-1798) (-)

Fluoranthene
Concen: 1.64 ng
RT: 19.156 min Scan# 1836
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Tgt Ion:	202	Resp:	148658
Ion Ratio	Lower	Upper	
202	100		
101	9.7	7.8	11.8
203	17.3	13.8	20.6






Abundance Scan 2127 (21.279 min): BN014249.D\data.ms (-2129) (-)

Chrysene-d12
Concen: 0.40 ng
RT: 21.279 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	240	Resp:	30282
Ion	Ratio	Lower	Upper
240	100		
120	12.3	9.4	14.2
236	26.8	21.7	32.5






Abundance Scan 1909 (19.518 min): BN014249.D\data.ms (-1839) (-)


Pyrene
Concen: 1.51 ng
RT: 19.518 min Scan# 1909
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19



Tgt Ion:	202	Resp:	146482
Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.5	24.7
203	17.4	13.9	20.9





Abundance Scan 1952 (19.732 min): BN014249.D\data.ms (-1934) (-)




Terphenyl-d14
Concen: 1.55 ng
RT: 19.732 min Scan# 1952
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.3	6.1	9.1
122	9.8	8.2	12.4



Abundance Scan 1952 (19.732 min): BN014251.D\data.ms (-1922) (-)





Abundance Scan 2126 (21.268 min): BN014249.D\data.ms (-2120) (-)




Benzo(a)anthracene
Concen: 1.44 ng
RT: 21.268 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.3	21.0	31.6
229	19.7	16.0	24.0



Abundance Scan 2126 (21.268 min): BN014251.D\data.ms (-2107) (-)






Abundance Scan 2131 (21.322 min): BN014249.D\data.ms (-2128) (-)




Chrysene
Concen: 1.62 ng
RT: 21.322 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.6	23.0	34.4
229	19.8	16.0	24.0






Abundance Scan 2121 (21.215 min): BN014249.D\data.ms (-2128) (-)



Bis(2-ethylhexyl)phthalate
Concen: 0.73 ng
RT: 21.215 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19


Tgt Ion	Ratio	Lower	Upper
149	100		
167	27.1	22.3	33.5
279	4.2	4.0	6.0






Abundance Scan 3216 (25.769 min): BN014249.D\data.ms (-3185) (-)

Indeno(1,2,3-cd)pyrene
Concen: 1.52 ng
RT: 25.769 min Scan# 3216
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.5	17.9	26.9
277	24.9	19.9	29.9






Abundance Scan 2558 (23.517 min): BN014249.D\data.ms (-2536) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.517 min Scan# 2558
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.8	19.8	29.8
265	49.1	35.3	52.9





Abundance Scan 2362 (22.846 min): BN014249.D\data.ms (-2362) (-)




Benzo(b)fluoranthene
Concen: 1.80 ng
RT: 22.850 min Scan# 2363
Delta R.T. 0.003 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Tgt Ion:	252	Resp:	149844
Ion	Ratio	Lower	Upper
252	100		
253	22.4	18.8	28.2
125	9.3	8.6	13.0


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Abundance Scan 2363 (22.850 min): BN014251.D\data.ms (-2304) (-)





Abundance Scan 2375 (22.891 min): BN014249.D\data.ms (-2375) (-)




Benzo(k)fluoranthene
Concen: 1.88 ng
RT: 22.894 min Scan# 2376
Delta R.T. 0.003 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Tgt Ion:	252	Resp:	150906
Ion	Ratio	Lower	Upper
252	100		
253	22.3	19.0	28.4
125	9.4	8.8	13.2




Abundance Scan 2376 (22.894 min): BN014251.D\data.ms (-2287) (-)






Abundance Scan 2530 (23.421 min): BN014249.D\data.ms (-3239) (-)

Benzo(a)pyrene
Concen: 1.67 ng
RT: 23.421 min Scan# 2530
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	252	Resp:	127184
Ion Ratio	Lower	Upper	
252	100		
253	22.7	19.9	29.9
125	10.9	10.4	15.6






Abundance Scan 3221 (25.787 min): BN014249.D\data.ms (-3240) (-)


Dibenzo(a,h)anthracene
Concen: 1.85 ng
RT: 25.787 min Scan# 3221
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19



Tgt Ion:	278	Resp:	146103
Ion Ratio	Lower	Upper	
278	100		
139	14.6	11.9	17.9
279	24.0	19.9	29.9




Abundance Scan 3416 (26.454 min): BN014249.D\data.ms (-3392) (-)



Benzo(g,h,i)perylene
Concen: 1.84 ng
RT: 26.454 min Scan# 3416
Delta R.T. 0.000 min
Lab File: BN014251.D
Acq: 09 Apr 2021 13:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	276	Resp:	147214
Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.3	28.9
138	19.2	15.7	23.5

