

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
 Data File : BN014252.D
 Acq On : 09 Apr 2021 13:55
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Apr 09 14:25:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 12:42:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	7620	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	26822	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	15094	0.40	ng	0.00
19) Phenanthrene-d10	17.091	188	32400	0.40	ng	0.00
29) Chrysene-d12	21.279	240	34647	0.40	ng	0.00
36) Perylene-d12	23.517	264	29349	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	53836	2.65	ng	0.00
5) Phenol-d6	6.879	99	70157	2.52	ng	0.00
8) Nitrobenzene-d5	8.857	82	53976	2.91	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	134265	2.95	ng	0.00
14) 2,4,6-Tribromophenol	15.831	330	20312	2.60	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	174475	3.32	ng	0.00
27) Fluoranthene-d10	19.126	212	301966	3.14	ng	0.00
31) Terphenyl-d14	19.732	244	230784	2.97	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.279	88	26796	3.18	ng	98
3) n-Nitrosodimethylamine	3.593	42	16788	3.61	ng	# 94
6) bis(2-Chloroethyl)ether	7.145	93	56270	3.01	ng	100
9) Naphthalene	10.543	128	208290	3.10	ng	99
10) Hexachlorobutadiene	10.834	225	40774	3.12	ng	# 99
12) 2-Methylnaphthalene	12.160	142	142343	2.93	ng	99
16) Acenaphthylene	14.054	152	199387	2.90	ng	100
17) Acenaphthene	14.410	154	137521	3.22	ng	99
18) Fluorene	15.399	166	174900	3.14	ng	99
20) 4,6-Dinitro-2-methylph...	15.463	198	13846	4.34	ng	# 56
21) 4-Bromophenyl-phenylether	16.288	248	53441	3.01	ng	97
22) Hexachlorobenzene	16.398	284	67873	3.51	ng	99
23) Atrazine	16.556	200	42436	2.43	ng	98
24) Pentachlorophenol	16.738	266	23250	2.89	ng	99
25) Phenanthrene	17.128	178	283094	3.29	ng	100
26) Anthracene	17.225	178	248955	2.97	ng	100
28) Fluoranthene	19.156	202	329317	3.27	ng	100
30) Pyrene	19.518	202	327909	2.96	ng	100
32) Benzo(a)anthracene	21.268	228	322013	2.93	ng	100
33) Chrysene	21.321	228	337627	3.16	ng	100
34) Bis(2-ethylhexyl)phtha...	21.215	149	107364	1.59	ng	98
35) Indeno(1,2,3-cd)pyrene	25.769	276	397033	3.04	ng	100
37) Benzo(b)fluoranthene	22.850	252	340937	3.56	ng	97
38) Benzo(k)fluoranthene	22.891	252	336743	3.64	ng	96
39) Benzo(a)pyrene	23.421	252	295469	3.37	ng	# 94
40) Dibenzo(a,h)anthracene	25.783	278	332487	3.66	ng	98
41) Benzo(g,h,i)perylene	26.457	276	334489	3.62	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
Data File : BN014252.D
Acq On : 09 Apr 2021 13:55
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Apr 09 14:25:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 12:42:56 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.716 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	152	Resp:	7620
Ion	Ratio	Lower	Upper
152	100		
150	155.0	124.8	187.2
115	60.7	48.6	73.0



1,4-Dioxane
Concen: 3.18 ng
RT: 3.279 min Scan# 62
Delta R.T. -0.008 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:	88	Resp:	26796
Ion	Ratio	Lower	Upper
88	100		
58	68.0	52.5	78.7
43	27.2	21.8	32.8









n-Nitrosodimethylamine
Concen: 3.61 ng
RT: 3.593 min Scan# 101
Delta R.T. -0.040 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 16788
Ion Ratio Lower Upper
42 100
74 133.7 111.2 166.8
44 5.2 12.5 18.7#



2-Fluorophenol
Concen: 2.65 ng
RT: 5.316 min Scan# 315
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion: 112 Resp: 53836
Ion Ratio Lower Upper
112 100
64 54.2 43.4 65.0
63 29.5 23.5 35.3









Phenol-d6
 Concen: 2.52 ng
 RT: 6.879 min Scan# 509
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	99	Resp:	70157
Ion	Ratio	Lower	Upper
99	100		
42	18.5	15.3	22.9
71	34.4	27.5	41.3




bis(2-Chloroethyl)ether
 Concen: 3.01 ng
 RT: 7.145 min Scan# 542
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55

Tgt Ion:	93	Resp:	56270
Ion	Ratio	Lower	Upper
93	100		
63	72.9	58.3	87.5
95	33.4	27.0	40.4






Abundance Scan 859 (10.492 min): BN014249.D\data.ms (-8547-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	136	Resp:	26822
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.0	4.8	7.2
68	5.5	4.4	6.6






Abundance Scan 730 (8.857 min): BN014249.D\data.ms (-72778)



Nitrobenzene-d5
Concen: 2.91 ng
RT: 8.857 min Scan# 730
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55


Tgt Ion:	82	Resp:	53976
Ion	Ratio	Lower	Upper
82	100		
128	39.4	29.7	44.5
54	42.3	36.5	54.7





Abundance Scan 863 (10.543 min): BN014249.D\data.ms (-857) (-)

Naphthalene
Concen: 3.10 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55


Instrument :
BNA_N
ClientSampled :
SSTDICC3.2



Tgt Ion:	128	Resp:	208290
Ion	Ratio	Lower	Upper
128	100		
129	10.9	9.0	13.6
127	12.7	10.5	15.7




Abundance Scan 863 (10.543 min): BN014252.D\data.ms (-855) (-)





Abundance Scan 886 (10.834 min): BN014249.D\data.ms (-886) (-)


Hexachlorobutadiene
Concen: 3.12 ng
RT: 10.834 min Scan# 886
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Tgt Ion:	225	Resp:	40774
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.1	51.9	77.9



Abundance Scan 886 (10.834 min): BN014252.D\data.ms (-878) (-)









2-Methylnaphthalene-d10
 Concen: 2.95 ng
 RT: 12.084 min Scan# 1069
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:152 Resp: 134265
 Ion Ratio Lower Upper
 152 100
 151 20.1 16.1 24.1



2-Methylnaphthalene
 Concen: 2.93 ng
 RT: 12.160 min Scan# 1086
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55

Tgt Ion:142 Resp: 142343
 Ion Ratio Lower Upper
 142 100
 141 88.1 70.6 106.0
 115 34.8 28.6 42.8









Scan 1309 (14.346 min): BN014249.D\data.ms (-1293) (-)
 Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.346 min Scan# 1309
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	164	Resp:	15094
Ion Ratio	Lower	Upper	
164	100		
162	100.5	79.8	119.8
160	45.7	36.6	54.8




Scan 1426 (15.831 min): BN014249.D\data.ms (-1418) (-)
 2,4,6-Tribromophenol
 Concen: 2.60 ng
 RT: 15.831 min Scan# 1426
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55

Tgt Ion:	330	Resp:	20312
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	32.2	25.4	38.0






Abundance Scan 1200 (12.963 min): BN014249.D\data.ms (-1195) (-)




2-Fluorobiphenyl
Concen: 3.32 ng
RT: 12.963 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	172	Resp:	174475
Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.6	42.8
170	24.3	19.7	29.5






Abundance Scan 1287 (14.067 min): BN014249.D\data.ms (-1216) (-)



Acenaphthylene
Concen: 2.90 ng
RT: 14.054 min Scan# 1286
Delta R.T. -0.013 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55


Tgt Ion:	152	Resp:	199387
Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.7	23.5
153	13.0	10.6	16.0






Abundance Scan 1314 (14.410 min): BN014249.D\data.ms (-1310) (-)

Acenaphthene
Concen: 3.22 ng
RT: 14.410 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	154	Resp:	137521
Ion Ratio	Lower	Upper	
154	100		
153	109.2	88.1	132.1
152	53.6	43.8	65.6






Abundance Scan 1392 (15.399 min): BN014249.D\data.ms (-1395) (-)

Fluorene
Concen: 3.14 ng
RT: 15.399 min Scan# 1392
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Tgt Ion:	166	Resp:	174900
Ion Ratio	Lower	Upper	
166	100		
165	98.4	78.0	117.0
167	13.5	10.6	16.0





Abundance Scan 1529 (17.091 min): BN014249.D\data.ms (-1514) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.091 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:	188	Resp:	32400
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	7.2	5.5	8.3





Abundance Scan 1529 (17.091 min): BN014252.D\data.ms (-1513) (-)




Abundance Scan 1397 (15.463 min): BN014249.D\data.ms (-1324) (-)

4,6-Dinitro-2-methylphenol
Concen: 4.34 ng
RT: 15.463 min Scan# 1397
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55


Tgt Ion:	198	Resp:	13846
Ion Ratio	Lower	Upper	
198	100		
51	40.7	76.0	114.0#
105	45.5	53.5	80.3#



Abundance Scan 1397 (15.463 min): BN014252.D\data.ms (-1389) (-)






Abundance Scan 1463 (16.288 min): BN014249.D\data.ms (-1427) (-)




4-Bromophenyl-phenylether
Concen: 3.01 ng
RT: 16.288 min Scan# 1463
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	248	Resp:	53441
Ion Ratio	Lower	Upper	
248	100		
250	97.7	76.9	115.3
141	48.4	41.4	62.0







Abundance Scan 1472 (16.398 min): BN014249.D\data.ms (-1422) (-)



Hexachlorobenzene
Concen: 3.51 ng
RT: 16.398 min Scan# 1472
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:	284	Resp:	67873
Ion Ratio	Lower	Upper	
284	100		
142	32.5	26.2	39.4
249	29.4	23.8	35.8









Scan 1485 (16.556 min): BN014249.D\data.ms (-1423)
 Atrazine
 Concen: 2.43 ng
 RT: 16.556 min Scan# 1485
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	200	Resp:	42436
Ion Ratio	Lower	Upper	
200	100		
173	25.2	20.9	31.3
215	48.0	39.2	58.8




Scan 1500 (16.738 min): BN014249.D\data.ms (-1424)
 Pentachlorophenol
 Concen: 2.89 ng
 RT: 16.738 min Scan# 1500
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55

Tgt Ion:	266	Resp:	23250
Ion Ratio	Lower	Upper	
266	100		
264	62.6	50.3	75.5
268	64.0	51.8	77.8





Abundance Scan 1532 (17.128 min): BN014249.D\data.ms (-1525) (-)




Phenanthrene
Concen: 3.29 ng
RT: 17.128 min Scan# 1532
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampled :
SSTDICC3.2


Tgt Ion:	178	Resp:	283094
Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.3	12.3	18.5



Abundance Scan 1532 (17.128 min): BN014252.D\data.ms (-1507) (-)





Abundance Scan 1540 (17.225 min): BN014249.D\data.ms (-1526) (-)




Anthracene
Concen: 2.97 ng
RT: 17.225 min Scan# 1540
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55


Tgt Ion:	178	Resp:	248955
Ion	Ratio	Lower	Upper
178	100		
176	18.5	15.0	22.4
179	15.3	12.3	18.5



Abundance Scan 1540 (17.225 min): BN014252.D\data.ms (-1532) (-)





Abundance Scan 1830 (19.126 min): BN014249.D\data.ms (-1797) (-)




Fluoranthene-d10
Concen: 3.14 ng
RT: 19.126 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	212	Resp:	301966
Ion	Ratio	Lower	Upper
212	100		
106	11.6	9.1	13.7
104	6.5	5.1	7.7



Abundance Scan 1830 (19.126 min): BN014252.D\data.ms (-1797) (-)





Abundance Scan 1836 (19.156 min): BN014249.D\data.ms (-1797) (-)




Fluoranthene
Concen: 3.27 ng
RT: 19.156 min Scan# 1836
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:	202	Resp:	329317
Ion	Ratio	Lower	Upper
202	100		
101	9.8	7.8	11.8
203	17.4	13.8	20.6




Abundance Scan 1836 (19.156 min): BN014252.D\data.ms (-1796) (-)






Abundance Scan 2127 (21.279 min): BN014249.D\data.ms (-2129) (-)

Chrysene-d12
Concen: 0.40 ng
RT: 21.279 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	240	Resp:	34647
Ion Ratio	Lower	Upper	
240	100		
120	11.7	9.4	14.2
236	26.8	21.7	32.5






Abundance Scan 1909 (19.518 min): BN014249.D\data.ms (-1839) (-)


Pyrene
Concen: 2.96 ng
RT: 19.518 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55



Tgt Ion:	202	Resp:	327909
Ion Ratio	Lower	Upper	
202	100		
200	20.7	16.5	24.7
203	17.5	13.9	20.9






Abundance Scan 1952 (19.732 min): BN014249.D\data.ms (-1934) (-)




Terphenyl-d14
Concen: 2.97 ng
RT: 19.732 min Scan# 1952
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.4	6.1	9.1
122	9.9	8.2	12.4






Abundance Scan 2126 (21.268 min): BN014249.D\data.ms (-2120) (-)




Benzo(a)anthracene
Concen: 2.93 ng
RT: 21.268 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.4	21.0	31.6
229	19.9	16.0	24.0






Abundance Scan 2131 (21.322 min): BN014249.D\data.ms (-2128) (-)




Chrysene
Concen: 3.16 ng
RT: 21.321 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	228	Resp:	337627
Ion	Ratio	Lower	Upper
228	100		
226	29.1	23.0	34.4
229	19.9	16.0	24.0






Abundance Scan 2121 (21.215 min): BN014249.D\data.ms (-2128) (-)



Bis(2-ethylhexyl)phthalate
Concen: 1.59 ng
RT: 21.215 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55


Tgt Ion:	149	Resp:	107364
Ion	Ratio	Lower	Upper
149	100		
167	27.1	22.3	33.5
279	4.0	4.0	6.0





Abundance Scan 3216 (25.769 min): BN014249.D\data.ms (-3185) (-)

Indeno(1,2,3-cd)pyrene
Concen: 3.04 ng
RT: 25.769 min Scan# 3216
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.5	17.9	26.9
277	24.9	19.9	29.9




Abundance Scan 3216 (25.769 min): BN014252.D\data.ms (-3187) (-)





Abundance Scan 2558 (23.517 min): BN014249.D\data.ms (-2539) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.517 min Scan# 2558
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.8	19.8	29.8
265	44.2	35.3	52.9



Abundance Scan 2558 (23.517 min): BN014252.D\data.ms (-2514) (-)





Abundance Scan 2362 (22.846 min): BN014249.D\data.ms (-2362) (-)




Benzo(b)fluoranthene
Concen: 3.56 ng
RT: 22.850 min Scan# 2363
Delta R.T. 0.003 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:	252	Resp:	340937
Ion	Ratio	Lower	Upper
252	100		
253	22.2	18.8	28.2
125	8.9	8.6	13.0


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Abundance Scan 2363 (22.850 min): BN014252.D\data.ms (-2304) (-)





Abundance Scan 2375 (22.891 min): BN014249.D\data.ms (-2375) (-)




Benzo(k)fluoranthene
Concen: 3.64 ng
RT: 22.891 min Scan# 2375
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:	252	Resp:	336743
Ion	Ratio	Lower	Upper
252	100		
253	21.8	19.0	28.4
125	9.1	8.8	13.2




Abundance Scan 2375 (22.891 min): BN014252.D\data.ms (-2287) (-)






Abundance Scan 2530 (23.421 min): BN014249.D\data.ms (-329) (-)

Benzo(a)pyrene
Concen: 3.37 ng
RT: 23.421 min Scan# 2530
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	19.9	29.9
125	10.0	10.4	15.6#






Abundance Scan 3221 (25.787 min): BN014249.D\data.ms (-320) (-)


Dibenzo(a,h)anthracene
Concen: 3.66 ng
RT: 25.783 min Scan# 3220
Delta R.T. -0.003 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55



Tgt Ion	Ratio	Lower	Upper
278	100		
139	14.8	11.9	17.9
279	23.6	19.9	29.9




Abundance Scan 3416 (26.454 min): BN014249.D\data.ms (-3352) (-)



Benzo(g,h,i)perylene
Concen: 3.62 ng
RT: 26.457 min Scan# 3417
Delta R.T. 0.003 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Abundance Scan 3417 (26.457 min): BN014252.D\data.ms



Tgt Ion:276 Resp: 334489
Ion Ratio Lower Upper
276 100
277 23.7 19.3 28.9
138 19.0 15.7 23.5

Abundance Scan 3417 (26.457 min): BN014252.D\data.ms (-3358) (-)

