

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
 Data File : BN014254.D
 Acq On : 09 Apr 2021 15:13
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN040921

Quant Time: Apr 09 15:49:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 15:06:44 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.510	0.552	-8.2	77	0.00
3	n-Nitrosodimethylamine	0.238	0.230	3.4	72	0.00
4 S	2-Fluorophenol	0.953	1.009	-5.9	70	0.00
5 S	Phenol-d6	1.190	1.239	-4.1	70	0.00
6	bis(2-Chloroethyl)ether	0.966	1.021	-5.7	70	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	71	0.00
8 S	Nitrobenzene-d5	0.254	0.254	0.0	69	0.00
9	Naphthalene	0.995	1.062	-6.7	71	0.00
10	Hexachlorobutadiene	0.199	0.215	-8.0	72	0.00
11 SURR	2-Methylnaphthalene-d10	0.624	0.658	-5.4	71	0.00
12	2-Methylnaphthalene	0.666	0.703	-5.6	71	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	71	0.00
14 S	2,4,6-Tribromophenol	0.156	0.131	16.0	57	0.00
15 S	2-Fluorobiphenyl	1.467	1.549	-5.6	72	0.00
16	Acenaphthylene	1.536	1.488	3.1	69	-0.01
17	Acenaphthene	1.125	1.164	-3.5	71	0.00
18	Fluorene	1.428	1.475	-3.3	70	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
20	4,6-Dinitro-2-methylphenol	0.047	0.043	8.5	65	0.00
21	4-Bromophenyl-phenylether	0.205	0.200	2.4	64	0.00
22	Hexachlorobenzene	0.261	0.280	-7.3	71	0.00
23	Atrazine	0.153	0.143	6.5	66	0.00
24	Pentachlorophenol	0.080	0.071	11.3	63	0.00
25	Phenanthrene	1.107	1.158	-4.6	69	0.00
26	Anthracene	0.917	0.925	-0.9	67	0.00
27 SURR	Fluoranthene-d10	1.160	1.208	-4.1	68	0.00
28	Fluoranthene	1.269	1.318	-3.9	68	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
30	Pyrene	1.260	1.331	-5.6	68	0.00
31 S	Terphenyl-d14	0.884	0.958	-8.4	69	0.00
32	Benzo(a)anthracene	1.140	1.210	-6.1	70	0.00
33	Chrysene	1.236	1.343	-8.7	69	0.00
34	Bis(2-ethylhexyl)phthalate	0.409	0.388	5.1	71	0.00
35	Indeno(1,2,3-cd)pyrene	1.457	1.456	0.1	69	0.00
36 I	Perylene-d12	1.000	1.000	0.0	67	0.00
37	Benzo(b)fluoranthene	1.423	1.572	-10.5	68	0.00
38	Benzo(k)fluoranthene	1.424	1.524	-7.0	69	0.00
39 C	Benzo(a)pyrene	1.200	1.243	-3.6	68	0.00
40	Dibenzo(a,h)anthracene	1.381	1.456	-5.4	69	0.00
41	Benzo(g,h,i)perylene	1.401	1.564	-11.6	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN040921