

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
 Data File : BN014247.D
 Acq On : 09 Apr 2021 10:56
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Apr 09 12:51:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 12:42:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	6952	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	25552	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	13776	0.40	ng	0.00
19) Phenanthrene-d10	17.092	188	31507	0.40	ng	0.00
29) Chrysene-d12	21.282	240	31908	0.40	ng	0.00
36) Perylene-d12	23.520	264	30420	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.317	112	1896	0.10	ng	0.00
5) Phenol-d6	6.879	99	2266	0.09	ng	0.00
8) Nitrobenzene-d5	8.857	82	1692	0.10	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	3799	0.09	ng	0.00
14) 2,4,6-Tribromophenol	15.831	330	534	0.08	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	4829	0.10	ng	0.00
27) Fluoranthene-d10	19.124	212	9059	0.10	ng	0.00
31) Terphenyl-d14	19.729	244	7121	0.10	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.279	88	1206	0.16	ng	# 81
6) bis(2-Chloroethyl)ether	7.145	93	1788	0.10	ng	98
9) Naphthalene	10.543	128	6373	0.10	ng	95
10) Hexachlorobutadiene	10.834	225	1279	0.10	ng	# 98
12) 2-Methylnaphthalene	12.155	142	4195	0.09	ng	# 95
16) Acenaphthylene	14.067	152	4936	0.08	ng	100
17) Acenaphthene	14.410	154	3777	0.10	ng	98
18) Fluorene	15.399	166	4830	0.10	ng	99
20) 4,6-Dinitro-2-methylph...	15.475	198	330	0.11	ng	# 1
21) 4-Bromophenyl-phenylether	16.289	248	1529	0.09	ng	# 92
22) Hexachlorobenzene	16.398	284	1958	0.10	ng	96
23) Atrazine	16.556	200	1232	0.07	ng	95
25) Phenanthrene	17.128	178	9041	0.11	ng	99
26) Anthracene	17.226	178	6854	0.08	ng	99
28) Fluoranthene	19.154	202	10177	0.10	ng	99
30) Pyrene	19.521	202	11180	0.11	ng	100
32) Benzo(a)anthracene	21.271	228	8720	0.09	ng	99
33) Chrysene	21.324	228	9576	0.10	ng	99
34) Bis(2-ethylhexyl)phtha...	21.218	149	4741	0.08	ng	96
35) Indeno(1,2,3-cd)pyrene	25.772	276	11648	0.10	ng	98
37) Benzo(b)fluoranthene	22.849	252	10215	0.10	ng	# 80
38) Benzo(k)fluoranthene	22.894	252	10230	0.11	ng	# 80
39) Benzo(a)pyrene	23.424	252	8527	0.09	ng	# 72
40) Dibenzo(a,h)anthracene	25.789	278	9718	0.10	ng	# 86
41) Benzo(g,h,i)perylene	26.457	276	9972	0.10	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
 Data File : BN014247.D
 Acq On : 09 Apr 2021 10:56
 Operator : CG/JU
 Sample : SSTDIC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.1


Quant Time: Apr 09 12:51:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 12:42:56 2021
 Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.716 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1





Tgt Ion: 152 Resp: 6952
Ion Ratio Lower Upper
152 100
150 155.2 124.8 187.2
115 61.8 48.6 73.0




#2
1,4-Dioxane
Concen: 0.16 ng
RT: 3.279 min Scan# 62
Delta R.T. -0.008 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56



Tgt Ion: 88 Resp: 1206
Ion Ratio Lower Upper
88 100
58 51.3 52.5 78.7#
43 39.2 21.8 32.8#









#4
2-Fluorophenol
Concen: 0.10 ng
RT: 5.317 min Scan# 315
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:112 Resp: 1896
Ion Ratio Lower Upper
112 100
64 53.7 43.4 65.0
63 31.1 23.5 35.3



#5
Phenol-d6
Concen: 0.09 ng
RT: 6.879 min Scan# 509
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion: 99 Resp: 2266
Ion Ratio Lower Upper
99 100
42 20.9 15.3 22.9
71 34.3 27.5 41.3









#6
bis(2-Chloroethyl)ether
Concen: 0.10 ng
RT: 7.145 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 93 Resp: 1788
Ion Ratio Lower Upper
93 100
63 71.0 58.3 87.5
95 34.2 27.0 40.4



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion: 136 Resp: 25552
Ion Ratio Lower Upper
136 100
137 11.4 9.0 13.4
54 6.1 4.8 7.2
68 5.7 4.4 6.6









#8
Nitrobenzene-d5
Concen: 0.10 ng
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 82 Resp: 1692
Ion Ratio Lower Upper
82 100
128 42.3 29.7 44.5
54 50.8 36.5 54.7



#9
Naphthalene
Concen: 0.10 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion: 128 Resp: 6373
Ion Ratio Lower Upper
128 100
129 13.2 9.0 13.6
127 14.9 10.5 15.7









#10
Hexachlorobutadiene
Concen: 0.10 ng
RT: 10.834 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:225 Resp: 1279
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 51.9 77.9



#11
2-Methylnaphthalene-d10
Concen: 0.09 ng
RT: 12.084 min Scan# 1069
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:152 Resp: 3799
Ion Ratio Lower Upper
152 100
151 20.4 16.1 24.1









#12
2-Methylnaphthalene
Concen: 0.09 ng
RT: 12.155 min Scan# 1085
Delta R.T. -0.004 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:142 Resp: 4195
Ion Ratio Lower Upper
142 100
141 90.6 70.6 106.0
115 43.0 28.6 42.8#



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:164 Resp: 13776
Ion Ratio Lower Upper
164 100
162 100.1 79.8 119.8
160 45.6 36.6 54.8









#14
2,4,6-Tribromophenol
Concen: 0.08 ng
RT: 15.831 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:330 Resp: 534
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.8 25.4 38.0



#15
2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.963 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:172 Resp: 4829
Ion Ratio Lower Upper
172 100
171 36.7 28.6 42.8
170 26.4 19.7 29.5









#16
Acenaphthylene
Concen: 0.08 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.7	23.5
153	13.5	10.6	16.0



#17
Acenaphthene
Concen: 0.10 ng
RT: 14.410 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion	Ratio	Lower	Upper
154	100		
153	111.6	88.1	132.1
152	56.0	43.8	65.6









#18
Fluorene
Concen: 0.10 ng
RT: 15.399 min Scan# 1392
Delta R.T. 0.000 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:166 Resp: 4830
Ion Ratio Lower Upper
166 100
165 96.2 78.0 117.0
167 13.8 10.6 16.0



#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.001 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:188 Resp: 31507
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.2 5.5 8.3









#20
4,6-Dinitro-2-methylphenol
Concen: 0.11 ng
RT: 15.475 min Scan# 1398
Delta R.T. 0.013 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:198 Resp: 330
Ion Ratio Lower Upper
198 100
51 248.4 76.0 114.0#
105 133.2 53.5 80.3#



#21
4-Bromophenyl-phenylether
Concen: 0.09 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.001 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:248 Resp: 1529
Ion Ratio Lower Upper
248 100
250 99.4 76.9 115.3
141 62.3 41.4 62.0#









#22
Hexachlorobenzene
Concen: 0.10 ng
RT: 16.398 min Scan# 1472
Delta R.T. 0.001 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:284 Resp: 1958
Ion Ratio Lower Upper
284 100
142 34.9 26.2 39.4
249 32.0 23.8 35.8



#23
Atrazine
Concen: 0.07 ng
RT: 16.556 min Scan# 1485
Delta R.T. 0.001 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:200 Resp: 1232
Ion Ratio Lower Upper
200 100
173 29.7 20.9 31.3
215 51.8 39.2 58.8









#25
Phenanthrene
Concen: 0.11 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.001 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 9041
Ion Ratio Lower Upper
178 100
176 19.2 15.3 22.9
179 15.9 12.3 18.5



#26
Anthracene
Concen: 0.08 ng
RT: 17.226 min Scan# 1540
Delta R.T. 0.001 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:178 Resp: 6854
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.4
179 14.8 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.10 ng

RT: 19.124 min Scan# 1828

Delta R.T. -0.002 min

Lab File: BN014247.D

Acq: 09 Apr 2021 10:56





Tgt Ion:	212	Resp:	9059
Ion Ratio	Lower	Upper	
212	100		
106	11.5	9.1	13.7
104	6.6	5.1	7.7

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1



#28

Fluoranthene

Concen: 0.10 ng




RT: 19.154 min Scan# 1834


Delta R.T. -0.002 min

Lab File: BN014247.D

Acq: 09 Apr 2021 10:56

Tgt Ion:	202	Resp:	10177
Ion Ratio	Lower	Upper	
202	100		
101	9.3	7.8	11.8
203	17.2	13.8	20.6





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.282 min Scan# 2126

Delta R.T. 0.003 min

Lab File: BN014247.D

Acq: 09 Apr 2021 10:56

Tgt Ion:240 Resp: 31908

Ion Ratio Lower Upper

240 100

120 10.7 9.4 14.2





236 26.9 21.7 32.5

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1



#30

Pyrene

Concen: 0.11 ng

RT: 19.521 min Scan# 1908

Delta R.T. 0.003 min

Lab File: BN014247.D

Acq: 09 Apr 2021 10:56




Tgt Ion:202 Resp: 11180


Ion Ratio Lower Upper

202 100

200 20.6 16.5 24.7

203 17.2 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.10 ng
 RT: 19.729 min Scan# 1950
 Delta R.T. -0.002 min
 Lab File: BN014247.D
 Acq: 09 Apr 2021 10:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 7121
 Ion Ratio Lower Upper
 244 100
 212 8.7 6.1 9.1
 122 12.8 8.2 12.4#



#32
 Benzo(a)anthracene
 Concen: 0.09 ng
 RT: 21.271 min Scan# 2125
 Delta R.T. 0.003 min
 Lab File: BN014247.D
 Acq: 09 Apr 2021 10:56

Tgt Ion:228 Resp: 8720
 Ion Ratio Lower Upper
 228 100
 226 26.8 21.0 31.6
 229 20.7 16.0 24.0









#33
Chrysene
Concen: 0.10 ng
RT: 21.324 min Scan# 2130
Delta R.T. 0.003 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:228	Resp:	9576
Ion Ratio	Lower	Upper
228	100	
226	29.4	23.0 34.4
229	20.8	16.0 24.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.08 ng
RT: 21.218 min Scan# 2120
Delta R.T. 0.003 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:149	Resp:	4741
Ion Ratio	Lower	Upper
149	100	
167	25.7	22.3 33.5
279	4.6	4.0 6.0









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.10 ng
RT: 25.772 min Scan# 3215
Delta R.T. 0.003 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	276	Resp:	11648
Ion Ratio	Lower	Upper	
276	100		
138	24.7	17.9	26.9
277	24.7	19.9	29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.520 min Scan# 2557
Delta R.T. 0.003 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:	264	Resp:	30420
Ion Ratio	Lower	Upper	
264	100		
260	24.9	19.8	29.8
265	46.2	35.3	52.9









#37
Benzo(b)fluoranthene
Concen: 0.10 ng
RT: 22.849 min Scan# 2361
Delta R.T. 0.003 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 10215
Ion Ratio Lower Upper
252 100
253 32.2 18.8 28.2#
125 20.3 8.6 13.0#



#38
Benzo(k)fluoranthene
Concen: 0.11 ng
RT: 22.894 min Scan# 2374
Delta R.T. 0.003 min
Lab File: BN014247.D
Acq: 09 Apr 2021 10:56

Tgt Ion:252 Resp: 10230
Ion Ratio Lower Upper
252 100
253 31.7 19.0 28.4#
125 21.2 8.8 13.2#





#39

Benzo(a)pyrene

Concen: 0.09 ng

RT: 23.424 min Scan# 2529

Delta R.T. 0.003 min

Lab File: BN014247.D

Acq: 09 Apr 2021 10:56





Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1

Tgt Ion:252	Resp:	8527
Ion Ratio	Lower	Upper
252	100	
253	37.3	19.9 29.9#
125	27.0	10.4 15.6#



#40

Dibenzo(a,h)anthracene

Concen: 0.10 ng




RT: 25.789 min Scan# 3220


Delta R.T. 0.003 min

Lab File: BN014247.D

Acq: 09 Apr 2021 10:56

Tgt Ion:278	Resp:	9718
Ion Ratio	Lower	Upper
278	100	
139	18.7	11.9 17.9#
279	33.0	19.9 29.9#





#41
 Benzo(g,h,i)perylene
 Concen: 0.10 ng
 RT: 26.457 min Scan# 3415
 Delta R.T. 0.003 min
 Lab File: BN014247.D
 Acq: 09 Apr 2021 10:56

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	9972
Ion	Ratio	Lower	Upper
276	100		
277	27.1	19.3	28.9
138	23.1	15.7	23.5

