

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
 Data File : BN014248.D
 Acq On : 09 Apr 2021 11:32
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Apr 09 12:50:22 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 12:42:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	7627	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	28100	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	15531	0.40	ng	0.00
19) Phenanthrene-d10	17.092	188	34557	0.40	ng	0.00
29) Chrysene-d12	21.282	240	34779	0.40	ng	0.00
36) Perylene-d12	23.520	264	33156	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	3676	0.18	ng	0.00
5) Phenol-d6	6.879	99	4499	0.16	ng	0.00
8) Nitrobenzene-d5	8.857	82	3292	0.17	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	8356	0.18	ng	0.00
14) 2,4,6-Tribromophenol	15.831	330	1107	0.14	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	10898	0.20	ng	0.00
27) Fluoranthene-d10	19.124	212	19279	0.19	ng	0.00
31) Terphenyl-d14	19.729	244	15060	0.19	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.279	88	2000	0.24	ng	94
3) n-Nitrosodimethylamine	3.633	42	661	0.14	ng	# 80
6) bis(2-Chloroethyl)ether	7.153	93	3635	0.19	ng	98
9) Naphthalene	10.543	128	13531	0.19	ng	99
10) Hexachlorobutadiene	10.834	225	2788	0.20	ng	# 98
12) 2-Methylnaphthalene	12.160	142	8837	0.17	ng	99
16) Acenaphthylene	14.067	152	10637	0.15	ng	99
17) Acenaphthene	14.409	154	8276	0.19	ng	100
18) Fluorene	15.399	166	10406	0.18	ng	100
20) 4,6-Dinitro-2-methylph...	15.463	198	706	0.21	ng	# 39
21) 4-Bromophenyl-phenylether	16.289	248	3406	0.18	ng	98
22) Hexachlorobenzene	16.398	284	4341	0.21	ng	100
23) Atrazine	16.556	200	2426	0.13	ng	99
24) Pentachlorophenol	16.739	266	1156	0.13	ng	98
25) Phenanthrene	17.128	178	18249	0.20	ng	100
26) Anthracene	17.226	178	14530	0.16	ng	99
28) Fluoranthene	19.153	202	20753	0.19	ng	99
30) Pyrene	19.521	202	21586	0.19	ng	100
32) Benzo(a)anthracene	21.271	228	18112	0.16	ng	99
33) Chrysene	21.314	228	20385	0.19	ng	96
34) Bis(2-ethylhexyl)phtha...	21.218	149	6329	0.09	ng	# 97
35) Indeno(1,2,3-cd)pyrene	25.772	276	25635	0.20	ng	99
37) Benzo(b)fluoranthene	22.849	252	20955	0.19	ng	# 94
38) Benzo(k)fluoranthene	22.894	252	22555	0.22	ng	# 93
39) Benzo(a)pyrene	23.421	252	17936	0.18	ng	# 90
40) Dibenzo(a,h)anthracene	25.786	278	21228	0.21	ng	95
41) Benzo(g,h,i)perylene	26.457	276	20884	0.20	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
Data File : BN014248.D
Acq On : 09 Apr 2021 11:32
Operator : CG/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Apr 09 12:50:22 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 12:42:56 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.716 min Scan# 613
 Delta R.T. 0.000 min
 Lab File: BN014248.D
 Acq: 09 Apr 2021 11:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:152 Resp: 7627
 Ion Ratio Lower Upper
 152 100
 150 157.1 124.8 187.2
 115 60.2 48.6 73.0



#2
 1,4-Dioxane
 Concen: 0.24 ng
 RT: 3.279 min Scan# 62
 Delta R.T. -0.008 min
 Lab File: BN014248.D
 Acq: 09 Apr 2021 11:32

Tgt Ion: 88 Resp: 2000
 Ion Ratio Lower Upper
 88 100
 58 58.4 52.5 78.7
 43 27.3 21.8 32.8









#3
n-Nitrosodimethylamine
Concen: 0.14 ng
RT: 3.633 min Scan# 106
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion: 42 Resp: 661
Ion Ratio Lower Upper
42 100
74 163.2 111.2 166.8
44 22.2 12.5 18.7#



#4
2-Fluorophenol
Concen: 0.18 ng
RT: 5.316 min Scan# 315
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion: 112 Resp: 3676
Ion Ratio Lower Upper
112 100
64 53.0 43.4 65.0
63 29.5 23.5 35.3









#5
Phenol-d6
Concen: 0.16 ng
RT: 6.879 min Scan# 509
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	99	Resp:	4499
Ion	Ratio	Lower	Upper
99	100		
42	18.5	15.3	22.9
71	34.2	27.5	41.3



#6
bis(2-Chloroethyl)ether
Concen: 0.19 ng
RT: 7.153 min Scan# 543
Delta R.T. 0.008 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:	93	Resp:	3635
Ion	Ratio	Lower	Upper
93	100		
63	73.6	58.3	87.5
95	35.4	27.0	40.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	136	Resp:	28100
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.4
54	6.6	4.8	7.2
68	5.8	4.4	6.6



#8
Nitrobenzene-d5
Concen: 0.17 ng
RT: 8.857 min Scan# 730
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:	82	Resp:	3292
Ion Ratio	Lower	Upper	
82	100		
128	39.8	29.7	44.5
54	47.6	36.5	54.7









#9
Naphthalene
Concen: 0.19 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	128	Resp:	13531
Ion Ratio	Lower	Upper	
128	100		
129	11.6	9.0	13.6
127	13.3	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.20 ng
RT: 10.834 min Scan# 886
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:	225	Resp:	2788
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.0	51.9	77.9









#11
2-Methylnaphthalene-d10
Concen: 0.18 ng
RT: 12.084 min Scan# 1069
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 8356
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.17 ng
RT: 12.160 min Scan# 1086
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:142 Resp: 8837
Ion Ratio Lower Upper
142 100
141 89.1 70.6 106.0
115 36.9 28.6 42.8









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:164 Resp: 15531
Ion Ratio Lower Upper
164 100
162 100.6 79.8 119.8
160 46.5 36.6 54.8



#14
2,4,6-Tribromophenol
Concen: 0.14 ng
RT: 15.831 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:330 Resp: 1107
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 31.8 25.4 38.0









#15
2-Fluorobiphenyl
Concen: 0.20 ng
RT: 12.963 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	172	Resp:	10898
Ion Ratio	Lower	Upper	
172	100		
171	35.6	28.6	42.8
170	24.8	19.7	29.5



#16
Acenaphthylene
Concen: 0.15 ng
RT: 14.067 min Scan# 1287
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:	152	Resp:	10637
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.7	23.5
153	13.0	10.6	16.0









#17
Acenaphthene
Concen: 0.19 ng
RT: 14.409 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:154 Resp: 8276
Ion Ratio Lower Upper
154 100
153 110.1 88.1 132.1
152 54.7 43.8 65.6



#18
Fluorene
Concen: 0.18 ng
RT: 15.399 min Scan# 1392
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:166 Resp: 10406
Ion Ratio Lower Upper
166 100
165 97.5 78.0 117.0
167 13.5 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:188	Resp:	34557
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	7.3	5.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.21 ng
RT: 15.463 min Scan# 1397
Delta R.T. -0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:198	Resp:	706
Ion Ratio	Lower	Upper
198	100	
51	165.7	76.0
105	101.7	53.5









#21
4-Bromophenyl-phenylether
Concen: 0.18 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:248 Resp: 3406
Ion Ratio Lower Upper
248 100
250 97.2 76.9 115.3
141 55.2 41.4 62.0



#22
Hexachlorobenzene
Concen: 0.21 ng
RT: 16.398 min Scan# 1472
Delta R.T. 0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:284 Resp: 4341
Ion Ratio Lower Upper
284 100
142 32.4 26.2 39.4
249 29.8 23.8 35.8









#23
Atrazine
Concen: 0.13 ng
RT: 16.556 min Scan# 1485
Delta R.T. 0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:200 Resp: 2426
Ion Ratio Lower Upper
200 100
173 26.7 20.9 31.3
215 48.6 39.2 58.8



#24
Pentachlorophenol
Concen: 0.13 ng
RT: 16.739 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:266 Resp: 1156
Ion Ratio Lower Upper
266 100
264 64.8 50.3 75.5
268 65.8 51.8 77.8









#25
Phenanthrene
Concen: 0.20 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:178 Resp: 18249
Ion Ratio Lower Upper
178 100
176 19.4 15.3 22.9
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.16 ng
RT: 17.226 min Scan# 1540
Delta R.T. 0.000 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:178 Resp: 14530
Ion Ratio Lower Upper
178 100
176 18.4 15.0 22.4
179 15.2 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.19 ng
RT: 19.124 min Scan# 1828
Delta R.T. -0.002 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:212 Resp: 19279
Ion Ratio Lower Upper
212 100
106 11.6 9.1 13.7
104 6.5 5.1 7.7



#28
Fluoranthene
Concen: 0.19 ng
RT: 19.153 min Scan# 1834
Delta R.T. -0.002 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:202 Resp: 20753
Ion Ratio Lower Upper
202 100
101 9.5 7.8 11.8
203 17.0 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.282 min Scan# 2126
Delta R.T. 0.003 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:240 Resp: 34779
Ion Ratio Lower Upper
240 100
120 9.9 9.4 14.2
236 26.8 21.7 32.5



#30
Pyrene
Concen: 0.19 ng
RT: 19.521 min Scan# 1908
Delta R.T. 0.003 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:202 Resp: 21586
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.6 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.19 ng
 RT: 19.729 min Scan# 1950
 Delta R.T. -0.002 min
 Lab File: BN014248.D
 Acq: 09 Apr 2021 11:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:244	Resp:	15060
Ion Ratio	Lower	Upper
244	100	
212	7.8	6.1 9.1
122	11.5	8.2 12.4



#32
 Benzo(a)anthracene
 Concen: 0.16 ng
 RT: 21.271 min Scan# 2125
 Delta R.T. 0.003 min
 Lab File: BN014248.D
 Acq: 09 Apr 2021 11:32

Tgt Ion:228	Resp:	18112
Ion Ratio	Lower	Upper
228	100	
226	26.1	21.0 31.6
229	20.5	16.0 24.0









#33
Chrysene
Concen: 0.19 ng
RT: 21.314 min Scan# 2129
Delta R.T. -0.008 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	228	Resp:	20385
Ion Ratio	Lower	Upper	
228	100		
226	31.3	23.0	34.4
229	19.2	16.0	24.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.09 ng
RT: 21.218 min Scan# 2120
Delta R.T. 0.003 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:	149	Resp:	6329
Ion Ratio	Lower	Upper	
149	100		
167	26.7	22.3	33.5
279	7.8	4.0	6.0









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.20 ng
RT: 25.772 min Scan# 3215
Delta R.T. 0.003 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:276 Resp: 25635
Ion Ratio Lower Upper
276 100
138 23.8 17.9 26.9
277 25.0 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.520 min Scan# 2557
Delta R.T. 0.003 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:264 Resp: 33156
Ion Ratio Lower Upper
264 100
260 24.5 19.8 29.8
265 43.8 35.3 52.9









#37
Benzo(b)fluoranthene
Concen: 0.19 ng
RT: 22.849 min Scan# 2361
Delta R.T. 0.003 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:252 Resp: 20955
Ion Ratio Lower Upper
252 100
253 26.1 18.8 28.2
125 13.9 8.6 13.0#



#38
Benzo(k)fluoranthene
Concen: 0.22 ng
RT: 22.894 min Scan# 2374
Delta R.T. 0.003 min
Lab File: BN014248.D
Acq: 09 Apr 2021 11:32

Tgt Ion:252 Resp: 22555
Ion Ratio Lower Upper
252 100
253 26.4 19.0 28.4
125 14.8 8.8 13.2#





#39

Benzo(a)pyrene

Concen: 0.18 ng

RT: 23.421 min Scan# 2528

Delta R.T. -0.001 min

Lab File: BN014248.D

Acq: 09 Apr 2021 11:32





Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2

Tgt Ion:252	Resp:	17936
Ion Ratio	Lower	Upper
252	100	
253	28.7	19.9 29.9
125	18.4	10.4 15.6#



#40

Dibenzo(a,h)anthracene

Concen: 0.21 ng




RT: 25.786 min Scan# 3219


Delta R.T. -0.001 min

Lab File: BN014248.D

Acq: 09 Apr 2021 11:32

Tgt Ion:278	Resp:	21228
Ion Ratio	Lower	Upper
278	100	
139	16.2	11.9 17.9
279	27.6	19.9 29.9





#41
 Benzo(g,h,i)perylene
 Concen: 0.20 ng
 RT: 26.457 min Scan# 3415
 Delta R.T. 0.003 min
 Lab File: BN014248.D
 Acq: 09 Apr 2021 11:32

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	276	Resp:	20884
Ion Ratio	Lower	Upper	
276	100		
277	25.8	19.3	28.9
138	20.4	15.7	23.5

