

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
 Data File : BN014252.D
 Acq On : 09 Apr 2021 13:55
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Apr 09 14:25:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 12:42:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	7620	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	26822	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	15094	0.40	ng	0.00
19) Phenanthrene-d10	17.091	188	32400	0.40	ng	0.00
29) Chrysene-d12	21.279	240	34647	0.40	ng	0.00
36) Perylene-d12	23.517	264	29349	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	53836	2.65	ng	0.00
5) Phenol-d6	6.879	99	70157	2.52	ng	0.00
8) Nitrobenzene-d5	8.857	82	53976	2.91	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	134265	2.95	ng	0.00
14) 2,4,6-Tribromophenol	15.831	330	20312	2.60	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	174475	3.32	ng	0.00
27) Fluoranthene-d10	19.126	212	301966	3.14	ng	0.00
31) Terphenyl-d14	19.732	244	230784	2.97	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.279	88	26796	3.18	ng	98
3) n-Nitrosodimethylamine	3.593	42	16788	3.61	ng	# 94
6) bis(2-Chloroethyl)ether	7.145	93	56270	3.01	ng	100
9) Naphthalene	10.543	128	208290	3.10	ng	99
10) Hexachlorobutadiene	10.834	225	40774	3.12	ng	# 99
12) 2-Methylnaphthalene	12.160	142	142343	2.93	ng	99
16) Acenaphthylene	14.054	152	199387	2.90	ng	100
17) Acenaphthene	14.410	154	137521	3.22	ng	99
18) Fluorene	15.399	166	174900	3.14	ng	99
20) 4,6-Dinitro-2-methylph...	15.463	198	13846	4.34	ng	# 56
21) 4-Bromophenyl-phenylether	16.288	248	53441	3.01	ng	97
22) Hexachlorobenzene	16.398	284	67873	3.51	ng	99
23) Atrazine	16.556	200	42436	2.43	ng	98
24) Pentachlorophenol	16.738	266	23250	2.89	ng	99
25) Phenanthrene	17.128	178	283094	3.29	ng	100
26) Anthracene	17.225	178	248955	2.97	ng	100
28) Fluoranthene	19.156	202	329317	3.27	ng	100
30) Pyrene	19.518	202	327909	2.96	ng	100
32) Benzo(a)anthracene	21.268	228	322013	2.93	ng	100
33) Chrysene	21.321	228	337627	3.16	ng	100
34) Bis(2-ethylhexyl)phtha...	21.215	149	107364	1.59	ng	98
35) Indeno(1,2,3-cd)pyrene	25.769	276	397033	3.04	ng	100
37) Benzo(b)fluoranthene	22.850	252	340937	3.56	ng	97
38) Benzo(k)fluoranthene	22.891	252	336743	3.64	ng	96
39) Benzo(a)pyrene	23.421	252	295469	3.37	ng	# 94
40) Dibenzo(a,h)anthracene	25.783	278	332487	3.66	ng	98
41) Benzo(g,h,i)perylene	26.457	276	334489	3.62	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040921\
Data File : BN014252.D
Acq On : 09 Apr 2021 13:55
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Apr 09 14:25:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 12:42:56 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.716 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 7620
Ion Ratio Lower Upper
152 100
150 155.0 124.8 187.2
115 60.7 48.6 73.0



#2
1,4-Dioxane
Concen: 3.18 ng
RT: 3.279 min Scan# 62
Delta R.T. -0.008 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion: 88 Resp: 26796
Ion Ratio Lower Upper
88 100
58 68.0 52.5 78.7
43 27.2 21.8 32.8









#3
n-Nitrosodimethylamine
Concen: 3.61 ng
RT: 3.593 min Scan# 101
Delta R.T. -0.040 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 16788
Ion Ratio Lower Upper
42 100
74 133.7 111.2 166.8
44 5.2 12.5 18.7#



#4
2-Fluorophenol
Concen: 2.65 ng
RT: 5.316 min Scan# 315
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion: 112 Resp: 53836
Ion Ratio Lower Upper
112 100
64 54.2 43.4 65.0
63 29.5 23.5 35.3









#5
Phenol-d6
Concen: 2.52 ng
RT: 6.879 min Scan# 509
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	99	Resp:	70157
Ion	Ratio	Lower	Upper
99	100		
42	18.5	15.3	22.9
71	34.4	27.5	41.3



#6
bis(2-Chloroethyl)ether
Concen: 3.01 ng
RT: 7.145 min Scan# 542
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:	93	Resp:	56270
Ion	Ratio	Lower	Upper
93	100		
63	72.9	58.3	87.5
95	33.4	27.0	40.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	26822
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.0	4.8	7.2
68	5.5	4.4	6.6



#8
Nitrobenzene-d5
Concen: 2.91 ng
RT: 8.857 min Scan# 730
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:	82	Resp:	53976
Ion	Ratio	Lower	Upper
82	100		
128	39.4	29.7	44.5
54	42.3	36.5	54.7









#9
Naphthalene
Concen: 3.10 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 208290
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 12.7 10.5 15.7



#10
Hexachlorobutadiene
Concen: 3.12 ng
RT: 10.834 min Scan# 886
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:225 Resp: 40774
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.9 77.9









#11
2-Methylnaphthalene-d10
Concen: 2.95 ng
RT: 12.084 min Scan# 1069
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 134265
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1



#12
2-Methylnaphthalene
Concen: 2.93 ng
RT: 12.160 min Scan# 1086
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:142 Resp: 142343
Ion Ratio Lower Upper
142 100
141 88.1 70.6 106.0
115 34.8 28.6 42.8









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 15094
Ion Ratio Lower Upper
164 100
162 100.5 79.8 119.8
160 45.7 36.6 54.8



#14
2,4,6-Tribromophenol
Concen: 2.60 ng
RT: 15.831 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:330 Resp: 20312
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 32.2 25.4 38.0









#15
2-Fluorobiphenyl
Concen: 3.32 ng
RT: 12.963 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 174475
Ion Ratio Lower Upper
172 100
171 35.2 28.6 42.8
170 24.3 19.7 29.5



#16
Acenaphthylene
Concen: 2.90 ng
RT: 14.054 min Scan# 1286
Delta R.T. -0.013 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:152 Resp: 199387
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 13.0 10.6 16.0









#17
Acenaphthene
Concen: 3.22 ng
RT: 14.410 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 137521
Ion Ratio Lower Upper
154 100
153 109.2 88.1 132.1
152 53.6 43.8 65.6



#18
Fluorene
Concen: 3.14 ng
RT: 15.399 min Scan# 1392
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:166 Resp: 174900
Ion Ratio Lower Upper
166 100
165 98.4 78.0 117.0
167 13.5 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.091 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 32400
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.2 5.5 8.3



#20
4,6-Dinitro-2-methylphenol
Concen: 4.34 ng
RT: 15.463 min Scan# 1397
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:198 Resp: 13846
Ion Ratio Lower Upper
198 100
51 40.7 76.0 114.0#
105 45.5 53.5 80.3#









#21
4-Bromophenyl-phenylether
Concen: 3.01 ng
RT: 16.288 min Scan# 1463
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampled :
SSTDICC3.2


Tgt Ion:248 Resp: 53441
Ion Ratio Lower Upper
248 100
250 97.7 76.9 115.3
141 48.4 41.4 62.0



#22
Hexachlorobenzene
Concen: 3.51 ng
RT: 16.398 min Scan# 1472
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:284 Resp: 67873
Ion Ratio Lower Upper
284 100
142 32.5 26.2 39.4
249 29.4 23.8 35.8









#23
Atrazine
Concen: 2.43 ng
RT: 16.556 min Scan# 1485
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 42436
Ion Ratio Lower Upper
200 100
173 25.2 20.9 31.3
215 48.0 39.2 58.8



#24
Pentachlorophenol
Concen: 2.89 ng
RT: 16.738 min Scan# 1500
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:266 Resp: 23250
Ion Ratio Lower Upper
266 100
264 62.6 50.3 75.5
268 64.0 51.8 77.8









#25
Phenanthrene
Concen: 3.29 ng
RT: 17.128 min Scan# 1532
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 283094
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.3 12.3 18.5



#26
Anthracene
Concen: 2.97 ng
RT: 17.225 min Scan# 1540
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:178 Resp: 248955
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.4
179 15.3 12.3 18.5









#27
Fluoranthene-d10
Concen: 3.14 ng
RT: 19.126 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 301966
Ion Ratio Lower Upper
212 100
106 11.6 9.1 13.7
104 6.5 5.1 7.7



#28
Fluoranthene
Concen: 3.27 ng
RT: 19.156 min Scan# 1836
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:202 Resp: 329317
Ion Ratio Lower Upper
202 100
101 9.8 7.8 11.8
203 17.4 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.279 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 34647
Ion Ratio Lower Upper
240 100
120 11.7 9.4 14.2
236 26.8 21.7 32.5



#30
Pyrene
Concen: 2.96 ng
RT: 19.518 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:202 Resp: 327909
Ion Ratio Lower Upper
202 100
200 20.7 16.5 24.7
203 17.5 13.9 20.9









#31
 Terphenyl-d14
 Concen: 2.97 ng
 RT: 19.732 min Scan# 1952
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 230784
 Ion Ratio Lower Upper
 244 100
 212 7.4 6.1 9.1
 122 9.9 8.2 12.4



#32
 Benzo(a)anthracene
 Concen: 2.93 ng
 RT: 21.268 min Scan# 2126
 Delta R.T. -0.000 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55

Tgt Ion:228 Resp: 322013
 Ion Ratio Lower Upper
 228 100
 226 26.4 21.0 31.6
 229 19.9 16.0 24.0









#33
Chrysene
Concen: 3.16 ng
RT: 21.321 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 337627
Ion Ratio Lower Upper
228 100
226 29.1 23.0 34.4
229 19.9 16.0 24.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.59 ng
RT: 21.215 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:149 Resp: 107364
Ion Ratio Lower Upper
149 100
167 27.1 22.3 33.5
279 4.0 4.0 6.0









#35
Indeno(1,2,3-cd)pyrene
Concen: 3.04 ng
RT: 25.769 min Scan# 3216
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:276 Resp: 397033
Ion Ratio Lower Upper
276 100
138 22.5 17.9 26.9
277 24.9 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.517 min Scan# 2558
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:264 Resp: 29349
Ion Ratio Lower Upper
264 100
260 24.8 19.8 29.8
265 44.2 35.3 52.9









#37
Benzo(b)fluoranthene
Concen: 3.56 ng
RT: 22.850 min Scan# 2363
Delta R.T. 0.003 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 340937
Ion Ratio Lower Upper
252 100
253 22.2 18.8 28.2
125 8.9 8.6 13.0



#38
Benzo(k)fluoranthene
Concen: 3.64 ng
RT: 22.891 min Scan# 2375
Delta R.T. -0.000 min
Lab File: BN014252.D
Acq: 09 Apr 2021 13:55

Tgt Ion:252 Resp: 336743
Ion Ratio Lower Upper
252 100
253 21.8 19.0 28.4
125 9.1 8.8 13.2





#39

Benzo(a)pyrene

Concen: 3.37 ng

RT: 23.421 min Scan# 2530

Delta R.T. -0.000 min

Lab File: BN014252.D

Acq: 09 Apr 2021 13:55





Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2

Tgt Ion:252	Resp:	295469
Ion Ratio	Lower	Upper
252	100	
253	22.1	19.9 29.9
125	10.0	10.4 15.6#



#40

Dibenzo(a,h)anthracene

Concen: 3.66 ng




RT: 25.783 min Scan# 3220


Delta R.T. -0.003 min

Lab File: BN014252.D

Acq: 09 Apr 2021 13:55


Tgt Ion:278	Resp:	332487
Ion Ratio	Lower	Upper
278	100	
139	14.8	11.9 17.9
279	23.6	19.9 29.9





#41
 Benzo(g,h,i)perylene
 Concen: 3.62 ng
 RT: 26.457 min Scan# 3417
 Delta R.T. 0.003 min
 Lab File: BN014252.D
 Acq: 09 Apr 2021 13:55

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2



Tgt Ion:276 Resp: 334489
 Ion Ratio Lower Upper
 276 100
 277 23.7 19.3 28.9
 138 19.0 15.7 23.5

