


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041221\
 Data File : BN014258.D
 Acq On : 12 Apr 2021 13:21
 Operator : CG/JU
 Sample : DFTPP30
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP030

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN041221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 12 17:52:31 2021



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.2	55842	PASS
68	69	0.00	2	1.9	1283	PASS
69	198	0.00	100	38.1	68161	PASS
70	69	0.00	2	0.8	567	PASS
127	198	10	80	47.9	85786	PASS
197	198	0.00	2	0.4	642	PASS
198	198	100	100	100.0	179008	PASS
199	198	5	9	6.8	12145	PASS
275	198	10	60	25.4	45408	PASS
365	198	1	100	3.5	6354	PASS
441	198	0.01	100	14.5	25997	PASS
442	198	50	100	93.3	167082	PASS
443	442	15	24	18.9	31581	PASS

Scan 2450 (17.498 min): BN014258.D
DFTPP30

m/z	Abundance
39.00	609.0
39.90	755.0
41.00	554.0
41.90	504.0
43.00	1185.0
44.00	2094.0
45.00	430.0
48.90	530.0
50.90	358.0
55.00	621.0
56.90	481.0
60.00	371.0
69.00	332.0
73.00	678.0
83.90	534.0
85.80	364.0
207.00	854.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

m/z	Abundance
36.90	393.0
38.00	981.0
39.00	5721.0
40.00	1127.0
41.00	902.0
42.00	501.0
43.00	1237.0
44.00	2261.0
45.00	583.0
49.00	1012.0
50.00	17984.0
51.00	68904.0
52.00	3590.0
53.00	453.0
55.00	1009.0
56.00	2392.0
57.00	5428.0
58.00	392.0
59.90	540.0
60.90	1000.0
61.90	1106.0
63.00	2972.0
63.90	426.0
65.00	1645.0
68.00	1508.0
68.90	81336.0
69.90	570.0
73.00	1492.0
74.00	8575.0
75.00	13811.0
76.00	4795.0
77.00	98912.0
78.00	7230.0
79.00	6612.0
80.00	4809.0
80.90	6769.0
82.00	1617.0
82.90	1530.0
83.80	913.0
84.90	1286.0
85.90	2079.0
86.90	1056.0
87.90	453.0
90.90	1482.0
92.00	1711.0
93.00	11119.0
93.90	949.0
94.90	363.0
96.00	671.0
97.00	360.0
97.90	8624.0
99.00	7691.0
99.90	743.0
101.00	3725.0
102.90	1228.0
103.90	2385.0
105.00	2404.0
105.90	740.0
107.00	28144.0
108.00	4450.0
108.90	811.0
109.90	49040.0
110.90	7971.0
112.00	976.0
112.90	352.0
116.00	1397.0
116.90	26872.0
118.00	2113.0
118.80	333.0
119.90	424.0
121.90	2004.0
123.00	2889.0
123.90	1409.0
124.90	1319.0
126.00	462.0
127.00	97768.0
128.00	7707.0
129.00	36752.0
130.00	3278.0
130.90	781.0
131.90	403.0
133.00	328.0
133.90	985.0
134.90	3238.0
135.90	1316.0
137.00	1977.0
137.80	388.0
140.00	566.0
140.90	4666.0
142.00	1572.0
142.90	1067.0
143.90	329.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

144.90	306.0
145.90	710.0
147.00	2331.0
147.90	6664.0
148.90	1678.0
149.90	437.0
151.10	693.0
151.90	388.0
152.90	1479.0
154.00	1064.0
155.00	2954.0
156.00	4307.0
157.00	1000.0
157.90	689.0
158.90	785.0
159.90	1260.0
161.00	2443.0
161.90	708.0
162.90	364.0
165.00	1840.0
166.00	1335.0
167.00	10379.0
168.00	8198.0
168.90	805.0
170.10	376.0
170.90	334.0
171.90	891.0
173.00	1098.0
174.00	2093.0
175.00	3689.0
175.90	807.0
177.00	1955.0
177.90	630.0
178.90	7311.0
179.90	5570.0
181.00	2565.0
182.00	480.0
183.80	594.0
185.00	3364.0
186.00	26496.0
187.00	7894.0
188.00	1050.0
188.90	1686.0
190.00	335.0
190.90	891.0
191.90	2312.0
193.00	2975.0
193.90	597.0
196.00	5001.0
196.80	1010.0
197.90	185984.0
199.00	12995.0
199.90	1072.0
201.50	780.0
202.90	1412.0
204.00	7255.0
205.00	12123.0
206.00	50776.0
207.00	7378.0
208.00	1856.0
208.90	739.0
210.10	827.0
211.00	2027.0
214.90	541.0
216.00	1002.0
217.00	14444.0
217.90	1751.0
221.00	7376.0
221.90	1946.0
222.90	3139.0
224.00	28528.0
225.00	7501.0
226.00	865.0
227.00	11918.0
227.90	1872.0
228.90	2456.0
231.00	946.0
234.00	693.0
234.90	773.0
235.90	584.0
237.00	1268.0
239.00	420.0
239.90	477.0
240.90	821.0
242.00	1446.0
243.00	1620.0
244.00	20952.0
245.00	2996.0
246.00	4062.0
247.00	1085.0
249.00	844.0
252.90	579.0
254.00	629.0
255.00	108144.0
256.00	16568.0
257.00	1110.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

258.00	6400.0
258.90	987.0
265.00	2411.0
265.90	391.0
271.90	358.0
273.00	3232.0
274.00	8209.0
275.00	47200.0
276.00	6541.0
277.00	3575.0
277.90	852.0
282.90	533.0
283.90	313.0
285.00	623.0
292.90	1063.0
296.00	14416.0
297.00	1757.0
303.00	1505.0
304.00	396.0
314.00	742.0
315.00	1479.0
316.00	844.0
320.90	441.0
323.00	4656.0
324.00	777.0
326.90	912.0
327.90	362.0
331.90	324.0
332.90	358.0
334.00	2622.0
335.00	739.0
341.10	471.0
346.00	934.0
352.00	1336.0
353.00	1019.0
354.00	1192.0
364.90	5983.0
365.90	1010.0
372.00	2507.0
372.90	474.0
383.00	721.0
401.90	798.0
403.00	1086.0
403.90	409.0
420.90	1072.0
422.00	971.0
423.00	7507.0
424.00	2563.0
425.00	363.0
441.00	21992.0
442.00	131328.0
443.00	24592.0
444.00	2900.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

m/z	Abundance
37.00	381.0
38.00	890.0
39.00	5236.0
39.90	1021.0
41.00	810.0
42.00	531.0
43.00	1251.0
44.00	2188.0
45.00	463.0
48.90	1044.0
50.00	17688.0
51.00	65320.0
52.00	3319.0
52.90	311.0
55.00	1020.0
56.00	2534.0
57.00	4982.0
58.00	349.0
59.90	483.0
60.90	1112.0
62.00	1071.0
63.00	2865.0
63.90	407.0
65.00	1703.0
67.00	331.0
68.00	1511.0
68.90	80664.0
69.90	648.0
73.00	1382.0
74.00	9131.0
75.00	13435.0
76.00	4835.0
77.00	98856.0
78.00	6842.0
79.00	6102.0
80.00	4589.0
81.00	6625.0
81.90	1932.0
83.00	1434.0
83.90	854.0
85.00	1361.0
85.90	2264.0
87.00	834.0
88.00	384.0
90.90	1689.0
91.90	1793.0
92.90	11015.0
93.90	813.0
95.90	651.0
96.90	321.0
98.00	9114.0
99.00	7764.0
99.90	752.0
101.00	3289.0
102.90	1140.0
103.90	2456.0
104.90	2393.0
105.90	598.0
107.00	29224.0
108.00	4505.0
109.00	798.0
109.90	51144.0
111.00	8494.0
111.90	1021.0
112.90	422.0
115.90	1499.0
117.00	27904.0
117.90	2121.0
118.90	355.0
120.00	449.0
121.90	1926.0
122.90	3277.0
123.90	1570.0
124.90	1311.0
126.00	522.0
127.00	100192.0
128.00	7590.0
129.00	38440.0
130.00	3439.0
130.90	681.0
132.00	528.0
132.90	416.0
133.90	1208.0
135.00	3333.0
135.90	1100.0
136.90	2052.0
137.90	397.0
139.90	444.0
140.90	4763.0
142.00	1437.0
143.00	1224.0
143.90	301.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

145.90	1009.0
147.00	3029.0
148.00	6694.0
148.90	1668.0
150.00	455.0
151.10	751.0
151.90	436.0
152.90	1684.0
153.90	1322.0
155.00	2963.0
156.00	4632.0
157.00	988.0
157.90	948.0
158.90	833.0
159.90	1704.0
160.90	2361.0
162.00	687.0
164.00	339.0
164.90	2025.0
165.90	1666.0
167.00	11964.0
168.00	8526.0
168.90	1029.0
169.90	338.0
170.90	562.0
171.90	938.0
172.90	1255.0
174.00	2249.0
175.00	4306.0
176.00	1071.0
176.90	1888.0
177.90	707.0
178.90	8400.0
180.00	6160.0
181.00	2505.0
181.90	538.0
183.90	531.0
185.00	3967.0
186.00	29832.0
187.00	8817.0
188.00	792.0
188.90	1870.0
189.90	381.0
190.90	982.0
192.00	2913.0
193.00	3136.0
193.90	763.0
196.00	5643.0
196.90	918.0
198.00	219904.0
198.90	14492.0
200.00	1245.0
201.40	831.0
203.00	1589.0
204.00	8218.0
205.00	13912.0
206.00	57728.0
207.00	8048.0
208.00	2135.0
208.90	736.0
210.00	869.0
211.00	2155.0
212.00	345.0
214.90	579.0
216.00	1293.0
216.90	16520.0
218.00	2203.0
221.00	9004.0
221.80	2302.0
223.00	3412.0
224.00	32232.0
225.00	8629.0
226.00	886.0
227.00	14205.0
227.90	1917.0
228.90	3003.0
229.90	428.0
231.00	1110.0
232.90	362.0
233.90	962.0
235.00	1087.0
235.90	670.0
237.00	1423.0
238.90	516.0
240.00	515.0
240.90	744.0
242.00	1734.0
243.00	1775.0
244.00	26992.0
245.00	3593.0
246.00	4603.0
247.00	1263.0
248.90	771.0
253.00	455.0
254.00	799.0
255.00	131648.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

256.00	19744.0
257.00	1679.0
258.00	7779.0
259.00	1197.0
264.90	3322.0
265.80	472.0
271.90	408.0
273.00	4016.0
274.00	9763.0
275.00	52384.0
276.00	7765.0
277.00	4869.0
278.00	769.0
281.00	375.0
282.90	510.0
283.90	449.0
285.00	898.0
292.00	314.0
293.00	1033.0
295.00	307.0
296.00	16984.0
297.00	2116.0
302.00	389.0
303.00	1646.0
303.90	504.0
314.00	797.0
315.00	1631.0
315.90	1035.0
320.90	499.0
323.00	5248.0
324.00	906.0
327.00	1254.0
328.00	674.0
332.00	429.0
332.90	712.0
334.00	3453.0
335.00	986.0
341.00	720.0
345.90	1656.0
352.00	1615.0
353.00	1078.0
354.00	1819.0
355.00	388.0
365.00	7370.0
366.00	944.0
370.90	494.0
372.00	2745.0
372.90	597.0
383.00	834.0
390.00	356.0
401.90	1212.0
403.00	1790.0
403.90	784.0
421.00	1251.0
422.00	1159.0
423.00	9610.0
424.00	2678.0
425.00	433.0
441.00	30488.0
442.00	206912.0
443.00	42456.0
444.00	4024.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

m/z	Abundance
36.90	319.0
38.00	636.0
39.00	3290.0
40.00	972.0
41.00	748.0
42.00	553.0
43.00	1229.0
44.00	2025.0
44.90	410.0
48.90	893.0
50.00	9188.0
51.00	34376.0
52.00	2006.0
55.00	889.0
56.00	1192.0
57.00	2999.0
58.00	311.0
60.00	466.0
61.00	555.0
61.90	787.0
63.00	1636.0
63.90	325.0
65.00	868.0
67.90	831.0
68.90	43480.0
69.90	484.0
73.00	1092.0
74.00	4595.0
75.00	7193.0
76.00	2733.0
77.00	51112.0
78.00	3748.0
79.00	3582.0
79.90	2765.0
81.00	3673.0
81.90	1154.0
82.90	980.0
83.90	673.0
84.90	1020.0
85.90	1403.0
87.00	520.0
87.80	307.0
91.00	1098.0
91.90	1082.0
93.00	5930.0
94.00	512.0
94.90	330.0
95.90	410.0
97.00	436.0
98.00	4755.0
98.90	4127.0
99.90	333.0
100.90	2188.0
102.90	790.0
103.90	1434.0
104.90	1494.0
106.00	439.0
107.00	17112.0
108.00	2464.0
108.90	425.0
109.90	28904.0
110.90	4633.0
111.90	608.0
115.90	699.0
116.90	15489.0
117.90	1131.0
121.90	1169.0
122.90	1551.0
123.90	898.0
125.00	623.0
126.00	373.0
127.00	59400.0
128.00	4272.0
129.00	22088.0
130.00	1894.0
130.90	514.0
133.90	671.0
134.90	2071.0
136.00	657.0
136.90	1166.0
139.90	322.0
140.90	2847.0
141.90	839.0
142.90	590.0
145.90	509.0
146.90	1775.0
147.90	4510.0
148.90	1084.0
150.90	419.0
152.90	1006.0
153.90	774.0
155.00	1790.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

156.00	2437.0
157.00	503.0
157.90	647.0
159.00	451.0
160.00	977.0
160.90	1483.0
161.90	530.0
164.90	1209.0
165.90	803.0
167.00	6851.0
167.90	4654.0
168.90	625.0
170.90	312.0
171.80	480.0
172.90	766.0
173.90	1196.0
175.00	2497.0
176.00	619.0
176.90	1130.0
177.90	429.0
178.90	4611.0
179.90	3611.0
180.90	1798.0
183.90	437.0
185.00	2427.0
186.00	18272.0
187.00	4962.0
188.00	550.0
188.90	1128.0
190.90	840.0
192.00	1643.0
192.90	1784.0
194.00	377.0
196.00	3030.0
197.90	131136.0
198.90	8949.0
199.90	737.0
201.50	537.0
202.90	1017.0
204.00	5316.0
205.00	8472.0
206.00	35224.0
207.00	5255.0
208.00	1431.0
209.00	609.0
210.20	585.0
210.90	1296.0
215.00	432.0
216.00	721.0
217.00	9806.0
217.90	1190.0
221.00	5139.0
221.80	1282.0
223.00	1989.0
224.00	19760.0
225.00	4909.0
226.00	582.0
227.00	8837.0
227.90	1232.0
228.90	1738.0
231.00	866.0
233.80	508.0
234.90	649.0
236.00	412.0
237.00	778.0
239.00	484.0
241.00	528.0
242.00	1251.0
243.00	1311.0
244.00	16640.0
245.00	2290.0
245.90	3302.0
246.90	704.0
249.00	636.0
254.00	378.0
255.00	81472.0
256.00	11612.0
257.00	990.0
258.00	4698.0
258.90	810.0
264.90	2031.0
266.00	320.0
272.00	305.0
273.00	2478.0
274.00	7021.0
275.00	36640.0
276.00	4879.0
276.90	3082.0
277.90	504.0
280.90	326.0
283.00	421.0
285.00	595.0
292.90	782.0
296.00	10745.0
296.90	1294.0
303.00	1213.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030

313.90	441.0
314.90	1181.0
316.00	864.0
320.90	406.0
323.00	3599.0
324.00	646.0
326.90	773.0
328.00	368.0
334.00	2450.0
335.00	571.0
341.00	414.0
345.90	1047.0
352.00	1272.0
353.00	787.0
354.00	1291.0
355.00	305.0
365.00	5710.0
365.90	957.0
370.90	307.0
372.00	1910.0
372.90	495.0
382.90	621.0
402.00	1042.0
403.00	1462.0
404.00	584.0
421.00	938.0
421.90	1088.0
423.00	8548.0
424.00	2041.0
424.90	308.0
441.00	25512.0
442.00	163008.0
443.00	27696.0
444.00	1966.0

Instrument :
BNA_N
ClientSampleId :
DFTPP030