

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041223\
 Data File : BN024842.D
 Acq On : 12 Apr 2023 11:24
 Operator : CG/JU
 Sample : PB152043BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK043

Quant Time: Apr 13 02:47:27 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 13 02:44:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	9742	0.400	ng/u1	0.00
4) Naphthalene-d8	10.759	136	30437	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.589	164	17177	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.339	188	35739	0.400	ng/u1	0.00
17) Chrysene-d12	21.533	240	27539	0.400	ng/u1	0.00
23) Perylene-d12	23.985	264	28981	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.314	96	65860	6.191	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.348	152	13599	0.348	ng/u1	0.00
18) Fluoranthene-d10	19.364	212	27942	0.409	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041223\
Data File : BN024842.D
Acq On : 12 Apr 2023 11:24
Operator : CG/JU
Sample : PB152043BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK043

Quant Time: Apr 13 02:47:27 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 13 02:44:00 2023
Response via : Initial Calibration

