

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041223\
 Data File : BN024847.D
 Acq On : 12 Apr 2023 14:31
 Operator : CG/JU
 Sample : 02250-13
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHDY3

Quant Time: Apr 13 02:47:57 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 13 02:44:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.950	152	8075	0.400	ng/ul	0.00
4) Naphthalene-d8	10.763	136	25788	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.592	164	15057	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.337	188	31691	0.400	ng/ul	0.00
17) Chrysene-d12	21.536	240	25128	0.400	ng/ul	0.00
23) Perylene-d12	23.977	264	26544	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.313	96	51470	5.837	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.352	152	13606	0.411	ng/ul	0.00
18) Fluoranthene-d10	19.367	212	33863	0.543	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.351	88	898	0.092	ng/ul#	72


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041223\
Data File : BN024847.D
Acq On : 12 Apr 2023 14:31
Operator : CG/JU
Sample : 02250-13
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BHDY3


Quant Time: Apr 13 02:47:57 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 13 02:44:00 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.092 ng/ul
 RT: 3.351 min Scan# 40
 Delta R.T. -0.000 min
 Lab File: BN024847.D
 Acq: 12 Apr 2023 14:31

Instrument :
 BNA_N
 ClientSampleId :
 BHDY3



Tgt Ion:	88	Resp:	898
Ion	Ratio	Lower	Upper
88	100		
43	81.7	39.1	58.7#
58	67.6	65.7	98.5

