

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041224\
 Data File : BN030859.D
 Acq On : 12 Apr 2024 12:55
 Operator : MA/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDICCC0.4

Quant Time: Apr 12 15:49:34 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 12 15:48:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.418	-4.5	100	0.00
3	n-Nitrosodimethylamine	0.400	0.400	0.0	100	0.00
4 S	2-Fluorophenol	0.400	0.382	4.5	100	0.00
5 S	Phenol-d6	0.400	0.372	7.0	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.383	4.3	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.374	6.5	100	0.00
9	Naphthalene	0.400	0.388	3.0	100	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.384	4.0	100	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	100	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.400	0.353	11.8	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.385	3.8	100	0.00
16	Acenaphthylene	0.400	0.366	8.5	100	0.00
17	Acenaphthene	0.400	0.381	4.8	100	0.00
18	Fluorene	0.400	0.380	5.0	100	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.332	17.0	100	0.00
21	4-Bromophenyl-phenylether	0.400	0.381	4.8	100	0.00
22	Hexachlorobenzene	0.400	0.383	4.3	100	0.00
23	Atrazine	0.400	0.365	8.8	100	0.00
24	Pentachlorophenol	0.400	0.352	12.0	100	0.00
25	Phenanthrene	0.400	0.384	4.0	100	0.00
26	Anthracene	0.400	0.368	8.0	100	0.00
27 SURR	Fluoranthene-d10	0.400	0.381	4.8	100	0.00
28	Fluoranthene	0.400	0.382	4.5	100	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
30	Pyrene	0.400	0.382	4.5	100	0.00
31 S	Terphenyl-d14	0.400	0.378	5.5	100	0.00
32	Benzo(a)anthracene	0.400	0.379	5.3	100	0.00
33	Chrysene	0.400	0.380	5.0	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.362	9.5	100	0.00
35 I	Perylene-d12	0.400	0.400	0.0	100	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.363	9.3	100	0.00
37	Benzo(b)fluoranthene	0.400	0.374	6.5	100	0.00
38	Benzo(k)fluoranthene	0.400	0.376	6.0	100	0.00
39 C	Benzo(a)pyrene	0.400	0.377	5.8	100	0.00
40	Dibenzo(a,h)anthracene	0.400	0.360	10.0	100	0.00
41	Benzo(g,h,i)perylene	0.400	0.365	8.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDICCC0.4