

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041224\
 Data File : BN030863.D
 Acq On : 12 Apr 2024 15:22
 Operator : MA/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Apr 12 15:51:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 12 15:48:12 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	6114	0.400	ng	0.00
7) Naphthalene-d8	10.453	136	17602	0.400	ng	0.00
13) Acenaphthene-d10	14.306	164	10068	0.400	ng	0.00
19) Phenanthrene-d10	17.065	188	21760	0.400	ng	0.00
29) Chrysene-d12	21.258	240	21764	0.400	ng	0.00
35) Perylene-d12	23.489	264	18041	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.270	112	74109	5.053	ng	0.00
5) Phenol-d6	6.844	99	95643	5.383	ng	0.00
8) Nitrobenzene-d5	8.819	82	68483	5.449	ng	0.00
11) 2-Methylnaphthalene-d10	12.043	152	138253	5.125	ng	0.00
14) 2,4,6-Tribromophenol	15.811	330	24059	5.901	ng	0.00
15) 2-Fluorobiphenyl	12.926	172	174585	4.892	ng	0.00
27) Fluoranthene-d10	19.098	212	319185	5.277	ng	0.00
31) Terphenyl-d14	19.702	244	244368	4.938	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.219	88	33335	4.512	ng	95
3) n-Nitrosodimethylamine	3.522	42	36011	5.073	ng	# 97
6) bis(2-Chloroethyl)ether	7.104	93	83819	5.054	ng	98
9) Naphthalene	10.496	128	244978	5.032	ng	99
10) Hexachlorobutadiene	10.784	225	47593	4.896	ng	# 100
12) 2-Methylnaphthalene	12.119	142	162118	5.090	ng	98
16) Acenaphthylene	14.028	152	248848	5.657	ng	100
17) Acenaphthene	14.370	154	159786	5.141	ng	97
18) Fluorene	15.364	166	208826	5.111	ng	100
21) 4-Bromophenyl-phenylether	16.258	248	67142	5.252	ng	99
22) Hexachlorobenzene	16.358	284	76029	5.093	ng	99
23) Atrazine	16.531	200	52953	5.553	ng	98
25) Phenanthrene	17.102	178	329435	5.174	ng	100
26) Anthracene	17.189	178	301965	5.674	ng	100
28) Fluoranthene	19.126	202	405904	5.150	ng	99
30) Pyrene	19.493	202	417452	5.102	ng	100
32) Benzo(a)anthracene	21.240	228	396019	5.249	ng	99
33) Chrysene	21.294	228	398190	4.925	ng	99
34) Bis(2-ethylhexyl)phtha...	21.178	149	190887	5.798	ng	99
36) Indeno(1,2,3-cd)pyrene	25.735	276	385589	5.465	ng	99
37) Benzo(b)fluoranthene	22.817	252	399067	4.996	ng	# 94
38) Benzo(k)fluoranthene	22.864	252	384648	5.105	ng	# 94
39) Benzo(a)pyrene	23.393	252	327542	5.325	ng	# 88
40) Dibenzo(a,h)anthracene	25.752	278	308332	5.501	ng	96
41) Benzo(g,h,i)perylene	26.425	276	323826	5.334	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041224\
Data File : BN030863.D
Acq On : 12 Apr 2024 15:22
Operator : MA/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Quant Time: Apr 12 15:51:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 12 15:48:12 2024
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.675 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN030863.D
 Acq: 12 Apr 2024 15:22

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0



Tgt Ion:152 Resp: 6114




Ion	Ratio	Lower	Upper
152	100		
150	153.8	121.6	182.4
115	59.1	46.4	69.6




#2
 1,4-Dioxane
 Concen: 4.512 ng
 RT: 3.219 min Scan# 33
 Delta R.T. -0.000 min
 Lab File: BN030863.D
 Acq: 12 Apr 2024 15:22

Tgt Ion: 88 Resp: 33335

Ion	Ratio	Lower	Upper
88	100		
43	27.3	22.6	33.8
58	69.1	51.0	76.4





#3

n-Nitrosodimethylamine

Concen: 5.073 ng

RT: 3.522 min Scan# 76

Delta R.T. -0.007 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion: 42 Resp: 36011

Ion Ratio Lower Upper

42 100

74 130.7 101.8 152.6

44 6.5 8.0 12.0#



#4

2-Fluorophenol

Concen: 5.053 ng

RT: 5.270 min Scan# 317

Delta R.T. -0.000 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22




Tgt Ion: 112 Resp: 74109


Ion Ratio Lower Upper

112 100

64 53.2 42.2 63.4

63 27.7 22.1 33.1









#5
Phenol-d6
Concen: 5.383 ng
RT: 6.844 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 95643
Ion Ratio Lower Upper
99 100
42 18.6 15.3 22.9
71 32.3 26.4 39.6



#6
bis(2-Chloroethyl)ether
Concen: 5.054 ng
RT: 7.104 min Scan# 571
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Tgt Ion: 93 Resp: 83819
Ion Ratio Lower Upper
93 100
63 72.2 59.7 89.5
95 32.9 27.1 40.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.453 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:136	Resp:	17602
Ion Ratio	Lower	Upper
136	100	
137	11.6	9.2 13.8
54	6.7	5.3 7.9
68	6.0	4.5 6.7



#8
Nitrobenzene-d5
Concen: 5.449 ng
RT: 8.819 min Scan# 786
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Tgt Ion: 82	Resp:	68483
Ion Ratio	Lower	Upper
82	100	
128	41.2	31.4 47.2
54	48.2	41.8 62.6









#9
Naphthalene
Concen: 5.032 ng
RT: 10.496 min Scan# 943
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 244978
Ion Ratio Lower Upper
128 100
129 10.9 9.1 13.7
127 13.1 10.7 16.1



#10
Hexachlorobutadiene
Concen: 4.896 ng
RT: 10.784 min Scan# 970
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Tgt Ion:225 Resp: 47593
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 5.125 ng
RT: 12.043 min Scan# 1180
Delta R.T. -0.004 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 138253
Ion Ratio Lower Upper
152 100
151 20.5 16.6 24.8



#12
2-Methylnaphthalene
Concen: 5.090 ng
RT: 12.119 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Tgt Ion:142 Resp: 162118
Ion Ratio Lower Upper
142 100
141 88.3 71.4 107.0
115 34.6 29.1 43.7









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.306 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 10068
Ion Ratio Lower Upper
164 100
162 104.8 83.7 125.5
160 50.2 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 5.901 ng
RT: 15.811 min Scan# 1675
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Tgt Ion:330 Resp: 24059
Ion Ratio Lower Upper
330 100
332 95.0 76.4 114.6
141 44.7 36.6 54.8









#15
 2-Fluorobiphenyl
 Concen: 4.892 ng
 RT: 12.926 min Scan# 1409
 Delta R.T. -0.000 min
 Lab File: BN030863.D
 Acq: 12 Apr 2024 15:22




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:	172	Resp:	174585
Ion Ratio	Lower	Upper	
172	100		
171	35.0	28.4	42.6
170	23.5	19.5	29.3



#16
 Acenaphthylene
 Concen: 5.657 ng
 RT: 14.028 min Scan# 1512
 Delta R.T. -0.000 min
 Lab File: BN030863.D
 Acq: 12 Apr 2024 15:22

Tgt Ion:	152	Resp:	248848
Ion Ratio	Lower	Upper	
152	100		
151	19.3	15.5	23.3
153	12.9	10.3	15.5





#17

Acenaphthene

Concen: 5.141 ng

RT: 14.370 min Scan# 1544

Delta R.T. -0.000 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:154 Resp: 159786

Ion Ratio Lower Upper

154 100

153 112.1 91.5 137.3

152 55.6 46.9 70.3



#18

Fluorene

Concen: 5.111 ng

RT: 15.364 min Scan# 1637

Delta R.T. -0.000 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22




Tgt Ion:166 Resp: 208826


Ion Ratio Lower Upper

166 100

165 97.3 78.0 117.0

167 13.5 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.065 min Scan# 1776

Delta R.T. -0.000 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 21760

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 7.8 5.8 8.6



#21

4-Bromophenyl-phenylether

Concen: 5.252 ng

RT: 16.258 min Scan# 1711

Delta R.T. -0.000 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22




Tgt Ion:248 Resp: 67142


Ion Ratio Lower Upper

248 100

250 98.7 79.0 118.4

141 50.9 42.6 64.0









#22
Hexachlorobenzene
Concen: 5.093 ng
RT: 16.358 min Scan# 1719
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 76029
Ion Ratio Lower Upper
284 100
142 39.2 31.6 47.4
249 32.1 26.0 39.0



#23
Atrazine
Concen: 5.553 ng
RT: 16.531 min Scan# 1733
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Tgt Ion:200 Resp: 52953
Ion Ratio Lower Upper
200 100
173 27.1 22.6 33.8
215 46.7 38.5 57.7









#25
 Phenanthrene
 Concen: 5.174 ng
 RT: 17.102 min Scan# 1779
 Delta R.T. -0.000 min
 Lab File: BN030863.D
 Acq: 12 Apr 2024 15:22




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:178 Resp: 329435
 Ion Ratio Lower Upper
 178 100
 176 19.0 15.3 22.9
 179 15.4 12.3 18.5



#26
 Anthracene
 Concen: 5.674 ng
 RT: 17.189 min Scan# 1786
 Delta R.T. -0.000 min
 Lab File: BN030863.D
 Acq: 12 Apr 2024 15:22

Tgt Ion:178 Resp: 301965
 Ion Ratio Lower Upper
 178 100
 176 18.4 14.7 22.1
 179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 5.277 ng
RT: 19.098 min Scan# 20
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:212 Resp: 319185
Ion Ratio Lower Upper
212 100
106 14.7 11.8 17.8
104 8.1 6.6 10.0



#28
Fluoranthene
Concen: 5.150 ng
RT: 19.126 min Scan# 2024
Delta R.T. -0.005 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Tgt Ion:202 Resp: 405904
Ion Ratio Lower Upper
202 100
101 11.7 9.4 14.2
203 17.4 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.258 min Scan# 2457
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:240 Resp: 21764
Ion Ratio Lower Upper
240 100
120 13.7 13.5 20.3
236 27.7 22.2 33.2



#30
Pyrene
Concen: 5.102 ng
RT: 19.493 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Tgt Ion:202 Resp: 417452
Ion Ratio Lower Upper
202 100
200 20.8 16.6 24.8
203 17.8 14.4 21.6









#31
 Terphenyl-d14
 Concen: 4.938 ng
 RT: 19.702 min Scan# 2148
 Delta R.T. -0.000 min
 Lab File: BN030863.D
 Acq: 12 Apr 2024 15:22




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:244 Resp: 244368
 Ion Ratio Lower Upper
 244 100
 212 8.1 6.7 10.1
 122 14.6 12.8 19.2



#32
 Benzo(a)anthracene
 Concen: 5.249 ng
 RT: 21.240 min Scan# 2455
 Delta R.T. -0.009 min
 Lab File: BN030863.D
 Acq: 12 Apr 2024 15:22

Tgt Ion:228 Resp: 396019
 Ion Ratio Lower Upper
 228 100
 226 27.6 21.3 31.9
 229 19.5 15.8 23.8





#33

Chrysene

Concen: 4.925 ng

RT: 21.294 min Scan# 2461

Delta R.T. -0.000 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:228 Resp: 398190

Ion Ratio Lower Upper

228 100

226 29.8 24.6 36.8

229 19.4 15.2 22.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 5.798 ng

RT: 21.178 min Scan# 2448

Delta R.T. -0.000 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22




Tgt Ion:149 Resp: 190887


Ion Ratio Lower Upper

149 100

167 27.5 21.4 32.0

279 3.1 2.6 4.0





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.489 min Scan# 2956

Delta R.T. -0.006 min

Lab File: BN030863.D


Acq: 12 Apr 2024 15:22

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0






Tgt Ion:264 Resp: 18041

Ion Ratio Lower Upper

264 100

260 25.1 19.8 29.8

265 60.7 42.5 63.7



#36

Indeno(1,2,3-cd)pyrene

Concen: 5.465 ng

RT: 25.735 min Scan# 3724

Delta R.T. -0.009 min

Lab File: BN030863.D

Acq: 12 Apr 2024 15:22




Tgt Ion:276 Resp: 385589


Ion Ratio Lower Upper

276 100

138 28.6 22.2 33.2


277 24.9 19.9 29.9








#37
Benzo(b)fluoranthene
Concen: 4.996 ng
RT: 22.817 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion:252 Resp: 399067
Ion Ratio Lower Upper
252 100
253 22.1 18.8 28.2
125 11.4 13.0 19.6#



#38
Benzo(k)fluoranthene
Concen: 5.105 ng
RT: 22.864 min Scan# 2742
Delta R.T. -0.003 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22


Tgt Ion:252 Resp: 384648
Ion Ratio Lower Upper
252 100
253 21.6 19.1 28.7
125 11.1 11.6 17.4#








#39
Benzo(a)pyrene
Concen: 5.325 ng
RT: 23.393 min Scan# 2923
Delta R.T. -0.003 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion:252 Resp: 327542
Ion Ratio Lower Upper
252 100
253 22.0 20.2 30.4
125 12.5 16.6 25.0#



#40
Dibenzo(a,h)anthracene
Concen: 5.501 ng
RT: 25.752 min Scan# 3730
Delta R.T. -0.003 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22


Tgt Ion:278 Resp: 308332
Ion Ratio Lower Upper
278 100
139 19.2 16.5 24.7
279 23.6 20.6 30.8





#41
Benzo(g,h,i)perylene
Concen: 5.334 ng
RT: 26.425 min Scan# 3960
Delta R.T. -0.003 min
Lab File: BN030863.D
Acq: 12 Apr 2024 15:22

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:276 Resp: 323826
Ion Ratio Lower Upper
276 100
277 23.6 19.8 29.8
138 24.7 21.2 31.8

