

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041320\
 Data File : BN010471.D
 Acq On : 13 Apr 2020 17:24
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN041320

Quant Time: Apr 13 18:06:59 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 18:04:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.294	0.328	-11.6	103	0.00
3	n-Nitrosodimethylamine	0.191	0.187	2.1	107	0.00
4 S	2-Fluorophenol	0.876	0.936	-6.8	98	0.00
5 S	Phenol-d6	1.145	1.225	-7.0	99	0.00
6	bis(2-Chloroethyl)ether	0.992	1.093	-10.2	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
8 S	Nitrobenzene-d5	0.280	0.305	-8.9	104	0.00
9	Naphthalene	0.993	1.090	-9.8	101	0.00
10	Hexachlorobutadiene	0.239	0.251	-5.0	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.670	-1.8	94	0.00
12	2-Methylnaphthalene	0.714	0.785	-9.9	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
14 S	2,4,6-Tribromophenol	0.189	0.178	5.8	91	0.00
15 S	2-Fluorobiphenyl	1.456	1.656	-13.7	104	0.00
16	Acenaphthylene	1.728	1.864	-7.9	102	0.00
17	Acenaphthene	1.110	1.172	-5.6	98	0.00
18	Fluorene	1.454	1.552	-6.7	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.222	0.233	-5.0	96	0.00
21	Hexachlorobenzene	0.254	0.281	-10.6	97	0.00
22	Pentachlorophenol	0.100	0.084	16.0	88	0.00
23	Phenanthrene	1.080	1.202	-11.3	99	0.00
24	Anthracene	0.988	1.083	-9.6	100	0.00
25 SURR	Fluoranthene-d10	1.213	1.230	-1.4	90	0.00
26	Fluoranthene	1.345	1.481	-10.1	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pvrene	1.346	1.453	-7.9	96	0.00
29 S	Terphenyl-d14	0.915	1.096	-19.8	108	0.00
30	Benzo(a)anthracene	1.305	1.378	-5.6	94	0.00
31	Chrysene	1.300	1.361	-4.7	94	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.641	4.3	89	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.225	-6.6	94	0.00
34 I	Perylene-d12	1.000	1.000	0.0	86	0.00
35	Benzo(b)fluoranthene	1.317	1.424	-8.1	93	0.00
36	Benzo(k)fluoranthene	1.320	1.428	-8.2	93	0.00
37 C	Benzo(a)pyrene	1.166	1.295	-11.1	96	0.00
38	Dibenzo(a,h)anthracene	1.066	1.179	-10.6	94	0.00
39	Benzo(a,h,i)perylene	1.073	1.214	-13.1	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041320\
Data File : BN010471.D
Acq On : 13 Apr 2020 17:24
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN041320

Quant Time: Apr 13 18:06:59 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 18:04:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			