

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041322\
 Data File : BN019447.D
 Acq On : 13 Apr 2022 19:11
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4338

Quant Time: Apr 14 01:58:10 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN041322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 13 14:35:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.818	152	4356	0.400	ng/ul	0.00
4) Naphthalene-d8	10.624	136	13249	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.461	164	8153	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.201	188	15512	0.400	ng/ul	0.00
17) Chrysene-d12	21.395	240	10011	0.400	ng/ul	0.00
23) Perylene-d12	23.745	264	7669	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.177	96	2181	0.389	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.219	152	7919	0.422	ng/ul	0.00
18) Fluoranthene-d10	19.235	212	17432	0.470	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.211	88	2201	0.408	ng/ul#	87
5) Naphthalene	10.673	128	16344	0.433	ng/ul	100
7) 2-Methylnaphthalene	12.290	142	10558	0.430	ng/ul#	100
8) 1-Methylnaphthalene	12.510	142	10512	0.408	ng/ul#	100
10) Acenaphthylene	14.179	152	14000	0.473	ng/ul	100
11) Acenaphthene	14.521	153	12549	0.425	ng/ul	98
12) Fluorene	15.511	166	13791	0.420	ng/ul	98
14) Pentachlorophenol	16.872	266	909	0.365	ng/ul	93
15) Phenanthrene	17.243	178	21630	0.420	ng/ul	100
16) Anthracene	17.340	178	16469	0.416	ng/ul	99
19) Fluoranthene	19.263	202	22591	0.454	ng/ul	99
20) Pyrene	19.626	202	23843	0.450	ng/ul	98
21) Benzo(a)anthracene	21.380	228	12410	0.403	ng/ul	99
22) Chrysene	21.430	228	17543	0.435	ng/ul	99
24) Benzo(b)fluoranthene	23.032	252	14450	0.407	ng/ul	96
25) Benzo(k)fluoranthene	23.075	252	17129	0.418	ng/ul	100
26) Benzo(a)pyrene	23.646	252	13650	0.450	ng/ul#	89
27) Indeno(1,2,3-cd)pyrene	26.172	276	13426	0.398	ng/ul#	45
28) Dibenzo(a,h)anthracene	26.202	278	9583	0.399	ng/ul#	90
29) Benzo(g,h,i)perylene	26.920	276	13272	0.407	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041322\
Data File : BN019447.D
Acq On : 13 Apr 2022 19:11
Operator : CG/JU
Sample : SSTDC00.4EC
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4338

Quant Time: Apr 14 01:58:10 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN041322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 13 14:35:39 2022
Response via : Initial Calibration

