


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041322\
 Data File : BN019432.D
 Acq On : 13 Apr 2022 09:52
 Operator : CG/JU
 Sample : DFTPP240
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP240

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN041322.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 13 14:35:39 2022



AutoFind: Scans 2492, 2493, 2494; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.6	169566	PASS
68	69	0.00	2	2.0	3374	PASS
69	198	0.00	100	40.6	169488	PASS
70	69	0.00	2	0.6	956	PASS
127	198	10	80	48.0	200725	PASS
197	198	0.00	2	0.2	953	PASS
198	198	100	100	100.0	417920	PASS
199	198	5	9	6.8	28339	PASS
275	198	10	60	27.9	116512	PASS
365	198	1	100	3.6	15061	PASS
441	198	0.01	100	11.5	47939	PASS
442	198	50	100	73.6	307797	PASS
443	442	15	24	20.3	62627	PASS

m/z	Abundance
39.00	219.0
40.00	674.0
41.00	370.0
42.00	207.0
43.00	592.0
44.00	1525.0
49.00	368.0
50.80	165.0
55.00	502.0
56.00	271.0
57.00	535.0
67.00	175.0
69.10	216.0
71.10	167.0
73.00	513.0
74.90	196.0
79.00	152.0
84.00	205.0
85.20	195.0
96.10	385.0
97.00	291.0
118.90	189.0
133.00	296.0
146.70	214.0
176.80	234.0
190.90	467.0
193.00	305.0
207.00	3580.0
208.10	955.0
208.90	430.0
281.00	578.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

m/z	Abundance
36.00	186.0
37.00	848.0
38.00	2100.0
39.10	14236.0
40.00	973.0
41.10	611.0
42.00	163.0
43.00	503.0
44.00	1272.0
44.90	556.0
48.00	159.0
49.10	1688.0
50.00	48920.0
51.00	188224.0
52.00	10048.0
53.00	537.0
55.00	1213.0
56.00	5606.0
57.00	13095.0
58.00	507.0
59.00	263.0
60.00	157.0
61.00	2506.0
62.00	3071.0
63.00	7491.0
64.00	1038.0
65.00	3774.0
65.90	255.0
67.00	368.0
68.10	3834.0
69.00	184704.0
70.00	995.0
70.90	224.0
73.00	2143.0
74.00	19464.0
75.00	29512.0
76.00	9834.0
77.00	209408.0
78.00	15441.0
79.00	13806.0
80.00	10487.0
81.00	14693.0
82.00	3730.0
83.00	3494.0
83.90	654.0
85.00	3298.0
86.00	4377.0
87.00	2167.0
88.00	743.0
89.00	416.0
91.00	3466.0
91.90	4439.0
93.00	23312.0
94.00	1672.0
94.90	491.0
96.00	1622.0
97.10	622.0
98.00	18504.0
99.00	15913.0
99.90	1289.0
101.00	8556.0
102.00	343.0
102.90	2355.0
104.00	5155.0
105.00	5098.0
106.00	1485.0
107.00	61864.0
108.00	10138.0
109.10	1983.0
110.00	110776.0
111.00	17776.0
112.00	2064.0
112.90	740.0
115.00	374.0
116.00	3612.0
117.00	49280.0
118.00	3708.0
119.00	701.0
120.00	730.0
121.00	326.0
122.00	4113.0
123.00	6604.0
124.00	3007.0
125.00	2577.0
127.00	211072.0
128.00	16016.0
129.00	78000.0
130.00	6692.0
131.00	1524.0
131.90	694.0
133.10	525.0
134.00	2457.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

135.00	6710.0
136.00	2703.0
137.00	3728.0
138.00	907.0
139.00	457.0
140.00	1046.0
141.00	10994.0
142.00	3412.0
143.00	2634.0
144.00	712.0
144.80	502.0
145.90	2067.0
147.00	5931.0
148.00	11583.0
149.00	2946.0
149.90	794.0
151.00	1723.0
152.90	2693.0
153.90	2401.0
155.00	5437.0
156.00	8849.0
157.10	1366.0
157.90	1888.0
159.00	1672.0
160.00	3552.0
161.00	5071.0
161.90	1358.0
163.00	395.0
163.90	617.0
165.00	4052.0
166.00	3254.0
167.00	21400.0
168.00	13437.0
169.00	2143.0
170.10	640.0
171.00	782.0
171.90	2020.0
173.00	2477.0
174.00	4392.0
175.00	8329.0
176.00	2173.0
177.00	4132.0
178.00	1559.0
178.90	15433.0
180.00	13025.0
181.00	5503.0
182.00	880.0
183.00	453.0
184.00	1517.0
185.00	7665.0
186.00	60288.0
187.00	17112.0
188.00	2269.0
189.00	3860.0
190.00	781.0
191.00	1828.0
192.00	4818.0
193.00	6050.0
193.90	1325.0
194.90	586.0
196.00	10154.0
197.90	405248.0
198.90	28552.0
200.00	2164.0
201.60	1943.0
202.90	3032.0
204.00	15640.0
205.10	25208.0
206.00	109296.0
207.00	17792.0
208.00	4295.0
208.90	1849.0
210.20	1671.0
211.00	4076.0
212.10	796.0
212.90	448.0
214.90	1008.0
216.00	2716.0
217.00	29824.0
217.90	3845.0
218.90	392.0
221.00	16115.0
221.80	4253.0
223.00	6476.0
224.00	62248.0
225.00	16096.0
226.00	2465.0
227.00	26080.0
227.90	4159.0
228.90	5304.0
230.00	702.0
231.00	2281.0
231.90	454.0
233.00	599.0
233.90	1522.0
235.00	1914.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

236.00	1288.0
237.10	3157.0
238.00	566.0
238.90	979.0
240.00	891.0
241.00	1549.0
242.00	3141.0
243.00	3660.0
244.00	45640.0
245.00	6852.0
246.00	8745.0
246.90	2098.0
247.80	479.0
248.90	1685.0
249.90	321.0
251.00	363.0
252.00	396.0
252.90	1021.0
254.10	2583.0
255.00	243712.0
256.00	35672.0
257.00	3679.0
258.00	13320.0
258.90	2294.0
259.80	524.0
260.90	307.0
264.10	354.0
265.00	5138.0
266.00	903.0
266.90	237.0
270.00	307.0
270.80	319.0
272.10	710.0
273.00	6853.0
274.00	18584.0
275.00	109504.0
276.00	14157.0
277.00	9334.0
278.00	1751.0
278.90	406.0
281.00	652.0
282.00	319.0
282.90	945.0
284.00	800.0
285.00	1343.0
286.10	251.0
289.00	275.0
289.90	292.0
291.00	401.0
292.00	396.0
293.00	2203.0
294.00	799.0
296.00	25272.0
297.00	4396.0
297.90	328.0
300.90	311.0
301.80	434.0
303.00	3289.0
304.00	872.0
304.90	225.0
307.90	407.0
308.80	293.0
310.00	456.0
313.00	160.0
314.00	1458.0
314.90	3474.0
316.00	1757.0
316.90	448.0
319.90	172.0
321.00	1057.0
322.00	459.0
323.00	8308.0
324.00	1692.0
324.90	156.0
325.90	223.0
326.90	1737.0
328.00	1068.0
328.90	202.0
332.00	653.0
333.00	852.0
334.00	5575.0
335.00	1556.0
336.00	230.0
339.00	217.0
339.90	230.0
340.90	1049.0
342.00	506.0
345.90	1692.0
346.90	528.0
350.90	181.0
352.00	3051.0
353.00	1793.0
354.00	3249.0
355.00	514.0
358.90	267.0
365.00	12571.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

366.00	1832.0
370.00	306.0
371.00	833.0
372.00	4361.0
373.00	1450.0
382.90	1397.0
384.00	444.0
384.90	210.0
389.90	681.0
390.90	398.0
391.90	348.0
401.00	253.0
402.00	2251.0
403.00	3001.0
403.90	1225.0
405.00	580.0
405.90	159.0
421.00	2068.0
422.00	2600.0
423.00	16544.0
424.00	7675.0
425.00	1216.0
441.10	32464.0
442.00	236608.0
443.00	49224.0
444.00	4786.0
445.10	198.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

m/z	Abundance
36.00	191.0
37.00	795.0
38.00	2401.0
39.00	14370.0
40.00	1218.0
41.00	775.0
42.00	176.0
43.00	459.0
43.90	1253.0
45.00	469.0
46.90	158.0
47.90	235.0
49.10	1845.0
50.00	51792.0
51.00	200128.0
52.00	10104.0
53.00	452.0
55.00	1102.0
56.00	6201.0
57.00	13902.0
58.00	601.0
59.00	177.0
59.90	196.0
61.00	2747.0
62.00	2928.0
63.00	7692.0
64.00	1281.0
65.00	4060.0
66.10	213.0
67.00	366.0
68.10	3865.0
69.00	200896.0
70.00	1246.0
71.00	414.0
73.00	1966.0
74.00	21056.0
75.00	31952.0
76.00	11291.0
77.00	229312.0
78.00	15467.0
79.00	15485.0
80.00	11449.0
81.00	16712.0
82.00	4278.0
83.00	4285.0
83.90	768.0
85.00	3845.0
86.00	4521.0
87.00	2150.0
87.90	848.0
89.00	427.0
91.00	3957.0
92.00	4703.0
93.00	26456.0
93.90	1726.0
95.00	646.0
96.00	1555.0
98.00	19808.0
99.00	17184.0
100.00	1317.0
101.00	8999.0
101.90	491.0
103.00	2918.0
104.00	5943.0
105.00	5712.0
106.00	1820.0
107.00	69656.0
108.00	11338.0
109.10	2142.0
110.00	124888.0
111.00	19184.0
112.00	2252.0
113.00	849.0
114.90	337.0
116.00	3739.0
117.00	56160.0
118.00	3859.0
119.00	585.0
120.00	774.0
121.10	339.0
122.00	4652.0
123.00	6941.0
124.00	3181.0
125.00	2971.0
127.00	236928.0
128.00	18464.0
129.00	91448.0
130.00	7722.0
131.00	1359.0
131.90	755.0
133.10	586.0
134.00	2902.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

135.00	7669.0
136.00	3214.0
137.00	4321.0
137.90	849.0
139.00	592.0
140.00	1337.0
141.00	11629.0
142.00	3882.0
143.00	2594.0
143.90	788.0
144.90	657.0
146.00	2215.0
146.90	6714.0
148.00	14526.0
149.00	3111.0
150.00	782.0
151.00	1559.0
152.00	843.0
153.00	4094.0
154.00	2822.0
155.00	6530.0
156.00	10457.0
157.00	2075.0
157.90	2340.0
159.00	1838.0
160.00	4060.0
161.00	5472.0
162.00	1492.0
163.00	551.0
163.90	667.0
165.00	4966.0
166.00	4400.0
167.00	25616.0
168.00	14947.0
169.00	2236.0
169.90	722.0
171.00	1124.0
172.00	2095.0
173.10	2921.0
174.00	5823.0
175.00	10025.0
176.00	2851.0
177.00	5355.0
178.00	1819.0
179.00	18936.0
179.90	14547.0
181.00	6010.0
182.00	907.0
182.80	461.0
183.20	406.0
184.00	1747.0
185.00	9538.0
186.00	72416.0
187.00	21232.0
188.00	2530.0
189.00	4305.0
190.00	705.0
191.00	2426.0
192.00	5785.0
193.00	7044.0
194.00	1468.0
194.90	828.0
196.00	14120.0
198.00	496320.0
198.90	33288.0
200.00	2946.0
201.50	2563.0
202.90	3588.0
204.00	18696.0
205.00	31608.0
206.00	142464.0
207.00	22464.0
208.00	5157.0
209.00	2188.0
210.00	2258.0
211.00	5206.0
211.90	1146.0
213.00	456.0
214.90	1748.0
216.00	2758.0
217.00	35584.0
218.00	4883.0
219.10	526.0
221.00	18256.0
221.80	5045.0
223.00	8628.0
224.00	74600.0
225.00	19344.0
226.00	2394.0
227.00	32040.0
227.90	4630.0
229.00	6431.0
230.00	819.0
231.00	2973.0
232.00	433.0
232.90	586.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

234.00	1925.0
235.00	2568.0
235.90	1338.0
237.00	3884.0
238.00	601.0
239.00	1147.0
240.00	1104.0
241.00	1867.0
242.00	4275.0
243.00	4437.0
244.00	60688.0
245.00	8213.0
246.00	11901.0
246.90	2375.0
247.90	639.0
249.00	2238.0
250.00	433.0
251.00	489.0
252.00	522.0
253.00	1277.0
255.00	299456.0
256.00	46552.0
257.00	3923.0
257.90	17400.0
259.00	2885.0
260.00	508.0
261.00	484.0
262.90	174.0
264.00	430.0
265.00	7338.0
265.90	1096.0
266.90	187.0
270.00	423.0
271.00	704.0
271.90	774.0
273.00	9224.0
274.00	24080.0
275.00	140032.0
276.00	18536.0
277.00	11339.0
278.00	2606.0
278.90	464.0
281.00	543.0
282.00	476.0
283.00	1459.0
284.00	862.0
285.00	1887.0
285.90	542.0
289.00	534.0
289.90	379.0
291.00	272.0
292.00	506.0
292.90	2845.0
293.90	845.0
295.00	1028.0
296.00	40456.0
296.90	5957.0
297.80	288.0
301.00	587.0
302.00	765.0
303.00	3979.0
304.00	1138.0
305.00	244.0
308.00	433.0
308.90	451.0
309.90	482.0
310.80	231.0
312.90	400.0
314.00	1747.0
315.00	4082.0
316.00	2437.0
317.00	511.0
321.00	1278.0
322.10	783.0
323.00	12809.0
324.00	2306.0
326.00	312.0
326.90	2387.0
328.00	1223.0
329.00	224.0
332.00	1030.0
333.00	1170.0
334.00	8687.0
335.00	2380.0
336.00	240.0
340.00	218.0
341.00	1353.0
341.90	346.0
345.90	2886.0
347.00	584.0
351.00	325.0
352.00	3999.0
352.90	2557.0
354.10	3722.0
354.90	1017.0
358.80	373.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

363.90	282.0
365.00	18544.0
365.90	2814.0
366.90	187.0
369.80	303.0
371.00	1054.0
372.00	6406.0
372.90	1708.0
374.00	279.0
377.10	171.0
382.90	1904.0
384.00	624.0
384.90	322.0
389.80	793.0
391.00	565.0
391.90	491.0
401.00	349.0
402.00	3120.0
403.00	5061.0
404.00	1607.0
405.00	550.0
415.00	245.0
421.00	3348.0
422.00	3642.0
423.00	25432.0
424.00	9025.0
425.00	1488.0
441.00	57832.0
442.00	344128.0
443.00	76944.0
444.00	6709.0
445.00	337.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

m/z	Abundance
36.00	190.0
37.00	569.0
38.00	1515.0
39.10	8771.0
40.00	1026.0
41.00	736.0
42.00	190.0
43.00	549.0
44.00	1373.0
44.90	293.0
48.00	180.0
49.00	1183.0
50.00	31920.0
51.00	120840.0
52.00	6189.0
52.90	428.0
55.00	976.0
56.00	3995.0
57.00	8822.0
58.00	522.0
58.90	158.0
61.00	1770.0
62.00	1896.0
63.00	5072.0
64.00	827.0
65.00	2264.0
66.00	187.0
67.00	247.0
68.00	2423.0
69.00	123512.0
69.90	628.0
71.00	459.0
73.00	1935.0
74.00	13193.0
75.00	20160.0
76.00	6669.0
77.00	144000.0
78.00	9415.0
79.00	9302.0
80.00	6881.0
81.00	10010.0
82.00	2350.0
83.00	2451.0
83.80	406.0
85.00	2438.0
86.00	3193.0
87.00	1417.0
88.00	608.0
88.90	312.0
91.00	2122.0
92.00	2634.0
93.00	17088.0
94.00	1210.0
95.00	422.0
96.00	1171.0
97.00	490.0
98.00	12666.0
99.00	10527.0
100.00	919.0
101.00	5466.0
101.90	394.0
102.90	1926.0
104.00	3715.0
105.00	3174.0
106.00	1205.0
107.00	44184.0
108.00	6728.0
109.00	1387.0
110.00	77704.0
111.00	12401.0
111.90	1522.0
112.90	597.0
115.00	247.0
116.00	2845.0
117.00	35176.0
117.90	2531.0
119.00	504.0
119.90	698.0
120.80	222.0
122.00	2896.0
123.00	4614.0
123.90	1990.0
125.00	2039.0
127.00	154176.0
128.00	12381.0
129.00	60680.0
130.00	5172.0
130.90	971.0
131.80	429.0
133.00	696.0
134.00	1962.0
135.00	4798.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

136.00	2023.0
137.00	2566.0
137.90	727.0
138.90	357.0
139.90	718.0
141.00	8392.0
142.00	2734.0
142.90	1538.0
144.00	525.0
144.90	379.0
145.90	1566.0
147.00	4311.0
148.00	9325.0
149.00	2187.0
150.00	667.0
151.10	1119.0
151.90	733.0
153.00	2545.0
154.00	2284.0
155.00	4583.0
156.00	7183.0
157.00	1226.0
157.90	1505.0
158.90	989.0
160.00	2511.0
161.00	4209.0
162.00	968.0
162.90	542.0
163.80	345.0
164.10	334.0
165.00	3183.0
166.00	2755.0
167.00	16352.0
168.00	9935.0
169.00	1281.0
169.90	812.0
171.00	513.0
171.90	1575.0
173.00	1704.0
174.00	3299.0
175.00	7106.0
176.00	1677.0
176.90	3131.0
177.90	1227.0
178.90	12419.0
180.00	9362.0
181.00	4150.0
182.00	660.0
183.00	400.0
183.90	1050.0
185.00	6567.0
186.00	49336.0
187.00	13256.0
188.00	2186.0
189.00	3210.0
189.90	534.0
191.00	1796.0
192.00	4472.0
193.00	5460.0
193.90	870.0
195.10	552.0
196.00	10375.0
197.10	2858.0
197.90	352192.0
199.00	23176.0
200.10	1623.0
201.60	1463.0
203.00	3261.0
204.00	12934.0
205.00	24040.0
206.00	99752.0
207.00	16073.0
208.00	4183.0
209.00	1814.0
210.10	1559.0
211.00	3594.0
213.00	325.0
214.90	1185.0
216.10	2093.0
217.00	26480.0
217.90	3007.0
219.00	355.0
221.00	14938.0
221.80	3496.0
222.90	5305.0
224.00	59000.0
225.00	13499.0
226.00	1772.0
227.00	23824.0
228.00	2884.0
229.00	4966.0
229.90	721.0
231.00	2002.0
232.00	275.0
233.00	538.0
233.90	1656.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

235.00	1663.0
235.90	1136.0
237.00	2764.0
237.90	324.0
239.00	893.0
239.90	754.0
240.90	1304.0
242.00	2768.0
243.00	3153.0
244.00	38920.0
245.00	5448.0
246.00	8925.0
247.00	1952.0
247.90	286.0
248.90	1540.0
249.80	345.0
250.80	495.0
252.00	351.0
252.90	991.0
255.00	227264.0
256.00	31072.0
257.00	2649.0
258.00	12881.0
259.00	2156.0
260.00	447.0
260.90	305.0
264.00	513.0
265.00	5549.0
266.00	744.0
266.80	186.0
269.90	377.0
270.90	489.0
272.00	527.0
273.00	7127.0
274.00	18736.0
275.00	100000.0
276.00	14456.0
277.00	8888.0
278.00	1544.0
279.00	375.0
281.00	408.0
282.00	306.0
283.00	1311.0
284.00	701.0
285.00	1776.0
286.00	292.0
288.90	329.0
290.00	265.0
291.00	215.0
291.90	449.0
293.00	2266.0
294.00	676.0
295.00	873.0
296.00	31168.0
297.00	4032.0
298.00	313.0
300.90	384.0
302.00	452.0
303.00	3278.0
304.10	888.0
305.00	247.0
308.00	370.0
308.90	308.0
309.90	500.0
312.80	175.0
313.10	184.0
314.00	1576.0
314.90	3257.0
316.00	1784.0
317.00	417.0
319.80	195.0
320.90	1102.0
322.00	533.0
323.00	9042.0
324.00	1744.0
324.90	247.0
326.10	325.0
326.90	2006.0
328.00	1275.0
329.00	219.0
331.90	740.0
332.90	930.0
334.00	5872.0
335.00	1598.0
335.90	251.0
338.80	198.0
341.00	1399.0
341.90	313.0
345.90	2309.0
346.90	482.0
351.00	210.0
351.90	3257.0
353.00	2163.0
354.00	3752.0
355.00	654.0
358.90	207.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240

364.90	14067.0
366.00	2074.0
369.90	337.0
371.00	884.0
372.00	6118.0
372.90	1290.0
373.90	217.0
383.00	1619.0
384.00	605.0
384.90	235.0
389.90	618.0
390.90	629.0
392.00	444.0
400.90	490.0
402.00	2315.0
403.00	3809.0
404.00	1292.0
404.90	429.0
414.90	223.0
421.00	3071.0
421.90	2987.0
423.00	22776.0
424.00	7329.0
425.00	1028.0
441.00	53520.0
442.00	342656.0
443.00	61712.0
444.00	5366.0
444.90	361.0

Instrument :
BNA_N
ClientSampleId :
DFTPP240