


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041323\
Data File : BN024854.D
Acq On : 13 Apr 2023 10:10
Operator : CG/JU
Sample : DFTPP321
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP321

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Apr 13 02:44:00 2023



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.8	553712	PASS
68	69	0.00	2	1.7	7797	PASS
69	198	0.00	100	48.4	455403	PASS
70	69	0.00	2	0.6	2647	PASS
127	198	10	80	51.7	486421	PASS
197	198	0.00	2	0.6	5463	PASS
198	198	100	100	100.0	941141	PASS
199	198	5	9	7.0	65536	PASS
275	198	10	60	24.3	228544	PASS
365	198	1	100	3.3	30632	PASS
441	198	0.01	100	11.3	106165	PASS
442	198	50	100	73.5	691435	PASS
443	442	15	24	20.1	138915	PASS

m/z	Abundance
36.00	314.0
37.90	162.0
39.10	381.0
40.00	2529.0
40.90	683.0
42.10	283.0
42.90	716.0
44.00	11263.0
44.90	378.0
49.00	657.0
49.90	236.0
51.10	357.0
55.00	378.0
57.10	209.0
58.90	226.0
64.90	155.0
67.00	181.0
73.00	3057.0
74.00	406.0
75.00	236.0
77.00	289.0
78.00	336.0
78.90	164.0
81.00	180.0
83.80	520.0
86.00	182.0
91.00	271.0
95.90	415.0
133.00	476.0
135.00	235.0
147.00	988.0
148.00	204.0
191.00	428.0
193.00	329.0
207.00	3494.0
208.00	637.0
208.90	400.0
221.00	820.0
221.90	172.0
222.90	366.0
266.90	263.0
281.00	1244.0
282.00	490.0
283.00	313.0
341.00	353.0
342.00	165.0
355.00	632.0
355.90	272.0
356.90	329.0
415.00	176.0
428.80	192.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

m/z	Abundance
36.00	708.0
37.00	2416.0
38.00	6242.0
39.10	37864.0
40.00	3788.0
41.00	1626.0
41.90	263.0
43.00	765.0
44.00	10304.0
45.00	1211.0
47.00	306.0
48.00	347.0
49.00	3818.0
50.00	126656.0
51.00	462528.0
52.00	25080.0
53.10	941.0
55.00	2667.0
56.00	12926.0
57.00	29560.0
58.00	1161.0
58.90	565.0
60.00	357.0
61.00	5568.0
62.00	5828.0
63.00	15403.0
64.00	2348.0
65.00	7729.0
66.00	468.0
66.90	507.0
68.00	6538.0
69.00	376960.0
70.00	2288.0
70.90	226.0
71.90	190.0
73.00	3911.0
74.00	38136.0
75.00	59168.0
76.00	22240.0
77.00	425472.0
78.00	31896.0
79.00	27816.0
80.00	21464.0
81.00	27560.0
82.00	7146.0
83.00	6997.0
84.00	1178.0
85.00	4285.0
86.00	8878.0
87.00	3833.0
87.90	1408.0
89.10	634.0
89.90	295.0
91.00	6653.0
92.00	7916.0
93.00	46984.0
94.00	3182.0
95.00	698.0
96.10	2161.0
96.90	1030.0
98.00	36544.0
99.00	32384.0
100.00	3109.0
101.00	15571.0
102.00	702.0
103.00	4818.0
104.00	9367.0
105.00	9273.0
106.00	3133.0
107.00	115312.0
108.00	17088.0
109.00	3377.0
110.00	194304.0
111.00	32712.0
112.00	3925.0
112.90	1251.0
115.00	656.0
116.00	5838.0
117.00	107944.0
118.00	8513.0
119.00	1064.0
119.90	1382.0
121.10	529.0
121.90	6972.0
123.00	11284.0
124.00	5380.0
125.00	5015.0
127.00	374592.0
128.00	28632.0
129.00	144704.0
129.90	12128.0
130.90	2655.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

132.00	1414.0
132.90	956.0
134.00	4136.0
135.00	11869.0
135.90	4546.0
137.00	8656.0
138.00	1384.0
139.00	638.0
140.00	1901.0
141.00	18152.0
142.00	5711.0
143.00	4564.0
144.00	1016.0
145.00	1094.0
146.00	3421.0
147.00	9736.0
148.00	24304.0
149.00	5476.0
149.90	1877.0
151.00	2527.0
152.10	1134.0
153.00	5065.0
154.00	3878.0
155.00	10636.0
156.00	14968.0
157.00	2655.0
158.00	3287.0
158.90	2546.0
160.00	6078.0
161.00	8572.0
162.00	2449.0
163.00	714.0
164.00	866.0
165.00	6492.0
166.00	5249.0
167.00	37760.0
168.00	30304.0
169.00	4320.0
170.10	1293.0
171.00	1452.0
171.90	3373.0
172.90	4437.0
174.00	7652.0
175.00	13907.0
176.00	3995.0
177.00	6493.0
178.00	1859.0
178.90	26224.0
180.00	21264.0
181.00	8744.0
181.90	1409.0
182.80	536.0
183.00	496.0
183.90	2183.0
185.00	11538.0
186.00	90288.0
187.00	28376.0
188.00	3376.0
188.90	6504.0
190.00	1087.0
191.00	3437.0
192.00	9422.0
193.00	9667.0
194.00	2282.0
195.00	1134.0
196.00	19616.0
196.90	4085.0
197.90	685824.0
198.90	47000.0
200.00	3749.0
201.40	2679.0
202.90	4935.0
204.00	24640.0
205.00	44416.0
206.00	173696.0
207.00	25672.0
208.00	6133.0
208.90	3276.0
210.10	2892.0
211.00	6452.0
211.90	1520.0
213.00	525.0
214.00	337.0
214.90	1959.0
216.00	3354.0
216.90	46496.0
218.00	6453.0
219.00	591.0
221.00	26056.0
223.00	9025.0
224.00	98600.0
225.00	24304.0
226.00	3530.0
227.00	40912.0
228.00	6073.0
229.00	8350.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

229.90	1015.0
231.00	3455.0
231.90	584.0
233.00	624.0
234.00	2034.0
235.00	2860.0
236.00	1902.0
237.00	4756.0
238.00	748.0
239.00	1471.0
239.90	1599.0
241.00	2046.0
242.00	5420.0
243.00	5024.0
244.00	68264.0
245.00	9229.0
246.00	14055.0
247.00	2994.0
247.90	584.0
249.00	2567.0
249.90	660.0
250.90	605.0
251.90	647.0
253.00	1561.0
254.00	2475.0
254.90	338240.0
256.00	57224.0
256.90	4360.0
258.00	21560.0
259.00	3719.0
259.90	500.0
261.00	458.0
262.80	165.0
264.00	637.0
264.90	8768.0
266.00	1241.0
267.00	274.0
270.00	389.0
271.00	1340.0
272.00	974.0
272.90	10690.0
274.00	27088.0
275.00	149184.0
276.00	20840.0
277.00	12922.0
277.90	2607.0
278.70	392.0
279.20	288.0
280.90	698.0
282.00	549.0
282.90	1602.0
284.00	1281.0
284.90	2699.0
286.00	521.0
287.90	157.0
289.00	655.0
290.00	546.0
290.90	391.0
291.90	657.0
293.00	3309.0
293.90	954.0
294.90	962.0
295.90	41232.0
296.90	6473.0
297.80	447.0
300.90	444.0
302.00	747.0
303.00	4514.0
304.00	1636.0
304.90	367.0
308.00	546.0
308.80	330.0
309.90	573.0
311.90	205.0
312.80	234.0
314.00	1961.0
314.90	3721.0
316.00	2454.0
316.90	543.0
319.00	179.0
319.70	224.0
321.00	1368.0
322.10	624.0
323.00	12899.0
324.00	2606.0
324.90	197.0
326.00	291.0
326.90	2880.0
327.90	1446.0
328.90	214.0
332.00	877.0
332.90	1149.0
334.00	8463.0
335.00	2311.0
335.90	284.0
339.90	174.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

340.90	1523.0
342.00	694.0
345.90	2920.0
346.90	849.0
350.80	273.0
351.90	3586.0
353.00	2810.0
354.00	4397.0
355.00	1022.0
355.90	153.0
358.00	195.0
359.00	315.0
363.80	220.0
364.90	17208.0
365.90	3060.0
367.10	204.0
370.00	314.0
370.90	954.0
371.90	6027.0
372.90	1183.0
373.90	180.0
376.90	186.0
382.90	1822.0
383.90	661.0
384.90	252.0
390.00	739.0
391.00	631.0
392.00	621.0
394.80	153.0
401.00	485.0
401.90	2549.0
403.00	4357.0
403.90	1529.0
405.00	375.0
409.80	160.0
414.90	224.0
420.90	3101.0
421.90	2994.0
422.90	19496.0
424.00	7349.0
424.90	1248.0
439.00	300.0
440.00	334.0
441.00	54272.0
442.00	372608.0
442.90	63208.0
444.00	6791.0
444.90	573.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

m/z	Abundance
36.00	771.0
37.00	2852.0
38.00	8776.0
39.10	51048.0
40.00	4492.0
41.00	1998.0
42.00	316.0
43.00	542.0
44.00	9977.0
45.00	1520.0
46.90	406.0
48.00	554.0
49.00	5166.0
50.00	176128.0
51.00	656384.0
52.00	32984.0
53.00	1309.0
55.00	3645.0
56.00	18496.0
57.00	39696.0
58.00	1941.0
59.00	560.0
59.90	634.0
61.00	7472.0
62.00	7928.0
63.00	22512.0
64.00	3060.0
65.00	11171.0
66.00	902.0
67.00	658.0
68.00	9607.0
69.00	541248.0
70.00	3404.0
71.00	318.0
71.80	253.0
73.10	5188.0
74.00	55864.0
75.00	87744.0
76.00	30512.0
77.00	623488.0
78.00	42768.0
79.00	40952.0
80.00	32320.0
81.00	42232.0
82.00	10471.0
83.00	9745.0
84.00	1603.0
85.00	6831.0
86.00	12803.0
87.00	5612.0
88.00	2023.0
88.90	945.0
90.00	192.0
91.00	9264.0
92.00	10813.0
93.00	70664.0
94.00	4872.0
95.00	876.0
96.10	2892.0
97.10	1353.0
98.00	53144.0
99.00	44536.0
100.00	3680.0
101.00	23304.0
101.90	1303.0
103.00	7454.0
104.00	13553.0
105.00	13773.0
106.00	4429.0
107.00	171968.0
108.00	27312.0
109.00	4359.0
110.00	305920.0
111.00	47616.0
112.00	5660.0
113.00	2006.0
114.10	701.0
115.00	697.0
116.00	8897.0
117.00	163776.0
118.00	12297.0
119.00	1322.0
120.00	2266.0
121.00	912.0
122.00	11566.0
123.00	16216.0
123.90	8485.0
125.00	7045.0
127.00	577920.0
128.00	43960.0
129.00	218368.0
130.00	19040.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

131.00	3553.0
131.90	1767.0
132.90	1058.0
134.00	6657.0
135.00	17512.0
136.00	7629.0
137.00	11313.0
138.00	2130.0
139.00	1146.0
140.00	3101.0
141.00	29552.0
142.00	9601.0
143.00	6570.0
144.00	2005.0
144.90	1616.0
146.00	5504.0
147.00	14733.0
148.00	38664.0
149.00	8727.0
150.00	2067.0
151.00	3412.0
151.80	2242.0
153.00	9142.0
154.00	6712.0
155.00	15879.0
156.00	23488.0
157.00	4594.0
157.90	5700.0
159.00	4120.0
160.00	8543.0
161.00	13092.0
162.00	3969.0
162.90	1173.0
164.00	1497.0
165.00	11209.0
166.00	8205.0
167.00	64096.0
168.00	39480.0
169.00	5545.0
170.00	1706.0
170.90	2359.0
172.00	5444.0
173.00	6403.0
174.00	12748.0
175.00	22040.0
176.00	6001.0
177.00	10008.0
178.00	3488.0
179.00	43080.0
180.00	31184.0
181.00	13711.0
182.10	1947.0
183.00	937.0
183.90	3028.0
185.00	20184.0
186.00	153152.0
187.00	43488.0
188.00	5498.0
189.00	9535.0
189.90	1629.0
191.00	5415.0
192.00	13910.0
193.00	15047.0
194.00	3365.0
195.10	1655.0
196.00	29328.0
197.00	6240.0
198.00	1081344.0
199.00	74944.0
200.00	6359.0
201.50	4894.0
202.90	8654.0
204.00	42400.0
205.00	70656.0
206.00	277760.0
207.00	40128.0
208.00	10246.0
209.00	3773.0
210.10	4416.0
210.90	10903.0
213.00	911.0
215.00	3469.0
216.00	6402.0
216.90	80872.0
218.00	10162.0
219.00	897.0
221.00	41536.0
221.80	11024.0
223.00	16880.0
224.00	163456.0
225.00	40112.0
226.00	5610.0
227.00	71312.0
228.00	9282.0
229.00	12598.0
230.00	1899.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

231.00	6155.0
232.00	1190.0
233.00	1124.0
234.00	4496.0
234.90	4310.0
235.90	2968.0
237.00	7558.0
237.90	1015.0
239.00	2982.0
240.00	2189.0
240.90	3642.0
242.00	8324.0
243.00	9954.0
244.00	121216.0
245.00	16166.0
245.90	24264.0
247.00	5007.0
248.00	1086.0
248.90	3598.0
250.00	801.0
251.00	1257.0
252.00	1085.0
253.00	2556.0
255.00	621632.0
256.00	81048.0
257.00	7886.0
258.00	36512.0
259.00	6091.0
259.90	1211.0
260.90	967.0
262.00	170.0
263.00	309.0
264.10	912.0
264.90	14813.0
266.00	1956.0
267.00	297.0
267.90	225.0
268.90	206.0
269.90	558.0
271.00	1926.0
272.00	1922.0
273.00	18496.0
274.00	45064.0
275.00	266688.0
276.00	37072.0
277.00	22704.0
278.00	4605.0
278.90	1015.0
281.00	1182.0
282.00	788.0
283.00	2516.0
284.00	2070.0
285.00	3998.0
285.90	657.0
287.90	248.0
289.00	900.0
290.00	640.0
290.90	718.0
292.00	1039.0
292.90	5436.0
294.00	1714.0
295.10	1553.0
296.00	78072.0
297.00	10391.0
298.00	905.0
298.80	181.0
301.00	1200.0
302.00	1398.0
303.00	8182.0
304.00	2960.0
305.00	574.0
307.90	1015.0
308.80	873.0
310.00	1144.0
310.90	179.0
311.90	316.0
313.00	906.0
314.00	3792.0
314.90	8416.0
316.00	5030.0
317.00	1214.0
318.80	270.0
320.00	385.0
321.00	2241.0
321.90	1204.0
323.00	23272.0
323.90	4016.0
324.80	505.0
326.00	401.0
326.90	4502.0
327.90	2271.0
328.90	365.0
329.90	161.0
330.80	164.0
332.00	1910.0
333.00	1992.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

334.00	15168.0
335.00	3630.0
336.00	632.0
338.80	401.0
339.80	301.0
341.00	2743.0
342.00	766.0
345.90	6333.0
346.90	1208.0
349.80	274.0
351.00	354.0
352.00	7065.0
352.90	5425.0
354.00	7916.0
355.00	1530.0
357.80	183.0
358.90	557.0
359.80	155.0
364.00	265.0
364.90	34080.0
365.90	5589.0
366.90	445.0
369.90	688.0
371.00	1693.0
372.00	11901.0
373.00	3622.0
374.00	260.0
376.80	327.0
382.90	3419.0
383.90	1098.0
385.10	289.0
389.90	1963.0
390.90	1148.0
391.90	836.0
396.80	173.0
400.80	673.0
402.00	4878.0
403.00	7158.0
404.00	2632.0
404.90	558.0
409.90	232.0
414.90	435.0
420.90	6441.0
422.00	6311.0
423.00	44200.0
423.90	12238.0
425.00	1749.0
425.90	201.0
438.00	277.0
438.60	302.0
440.00	633.0
441.00	114400.0
442.00	775744.0
443.00	157952.0
444.00	14974.0
445.00	916.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

m/z	Abundance
35.90	546.0
37.00	3101.0
38.00	7113.0
39.00	44496.0
40.00	4254.0
41.00	1522.0
42.20	279.0
42.90	811.0
44.00	10980.0
45.00	1254.0
46.00	211.0
47.00	363.0
48.00	438.0
49.00	4383.0
50.00	141632.0
51.00	543296.0
52.00	28256.0
53.00	1418.0
54.00	179.0
55.00	2844.0
56.00	15525.0
57.00	34576.0
58.00	1486.0
58.90	443.0
60.00	386.0
61.00	6425.0
62.00	6652.0
63.00	19296.0
64.00	2758.0
65.00	8822.0
65.90	726.0
67.10	895.0
68.00	7246.0
69.00	448000.0
70.00	2249.0
71.10	290.0
72.00	239.0
73.00	5060.0
74.00	48080.0
75.00	72632.0
76.00	26824.0
77.00	523776.0
78.00	36280.0
79.00	34904.0
80.00	24784.0
81.00	35816.0
82.00	8473.0
83.00	8335.0
84.00	1293.0
85.00	6245.0
86.00	10137.0
87.00	4383.0
87.90	1749.0
89.00	755.0
90.00	251.0
91.00	7538.0
92.00	9278.0
93.00	59608.0
94.00	3859.0
95.00	1058.0
96.00	2935.0
97.10	1005.0
98.00	44808.0
99.00	38168.0
100.00	2905.0
101.00	20008.0
101.90	1027.0
102.90	5822.0
103.90	12212.0
105.00	11526.0
106.00	3918.0
107.00	149760.0
108.00	22784.0
109.00	4200.0
110.00	266880.0
111.00	40488.0
112.00	5115.0
113.00	1772.0
114.00	390.0
115.00	887.0
116.00	8153.0
117.00	140224.0
118.00	10117.0
119.00	1114.0
120.00	2014.0
120.90	708.0
121.90	10076.0
123.00	15831.0
124.00	7515.0
124.90	5823.0
126.10	2153.0
127.00	506752.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

128.00	37728.0
129.00	198592.0
130.00	17272.0
131.00	3540.0
132.00	1539.0
133.00	793.0
134.00	5920.0
135.00	16968.0
136.00	6525.0
137.00	9856.0
138.00	1840.0
138.90	1129.0
140.00	2678.0
140.90	25528.0
142.00	9017.0
143.00	5539.0
144.00	1788.0
144.90	1531.0
146.00	4648.0
147.00	13612.0
148.00	34800.0
148.90	7098.0
150.00	2130.0
151.10	3374.0
151.70	2534.0
153.00	8167.0
154.00	6324.0
155.00	14789.0
156.00	23368.0
157.10	4796.0
158.00	4312.0
159.00	3620.0
160.00	7526.0
161.00	12572.0
162.00	3193.0
163.00	1152.0
163.90	1404.0
164.90	10232.0
166.00	8324.0
167.00	56368.0
168.00	33712.0
169.00	5643.0
170.00	1814.0
170.90	1961.0
172.00	5460.0
173.00	6792.0
174.00	11472.0
175.00	22360.0
176.00	5575.0
177.00	9657.0
178.00	3085.0
179.00	40584.0
180.00	28800.0
181.00	12362.0
182.00	2263.0
182.80	1162.0
184.00	3556.0
185.00	20592.0
186.00	151744.0
187.00	41640.0
188.00	4780.0
189.00	9586.0
189.90	1543.0
191.00	4607.0
192.00	12076.0
193.00	15178.0
194.00	2710.0
195.00	1751.0
196.00	30256.0
196.80	6064.0
198.00	1056256.0
199.00	74664.0
200.00	5720.0
201.40	4289.0
202.90	6900.0
204.00	39424.0
205.00	65104.0
206.00	280512.0
207.00	40752.0
208.00	10440.0
209.00	3626.0
210.20	4687.0
211.00	11753.0
212.90	553.0
214.00	384.0
215.00	3504.0
216.00	6308.0
216.90	79152.0
217.90	9382.0
219.10	773.0
220.00	1350.0
221.00	41640.0
223.00	16317.0
224.00	159552.0
225.00	40880.0
226.00	4397.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

227.00	69456.0
228.00	8799.0
229.00	14492.0
230.10	1577.0
231.00	5374.0
231.90	1200.0
233.00	780.0
234.00	3776.0
234.90	5252.0
236.00	3612.0
237.00	7193.0
238.00	853.0
239.00	2402.0
239.90	2187.0
241.00	3681.0
242.00	8160.0
243.00	9223.0
244.00	121704.0
245.00	17312.0
246.00	24296.0
247.00	5125.0
247.90	1197.0
248.90	4197.0
250.00	840.0
251.00	1280.0
252.00	1189.0
252.90	2848.0
255.00	634752.0
256.00	95344.0
257.00	6274.0
258.00	34824.0
259.00	5636.0
260.10	1098.0
261.00	974.0
262.00	167.0
262.90	404.0
263.90	1131.0
265.00	16062.0
266.00	2001.0
266.90	356.0
269.90	595.0
271.00	1650.0
271.90	1702.0
273.00	18904.0
274.00	49528.0
275.00	269760.0
276.00	38320.0
276.90	25864.0
278.00	3749.0
278.90	917.0
281.00	747.0
282.00	884.0
283.00	2590.0
284.00	1999.0
284.90	4508.0
286.00	845.0
287.90	268.0
288.90	990.0
290.00	1000.0
290.90	995.0
291.90	1221.0
292.90	5061.0
294.00	1745.0
295.00	1722.0
296.00	88520.0
297.00	11745.0
298.10	769.0
298.90	245.0
301.00	1220.0
301.90	1371.0
303.00	9437.0
304.00	2753.0
304.90	362.0
308.00	1184.0
308.80	897.0
309.90	1124.0
310.90	223.0
311.90	273.0
313.00	596.0
314.00	3575.0
314.90	9112.0
316.00	6050.0
316.90	922.0
317.90	196.0
319.90	394.0
320.90	2796.0
322.00	1306.0
323.00	25656.0
324.00	5193.0
324.90	574.0
326.00	770.0
327.00	4962.0
327.90	2987.0
329.00	455.0
330.90	198.0
332.00	2017.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321

333.00	2652.0
334.00	17680.0
335.00	4443.0
335.90	425.0
339.00	394.0
339.90	460.0
341.00	3555.0
342.10	782.0
342.80	166.0
345.90	6491.0
346.80	1062.0
347.90	254.0
350.20	217.0
352.00	8652.0
353.00	5310.0
354.00	8614.0
355.00	1751.0
356.00	263.0
356.90	153.0
357.90	185.0
358.90	862.0
359.90	258.0
361.00	351.0
364.90	40608.0
365.90	5843.0
366.90	345.0
369.90	889.0
370.90	2498.0
372.00	14246.0
373.00	3809.0
374.00	411.0
377.00	393.0
382.90	4412.0
383.90	1133.0
385.00	273.0
390.00	1869.0
390.90	1282.0
392.00	1004.0
393.00	204.0
394.90	177.0
396.70	217.0
400.90	924.0
401.90	7123.0
402.90	8698.0
404.00	2970.0
405.00	605.0
410.00	271.0
414.90	415.0
421.00	7548.0
422.00	8451.0
423.00	50832.0
424.00	13837.0
424.90	1440.0
438.00	338.0
439.00	393.0
439.90	531.0
441.00	149824.0
442.00	925952.0
443.00	195584.0
444.00	16560.0
445.10	794.0

Instrument :
BNA_N
ClientSampleId :
DFTPP321