

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041324\
 Data File : BN030905.D
 Acq On : 14 Apr 2024 07:48
 Operator : MA/JU
 Sample : PB160192BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS192

Quant Time: Apr 14 22:21:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 14 22:12:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	145367	20.000	ng/u1	0.00
20) Naphthalene-d8	10.446	136	649822	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.310	164	433858	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.057	188	949982	20.000	ng/u1	0.00
79) Chrysene-d12	21.298	240	1067017	20.000	ng/u1	0.00
88) Perylene-d12	24.215	264	1198018	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	28844	9.303	ng/uL	0.00
4) Pyridine-d5	3.581	84	339530	36.524	ng/u1	0.00
7) Phenol-d5	6.834	99	456319	38.804	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.005	67	260933	35.925	ng/u1	0.00
11) 2-Chlorophenol-d4	7.199	132	360877	40.453	ng/u1	0.00
15) 4-Methylphenol-d8	8.369	113	362230	36.754	ng/u1	0.00
21) Nitrobenzene-d5	8.822	128	179047	39.720	ng/u1	0.00
24) 2-Nitrophenol-d4	9.534	143	179950	43.637	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.069	165	362614	40.166	ng/u1	0.00
31) 4-Chloroaniline-d4	10.587	131	526732	35.440	ng/u1	0.00
46) Dimethylphthalate-d6	13.722	166	1063197	36.704	ng/u1	0.00
49) Acenaphthylene-d8	13.998	160	1302536	39.109	ng/u1	0.00
54) 4-Nitrophenol-d4	14.516	143	240450	43.397	ng/u1	0.00
60) Fluorene-d10	15.304	176	990480	39.075	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.428	200	152831	36.448	ng/u1	0.00
73) Anthracene-d10	17.157	188	1613113	39.410	ng/u1	0.00
81) Pyrene-d10	19.457	212	2064546	35.988	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.015	264	2343422	42.452	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.211	88	40353	12.150	ng/uL	96
5) Pyridine	3.599	79	343598	36.630	ng/u1	98
6) Benzaldehyde	6.816	77	239085	43.658	ng/u1	97
8) Phenol	6.863	94	477456	38.974	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.099	93	351028	35.139	ng/u1	99
12) 2-Chlorophenol	7.228	128	380614	39.773	ng/u1	97
13) 2-Methylphenol	8.105	108	346937	36.752	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.181	45	459640	32.676	ng/u1	100
16) Acetophenone	8.487	105	552882	35.530	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.475	70	265480	33.874	ng/u1	98
18) 4-Methylphenol	8.434	108	393365	37.223	ng/u1	99
19) Hexachloroethane	8.734	117	146255	36.453	ng/u1	97
22) Nitrobenzene	8.863	77	411299	37.377	ng/u1	100
23) Isophorone	9.387	82	759055	34.588	ng/u1	99
25) 2-Nitrophenol	9.569	139	201903	41.659	ng/u1	98
26) 2,4-Dimethylphenol	9.628	107	356728	33.117	ng/u1	100
27) Bis(2-Chloroethoxy)met...	9.869	93	482588	34.587	ng/u1	98
29) 2,4-Dichlorophenol	10.093	162	365410	39.654	ng/u1	98
30) Naphthalene	10.493	128	1213323	36.405	ng/u1	99
32) 4-Chloroaniline	10.610	127	498002	34.768	ng/u1	97
33) Hexachlorobutadiene	10.775	225	203378	35.064	ng/u1	98
34) Caprolactam	11.393	113	78055	22.674	ng/u1	99
35) 4-Chloro-3-methylphenol	11.740	107	400847	38.622	ng/u1	100
36) 2-Methylnaphthalene	12.110	142	838230	36.014	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041324\
 Data File : BN030905.D
 Acq On : 14 Apr 2024 07:48
 Operator : MA/JU
 Sample : PB160192BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SLCS192


Quant Time: Apr 14 22:21:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 14 22:12:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.334	142	864232	36.578	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.481	216	423819	37.853	ng/ul	96
40) Hexachlorocyclopentadiene	12.457	237	155342	24.431	ng/ul	97
41) 2,4,6-Trichlorophenol	12.728	196	280011	40.021	ng/ul	95
42) 2,4,5-Trichlorophenol	12.798	196	320182	40.521	ng/ul	98
43) 1,1'-Biphenyl	13.140	154	1109153	37.507	ng/ul	99
44) 2-Chloronaphthalene	13.175	162	863722	37.103	ng/ul	98
45) 2-Nitroaniline	13.392	65	256575	41.416	ng/ul	99
47) Dimethylphthalate	13.775	163	1093237	36.536	ng/ul	99
48) 2,6-Dinitrotoluene	13.892	165	219866	39.631	ng/ul	98
50) Acenaphthylene	14.028	152	1441245	37.453	ng/ul	100
51) 3-Nitroaniline	14.222	138	259105	44.002	ng/ul	98
52) Acenaphthene	14.375	153	935041	36.462	ng/ul	99
53) 2,4-Dinitrophenol	14.428	184	88550	34.751	ng/ul	95
55) 4-Nitrophenol	14.534	109	185754	42.315	ng/ul	95
56) Dibenzofuran	14.710	168	1336887	37.504	ng/ul	97
57) 2,4-Dinitrotoluene	14.681	165	341467	42.158	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.934	232	273462	41.671	ng/ul	95
59) Diethylphthalate	15.145	149	1105120	36.355	ng/ul	100
61) Fluorene	15.363	166	1069537	37.748	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.357	204	504898	36.321	ng/ul	97
63) 4-Nitroaniline	15.392	138	297705	52.003	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.445	198	171500	38.064	ng/ul	95
67) N-Nitrosodiphenylamine	15.575	169	969610	37.478	ng/ul	98
68) 4-Bromophenyl-phenylether	16.251	248	320173	35.493	ng/ul	92
69) Hexachlorobenzene	16.357	284	380827	35.852	ng/ul	98
70) Atrazine	16.533	200	322946	36.014	ng/ul	98
71) Pentachlorophenol	16.704	266	247687	39.548	ng/ul	98
72) Phenanthrene	17.098	178	1874029	38.396	ng/ul	100
74) Anthracene	17.192	178	1860053	37.698	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.098	216	431925	37.009	ng/ul	97
76) Pentachlorobenzene	14.622	250	410679	34.643	ng/ul	92
77) Carbazole	17.463	167	1857946	42.938	ng/ul	99
78) Di-n-butylphthalate	18.033	149	2103059	39.841	ng/ul	100
80) Fluoranthene	19.122	202	2347482	34.423	ng/ul	97
82) Pyrene	19.486	202	2572941	35.725	ng/ul	100
83) Butylbenzylphthalate	20.398	149	1077451	39.174	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.210	252	775744	39.567	ng/ul	94
85) Benzo(a)anthracene	21.274	228	2704115	37.973	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.204	149	1515058	37.353	ng/ul	98
87) Chrysene	21.339	228	2602821	38.724	ng/ul	99
89) Di-n-octyl phthalate	22.316	149	2869149	37.818	ng/ul	100
90) Benzo(b)fluoranthene	23.292	252	2822204	40.028	ng/ul	98
91) Benzo(k)fluoranthene	23.357	252	2876963	40.180	ng/ul	98
93) Benzo(a)pyrene	24.080	252	2351638	35.309	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	27.498	276	3036071	42.170	ng/ul	99
95) Dibenzo(a,h)anthracene	27.562	278	2492178	41.119	ng/ul	96
96) Benzo(g,h,i)perylene	28.509	276	1149873	20.344	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 14 22:21:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Apr 14 22:12:55 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 12.150 ng/uL
 RT: 3.211 min Scan# 38
 Delta R.T. -0.000 min
 Lab File: BN030905.D
 Acq: 14 Apr 2024 07:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS192


Tgt Ion:	88	Resp:	40353
Ion	Ratio	Lower	Upper
88	100		
43	31.7	23.8	35.8
58	66.9	56.6	84.8



#5
 Pyridine
 Concen: 36.630 ng/uL
 RT: 3.599 min Scan# 104
 Delta R.T. -0.000 min
 Lab File: BN030905.D
 Acq: 14 Apr 2024 07:48

Tgt Ion:	79	Resp:	343598
Ion	Ratio	Lower	Upper
79	100		
52	72.6	56.2	84.4
51	32.8	27.0	40.4









#6
Benzaldehyde
Concen: 43.658 ng/ul
RT: 6.816 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion: 77 Resp: 239085
Ion Ratio Lower Upper
77 100
105 100.7 78.1 117.1
106 97.0 75.6 113.4



#8
Phenol
Concen: 38.974 ng/ul
RT: 6.863 min Scan# 659
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion: 94 Resp: 477456
Ion Ratio Lower Upper
94 100
65 26.8 21.4 32.0
66 35.4 28.5 42.7









#10
Bis(2-Chloroethyl)ether
Concen: 35.139 ng/ul
RT: 7.099 min Scan# 699
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion: 93 Resp: 351028
Ion Ratio Lower Upper
93 100
63 71.1 58.2 87.4
95 32.4 26.0 39.0



#12
2-Chlorophenol
Concen: 39.773 ng/ul
RT: 7.228 min Scan# 721
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:128 Resp: 380614
Ion Ratio Lower Upper
128 100
64 44.0 37.0 55.4
130 32.0 26.3 39.5









#13
2-Methylphenol
Concen: 36.752 ng/ul
RT: 8.105 min Scan# 81
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:108 Resp: 346937
Ion Ratio Lower Upper
108 100
107 87.8 71.8 107.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.676 ng/ul
RT: 8.181 min Scan# 883
Delta R.T. -0.006 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion: 45 Resp: 459640
Ion Ratio Lower Upper
45 100
77 15.1 12.1 18.1
79 10.7 8.6 13.0









#16
 Acetophenone
 Concen: 35.530 ng/ul
 RT: 8.487 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: BN030905.D
 Acq: 14 Apr 2024 07:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS192


Tgt Ion:105 Resp: 552882
 Ion Ratio Lower Upper
 105 100
 77 75.6 61.2 91.8
 51 25.3 21.0 31.4



#17
 N-Nitroso-di-n-propylamine
 Concen: 33.874 ng/ul
 RT: 8.475 min Scan# 933
 Delta R.T. 0.006 min
 Lab File: BN030905.D
 Acq: 14 Apr 2024 07:48

Tgt Ion: 70 Resp: 265480
 Ion Ratio Lower Upper
 70 100
 42 57.0 47.4 71.0
 101 10.1 7.9 11.9
 130 23.9 19.0 28.4









#18
4-Methylphenol
Concen: 37.223 ng/ul
RT: 8.434 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:108 Resp: 393365
Ion Ratio Lower Upper
108 100
107 109.1 88.1 132.1



#19
Hexachloroethane
Concen: 36.453 ng/ul
RT: 8.734 min Scan# 977
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:117 Resp: 146255
Ion Ratio Lower Upper
117 100
201 99.5 76.5 114.7
199 59.3 48.5 72.7









#22
Nitrobenzene
Concen: 37.377 ng/ul
RT: 8.863 min Scan# 999
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion: 77 Resp: 411299
Ion Ratio Lower Upper
77 100
123 46.8 37.5 56.3
65 13.3 10.6 15.8



#23
Isophorone
Concen: 34.588 ng/ul
RT: 9.387 min Scan# 1088
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion: 82 Resp: 759055
Ion Ratio Lower Upper
82 100
95 7.1 5.7 8.5
138 18.6 14.6 21.8









#25
2-Nitrophenol
Concen: 41.659 ng/ul
RT: 9.569 min Scan# 1119
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:139 Resp: 201903
Ion Ratio Lower Upper
139 100
65 47.5 39.5 59.3
109 29.7 23.8 35.8



#26
2,4-Dimethylphenol
Concen: 33.117 ng/ul
RT: 9.628 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:107 Resp: 356728
Ion Ratio Lower Upper
107 100
121 52.4 41.6 62.4
122 88.9 71.3 106.9





#27

Bis(2-Chloroethoxy)methane

Concen: 34.587 ng/ul

RT: 9.869 min Scan# 1170

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion: 93 Resp: 482588

Ion Ratio Lower Upper

93 100

95 33.0 25.4 38.0

123 12.1 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 39.654 ng/ul

RT: 10.093 min Scan# 1208

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion: 162 Resp: 365410


Ion Ratio Lower Upper

162 100

164 66.6 51.8 77.8

98 36.0 28.0 42.0









#30
Naphthalene
Concen: 36.405 ng/ul
RT: 10.493 min Scan# 1277
Delta R.T. -0.006 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:128 Resp: 1213323
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.6 10.5 15.7



#32
4-Chloroaniline
Concen: 34.768 ng/ul
RT: 10.610 min Scan# 1296
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:127 Resp: 498002
Ion Ratio Lower Upper
127 100
129 33.1 25.3 37.9









#33
Hexachlorobutadiene
Concen: 35.064 ng/ul
RT: 10.775 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:225 Resp: 203378
Ion Ratio Lower Upper
225 100
223 64.3 50.9 76.3
227 62.6 51.6 77.4



#34
Caprolactam
Concen: 22.674 ng/ul
RT: 11.393 min Scan# 1429
Delta R.T. 0.006 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:113 Resp: 78055
Ion Ratio Lower Upper
113 100
55 155.7 124.7 187.1
56 116.0 90.6 135.8





#35

4-Chloro-3-methylphenol

Concen: 38.622 ng/ul

RT: 11.740 min Scan# 1488

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:107 Resp: 400847

Ion Ratio Lower Upper

107 100

144 26.3 21.4 32.0

142 82.3 65.9 98.9



#36

2-Methylnaphthalene

Concen: 36.014 ng/ul

RT: 12.110 min Scan# 1551

Delta R.T. -0.006 min

Lab File: BN030905.D




Acq: 14 Apr 2024 07:48


Tgt Ion:142 Resp: 838230

Ion Ratio Lower Upper

142 100

141 87.5 70.8 106.2





#37

1-Methylnaphthalene

Concen: 36.578 ng/ul

RT: 12.334 min Scan# 1

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:142 Resp: 864232

Ion Ratio Lower Upper

142 100

141 89.0 73.0 109.4

116 3.9 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 37.853 ng/ul

RT: 12.481 min Scan# 1614

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Tgt Ion:216 Resp: 423819




Ion Ratio Lower Upper


216 100

214 80.6 60.9 91.3

179 23.4 17.4 26.0

108 18.2 14.2 21.4





#40

Hexachlorocyclopentadiene

Concen: 24.431 ng/ul

RT: 12.457 min Scan# 1

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:237 Resp: 155342

Ion Ratio Lower Upper

237 100

235 65.2 50.6 76.0

272 13.9 10.2 15.4



#41

2,4,6-Trichlorophenol

Concen: 40.021 ng/ul

RT: 12.728 min Scan# 1656

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion:196 Resp: 280011


Ion Ratio Lower Upper

196 100

198 99.7 74.6 111.8

200 31.9 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 40.521 ng/ul
RT: 12.798 min Scan# 1668
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:196 Resp: 320182
Ion Ratio Lower Upper
196 100
198 89.5 73.0 109.6
200 29.9 25.2 37.8



#43
1,1'-Biphenyl
Concen: 37.507 ng/ul
RT: 13.140 min Scan# 1726
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:154 Resp: 1109153
Ion Ratio Lower Upper
154 100
153 38.7 31.5 47.3
76 12.4 9.8 14.6





#44

2-Chloronaphthalene

Concen: 37.103 ng/ul

RT: 13.175 min Scan# 1

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:162 Resp: 863722

Ion Ratio Lower Upper

162 100

164 33.7 25.7 38.5

127 40.7 32.2 48.2



#45

2-Nitroaniline

Concen: 41.416 ng/ul

RT: 13.392 min Scan# 1769

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion: 65 Resp: 256575


Ion Ratio Lower Upper

65 100

92 71.4 57.7 86.5

138 114.5 90.6 135.8









#47
Dimethylphthalate
Concen: 36.536 ng/ul
RT: 13.775 min Scan# 1834
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:163 Resp: 1093237
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 9.8 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 39.631 ng/ul
RT: 13.892 min Scan# 1854
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:165 Resp: 219866
Ion Ratio Lower Upper
165 100
89 55.7 43.5 65.3
121 20.8 15.7 23.5





#50

Acenaphthylene

Concen: 37.453 ng/ul

RT: 14.028 min Scan# 1877

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:152 Resp: 1441245

Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.8

153 13.1 10.7 16.1



#51

3-Nitroaniline

Concen: 44.002 ng/ul

RT: 14.222 min Scan# 1910

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion:138 Resp: 259105


Ion Ratio Lower Upper

138 100

108 11.8 9.6 14.4

92 113.8 89.6 134.4









#52
Acenaphthene
Concen: 36.462 ng/ul
RT: 14.375 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:153 Resp: 935041
Ion Ratio Lower Upper
153 100
152 47.7 37.6 56.4
154 86.4 70.3 105.5



#53
2,4-Dinitrophenol
Concen: 34.751 ng/ul
RT: 14.428 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:184 Resp: 88550
Ion Ratio Lower Upper
184 100
63 65.4 49.9 74.9
154 65.9 49.5 74.3









#55
4-Nitrophenol
Concen: 42.315 ng/ul
RT: 14.534 min Scan# 1963
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:109 Resp: 185754
Ion Ratio Lower Upper
109 100
139 144.0 111.8 167.6
65 131.2 98.7 148.1



#56
Dibenzofuran
Concen: 37.504 ng/ul
RT: 14.710 min Scan# 1993
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:168 Resp: 1336887
Ion Ratio Lower Upper
168 100
139 36.3 30.2 45.4









#57
2,4-Dinitrotoluene
Concen: 42.158 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:165 Resp: 341467
Ion Ratio Lower Upper
165 100
63 40.4 34.7 52.1
182 2.8 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.671 ng/ul
RT: 14.934 min Scan# 2031
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:232 Resp: 273462
Ion Ratio Lower Upper
232 100
131 50.0 36.6 55.0
130 2.5 2.2 3.2
166 34.5 26.0 39.0









#59
Diethylphthalate
Concen: 36.355 ng/ul
RT: 15.145 min Scan# 2067
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:149 Resp: 1105120
Ion Ratio Lower Upper
149 100
177 21.7 17.4 26.2
150 12.2 9.8 14.8



#61
Fluorene
Concen: 37.748 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:166 Resp: 1069537
Ion Ratio Lower Upper
166 100
165 97.9 78.2 117.4
167 13.7 11.0 16.4









#62
 4-Chlorophenyl-phenylether
 Concen: 36.321 ng/ul
 RT: 15.357 min Scan# 2103
 Delta R.T. -0.006 min
 Lab File: BN030905.D
 Acq: 14 Apr 2024 07:48

Instrument :
 BNA_N
 ClientSampleId :
 SLCS192

Tgt Ion:204 Resp: 504898




Ion	Ratio	Lower	Upper
204	100		
206	32.9	25.8	38.6
141	64.8	49.1	73.7




#63
 4-Nitroaniline
 Concen: 52.003 ng/ul
 RT: 15.392 min Scan# 2109
 Delta R.T. -0.000 min
 Lab File: BN030905.D
 Acq: 14 Apr 2024 07:48

Tgt Ion:138 Resp: 297705

Ion	Ratio	Lower	Upper
138	100		
92	50.6	39.7	59.5
108	76.2	62.6	94.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 38.064 ng/ul
 RT: 15.445 min Scan# 2118
 Delta R.T. 0.006 min
 Lab File: BN030905.D
 Acq: 14 Apr 2024 07:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS192


Tgt Ion:198 Resp: 171500
 Ion Ratio Lower Upper
 198 100
 182 3.9 3.6 5.4
 77 24.6 22.2 33.4



#67
 N-Nitrosodiphenylamine
 Concen: 37.478 ng/ul
 RT: 15.575 min Scan# 2140
 Delta R.T. -0.000 min
 Lab File: BN030905.D
 Acq: 14 Apr 2024 07:48

Tgt Ion:169 Resp: 969610
 Ion Ratio Lower Upper
 169 100
 168 65.9 51.9 77.9
 167 33.9 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 35.493 ng/ul
RT: 16.251 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:248 Resp: 320173
Ion Ratio Lower Upper
248 100
250 103.7 80.0 120.0
141 86.5 60.5 90.7



#69
Hexachlorobenzene
Concen: 35.852 ng/ul
RT: 16.357 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:284 Resp: 380827
Ion Ratio Lower Upper
284 100
142 36.7 28.6 43.0
249 29.5 25.0 37.4





#70

Atrazine

Concen: 36.014 ng/ul

RT: 16.533 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:200 Resp: 322946

Ion Ratio Lower Upper

200 100

173 27.3 21.5 32.3

215 49.5 37.8 56.8



#71

Pentachlorophenol

Concen: 39.548 ng/ul

RT: 16.704 min Scan# 2332

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion:266 Resp: 247687


Ion Ratio Lower Upper

266 100

264 62.1 51.0 76.6

268 63.5 52.5 78.7









#72
Phenanthrene
Concen: 38.396 ng/ul
RT: 17.098 min Scan# 2400
Delta R.T. -0.006 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:178 Resp: 1874029
Ion Ratio Lower Upper
178 100
179 15.3 12.4 18.6
176 18.9 15.0 22.6



#74
Anthracene
Concen: 37.698 ng/ul
RT: 17.192 min Scan# 2415
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:178 Resp: 1860053
Ion Ratio Lower Upper
178 100
179 15.5 12.0 18.0
176 19.0 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 37.009 ng/uL

RT: 13.098 min Scan# 1719

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:216 Resp: 431925

Ion Ratio Lower Upper

216 100

214 76.4 63.1 94.7

218 47.7 39.5 59.3



#76

Pentachlorobenzene

Concen: 34.643 ng/uL

RT: 14.622 min Scan# 1978

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion:250 Resp: 410679


Ion Ratio Lower Upper

250 100

248 64.7 45.8 68.6

252 65.9 48.7 73.1





#77

Carbazole

Concen: 42.938 ng/ul

RT: 17.463 min Scan# 2461

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:167 Resp: 1857946

Ion Ratio Lower Upper

167 100

166 21.0 16.6 24.8

139 13.6 10.5 15.7



#78

Di-n-butylphthalate

Concen: 39.841 ng/ul

RT: 18.033 min Scan# 2558

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion:149 Resp: 2103059


Ion Ratio Lower Upper

149 100

150 9.0 7.2 10.8

104 5.1 4.1 6.1









#80
Fluoranthene
Concen: 34.423 ng/ul
RT: 19.122 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:202 Resp: 2347482
Ion Ratio Lower Upper
202 100
101 12.5 9.0 13.6
100 9.4 6.8 10.2



#82
Pyrene
Concen: 35.725 ng/ul
RT: 19.486 min Scan# 2805
Delta R.T. -0.006 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:202 Resp: 2572941
Ion Ratio Lower Upper
202 100
101 14.1 11.3 16.9
100 11.4 9.2 13.8









#83
Butylbenzylphthalate
Concen: 39.174 ng/ul
RT: 20.398 min Scan# 2960
Delta R.T. -0.006 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:149 Resp: 1077451
Ion Ratio Lower Upper
149 100
91 70.2 58.2 87.2
206 20.8 15.4 23.0



#84
3,3'-Dichlorobenzidine
Concen: 39.567 ng/ul
RT: 21.210 min Scan# 3098
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:252 Resp: 775744
Ion Ratio Lower Upper
252 100
254 69.0 51.1 76.7
126 13.0 9.3 13.9





#85

Benzo(a)anthracene

Concen: 37.973 ng/ul

RT: 21.274 min Scan# 3110

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:228 Resp: 2704115

Ion Ratio Lower Upper

228 100

229 19.1 15.8 23.6

226 26.1 20.5 30.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 37.353 ng/ul

RT: 21.204 min Scan# 3097

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion:149 Resp: 1515058


Ion Ratio Lower Upper

149 100

167 27.7 21.4 32.0

279 3.9 3.5 5.3





#87

Chrysene

Concen: 38.724 ng/ul

RT: 21.339 min Scan# 3120

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:228 Resp: 2602821

Ion Ratio Lower Upper

228 100

226 31.3 24.3 36.5

229 19.6 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 37.818 ng/ul




RT: 22.316 min Scan# 3286


Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Tgt Ion:149 Resp: 2869149









#90
Benzo(b)fluoranthene
Concen: 40.028 ng/ul
RT: 23.292 min Scan# 3453
Delta R.T. -0.006 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48




Instrument :
BNA_N
ClientSampleId :
SLCS192


Tgt Ion:252 Resp: 2822204
Ion Ratio Lower Upper
252 100
253 21.8 16.7 25.1
125 11.7 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 40.180 ng/ul
RT: 23.357 min Scan# 3463
Delta R.T. -0.000 min
Lab File: BN030905.D
Acq: 14 Apr 2024 07:48

Tgt Ion:252 Resp: 2876963
Ion Ratio Lower Upper
252 100
253 20.0 16.9 25.3
125 11.3 8.5 12.7





#93

Benzo(a)pyrene

Concen: 35.309 ng/ul

RT: 24.080 min Scan# 3

Delta R.T. -0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:252 Resp: 2351638

Ion Ratio Lower Upper

252 100

253 21.1 16.7 25.1

125 12.2 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 42.170 ng/ul

RT: 27.498 min Scan# 4167

Delta R.T. -0.000 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48




Tgt Ion:276 Resp: 3036071


Ion Ratio Lower Upper

276 100

138 22.9 17.8 26.8

277 22.8 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 41.119 ng/ul

RT: 27.562 min Scan# 41

Delta R.T. 0.006 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Instrument :

BNA_N

ClientSampleId :

SLCS192





Tgt Ion:278 Resp: 2492178

Ion Ratio Lower Upper

278 100

139 17.8 13.2 19.8

279 24.1 17.4 26.2



#96

Benzo(g,h,i)perylene

Concen: 20.344 ng/ul

RT: 28.509 min Scan# 4339

Delta R.T. -0.018 min

Lab File: BN030905.D

Acq: 14 Apr 2024 07:48

Tgt Ion:276 Resp: 1149873

Ion Ratio Lower Upper

276 100

138 25.9 19.3 28.9

277 24.4 20.1 30.1

