

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041524\
 Data File : BN030943.D
 Acq On : 15 Apr 2024 17:44
 Operator : MA/JU
 Sample : PB159883BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB159883BL

Quant Time: Apr 15 19:22:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 15 17:03:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.653	152	3444	0.400	ng	0.00
7) Naphthalene-d8	10.432	136	9707	0.400	ng	0.00
13) Acenaphthene-d10	14.295	164	5071	0.400	ng	0.00
19) Phenanthrene-d10	17.040	188	12417	0.400	ng	0.00
29) Chrysene-d12	21.240	240	10815	0.400	ng	0.00
35) Perylene-d12	23.457	264	11022	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.248	112	3424	0.446	ng	0.00
5) Phenol-d6	6.830	99	3878	0.417	ng	0.00
8) Nitrobenzene-d5	8.809	82	2678	0.399	ng	0.01
11) 2-Methylnaphthalene-d10	12.028	152	5613	0.376	ng	0.00
14) 2,4,6-Tribromophenol	15.787	330	556	0.283	ng	0.00
15) 2-Fluorobiphenyl	12.916	172	8252	0.484	ng	0.01
27) Fluoranthene-d10	19.079	212	13969	0.418	ng	0.00
31) Terphenyl-d14	19.683	244	12195	0.486	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041524\
Data File : BN030943.D
Acq On : 15 Apr 2024 17:44
Operator : MA/JU
Sample : PB159883BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB159883BL

Quant Time: Apr 15 19:22:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 15 17:03:47 2024
Response via : Initial Calibration

