

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041524\
 Data File : BN030950.D
 Acq On : 15 Apr 2024 23:58
 Operator : MA/JU
 Sample : PB160202BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB160202BL

Quant Time: Apr 16 00:26:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 15 17:03:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.653	152	3953	0.400	ng	0.00
7) Naphthalene-d8	10.432	136	11347	0.400	ng	0.00
13) Acenaphthene-d10	14.295	164	5967	0.400	ng	0.00
19) Phenanthrene-d10	17.040	188	14502	0.400	ng	0.00
29) Chrysene-d12	21.231	240	11765	0.400	ng	# 0.00
35) Perylene-d12	23.451	264	8850	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.248	112	3362	0.382	ng	0.00
5) Phenol-d6	6.830	99	3738	0.350	ng	0.00
8) Nitrobenzene-d5	8.809	82	2437	0.310	ng	0.01
11) 2-Methylnaphthalene-d10	12.024	152	5719	0.327	ng	0.00
14) 2,4,6-Tribromophenol	15.787	330	682	0.295	ng	0.00
15) 2-Fluorobiphenyl	12.916	172	7313	0.365	ng	0.01
27) Fluoranthene-d10	19.075	212	13511	0.346	ng	0.00
31) Terphenyl-d14	19.683	244	9960	0.365	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041524\
Data File : BN030950.D
Acq On : 15 Apr 2024 23:58
Operator : MA/JU
Sample : PB160202BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB160202BL

Quant Time: Apr 16 00:26:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 15 17:03:47 2024
Response via : Initial Calibration

