

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041524\
 Data File : BN030959.D
 Acq On : 16 Apr 2024 05:25
 Operator : MA/JU
 Sample : P2073-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RSB-01A-17-041024

Quant Time: Apr 16 06:34:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 15 17:03:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.646	152	5565	0.400	ng	0.00
7) Naphthalene-d8	10.421	136	17746	0.400	ng	#-0.01
13) Acenaphthene-d10	14.284	164	10721	0.400	ng	-0.01
19) Phenanthrene-d10	17.040	188	25677	0.400	ng	0.00
29) Chrysene-d12	21.231	240	24426	0.400	ng	# 0.00
35) Perylene-d12	23.454	264	24870	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.241	112	1955	0.158	ng	0.00
5) Phenol-d6	6.823	99	1677	0.112	ng	0.00
8) Nitrobenzene-d5	8.798	82	3490	0.284	ng	0.00
11) 2-Methylnaphthalene-d10	12.021	152	8586	0.314	ng	0.00
14) 2,4,6-Tribromophenol	15.787	330	1590	0.382	ng	0.00
15) 2-Fluorobiphenyl	12.905	172	10056	0.279	ng	0.00
27) Fluoranthene-d10	19.070	212	26560	0.384	ng	0.00
31) Terphenyl-d14	19.679	244	22251	0.392	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.475	128	2040	0.042	ng	# 89
12) 2-Methylnaphthalene	12.096	142	1009	0.031	ng	# 92
17) Acenaphthene	14.348	154	1008	0.030	ng	99
18) Fluorene	15.343	166	909	0.021	ng	# 98
25) Phenanthrene	17.078	178	2435	0.032	ng	98
34) Bis(2-ethylhexyl)phtha...	21.160	149	9318	0.239	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041524\
Data File : BN030959.D
Acq On : 16 Apr 2024 05:25
Operator : MA/JU
Sample : P2073-02
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RSB-01A-17-041024

Quant Time: Apr 16 06:34:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 15 17:03:47 2024
Response via : Initial Calibration

