

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041524\
 Data File : BN030942.D
 Acq On : 15 Apr 2024 17:07
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN041524

Quant Time: Apr 15 17:33:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 15 17:03:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.391	2.3	102	0.00
3	n-Nitrosodimethylamine	0.400	0.377	5.8	100	0.00
4 S	2-Fluorophenol	0.400	0.376	6.0	99	0.00
5 S	Phenol-d6	0.400	0.359	10.3	98	0.00
6	bis(2-Chloroethyl)ether	0.400	0.368	8.0	98	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	99	0.00
8 S	Nitrobenzene-d5	0.400	0.367	8.3	98	0.00
9	Naphthalene	0.400	0.389	2.8	100	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	102	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	100	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.400	0.341	14.8	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.401	-0.3	101	0.00
16	Acenaphthylene	0.400	0.361	9.8	97	0.00
17	Acenaphthene	0.400	0.384	4.0	98	0.00
18	Fluorene	0.400	0.381	4.8	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.384	4.0	96	0.00
21	4-Bromophenyl-phenylether	0.400	0.372	7.0	99	0.00
22	Hexachlorobenzene	0.400	0.379	5.3	98	0.00
23	Atrazine	0.400	0.341	14.8	97	0.01
24	Pentachlorophenol	0.400	0.385	3.8	95	0.00
25	Phenanthrene	0.400	0.372	7.0	98	0.00
26	Anthracene	0.400	0.347	13.3	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.362	9.5	97	0.00
28	Fluoranthene	0.400	0.358	10.5	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
30	Pyrene	0.400	0.383	4.3	97	0.00
31 S	Terphenyl-d14	0.400	0.381	4.8	99	0.00
32	Benzo(a)anthracene	0.400	0.356	11.0	96	0.00
33	Chrysene	0.400	0.385	3.8	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.334	16.5	93	0.00
35 I	Perylene-d12	0.400	0.400	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.364	9.0	96	0.00
37	Benzo(b)fluoranthene	0.400	0.380	5.0	99	0.00
38	Benzo(k)fluoranthene	0.400	0.377	5.8	98	0.00
39 C	Benzo(a)pyrene	0.400	0.347	13.3	94	0.00
40	Dibenzo(a,h)anthracene	0.400	0.367	8.3	97	0.00
41	Benzo(g,h,i)perylene	0.400	0.379	5.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN041524