

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041524\
 Data File : BN030968.D
 Acq On : 16 Apr 2024 13:00
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 16 14:14:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 15 17:03:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	-0.01
2	1,4-Dioxane	0.400	0.361	9.8	96	0.00
3	n-Nitrosodimethylamine	0.400	0.375	6.3	102	-0.01
4 S	2-Fluorophenol	0.400	0.390	2.5	104	-0.01
5 S	Phenol-d6	0.400	0.382	4.5	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	-0.01
8 S	Nitrobenzene-d5	0.400	0.369	7.8	101	-0.01
9	Naphthalene	0.400	0.386	3.5	101	-0.01
10	Hexachlorobutadiene	0.400	0.385	3.8	100	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.381	4.8	102	0.00
12	2-Methylnaphthalene	0.400	0.380	5.0	101	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.346	13.5	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.296	26.0#	75	0.00
16	Acenaphthylene	0.400	0.346	13.5	94	0.00
17	Acenaphthene	0.400	0.383	4.3	99	0.00
18	Fluorene	0.400	0.385	3.8	101	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.377	5.8	97	-0.01
21	4-Bromophenyl-phenylether	0.400	0.359	10.3	99	0.00
22	Hexachlorobenzene	0.400	0.365	8.8	98	-0.01
23	Atrazine	0.400	0.311	22.3	92	0.04
24	Pentachlorophenol	0.400	0.426	-6.5	115	0.00
25	Phenanthrene	0.400	0.373	6.8	102	0.00
26	Anthracene	0.400	0.347	13.3	100	-0.01
27 SURR	Fluoranthene-d10	0.400	0.379	5.3	106	0.00
28	Fluoranthene	0.400	0.373	6.8	105	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	109	0.00
30	Pyrene	0.400	0.383	4.3	107	0.00
31 S	Terphenyl-d14	0.400	0.376	6.0	108	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	112	0.00
33	Chrysene	0.400	0.388	3.0	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.349	12.8	107	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.340	15.0	102	0.00
37	Benzo(b)fluoranthene	0.400	0.388	3.0	115	0.00
38	Benzo(k)fluoranthene	0.400	0.383	4.3	114	0.00
39 C	Benzo(a)pyrene	0.400	0.368	8.0	114	0.00
40	Dibenzo(a,h)anthracene	0.400	0.337	15.8	101	0.00
41	Benzo(g,h,i)perylene	0.400	0.350	12.5	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4