

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041724\
 Data File : BN030982.D
 Acq On : 17 Apr 2024 18:06
 Operator : MA/JU
 Sample : PB160253BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK253

Quant Time: Apr 17 22:31:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 17 22:24:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	118687	20.000	ng/u1	0.00
20) Naphthalene-d8	10.422	136	530638	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.287	164	394727	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.034	188	927537	20.000	ng/u1	0.00
79) Chrysene-d12	21.269	240	968086	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	1195220	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	11185	4.418	ng/uL	0.00
4) Pyridine-d5	3.558	84	128094	16.877	ng/u1	0.00
7) Phenol-d5	6.816	99	472	0.049	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.981	67	193501	32.630	ng/u1	0.00
11) 2-Chlorophenol-d4	0.000	132	0	0.000	ng/u1	
15) 4-Methylphenol-d8	8.352	113	2905	0.361	ng/u1	0.00
21) Nitrobenzene-d5	8.799	128	121837	33.099	ng/u1	0.00
24) 2-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
28) 2,4-Dichlorophenol-d3	0.000	165	0	0.000	ng/u1	
31) 4-Chloroaniline-d4	10.563	131	414470	34.150	ng/u1	0.00
46) Dimethylphthalate-d6	13.704	166	1011565	38.383	ng/u1	0.00
49) Acenaphthylene-d8	13.981	160	1049317	34.630	ng/u1	0.00
54) 4-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
60) Fluorene-d10	15.281	176	899747	39.014	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	0.000	200	0	0.000	ng/u1	
73) Anthracene-d10	17.134	188	1622006	40.587	ng/u1	0.00
81) Pyrene-d10	19.439	212	2021119	38.832	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.980	264	2389521	43.388	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041724\
Data File : BN030982.D
Acq On : 17 Apr 2024 18:06
Operator : MA/JU
Sample : PB160253BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK253

Quant Time: Apr 17 22:31:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 17 22:24:16 2024
Response via : Initial Calibration

