

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041724\
 Data File : BN031005.D
 Acq On : 18 Apr 2024 11:14
 Operator : MA/JU
 Sample : P2077-04
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AK5

Quant Time: Apr 18 12:48:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 17 22:24:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	153086	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	697660	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	493687	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.034	188	1116472	20.000	ng/u1	0.00
79) Chrysene-d12	21.269	240	1263681	20.000	ng/u1	0.00
88) Perylene-d12	24.169	264	1624178	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.152	96	11445	3.505	ng/uL	0.00
4) Pyridine-d5	3.570	84	22406	2.289	ng/u1	0.01
7) Phenol-d5	6.811	99	57442	4.638	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.981	67	200547	26.219	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	188524	20.068	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	115419	11.120	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	129377	26.733	ng/u1	0.00
24) 2-Nitrophenol-d4	9.511	143	119318	26.950	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	230721	23.804	ng/u1	0.00
31) 4-Chloroaniline-d4	10.563	131	265940	16.666	ng/u1	0.00
46) Dimethylphthalate-d6	13.699	166	981443	29.776	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	1035779	27.331	ng/u1	0.00
54) 4-Nitrophenol-d4	14.498	143	24204	3.839	ng/u1	0.00
60) Fluorene-d10	15.281	176	852567	29.558	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	115257	23.388	ng/u1	0.00
73) Anthracene-d10	17.128	188	1455886	30.265	ng/u1	0.00
81) Pyrene-d10	19.434	212	1954780	28.772	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	2493083	33.313	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041724\
Data File : BN031005.D
Acq On : 18 Apr 2024 11:14
Operator : MA/JU
Sample : P2077-04
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AK5

Quant Time: Apr 18 12:48:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 17 22:24:16 2024
Response via : Initial Calibration

