

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041724\
 Data File : BN031029.D
 Acq On : 19 Apr 2024 05:05
 Operator : MA/JU
 Sample : PB160251BL
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK251

Quant Time: Apr 19 05:41:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 17 22:24:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	100583	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	471122	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	339657	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.034	188	790944	20.000	ng/u1	0.00
79) Chrysene-d12	21.263	240	877399	20.000	ng/u1	-0.01
88) Perylene-d12	24.168	264	1140473	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	11652	5.431	ng/uL	0.00
4) Pyridine-d5	3.552	84	162800	25.310	ng/u1	0.00
7) Phenol-d5	6.811	99	222640	27.362	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.981	67	145165	28.885	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	178926	28.988	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	189443	27.780	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	92395	28.271	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	90768	30.360	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	179158	27.372	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	298185	27.673	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	709061	31.267	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	757628	29.057	ng/u1	0.00
54) 4-Nitrophenol-d4	14.492	143	113064	26.065	ng/u1	0.00
60) Fluorene-d10	15.281	176	599920	30.231	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.398	200	77226	22.120	ng/u1	-0.01
73) Anthracene-d10	17.128	188	1053655	30.918	ng/u1	0.00
81) Pyrene-d10	19.433	212	1299123	27.540	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.963	264	1673856	31.852	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041724\
Data File : BN031029.D
Acq On : 19 Apr 2024 05:05
Operator : MA/JU
Sample : PB160251BL
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK251

Quant Time: Apr 19 05:41:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 17 22:24:16 2024
Response via : Initial Calibration

