

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041724\
 Data File : BN031039.D
 Acq On : 19 Apr 2024 11:37
 Operator : MA/JU
 Sample : PB160251BS
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS251

Quant Time: Apr 19 12:37:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 17 22:24:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	97305	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	471061	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.287	164	338445	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.033	188	780124	20.000	ng/u1	0.00
79) Chrysene-d12	21.269	240	855984	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	1146869	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	15670	7.550	ng/uL	0.00
4) Pyridine-d5	3.552	84	194503	31.257	ng/u1	0.00
7) Phenol-d5	6.810	99	292554	37.166	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.981	67	160603	33.034	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	227351	38.074	ng/u1	0.00
15) 4-Methylphenol-d8	8.346	113	243313	36.882	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	119366	36.529	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	128926	43.128	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	249099	38.063	ng/u1	0.00
31) 4-Chloroaniline-d4	10.563	131	366645	34.030	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	824342	36.481	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	931735	35.863	ng/u1	0.00
54) 4-Nitrophenol-d4	14.498	143	182748	42.281	ng/u1	0.00
60) Fluorene-d10	15.281	176	722012	36.514	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	137858	40.035	ng/u1	0.00
73) Anthracene-d10	17.133	188	1164573	34.647	ng/u1	0.00
81) Pyrene-d10	19.433	212	1496988	32.528	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.980	264	1904327	36.036	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	24047	10.817	ng/uL	92
5) Pyridine	3.569	79	195378	31.116	ng/u1	95
6) Benzaldehyde	6.787	77	146035	39.838	ng/u1	96
8) Phenol	6.840	94	302090	36.840	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.069	93	221949	33.191	ng/u1	100
12) 2-Chlorophenol	7.205	128	236409	36.906	ng/u1	95
13) 2-Methylphenol	8.081	108	224951	35.600	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.163	45	290982	30.903	ng/u1	98
16) Acetophenone	8.457	105	373929	35.899	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.440	70	189136	36.053	ng/u1	94
18) 4-Methylphenol	8.405	108	258345	36.521	ng/u1	98
19) Hexachloroethane	8.704	117	90839	33.824	ng/u1	98
22) Nitrobenzene	8.834	77	268216	33.624	ng/u1	99
23) Isophorone	9.357	82	561531	35.298	ng/u1	98
25) 2-Nitrophenol	9.540	139	139675	39.756	ng/u1	97
26) 2,4-Dimethylphenol	9.604	107	233580	29.913	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.840	93	329857	32.612	ng/u1	99
29) 2,4-Dichlorophenol	10.069	162	248401	37.186	ng/u1	99
30) Naphthalene	10.469	128	813947	33.690	ng/u1	99
32) 4-Chloroaniline	10.587	127	340040	32.748	ng/u1	99
33) Hexachlorobutadiene	10.746	225	133382	31.723	ng/u1	98
34) Caprolactam	11.357	113	102015	40.880	ng/u1	92
35) 4-Chloro-3-methylphenol	11.710	107	292342	38.856	ng/u1	97
36) 2-Methylnaphthalene	12.087	142	595401	35.289	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041724\
 Data File : BN031039.D
 Acq On : 19 Apr 2024 11:37
 Operator : MA/JU
 Sample : PB160251BS
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS251

Quant Time: Apr 19 12:37:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 17 22:24:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.310	142	601393	35.112	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.457	216	302057	34.583	ng/ul	99
40) Hexachlorocyclopentadiene	12.434	237	130223	26.254	ng/ul	99
41) 2,4,6-Trichlorophenol	12.704	196	210728	38.609	ng/ul	98
42) 2,4,5-Trichlorophenol	12.775	196	231232	37.514	ng/ul	97
43) 1,1'-Biphenyl	13.116	154	784041	33.988	ng/ul	100
44) 2-Chloronaphthalene	13.151	162	614370	33.832	ng/ul	100
45) 2-Nitroaniline	13.369	65	189807	39.275	ng/ul	98
47) Dimethylphthalate	13.751	163	818053	35.046	ng/ul	99
48) 2,6-Dinitrotoluene	13.869	165	164576	38.028	ng/ul	99
50) Acenaphthylene	14.004	152	1044210	34.785	ng/ul	99
51) 3-Nitroaniline	14.198	138	193168	42.052	ng/ul	98
52) Acenaphthene	14.351	153	675004	33.743	ng/ul	100
53) 2,4-Dinitrophenol	14.404	184	90633	45.595	ng/ul	92
55) 4-Nitrophenol	14.510	109	141932	41.447	ng/ul	96
56) Dibenzofuran	14.687	168	966878	34.770	ng/ul	98
57) 2,4-Dinitrotoluene	14.657	165	254726	40.315	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.916	232	210769	41.172	ng/ul	98
59) Diethylphthalate	15.122	149	851627	35.914	ng/ul	99
61) Fluorene	15.339	166	804200	36.384	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.339	204	366921	33.836	ng/ul	97
63) 4-Nitroaniline	15.369	138	215313	48.214	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.422	198	147571	39.885	ng/ul	95
67) N-Nitrosodiphenylamine	15.551	169	732987	34.501	ng/ul	99
68) 4-Bromophenyl-phenylether	16.234	248	243746	32.904	ng/ul	97
69) Hexachlorobenzene	16.334	284	289063	33.138	ng/ul	97
70) Atrazine	16.510	200	272748	37.039	ng/ul	98
71) Pentachlorophenol	16.681	266	193562	37.635	ng/ul	96
72) Phenanthrene	17.075	178	1357841	33.878	ng/ul	100
74) Anthracene	17.169	178	1389657	34.297	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.075	216	306987	32.031	ng/ul	94
76) Pentachlorobenzene	14.604	250	298337	30.646	ng/ul	91
77) Carbazole	17.445	167	1353734	38.098	ng/ul	100
78) Di-n-butylphthalate	18.010	149	1641229	37.862	ng/ul	100
80) Fluoranthene	19.098	202	1712322	31.300	ng/ul	98
82) Pyrene	19.463	202	1841293	31.869	ng/ul	99
83) Butylbenzylphthalate	20.374	149	864201	39.167	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.186	252	651282	41.408	ng/ul	96
85) Benzo(a)anthracene	21.251	228	1966903	34.430	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.180	149	1280131	39.342	ng/ul	99
87) Chrysene	21.316	228	1898711	35.213	ng/ul	99
89) Di-n-octyl phthalate	22.280	149	2457021	33.830	ng/ul	100
90) Benzo(b)fluoranthene	23.257	252	2256806	33.437	ng/ul	98
91) Benzo(k)fluoranthene	23.321	252	2223051	32.432	ng/ul	98
93) Benzo(a)pyrene	24.039	252	2002561	31.408	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.427	276	2827350	41.022	ng/ul	99
95) Dibenzo(a,h)anthracene	27.486	278	2305871	39.741	ng/ul	94
96) Benzo(g,h,i)perylene	28.451	276	2197047	40.605	ng/ul	95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 10.817 ng/uL
 RT: 3.181 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37




Instrument :
 BNA_N
 ClientSampleId :
 SLCS251


Tgt Ion: 88 Resp: 24047
 Ion Ratio Lower Upper
 88 100
 43 32.9 23.8 35.8
 58 63.8 56.6 84.8



#5
 Pyridine
 Concen: 31.116 ng/uL
 RT: 3.569 min Scan# 99
 Delta R.T. -0.006 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37

Tgt Ion: 79 Resp: 195378
 Ion Ratio Lower Upper
 79 100
 52 74.9 56.2 84.4
 51 36.2 27.0 40.4





#6

Benzaldehyde

Concen: 39.838 ng/ul

RT: 6.787 min Scan# 646

Delta R.T. -0.006 min

Lab File: BN031039.D


Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251






Tgt Ion: 77 Resp: 146035

Ion Ratio Lower Upper

77 100

105 104.3 78.1 117.1

106 92.5 75.6 113.4



#8

Phenol

Concen: 36.840 ng/ul

RT: 6.840 min Scan# 655

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion: 94 Resp: 302090


Ion Ratio Lower Upper

94 100

65 26.0 21.4 32.0

66 34.7 28.5 42.7









#10
Bis(2-Chloroethyl)ether
Concen: 33.191 ng/ul
RT: 7.069 min Scan# 694
Delta R.T. -0.006 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion: 93 Resp: 221949
Ion Ratio Lower Upper
93 100
63 73.2 58.2 87.4
95 32.3 26.0 39.0



#12
2-Chlorophenol
Concen: 36.906 ng/ul
RT: 7.205 min Scan# 717
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:128 Resp: 236409
Ion Ratio Lower Upper
128 100
64 41.7 37.0 55.4
130 30.9 26.3 39.5









#13
2-Methylphenol
Concen: 35.600 ng/ul
RT: 8.081 min Scan# 80
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:108 Resp: 224951
Ion Ratio Lower Upper
108 100
107 88.4 71.8 107.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 30.903 ng/ul
RT: 8.163 min Scan# 880
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion: 45 Resp: 290982
Ion Ratio Lower Upper
45 100
77 16.0 12.1 18.1
79 11.9 8.6 13.0









#16
 Acetophenone
 Concen: 35.899 ng/ul
 RT: 8.457 min Scan# 910
 Delta R.T. -0.006 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37




Instrument :
 BNA_N
 ClientSampleId :
 SLCS251


Tgt Ion:105 Resp: 373929
 Ion Ratio Lower Upper
 105 100
 77 77.1 61.2 91.8
 51 25.8 21.0 31.4



#17
 N-Nitroso-di-n-propylamine
 Concen: 36.053 ng/ul
 RT: 8.440 min Scan# 927
 Delta R.T. -0.006 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37

Tgt Ion: 70 Resp: 189136
 Ion Ratio Lower Upper
 70 100
 42 53.9 47.4 71.0
 101 9.9 7.9 11.9
 130 21.6 19.0 28.4









#18
4-Methylphenol
Concen: 36.521 ng/ul
RT: 8.405 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:108 Resp: 258345
Ion Ratio Lower Upper
108 100
107 112.1 88.1 132.1



#19
Hexachloroethane
Concen: 33.824 ng/ul
RT: 8.704 min Scan# 972
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:117 Resp: 90839
Ion Ratio Lower Upper
117 100
201 98.9 76.5 114.7
199 60.9 48.5 72.7









#22
Nitrobenzene
Concen: 33.624 ng/ul
RT: 8.834 min Scan# 994
Delta R.T. -0.006 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion: 77 Resp: 268216
Ion Ratio Lower Upper
77 100
123 46.5 37.5 56.3
65 12.8 10.6 15.8



#23
Isophorone
Concen: 35.298 ng/ul
RT: 9.357 min Scan# 1083
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion: 82 Resp: 561531
Ion Ratio Lower Upper
82 100
95 7.0 5.7 8.5
138 19.1 14.6 21.8









#25
2-Nitrophenol
Concen: 39.756 ng/ul
RT: 9.540 min Scan# 1114
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:139 Resp: 139675
Ion Ratio Lower Upper
139 100
65 48.1 39.5 59.3
109 32.0 23.8 35.8



#26
2,4-Dimethylphenol
Concen: 29.913 ng/ul
RT: 9.604 min Scan# 1125
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:107 Resp: 233580
Ion Ratio Lower Upper
107 100
121 50.2 41.6 62.4
122 86.5 71.3 106.9









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.612 ng/ul
 RT: 9.840 min Scan# 1165
 Delta R.T. -0.006 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37




Instrument :
 BNA_N
 ClientSampleId :
 SLCS251


Tgt Ion: 93 Resp: 329857
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.4 38.0
 123 12.1 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 37.186 ng/ul
 RT: 10.069 min Scan# 1204
 Delta R.T. 0.000 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37

Tgt Ion: 162 Resp: 248401
 Ion Ratio Lower Upper
 162 100
 164 64.4 51.8 77.8
 98 35.5 28.0 42.0









#30
Naphthalene
Concen: 33.690 ng/ul
RT: 10.469 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:128 Resp: 813947
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 12.8 10.5 15.7



#32
4-Chloroaniline
Concen: 32.748 ng/ul
RT: 10.587 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:127 Resp: 340040
Ion Ratio Lower Upper
127 100
129 32.2 25.3 37.9





#33

Hexachlorobutadiene

Concen: 31.723 ng/ul

RT: 10.746 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:225 Resp: 133382

Ion Ratio Lower Upper

225 100

223 63.6 50.9 76.3

227 67.8 51.6 77.4



#34

Caprolactam

Concen: 40.880 ng/ul

RT: 11.357 min Scan# 1423

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:113 Resp: 102015


Ion Ratio Lower Upper

113 100

55 141.9 124.7 187.1

56 107.8 90.6 135.8









#35
4-Chloro-3-methylphenol
Concen: 38.856 ng/ul
RT: 11.710 min Scan# 1483
Delta R.T. -0.006 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:107 Resp: 292342
Ion Ratio Lower Upper
107 100
144 25.9 21.4 32.0
142 79.4 65.9 98.9



#36
2-Methylnaphthalene
Concen: 35.289 ng/ul
RT: 12.087 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:142 Resp: 595401
Ion Ratio Lower Upper
142 100
141 86.0 70.8 106.2









#37
1-Methylnaphthalene
Concen: 35.112 ng/ul
RT: 12.310 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:142 Resp: 601393
Ion Ratio Lower Upper
142 100
141 92.6 73.0 109.4
116 3.7 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.583 ng/ul
RT: 12.457 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:216 Resp: 302057
Ion Ratio Lower Upper
216 100
214 76.6 60.9 91.3
179 21.3 17.4 26.0
108 17.6 14.2 21.4





#40

Hexachlorocyclopentadiene

Concen: 26.254 ng/ul

RT: 12.434 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:237 Resp: 130223

Ion Ratio Lower Upper

237 100

235 64.3 50.6 76.0

272 13.0 10.2 15.4



#41

2,4,6-Trichlorophenol

Concen: 38.609 ng/ul

RT: 12.704 min Scan# 1652

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:196 Resp: 210728


Ion Ratio Lower Upper

196 100

198 95.2 74.6 111.8


200 29.3 24.8 37.2








#42
2,4,5-Trichlorophenol
Concen: 37.514 ng/ul
RT: 12.775 min Scan# 1664
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Instrument :
BNA_N
ClientSampleId :
SLCS251







Tgt Ion:196 Resp: 231232
Ion Ratio Lower Upper
196 100
198 93.6 73.0 109.6
200 29.6 25.2 37.8



#43
1,1'-Biphenyl
Concen: 33.988 ng/ul
RT: 13.116 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:154 Resp: 784041
Ion Ratio Lower Upper
154 100
153 39.2 31.5 47.3
76 12.0 9.8 14.6









#44
2-Chloronaphthalene
Concen: 33.832 ng/ul
RT: 13.151 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:162 Resp: 614370
Ion Ratio Lower Upper
162 100
164 31.9 25.7 38.5
127 40.6 32.2 48.2



#45
2-Nitroaniline
Concen: 39.275 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion: 65 Resp: 189807
Ion Ratio Lower Upper
65 100
92 72.6 57.7 86.5
138 116.8 90.6 135.8





#47

Dimethylphthalate

Concen: 35.046 ng/ul

RT: 13.751 min Scan# 1830

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:163 Resp: 818053

Ion Ratio Lower Upper

163 100

194 4.2 3.3 4.9

164 10.0 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 38.028 ng/ul

RT: 13.869 min Scan# 1850

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:165 Resp: 164576


Ion Ratio Lower Upper

165 100

89 54.7 43.5 65.3

121 20.6 15.7 23.5









#50
Acenaphthylene
Concen: 34.785 ng/ul
RT: 14.004 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:152 Resp: 1044210
Ion Ratio Lower Upper
152 100
151 19.3 15.8 23.8
153 13.2 10.7 16.1



#51
3-Nitroaniline
Concen: 42.052 ng/ul
RT: 14.198 min Scan# 1906
Delta R.T. -0.006 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:138 Resp: 193168
Ion Ratio Lower Upper
138 100
108 12.1 9.6 14.4
92 110.2 89.6 134.4









#52
 Acenaphthene
 Concen: 33.743 ng/ul
 RT: 14.351 min Scan# 1932
 Delta R.T. 0.000 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37

Instrument :
 BNA_N
 ClientSampleId :
 SLCS251

Tgt Ion:153 Resp: 675004




Ion	Ratio	Lower	Upper
153	100		
152	46.4	37.6	56.4
154	87.8	70.3	105.5




#53
 2,4-Dinitrophenol
 Concen: 45.595 ng/ul
 RT: 14.404 min Scan# 1941
 Delta R.T. 0.000 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37

Tgt Ion:184 Resp: 90633

Ion	Ratio	Lower	Upper
184	100		
63	65.7	49.9	74.9
154	70.6	49.5	74.3









#55
4-Nitrophenol
Concen: 41.447 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:109 Resp: 141932
Ion Ratio Lower Upper
109 100
139 135.2 111.8 167.6
65 127.2 98.7 148.1



#56
Dibenzofuran
Concen: 34.770 ng/ul
RT: 14.687 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:168 Resp: 966878
Ion Ratio Lower Upper
168 100
139 39.0 30.2 45.4









#57
2,4-Dinitrotoluene
Concen: 40.315 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:165 Resp: 254726
Ion Ratio Lower Upper
165 100
63 42.9 34.7 52.1
182 2.9 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.172 ng/ul
RT: 14.916 min Scan# 2028
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:232 Resp: 210769
Ion Ratio Lower Upper
232 100
131 44.7 36.6 55.0
130 2.4 2.2 3.2
166 30.6 26.0 39.0





#59

Diethylphthalate

Concen: 35.914 ng/ul

RT: 15.122 min Scan# 2100

Delta R.T. -0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:149 Resp: 851627

Ion Ratio Lower Upper

149 100

177 22.0 17.4 26.2

150 12.8 9.8 14.8



#61

Fluorene

Concen: 36.384 ng/ul

RT: 15.339 min Scan# 2100

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:166 Resp: 804200


Ion Ratio Lower Upper

166 100

165 95.8 78.2 117.4

167 13.5 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 33.836 ng/ul
RT: 15.339 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:204 Resp: 366921
Ion Ratio Lower Upper
204 100
206 33.6 25.8 38.6
141 63.7 49.1 73.7



#63
4-Nitroaniline
Concen: 48.214 ng/ul
RT: 15.369 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:138 Resp: 215313
Ion Ratio Lower Upper
138 100
92 48.8 39.7 59.5
108 75.1 62.6 94.0









#66
4,6-Dinitro-2-methylphenol
Concen: 39.885 ng/ul
RT: 15.422 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:198 Resp: 147571
Ion Ratio Lower Upper
198 100
182 4.2 3.6 5.4
77 25.0 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 34.501 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:169 Resp: 732987
Ion Ratio Lower Upper
169 100
168 66.6 51.9 77.9
167 36.1 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 32.904 ng/ul
RT: 16.234 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:248 Resp: 243746
Ion Ratio Lower Upper
248 100
250 103.3 80.0 120.0
141 78.7 60.5 90.7



#69
Hexachlorobenzene
Concen: 33.138 ng/ul
RT: 16.334 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:284 Resp: 289063
Ion Ratio Lower Upper
284 100
142 37.6 28.6 43.0
249 32.3 25.0 37.4





#70

Atrazine

Concen: 37.039 ng/ul

RT: 16.510 min Scan# 2299

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:200 Resp: 272748

Ion Ratio Lower Upper

200 100

173 26.3 21.5 32.3

215 45.8 37.8 56.8



#71

Pentachlorophenol

Concen: 37.635 ng/ul

RT: 16.681 min Scan# 2328

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:266 Resp: 193562


Ion Ratio Lower Upper

266 100

264 58.0 51.0 76.6

268 65.2 52.5 78.7





#72

Phenanthrene

Concen: 33.878 ng/u1

RT: 17.075 min Scan# 2410

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:178 Resp: 1357841

Ion Ratio Lower Upper

178 100

179 15.8 12.4 18.6

176 18.9 15.0 22.6



#74

Anthracene

Concen: 34.297 ng/u1

RT: 17.169 min Scan# 2411

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:178 Resp: 1389657


Ion Ratio Lower Upper

178 100

179 15.6 12.0 18.0


176 18.4 15.0 22.4








#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.031 ng/uL
RT: 13.075 min Scan# 1715
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37


Instrument :
BNA_N
ClientSampleId :
SLCS251





Tgt Ion:216 Resp: 306987
Ion Ratio Lower Upper
216 100
214 73.4 63.1 94.7
218 46.0 39.5 59.3




#76
Pentachlorobenzene
Concen: 30.646 ng/uL
RT: 14.604 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37



Tgt Ion:250 Resp: 298337
Ion Ratio Lower Upper
250 100
248 63.5 45.8 68.6
252 68.2 48.7 73.1









#77
 Carbazole
 Concen: 38.098 ng/ul
 RT: 17.445 min Scan# 2457
 Delta R.T. 0.000 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37

Instrument :
 BNA_N
 ClientSampleId :
 SLCS251

Tgt Ion:167 Resp: 1353734




Ion	Ratio	Lower	Upper
167	100		
166	20.8	16.6	24.8
139	13.4	10.5	15.7




#78
 Di-n-butylphthalate
 Concen: 37.862 ng/ul
 RT: 18.010 min Scan# 2554
 Delta R.T. -0.000 min
 Lab File: BN031039.D
 Acq: 19 Apr 2024 11:37

Tgt Ion:149 Resp: 1641229

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.2	10.8
104	5.0	4.1	6.1





#80

Fluoranthene

Concen: 31.300 ng/u1

RT: 19.098 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:202 Resp: 1712322

Ion Ratio Lower Upper

202 100

101 11.8 9.0 13.6

100 9.2 6.8 10.2



#82

Pyrene

Concen: 31.869 ng/u1

RT: 19.463 min Scan# 2801

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:202 Resp: 1841293


Ion Ratio Lower Upper

202 100

101 13.9 11.3 16.9

100 11.0 9.2 13.8





#83

Butylbenzylphthalate

Concen: 39.167 ng/ul

RT: 20.374 min Scan# 2956

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:149 Resp: 864201

Ion Ratio Lower Upper

149 100

91 69.7 58.2 87.2

206 19.0 15.4 23.0



#84

3,3'-Dichlorobenzidine

Concen: 41.408 ng/ul

RT: 21.186 min Scan# 3094

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:252 Resp: 651282


Ion Ratio Lower Upper

252 100

254 67.0 51.1 76.7

126 12.5 9.3 13.9









#85
Benzo(a)anthracene
Concen: 34.430 ng/ul
RT: 21.251 min Scan# 3105
Delta R.T. -0.006 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37




Instrument :
BNA_N
ClientSampleId :
SLCS251


Tgt Ion:228 Resp: 1966903
Ion Ratio Lower Upper
228 100
229 20.6 15.8 23.6
226 28.6 20.5 30.7



#86
Bis(2-ethylhexyl)phthalate
Concen: 39.342 ng/ul
RT: 21.180 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BN031039.D
Acq: 19 Apr 2024 11:37

Tgt Ion:149 Resp: 1280131
Ion Ratio Lower Upper
149 100
167 27.0 21.4 32.0
279 3.9 3.5 5.3





#87

Chrysene

Concen: 35.213 ng/ul

RT: 21.316 min Scan# 3116

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:228 Resp: 1898711

Ion Ratio Lower Upper

228 100

226 29.8 24.3 36.5

229 19.8 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 33.830 ng/ul




RT: 22.280 min Scan# 3280


Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Tgt Ion:149 Resp: 2457021





#90

Benzo(b)fluoranthene

Concen: 33.437 ng/ul

RT: 23.257 min Scan# 3445

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:252 Resp: 2256806

Ion Ratio Lower Upper

252 100

253 19.8 16.7 25.1

125 10.7 8.5 12.7



#91

Benzo(k)fluoranthene

Concen: 32.432 ng/ul

RT: 23.321 min Scan# 3457

Delta R.T. 0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:252 Resp: 2223051


Ion Ratio Lower Upper

252 100

253 22.1 16.9 25.3

125 10.4 8.5 12.7





#93

Benzo(a)pyrene

Concen: 31.408 ng/ul

RT: 24.039 min Scan# 31

Delta R.T. 0.000 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:252 Resp: 2002561

Ion Ratio Lower Upper

252 100

253 21.8 16.7 25.1

125 12.0 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 41.022 ng/ul

RT: 27.427 min Scan# 4155

Delta R.T. -0.012 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37




Tgt Ion:276 Resp: 2827350


Ion Ratio Lower Upper

276 100

138 21.8 17.8 26.8

277 22.5 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 39.741 ng/ul

RT: 27.486 min Scan# 4163

Delta R.T. -0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Instrument :

BNA_N

ClientSampleId :

SLCS251





Tgt Ion:278 Resp: 2305871

Ion Ratio Lower Upper

278 100

139 18.2 13.2 19.8

279 25.6 17.4 26.2



#96

Benzo(g,h,i)perylene

Concen: 40.605 ng/ul

RT: 28.451 min Scan# 4329

Delta R.T. 0.006 min

Lab File: BN031039.D

Acq: 19 Apr 2024 11:37

Tgt Ion:276 Resp: 2197047

Ion Ratio Lower Upper

276 100

138 23.1 19.3 28.9

277 20.8 20.1 30.1

