

Data Path : Z:\HPCHEM1\BNA N\Data\BN041818\
 Data File : BN000765.D
 Acq On : 19 Apr 2018 06:02
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 19 07:20:13 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-BN041818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 18 16:58:01 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.618	0.595	3.7	92	0.00
3	Pyridine	1.627	1.719	-5.7	92	0.00
4	n-Nitrosodimethylamine	0.447	0.464	-3.8	92	0.00
5 S	2-Fluorophenol	1.346	1.378	-2.4	94	0.00
6	Aniline	2.434	2.502	-2.8	94	0.00
7 S	Phenol-d6	1.835	1.897	-3.4	94	0.00
8	2-Chlorophenol	1.392	1.430	-2.7	93	0.00
9	Benzaldehyde	0.846	0.881	-4.1	94	0.00
10 C	Phenol	1.934	1.981	-2.4	94	0.00
11	bis(2-Chloroethyl)ether	1.582	1.615	-2.1	95	0.00
12	1,3-Dichlorobenzene	1.617	1.640	-1.4	96	0.00
13 C	1,4-Dichlorobenzene	1.642	1.658	-1.0	94	0.00
14	1,2-Dichlorobenzene	1.567	1.591	-1.5	96	0.00
15	Benzyl Alcohol	1.262	1.372	-8.7	95	0.00
16	2,2'-oxybis(1-Chloropropane	1.680	1.675	0.3	92	0.00
17	2-Methylphenol	1.300	1.364	-4.9	95	0.00
18	Hexachloroethane	0.605	0.622	-2.8	94	0.00
19 P	n-Nitroso-di-n-propylamine	1.141	1.203	-5.4	93	0.00
20	3+4-Methylphenols	1.787	1.894	-6.0	94	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
22	Acetophenone	0.573	0.579	-1.0	93	0.00
23 S	Nitrobenzene-d5	0.384	0.403	-4.9	93	0.00
24	Nitrobenzene	0.411	0.423	-2.9	92	0.00
25	Isophorone	0.695	0.730	-5.0	93	0.00
26 C	2-Nitrophenol	0.169	0.176	-4.1	96	0.00
27	2,4-Dimethylphenol	0.275	0.283	-2.9	95	0.00
28	bis(2-Chloroethoxy)methane	0.453	0.458	-1.1	91	0.00
29 C	2,4-Dichlorophenol	0.271	0.286	-5.5	97	0.00
30	1,2,4-Trichlorobenzene	0.327	0.332	-1.5	96	0.00
31	Naphthalene	1.075	1.076	-0.1	96	0.00
32	Benzoic acid	0.238	0.254	-6.7	94	0.00
33	4-Chloroaniline	0.439	0.456	-3.9	95	0.00
34 C	Hexachlorobutadiene	0.177	0.183	-3.4	100	0.00
35	Caprolactam	0.094	0.099	-5.3	94	0.00
36 C	4-Chloro-3-methylphenol	0.320	0.342	-6.9	95	0.00
37	2-Methylnaphthalene	0.704	0.719	-2.1	97	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
39	1,2,4,5-Tetrachlorobenzene	0.620	0.640	-3.2	98	0.00
40 P	Hexachlorocyclopentadiene	0.364	0.390	-7.1	102	0.00
41 S	2,4,6-Tribromophenol	0.201	0.220	-9.5	98	0.00
42 C	2,4,6-Trichlorophenol	0.376	0.410	-9.0	98	0.00
43	2,4,5-Trichlorophenol	0.435	0.473	-8.7	100	0.00
44 S	2-Fluorobiphenyl	1.449	1.473	-1.7	99	0.00

Data Path : Z:\HPCHEM1\BNA N\Data\BN041818\
 Data File : BN000765.D
 Acq On : 19 Apr 2018 06:02
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 19 07:20:13 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-BN041818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 18 16:58:01 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.790	1.829	-2.2	96	0.00
46	2-Chloronaphthalene	1.342	1.379	-2.8	98	0.00
47	2-Nitroaniline	0.389	0.414	-6.4	97	0.00
48	Acenaphthylene	2.128	2.224	-4.5	97	0.00
49	Dimethylphthalate	1.598	1.641	-2.7	96	0.00
50	2,6-Dinitrotoluene	0.339	0.352	-3.8	95	0.00
51 C	Acenaphthene	1.297	1.311	-1.1	95	0.00
52	3-Nitroaniline	0.385	0.397	-3.1	93	0.00
53 P	2,4-Dinitrophenol	0.154	0.169	-9.7	100	0.00
54	Dibenzofuran	1.888	1.900	-0.6	93	0.00
55 P	4-Nitrophenol	0.286	0.295	-3.1	95	0.00
56	2,4-Dinitrotoluene	0.443	0.465	-5.0	95	0.00
57	Fluorene	1.493	1.495	-0.1	92	0.00
58	2,3,4,6-Tetrachlorophenol	0.334	0.354	-6.0	94	0.00
59	Diethylphthalate	1.584	1.632	-3.0	93	0.00
60	4-Chlorophenyl-phenylether	0.698	0.696	0.3	93	0.00
61	4-Nitroaniline	0.379	0.387	-2.1	92	0.00
62	Azobenzene	1.696	1.715	-1.1	90	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
64	4,6-Dinitro-2-methylphenol	0.124	0.129	-4.0	95	0.00
65 c	n-Nitrosodiphenylamine	0.671	0.686	-2.2	93	0.00
66	4-Bromophenyl-phenylether	0.203	0.210	-3.4	95	0.00
67	Hexachlorobenzene	0.235	0.242	-3.0	97	0.00
68	Atrazine	0.182	0.199	-9.3	94	0.00
69 C	Pentachlorophenol	0.152	0.159	-4.6	96	0.00
70	Phenanthrene	1.158	1.156	0.2	96	0.00
71	Anthracene	1.134	1.143	-0.8	97	0.00
72	Carbazole	1.066	1.060	0.6	96	0.00
73	Di-n-butylphthalate	1.215	1.276	-5.0	98	0.00
74 C	Fluoranthene	1.167	1.152	1.3	97	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
76	Benzidine	0.542	0.601	-10.9	105	0.00
77	Pyrene	1.441	1.515	-5.1	94	0.00
78 S	Terphenyl-d14	0.928	0.960	-3.4	94	0.00
79	Butylbenzylphthalate	0.607	0.643	-5.9	98	-0.01
80	Benzo(a)anthracene	1.260	1.290	-2.4	98	0.00
81	3,3'-Dichlorobenzidine	0.397	0.427	-7.6	101	0.00
82	Chrysene	1.220	1.235	-1.2	98	0.00
83	Bis(2-ethylhexyl)phthalate	0.927	0.950	-2.5	96	0.00
84 c	Di-n-octyl phthalate	1.326	1.378	-3.9	98	-0.01
85	Indeno(1,2,3-cd)pyrene	1.049	1.148	-9.4	113	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	109	-0.01
87	Benzo(b)fluoranthene	1.215	1.241	-2.1	109	-0.01

Data Path : Z:\HPCHEM1\BNA N\Data\BN041818\
Data File : BN000765.D
Acq On : 19 Apr 2018 06:02
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC040

Quant Time: Apr 19 07:20:13 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-BN041818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 18 16:58:01 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.210	1.232	-1.8	107	0.00
89 C	Benzo(a)pyrene	1.109	1.172	-5.7	109	0.00
90	Dibenzo(a,h)anthracene	1.093	1.187	-8.6	113	-0.01
91	Benzo(a,h,i)perylene	1.067	1.167	-9.4	113	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0