

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041822\
 Data File : BM034716.D
 Acq On : 18 Apr 2022 13:27
 Operator : CG/JU
 Sample : SSTD02020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Apr 18 15:42:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 18 15:40:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	101924	20.000	ng/ul	# 0.00
20) Naphthalene-d8	10.616	136	425683	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.463	164	248244	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.204	188	484998	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.386	240	467954	20.000	ng/ul	# 0.00
88) Perylene-d12	23.733	264	481109	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.169	96	23122	8.090	ng/uL	0.00
4) Pyridine-d5	3.599	84	131251	19.028	ng/ul	0.00
7) Phenol-d5	6.975	99	174439	20.563	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	119966	21.222	ng/ul	0.00
11) 2-Chlorophenol-d4	7.340	132	139759	21.312	ng/ul	0.00
15) 4-Methylphenol-d8	8.528	113	138754	20.939	ng/ul	0.00
21) Nitrobenzene-d5	8.969	128	66115	21.133	ng/ul	0.00
24) 2-Nitrophenol-d4	9.698	143	71150	21.844	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.239	165	131194	21.390	ng/ul	0.00
31) 4-Chloroaniline-d4	10.757	131	167606	19.808	ng/ul	0.00
46) Dimethylphthalate-d6	13.874	166	381196	21.251	ng/ul	0.00
49) Acenaphthylene-d8	14.157	160	498092	21.605	ng/ul	0.00
54) 4-Nitrophenol-d4	14.674	143	42766	15.636	ng/ul	0.00
60) Fluorene-d10	15.457	176	322198	20.786	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.580	200	47680	16.104	ng/ul	0.00
73) Anthracene-d10	17.304	188	476331	20.600	ng/ul	0.00
81) Pyrene-d10	19.598	212	544147	21.305	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.580	264	504366	20.841	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.205	88	24196	8.407	ng/uL#	81
5) Pyridine	3.616	79	146045	19.376	ng/ul#	67
6) Benzaldehyde	6.940	77	85675	16.473	ng/ul	90
8) Phenol	7.004	94	174307	20.339	ng/ul	96
10) Bis(2-Chloroethyl)ether	7.228	93	146254	21.146	ng/ul#	85
12) 2-Chlorophenol	7.369	128	143174	21.418	ng/ul#	85
13) 2-Methylphenol	8.263	108	131317	20.727	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.345	45	241167	23.322	ng/ul#	85
16) Acetophenone	8.634	105	221636	21.428	ng/ul#	84
17) N-Nitroso-di-n-propyla...	8.616	70	124702	22.650	ng/ul#	84
18) 4-Methylphenol	8.592	108	142904	21.204	ng/ul	98
19) Hexachloroethane	8.892	117	64314	21.390	ng/ul	82
22) Nitrobenzene	9.010	77	176515	21.522	ng/ul	89
23) Isophorone	9.539	82	322198	22.122	ng/ul#	92
25) 2-Nitrophenol	9.728	139	76159	21.830	ng/ul#	85
26) 2,4-Dimethylphenol	9.798	107	166374	21.485	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.028	93	193020	22.027	ng/ul	99
29) 2,4-Dichlorophenol	10.269	162	126117	20.834	ng/ul#	95
30) Naphthalene	10.663	128	461602	20.829	ng/ul	99
32) 4-Chloroaniline	10.781	127	167425	19.750	ng/ul	99
33) Hexachlorobutadiene	10.957	225	84240	19.794	ng/ul	98
34) Caprolactam	11.539	113	5803	2.951	ng/ul#	65
35) 4-Chloro-3-methylphenol	11.916	107	139666	22.226	ng/ul	90
36) 2-Methylnaphthalene	12.286	142	297453	20.943	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041822\
 Data File : BM034716.D
 Acq On : 18 Apr 2022 13:27
 Operator : CG/JU
 Sample : SST02020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Apr 18 15:42:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 18 15:40:19 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.504	142	307346	21.092	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.657	216	151656	20.410	ng/ul#	97
40) Hexachlorocyclopentadiene	12.639	237	84357	17.300	ng/ul#	94
41) 2,4,6-Trichlorophenol	12.898	196	92471	20.486	ng/ul	93
42) 2,4,5-Trichlorophenol	12.975	196	93714	20.035	ng/ul	99
43) 1,1'-Biphenyl	13.298	154	400357	21.290	ng/ul#	97
44) 2-Chloronaphthalene	13.339	162	303095	20.994	ng/ul	97
45) 2-Nitroaniline	13.545	65	88641	22.517	ng/ul#	77
47) Dimethylphthalate	13.922	163	369819	20.950	ng/ul	99
48) 2,6-Dinitrotoluene	14.039	165	71862	21.191	ng/ul#	83
50) Acenaphthylene	14.180	152	502403	21.465	ng/ul	99
51) 3-Nitroaniline	14.374	138	35978	12.135	ng/ul	85
52) Acenaphthene	14.527	153	329044	21.116	ng/ul	98
53) 2,4-Dinitrophenol	14.580	184	15837	8.287	ng/ul#	62
55) 4-Nitrophenol	14.692	109	39674	16.986	ng/ul#	83
56) Dibenzofuran	14.863	168	447690	21.139	ng/ul	95
57) 2,4-Dinitrotoluene	14.833	165	101459	21.496	ng/ul#	73
58) 2,3,4,6-Tetrachlorophenol	15.092	232	79025	19.243	ng/ul#	87
59) Diethylphthalate	15.286	149	377312	21.044	ng/ul	98
61) Fluorene	15.510	166	370881	20.756	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.504	204	178469	20.692	ng/ul	91
63) 4-Nitroaniline	15.545	138	8737	3.347	ng/ul#	91
66) 4,6-Dinitro-2-methylph...	15.592	198	47769	16.181	ng/ul#	80
67) N-Nitrosodiphenylamine	15.721	169	308951	21.381	ng/ul	99
68) 4-Bromophenyl-phenylether	16.398	248	104315	20.807	ng/ul#	90
69) Hexachlorobenzene	16.515	284	122817	20.398	ng/ul#	90
70) Atrazine	16.668	200	108160	20.002	ng/ul	98
71) Pentachlorophenol	16.863	266	42190	11.444	ng/ul	99
72) Phenanthrene	17.245	178	563650	20.827	ng/ul	99
74) Anthracene	17.339	178	557211	20.778	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.263	216	158230	20.773	ng/ul	99
76) Pentachlorobenzene	14.786	250	147891	20.689	ng/ul	99
77) Carbazole	17.615	167	443000	19.969	ng/ul	98
78) Di-n-butylphthalate	18.180	149	595455	21.148	ng/ul	99
80) Fluoranthene	19.268	202	609454	21.562	ng/ul#	90
82) Pyrene	19.627	202	659251	21.459	ng/ul#	87
83) Butylbenzylphthalate	20.527	149	263176	21.882	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.309	252	139601	17.022	ng/ul#	97
85) Benzo(a)anthracene	21.374	228	571246	20.994	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.309	149	381600	21.200	ng/ul	98
87) Chrysene	21.427	228	593359	20.223	ng/ul	100
89) Di-n-octyl phthalate	22.209	149	651902	20.353	ng/ul	100
90) Benzo(b)fluoranthene	23.021	252	612694	21.432	ng/ul#	97
91) Benzo(k)fluoranthene	23.068	252	611615	20.373	ng/ul#	97
93) Benzo(a)pyrene	23.627	252	585202	20.712	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.162	276	646575	21.384	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.180	278	512035	20.656	ng/ul#	96
96) Benzo(g,h,i)perylene	26.909	276	601838	20.737	ng/ul#	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041822\
 Data File : BM034716.D
 Acq On : 18 Apr 2022 13:27
 Operator : CG/JU
 Sample : SST02020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Apr 18 15:42:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 18 15:40:19 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.407 ng/uL
RT: 3.205 min Scan# 31
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 88 Resp: 24196
Ion Ratio Lower Upper
88 100
43 34.3 21.0 31.6#
58 75.3 47.8 71.8#



#5
Pyridine
Concen: 19.376 ng/uL
RT: 3.616 min Scan# 107
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion: 79 Resp: 146045
Ion Ratio Lower Upper
79 100
52 85.6 47.7 71.5#
51 48.5 25.6 38.4#









#6
Benzaldehyde
Concen: 16.473 ng/ul
RT: 6.940 min Scan# 672
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 77 Resp: 85675
Ion Ratio Lower Upper
77 100
105 89.6 80.8 121.2
106 87.0 76.6 115.0



#8
Phenol
Concen: 20.339 ng/ul
RT: 7.004 min Scan# 683
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion: 94 Resp: 174307
Ion Ratio Lower Upper
94 100
65 32.9 22.9 34.3
66 43.1 33.7 50.5









#10
 Bis(2-Chloroethyl)ether
 Concen: 21.146 ng/ul
 RT: 7.228 min Scan# 721
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27

Instrument :
 BNA_N
 ClientSampleId :

Tgt Ion: 93 Resp: 146254




Ion	Ratio	Lower	Upper
93	100		
63	85.3	54.7	82.1#
95	32.2	26.6	39.8




#12
 2-Chlorophenol
 Concen: 21.418 ng/ul
 RT: 7.369 min Scan# 745
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27

Tgt Ion:128 Resp: 143174

Ion	Ratio	Lower	Upper
128	100		
64	54.6	30.6	46.0#
130	31.6	25.8	38.6









#13
2-Methylphenol
Concen: 20.727 ng/ul
RT: 8.263 min Scan# 897
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:108 Resp: 131317
Ion Ratio Lower Upper
108 100
107 92.5 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 23.322 ng/ul
RT: 8.345 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27


Tgt Ion: 45 Resp: 241167
Ion Ratio Lower Upper
45 100
77 13.2 17.1 25.7#
79 10.7 12.8 19.2#








#16
 Acetophenone
 Concen: 21.428 ng/ul
 RT: 8.634 min Scan# 960
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27


Instrument :
 BNA_N
 ClientSampleId :





Tgt Ion: 105 Resp: 221636
 Ion Ratio Lower Upper
 105 100
 77 86.1 59.4 89.2
 51 34.7 19.3 28.9#




#17
 N-Nitroso-di-n-propylamine
 Concen: 22.650 ng/ul
 RT: 8.616 min Scan# 957
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27



Tgt Ion: 70 Resp: 124702
 Ion Ratio Lower Upper
 70 100
 42 63.8 40.6 60.8#
 101 8.9 8.2 12.2
 130 16.9 19.9 29.9#





#18

4-Methylphenol

Concen: 21.204 ng/ul

RT: 8.592 min Scan# 91

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N





ClientSampleId :

Tgt Ion:108 Resp: 142904

Ion Ratio Lower Upper

108 100

107 108.7 89.0 133.4



#19

Hexachloroethane

Concen: 21.390 ng/ul

RT: 8.892 min Scan# 1004

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion:117 Resp: 64314


Ion Ratio Lower Upper

117 100

201 91.6 89.9 134.9

199 58.0 57.2 85.8





#22

Nitrobenzene

Concen: 21.522 ng/ul

RT: 9.010 min Scan# 1024

Delta R.T. 0.000 min


Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :






Tgt Ion: 77 Resp: 176515

Ion Ratio Lower Upper

77 100

123 38.4 38.2 57.2

65 14.0 10.6 15.8



#23

Isophorone

Concen: 22.122 ng/ul

RT: 9.539 min Scan# 1114

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion: 82 Resp: 322198


Ion Ratio Lower Upper

82 100

95 7.0 6.8 10.2

138 16.1 16.5 24.7#









#25
2-Nitrophenol
Concen: 21.830 ng/ul
RT: 9.728 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:139 Resp: 76159
Ion Ratio Lower Upper
139 100
65 62.9 37.8 56.6#
109 31.4 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 21.485 ng/ul
RT: 9.798 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:107 Resp: 166374
Ion Ratio Lower Upper
107 100
121 49.8 40.8 61.2
122 81.8 69.0 103.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.027 ng/ul
 RT: 10.028 min Scan# 1197
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion: 93 Resp: 193020
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.4 38.2
 123 10.5 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 20.834 ng/ul
 RT: 10.269 min Scan# 1238
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27

Tgt Ion: 162 Resp: 126117
 Ion Ratio Lower Upper
 162 100
 164 65.1 52.8 79.2
 98 39.1 26.0 39.0#









#30
Naphthalene
Concen: 20.829 ng/u1
RT: 10.663 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:128 Resp: 461602
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.0 9.9 14.9



#32
4-Chloroaniline
Concen: 19.750 ng/u1
RT: 10.781 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:127 Resp: 167425
Ion Ratio Lower Upper
127 100
129 32.1 25.1 37.7





#33

Hexachlorobutadiene

Concen: 19.794 ng/ul

RT: 10.957 min Scan# 1

Delta R.T. 0.000 min


Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:225 Resp: 84240

Ion Ratio Lower Upper

225 100

223 64.6 50.6 75.8

227 61.5 50.3 75.5



#34

Caprolactam

Concen: 2.951 ng/ul

RT: 11.539 min Scan# 1454

Delta R.T. -0.012 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion:113 Resp: 5803


Ion Ratio Lower Upper

113 100

55 173.1 100.9 151.3#

56 129.8 81.0 121.4#









#35
4-Chloro-3-methylphenol
Concen: 22.226 ng/ul
RT: 11.916 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:107 Resp: 139666
Ion Ratio Lower Upper
107 100
144 24.4 23.0 34.4
142 77.2 69.8 104.8



#36
2-Methylnaphthalene
Concen: 20.943 ng/ul
RT: 12.286 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:142 Resp: 297453
Ion Ratio Lower Upper
142 100
141 88.5 71.4 107.0





#37

1-Methylnaphthalene

Concen: 21.092 ng/ul

RT: 12.504 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:142 Resp: 307346

Ion Ratio Lower Upper

142 100

141 92.3 74.5 111.7

116 4.3 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.410 ng/ul

RT: 12.657 min Scan# 1644

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Tgt Ion:216 Resp: 151656




Ion Ratio Lower Upper


216 100

214 78.9 62.2 93.2

179 21.6 15.4 23.2

108 19.5 11.8 17.6#









#40
Hexachlorocyclopentadiene
Concen: 17.300 ng/ul
RT: 12.639 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:237 Resp: 84357
Ion Ratio Lower Upper
237 100
235 61.8 52.9 79.3
272 11.8 11.8 17.8#



#41
2,4,6-Trichlorophenol
Concen: 20.486 ng/ul
RT: 12.898 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27


Tgt Ion:196 Resp: 92471
Ion Ratio Lower Upper
196 100
198 94.1 80.8 121.2
200 29.9 26.5 39.7








#42
2,4,5-Trichlorophenol
Concen: 20.035 ng/u1
RT: 12.975 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27


Instrument :
BNA_N
ClientSampleId :





Tgt Ion:196 Resp: 93714
Ion Ratio Lower Upper
196 100
198 98.2 78.0 117.0
200 30.4 24.5 36.7




#43
1,1'-Biphenyl
Concen: 21.290 ng/u1
RT: 13.298 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Tgt Ion:154 Resp: 400357
Ion Ratio Lower Upper
154 100
153 39.8 32.6 49.0
76 13.8 7.7 11.5#








#44
2-Chloronaphthalene
Concen: 20.994 ng/ul
RT: 13.339 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:162 Resp: 303095
Ion Ratio Lower Upper
162 100
164 33.4 26.6 40.0
127 40.2 29.6 44.4



#45
2-Nitroaniline
Concen: 22.517 ng/ul
RT: 13.545 min Scan# 1795
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion: 65 Resp: 88641
Ion Ratio Lower Upper
65 100
92 62.8 57.4 86.2
138 84.8 94.8 142.2#









#47
Dimethylphthalate
Concen: 20.950 ng/ul
RT: 13.922 min Scan# 1859
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:163 Resp: 369819
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.2
164 9.8 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 21.191 ng/ul
RT: 14.039 min Scan# 1879
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:165 Resp: 71862
Ion Ratio Lower Upper
165 100
89 70.1 44.2 66.4#
121 24.6 16.7 25.1





#50

Acenaphthylene

Concen: 21.465 ng/ul

RT: 14.180 min Scan# 1903

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:152 Resp: 502403

Ion Ratio Lower Upper

152 100

151 20.1 15.8 23.8

153 13.2 10.2 15.2



#51

3-Nitroaniline

Concen: 12.135 ng/ul

RT: 14.374 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion:138 Resp: 35978


Ion Ratio Lower Upper

138 100

108 10.3 8.3 12.5

92 125.6 87.0 130.4









#52
Acenaphthene
Concen: 21.116 ng/ul
RT: 14.527 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:153 Resp: 329044
Ion Ratio Lower Upper
153 100
152 46.2 37.4 56.0
154 86.9 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 8.287 ng/ul
RT: 14.580 min Scan# 1971
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:184 Resp: 15837
Ion Ratio Lower Upper
184 100
63 98.9 43.8 65.6#
154 75.8 49.9 74.9#









#55
4-Nitrophenol
Concen: 16.986 ng/ul
RT: 14.692 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:109 Resp: 39674
Ion Ratio Lower Upper
109 100
139 116.5 94.6 142.0
65 140.4 83.5 125.3#



#56
Dibenzofuran
Concen: 21.139 ng/ul
RT: 14.863 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:168 Resp: 447690
Ion Ratio Lower Upper
168 100
139 40.0 29.8 44.6









#57
 2,4-Dinitrotoluene
 Concen: 21.496 ng/ul
 RT: 14.833 min Scan# 2014
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:165 Resp: 101459
 Ion Ratio Lower Upper
 165 100
 63 55.1 30.2 45.4#
 182 3.5 2.2 3.2#



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 19.243 ng/ul
 RT: 15.092 min Scan# 2058
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27

Tgt Ion:232 Resp: 79025
 Ion Ratio Lower Upper
 232 100
 131 51.2 32.2 48.4#
 130 3.4 1.9 2.9#
 166 33.3 23.0 34.4









#59
Diethylphthalate
Concen: 21.044 ng/ul
RT: 15.286 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:149 Resp: 377312
Ion Ratio Lower Upper
149 100
177 21.1 17.9 26.9
150 12.4 10.2 15.2



#61
Fluorene
Concen: 20.756 ng/ul
RT: 15.510 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:166 Resp: 370881
Ion Ratio Lower Upper
166 100
165 98.6 76.9 115.3
167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 20.692 ng/ul
RT: 15.504 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:204 Resp: 178469
Ion Ratio Lower Upper
204 100
206 31.2 25.5 38.3
141 65.5 44.2 66.2



#63
4-Nitroaniline
Concen: 3.347 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:138 Resp: 8737
Ion Ratio Lower Upper
138 100
92 58.3 36.3 54.5#
108 82.7 68.9 103.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 16.181 ng/ul
 RT: 15.592 min Scan# 2143
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:198 Resp: 47769
 Ion Ratio Lower Upper
 198 100
 182 5.6 3.8 5.6
 77 34.4 18.5 27.7#



#67
 N-Nitrosodiphenylamine
 Concen: 21.381 ng/ul
 RT: 15.721 min Scan# 2165
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27

Tgt Ion:169 Resp: 308951
 Ion Ratio Lower Upper
 169 100
 168 66.7 52.8 79.2
 167 35.8 28.1 42.1









#68
4-Bromophenyl-phenylether
Concen: 20.807 ng/ul
RT: 16.398 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:248 Resp: 104315
Ion Ratio Lower Upper
248 100
250 98.9 78.4 117.6
141 82.6 51.4 77.2#



#69
Hexachlorobenzene
Concen: 20.398 ng/ul
RT: 16.515 min Scan# 2300
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:284 Resp: 122817
Ion Ratio Lower Upper
284 100
142 40.9 24.6 36.8#
249 30.6 24.1 36.1





#70

Atrazine

Concen: 20.002 ng/ul

RT: 16.668 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:200 Resp: 108160

Ion Ratio Lower Upper

200 100

173 26.6 20.2 30.4

215 47.0 38.4 57.6



#71

Pentachlorophenol

Concen: 11.444 ng/ul

RT: 16.863 min Scan# 2359

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion:266 Resp: 42190


Ion Ratio Lower Upper

266 100

264 60.4 49.3 73.9

268 63.3 50.3 75.5









#72
Phenanthrene
Concen: 20.827 ng/u1
RT: 17.245 min Scan# 2424
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:178 Resp: 563650
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 18.7 15.0 22.4



#74
Anthracene
Concen: 20.778 ng/u1
RT: 17.339 min Scan# 2440
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:178 Resp: 557211
Ion Ratio Lower Upper
178 100
179 15.5 11.9 17.9
176 18.3 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.773 ng/uL

RT: 13.263 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:216 Resp: 158230

Ion Ratio Lower Upper

216 100

214 78.0 62.0 93.0

218 48.2 37.8 56.8



#76

Pentachlorobenzene

Concen: 20.689 ng/uL

RT: 14.786 min Scan# 2006

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion:250 Resp: 147891


Ion Ratio Lower Upper

250 100

248 62.0 50.6 75.8

252 62.1 50.2 75.4





#77

Carbazole

Concen: 19.969 ng/ul

RT: 17.615 min Scan# 2487

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:167 Resp: 443000

Ion Ratio Lower Upper

167 100

166 22.0 16.6 25.0

139 13.2 9.9 14.9



#78

Di-n-butylphthalate

Concen: 21.148 ng/ul

RT: 18.180 min Scan# 2583

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion:149 Resp: 595455


Ion Ratio Lower Upper

149 100

150 8.8 7.4 11.0

104 5.8 4.2 6.2









#80
 Fluoranthene
 Concen: 21.562 ng/u1
 RT: 19.268 min Scan# 2768
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:202 Resp: 609454
 Ion Ratio Lower Upper
 202 100
 101 11.2 6.2 9.2#
 100 9.2 4.7 7.1#



#82
 Pyrene
 Concen: 21.459 ng/u1
 RT: 19.627 min Scan# 2829
 Delta R.T. 0.000 min
 Lab File: BM034716.D
 Acq: 18 Apr 2022 13:27

Tgt Ion:202 Resp: 659251
 Ion Ratio Lower Upper
 202 100
 101 13.3 7.0 10.4#
 100 12.2 5.8 8.6#





#83

Butylbenzylphthalate

Concen: 21.882 ng/ul

RT: 20.527 min Scan# 2982

Delta R.T. 0.000 min

Lab File: BM034716.D





Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :

Tgt Ion:	149	Resp:	263176
Ion Ratio	Lower	Upper	
149	100		
91	80.5	50.9	76.3#
206	19.4	19.3	28.9



#84

3,3'-Dichlorobenzidine

Concen: 17.022 ng/ul




RT: 21.309 min Scan# 3115


Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Tgt Ion:	252	Resp:	139601
Ion Ratio	Lower	Upper	
252	100		
254	64.9	50.7	76.1
126	11.8	6.6	9.8#









#85
Benzo(a)anthracene
Concen: 20.994 ng/ul
RT: 21.374 min Scan# 3126
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:228 Resp: 571246
Ion Ratio Lower Upper
228 100
229 19.2 15.4 23.2
226 27.1 21.0 31.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.200 ng/ul
RT: 21.309 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:149 Resp: 381600
Ion Ratio Lower Upper
149 100
167 26.4 22.1 33.1
279 4.2 4.2 6.2









#87
Chrysene
Concen: 20.223 ng/ul
RT: 21.427 min Scan# 3135
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:228 Resp: 593359
Ion Ratio Lower Upper
228 100
226 29.1 23.4 35.2
229 19.5 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 20.353 ng/ul
RT: 22.209 min Scan# 3268
Delta R.T. 0.000 min
Lab File: BM034716.D
Acq: 18 Apr 2022 13:27

Tgt Ion:149 Resp: 651902





#90

Benzo(b)fluoranthene

Concen: 21.432 ng/ul

RT: 23.021 min Scan# 3406

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:252 Resp: 612694

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 10.2 5.8 8.6#



#91

Benzo(k)fluoranthene

Concen: 20.373 ng/ul

RT: 23.068 min Scan# 3414

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion:252 Resp: 611615


Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3

125 9.9 5.1 7.7#





#93

Benzo(a)pyrene

Concen: 20.712 ng/ul

RT: 23.627 min Scan# 3509

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:252 Resp: 585202

Ion Ratio Lower Upper

252 100

253 22.5 17.5 26.3

125 11.4 6.2 9.4#



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.384 ng/ul

RT: 26.162 min Scan# 3940

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27




Tgt Ion:276 Resp: 646575


Ion Ratio Lower Upper

276 100

138 23.6 13.6 20.4#

277 24.4 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 20.656 ng/ul

RT: 26.180 min Scan# 3943

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:278 Resp: 512035

Ion Ratio Lower Upper

278 100

139 15.2 9.6 14.4#

279 23.3 19.3 28.9



#96

Benzo(g,h,i)perylene

Concen: 20.737 ng/ul

RT: 26.909 min Scan# 4067

Delta R.T. 0.000 min

Lab File: BM034716.D

Acq: 18 Apr 2022 13:27

Tgt Ion:276 Resp: 601838

Ion Ratio Lower Upper

276 100

138 20.1 11.8 17.6#

277 24.2 18.8 28.2

